

# Hach Digital Titrator



## Accurate, convenient titrations without the bulk, fragility, or waste of a conventional buret.

The Digital Titrator is a unique device that, despite its name, relies on mechanical components and human operation, eliminating the need for an external power source. This innovative tool is designed to accommodate interchangeable titrant cartridges, allowing for multiple titrations by simply changing the cartridge and delivery tube. This feature significantly reduces the time spent on cleaning and assembling cumbersome, fragile glass burets, and virtually eliminates the risk of cross-contamination and over-titration.

The Digital Titrator offers high precision and accuracy comparable to traditional buret titrations, but with reduced sensitivity to technique. Its design ensures it is both portable and durable, making it suitable for both laboratory and field use. This precision dispensing device performs titrations quickly and economically. Designed and built for durability, the Digital Titrator comes with a lifetime warranty: Hach<sup>®</sup> Company will repair or replace it free of charge, provided it has not been abused. Whether you carry it to sample sites within the water or wastewater plant or take it to the field for ecology or water quality studies, the Digital Titrator is built to withstand heavy use.

Weighing just 132 grams (less than 4 oz.), and with a variety of titrant cartridges available, the Digital Titrator is a versatile and valuable addition to any lab.

### Accurate results - when and where you need them

Digital Titrator methods use a high-precision dispensation device and cartridges of concentrated titrant to allow analyses in the lab, the plant or the field.

### Fast and Flexible

Interchangeable cartridges allow you to change titrants in seconds. No cleaning or rinsing burets.

### Versatile

More than 50 parameter/range combinations are available

### Economical

Concentrated titrants allow 40-100 typical titrations per cartridge.

### End-Point easier to determine

The number used for calculation of sample concentration is recorded directly from the Digital Titrator's digital counter. For many methods, the displayed value is the concentration.

## Technical Data\*

### Digital Titrator

<b>Delivery</b>	800 digits/mL
<b>Accuracy</b>	± 1% (Uncertainty of readings is 1 digit. Most samples require more than 100 digits)
<b>Weight</b>	132 g (4 oz.)

### Cartridges

<b>Volume</b>	13 mL
<b>Number of tests</b>	50 - 100. Most reagents are formulated to provide 100 typical titrations. Number of tests may vary depending on sample strength.
<b>Weight</b>	56.75 g (2 oz.)

*\*Subject to change without notice.*

## Titration Cartridges Available\*

Parameter	Part-Number	Application†	Description	Ranges (vary with sample volume)
Acid-Base	1439001	4	8.00 N HCl	10 - 4000 meq/L Base
	1439101		8.00 N H <sub>2</sub> SO <sub>4</sub>	10 - 4000 meq/L Base
	1438101		8 N NaOH	10 - 4000 meq/L Acid
Acidity	1437701	1, 2, 3	0.1600 N NaOH	10 - 160 mg/L CaCO <sub>3</sub>
	1437901		1.600 N NaOH	1 - 4000 mg/L CaCO <sub>3</sub>
Alkalinity	1438801	1, 2, 3	0.1600 N H <sub>2</sub> SO <sub>4</sub>	10 - 160 mg/L CaCO <sub>3</sub>
	1438901		1.600 N H <sub>2</sub> SO <sub>4</sub>	100 - 4000 mg/L CaCO <sub>3</sub>
Hardness, Calcium and Total	1436401	1, 2, 3	0.0800 M EDTA	10 - 160 mg/L CaCO <sub>3</sub>
	1439901		0.800 M EDTA	100 - 4000 mg/L CaCO <sub>3</sub>
	1440201		CDTA‡, 0.0800 M	10 - 160 mg/L CaCO <sub>3</sub>
	1440301		CDTA‡, 0.800 M	100 - 4000 mg/L CaCO <sub>3</sub>
	1496001		0.1428 M EDTA	1 - 16 G.d.h.
	1495901		0.714 M EDTA	10 - 200 G.d.h.
Carbon Dioxide	1437801	1, 2	0.3636 N NaOH	10 - 100 mg/L CO <sub>2</sub>
	1438001		3.636 N NaOH	100 - 1000 mg/L CO <sub>2</sub>
Chloride	1439601	1, 2, 3	0.2256 N AgNO <sub>3</sub>	10 - 100 mg/L Cl <sup>-</sup>
	1439701		1.128 N AgNO <sub>3</sub>	100 - 10000 mg/L Cl <sup>-</sup>
	1439301		0.2256 N Hg(NO <sub>3</sub> ) <sub>2</sub>	10 - 160 mg/L Cl <sup>-</sup>
	92101		2.256 N Hg(NO <sub>3</sub> ) <sub>2</sub>	100 - 8000 mg/L Cl <sup>-</sup>
Chlorine	2259901	1, 2, 3	PAO**, 0.00451 N	0.001 - 1 mg/L as Cl <sub>2</sub>
	2142001		PAO**, 0.00451 N	0.01 - 10 mg/L as Cl <sub>2</sub>
	199901	1, 2	0.00564 N PAO	0 - 1000 µg/L Cl <sub>2</sub>
	2292301		0.00564 N FEAS	0 - 3.00 mg/L Cl <sub>2</sub>
	2267301	4	0.113 N Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub>	20 - 2000 mg/L Cl <sub>2</sub>
	1440101		2.00 N Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub>	2000 - 70000 mg/L Cl <sub>2</sub>
Chromate	2267601	4	0.2068 N Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub>	20 - 400 mg/L CrO <sub>4</sub>
Chelant, Free	2062501	4	0.0800 M MgCl <sub>2</sub>	0 - 20.0 mg/L CaCO <sub>3</sub>
EDTA	2081701	1, 4	0.0716 M EDTA	10 - 100 mg/L Fe
	2081801		0.716 M EDTA	100 - 1000 mg/L Fe
Oxygen, Dissolved	2267501	1, 2, 3	0.200 N Na <sub>2</sub> S <sub>2</sub> O <sub>3</sub>	1 - 10 mg/L DO
Sulfite	2393701	3	2.570 N Hg(NO <sub>3</sub> ) <sub>2</sub>	0 - 100 ppt Salinity
Sulfite	1496101	1, 2, 3	0.3998 N KIO <sub>3</sub> -KI	0 - 800 mg/L SO <sub>3</sub> <sup>2-</sup>
Volatile Acids	1484201	4	0.9274 N NaOH	100 - 2400 mg/L CH <sub>3</sub> COOH

Please note: not all items are available in all geographies.

\*Other cartridges are available. Note that other reagents are required in addition to these titration cartridges to complete a test procedure.

† Suitable for but not limited to analysis in the following application: 1 = water; 2 = wastewater; 3 = seawater; 4 = other.

‡ CDTA is substituted for EDTA when samples contain iron.

## Order Information

<b>1690001</b>	Digital Titrator with plastic case, manual and 5 straight delivery tubes
<b>1720500</b>	Delivery Tubes, straight with J-hook, 5/pk
<b>4157800</b>	Delivery tubes for digital titrator, 90° bend
<b>1449501</b>	Empty Titration Cartridge with blank label
<b>1940100</b>	Electrode and Digital Titrator Holder
<b>2270900</b>	Digital Titrator Kit

## Digital Titrator Kits and Reagent Sets

Many single- and multiple-parameter test kits are available with the Digital Titrator for rapid, convenient on-site titrimetric testing. Test kits contain the Digital Titrator, appropriate titration cartridge(s), buffers and/or indicators and necessary beakers, flasks or other labware.

The Universal Digital Titrator Kit can be customized to fit your analysis needs by adding the reagent sets of your choice. Order Catalog Number 2270900.

<b>2064000</b>	Acidity Test Kit, Model AC-DT
<b>2063700</b>	Alkalinity Test Kit, Model AL-DT
<b>2064100</b>	Carbon Dioxide Test Kit, Model CA-DT
<b>2063500</b>	Chloride Test Kit, Model CD-DT
<b>2580600</b>	Chloride Test Kit, Model CDS-DT
<b>2063600</b>	Hardness Test Kit, Model HA-DT
<b>2063900</b>	Hardness Test Kit, Model HAC-DT
<b>2687000</b>	Hypochlorite Test Kit, Model CN-HRDT
<b>2063300</b>	Sulfite Test Kit, Model SU-DT
<b>2471100</b>	Total Chlorine Test Kit, Model CN-DT

## Multiple-Parameter Kits

Kits are equipped with reagents for tests listed. Order replacement reagents as identified in each kit.

<b>2350700</b>	Professional Boiler and Cooling Water Test Kit, Model PBC-DT
<b>243001</b>	Ten-Parameter Aquaculture Test Kit, Model FF-2
<b>243003</b>	Saltwater Aquaculture Test Kit, Model FF-3



**World Headquarters: Loveland, Colorado USA | [hach.com](http://hach.com)**

**United States** 800-227-4224 fax: 970-669-2932 email: [orders@hach.com](mailto:orders@hach.com)  
**Outside United States** 970-669-3050 fax: 970-461-3939 email: [intl@hach.com](mailto:intl@hach.com)

©Hach Company, 2025. All rights reserved.

In the interest of improving and updating its equipment, Hach Company reserves the right to alter specifications to equipment at any time.

**DOC052.53.25038.Jul25**