

REFERENCE MATERIAL CERTIFICATE

Reference Material

This certificate is designed in accordance with ISO Guide 31. This reference material (RM) was designed, produced and verified in accordance with a registered quality management system ISO 9001. All measurements were performed according to ISO/IEC 17025 by a DAkkS accredited laboratory (D-PL-19883-01-00).

Product Name

Ochratoxin A 2 µg/mL in Methanol

Product Code

DRE-A15670000ME-2

Lot Number

1-1565768ME

CAS No.

303-47-9

Format

Solution

Mol. Weight

403.81

Expiry Date


01 Oct 2027

Mol. Formula

C₂₀H₁₈ClNO₆

Storage Temp

-18°C ± 4°C



CERTIFIED

Concentration
2.0 µg/mL

CERTIFIED

Expanded Uncertainty (U)
0.1 µg/mL

Uncertainty

The certified value(s) and uncertainty(ies) are determined in accordance with EURACHEM/CITAC Guide for "Quantifying Uncertainty in Analytical Measurement, 3rd edition", with an 95% confidence level (k=2). Uncertainty is based on the Total Combined Uncertainty, including uncertainties of characterisation and stability testing. Stability values are based on real evidence opposed to simulation.


The producer certifies that this reference material meets the specification stated in this certificate until the expiry date, provided it is stored unopened at the recommended temperature herein. Product warranties for this reference material are set out in the terms and conditions of purchase.

CERTIFIED BY	CERTIFIED ON		
I. Milusic	01 Oct 2025	<i>J. Milusic</i>	RM Release



REFERENCE MATERIAL CERTIFICATE

CHROMATOGRAM



Signal:	*DAD1A, Sig=205.0, 2.0 Ref=off [b]			
Nr.	RT [min]	Area	Height	Area%
1	5.80	87.78736	31.40	100.00

Instrument
UHPLC/DAD

Column
LUNA Omega C18 1.6 µm 100 x 2.1 mm

Method Details

Eluent A: 100% Acetonitrile
Eluent B: Water + H3PO4

Time [min]	Eluent A [%]	Eluent B [%]
0	10	90
0.3	10	90
8	98	2
9.5	98	2

Flow
0.5 mL/min

Method of Preparation

The certified value is based on gravimetric and volumetric preparation of this RM. This RM has been confirmed by the appropriate analytical techniques.

Batch Information

Solvent: Methanol, Lot No. V2N046033A34, 500.00 mL.

Gravimetric Data

Compound Name	Lot No.	Weight (mg)	Purity (%)
Ochratoxin A	1536821	1.008	99.5

Intended Use

This RM is intended for use in a laboratory as a calibration and quality control standard or in method development for analytical techniques.

Safety

Proper precautions should be observed while handling. See Safety Data Sheet.

Storage

The RM should be stored in the original sealed container at the indicated temperature.

Traceability

The balances used for gravimetric measurements are calibrated with weights traceable to the national standards (DKD). The calibration of the balances is verified daily internally and annually by an external accredited calibration service. Characterisation methods accredited to ISO 17025 are traceable through an unbroken chain of comparisons with national and international standards to the International System of Units (SI).

Instructions for use

The RM should be used shortly after opening to avoid concentration changes due to evaporation. It is recommended to use 1 mL as the minimum sample size and if less material is used, to increase the certified uncertainty by a factor of two for half sample and four for a quarter of sample. If storage after opening is necessary, the RM should be tightly closed and kept from light and moisture. If the RM was in a sealed ampoule, it should be transferred to a vial with minimum head space. Visit the support section of our website lgcstandards.com for a series of Dr. Ehrenstorfer Tech Tip videos and frequently asked questions.

LGC Labor GmbH
Bgm.-Schlosser-Straße 6A
86199 Augsburg, Germany
T | +49 821 906080
F | +49 821 9060888
E | dr.ehrenstorfer@lgcgroup.com