



thermo scientific

NanoDrop Micro Volume UV/Vis
Spectrophotometers and Fluorometers

NanoDrop Ultra

User Guide

M022 NanoDrop Ultra User Guide

Revision C August 2025

ThermoFisher
SCIENTIFIC

© 2025 Thermo Fisher Scientific Inc. All rights reserved.

Wi-Fi is either a trademark or a registered trademark of Wi-Fi Alliance in the United States and/or other countries. Bluetooth is either a trademark or a registered trademark of Bluetooth Special Interest Group. Windows and Microsoft OneDrive are either trademarks or registered trademarks of Microsoft Corporation in the United States and/or other countries. Google Drive is a registered trademark of Google LLC. All other trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries.

For U.S. Technical Support, please contact:

Thermo Fisher Scientific
3411 Silverside Road
Tatnall Building, Suite 100
Wilmington, DE 19810 U.S.A.

Telephone: 302 479 7707
Toll Free: 1 877 724 7690 (U.S. & Canada only)
E-mail: nanodrop@thermofisher.com

For International Support, please visit:

[http://www.thermofisher.com/
NanoDropSupport](http://www.thermofisher.com/NanoDropSupport)

Contact your local distributor. For contact information go to:

[http://www.thermofisher.com/
NanoDropDistributors](http://www.thermofisher.com/NanoDropDistributors)

Thermo Fisher Scientific Inc. provides this document to its customers with a product purchase to use in the product operation. This document is copyright protected and any reproduction of the whole or any part of this document is strictly prohibited, except with the written authorization of Thermo Fisher Scientific Inc.

The contents of this document are subject to change without notice. All technical information in this document is for reference purposes only. System configurations and specifications in this document supersede all previous information received by the purchaser.

Thermo Fisher Scientific Inc. makes no representations that this document is complete, accurate or error-free and assumes no responsibility and will not be liable for any errors, omissions, damage or loss that might result from any use of this document, even if the information in the document is followed properly.

This document is not part of any sales contract between Thermo Fisher Scientific Inc. and a purchaser. This document shall in no way govern or modify any Terms and Conditions of Sale, which Terms and Conditions of Sale shall govern all conflicting information between the two documents.

For Research Use Only. This instrument or accessory is not a medical device and is not intended to be used for the prevention, diagnosis, treatment or cure of disease.



WARNING Avoid an explosion or fire hazard. This instrument or accessory is not designed for use in an explosive atmosphere.

Contents

Chapter 1 About the Spectrophotometer	11
Features	12
Touchscreen	12
Cuvette Holder	13
Fluorescence LEDs	14
USB-A port	14
USB-B port	14
USB-C ports	15
Wi-Fi™ and Bluetooth® Dongle port	15
Instrument Status Indicator	15
Accessories	16
NanoDrop Ultra USB Wi-Fi and Bluetooth Dongle	16
DYMO™ LabelWriter™ 550 USB Label Printer	16
PR-1 Pedestal Reconditioning Kit	17
PV-1 Performance Verification Solution	17
FL-1 Fluorescence Verification Kit	17
NanoDrop Ultra dsDNA BR Fluorescence Assay	17
NanoDrop Ultra RNA BR Fluorescence Assay	17
NanoDrop Ultra Instrument Dust Cover	18
NanoDrop Ultra Microfiber Screen Wipe	18
USB-C Battery Pack	18
Instrument Detection Limits	18
Absorbance-Based Detection Limits	19
Fluorescence-Based Detection Limits	19
Chapter 2 Instrument Set up	21
Register Your Instrument	21

Contents

Set Up Your Instrument	21
Connect Power	22
Connect an Accessory	22
Connect the USB Wi-Fi and Bluetooth Dongle	22
Out-of-Box Experience	22
Set Up Bluetooth Connections	24
Set up Wireless Connection	25
Set Up Cloud Connection	25
Set Up Ethernet Connection	25
Set Up USB Connection	26
Assess Instrument Connectivity	27
Operating Specifications	27
Computer Requirements	28
Update Software	28
Technical Support	29
For U.S./Canada Support, please contact:	29
For International Support, please contact:	29
Chapter 3 Application Measurement Ranges	31
Detection Limits for All Applications	31
Detection limits for absorbance-based applications	31
Detection limits (absorbance) for pre-defined fluorescent dyes	33
Detection Limits for Fluorescence-based Applications	33
Chapter 4 Nucleic Acid Applications	35
dsDNA, ssDNA or RNA	35
Best practices for nucleic acid measurements	35
Measure dsDNA, ssDNA or RNA	36
Reported Results	39
Settings	42
Detection Limits	43
Kit editor	44
Calculations	46
Microarray	51
Measure Microarray	51
Reported Results	54
Settings	56
Detection Limits	60
Calculations for Microarray Measurements	61
Custom Factor	65
Measure Custom Factor	65
Reported Results	68
Settings	70
Detection Limits	70

Oligo DNA or Oligo RNA	73
Measure Oligo DNA or Oligo RNA	73
Reported Results	76
Settings	78
Detection Limits	80
Calculations	81
Chapter 5 Protein Applications	85
Protein A280	85
Best practices for protein measurements	85
Measure Protein A280	87
Reported Results	90
Settings	92
Protein editor	95
Detection Limits	97
Calculations	97
Protein A205	103
Measure Protein A205	103
Protein A205 Reported Results	105
Settings	108
Detection Limits	109
Calculations	110
Proteins and Labels	113
Measure Proteins and Labels	113
Reported Results	116
Settings	118
Detection Limits	121
Calculations	121
Protein BCA	125
Measure Protein BCA	129
Reported Results	132
Settings	135
Protein Bradford	137
Measure Protein Bradford	138
Reported Results	141
Settings	144
Protein Lowry	147
Measure Protein Lowry	148
Reported Results	151
Settings	154
Protein Pierce 660	157
Measure Protein Pierce 660	158
Reported Results	161
Settings	164

Chapter 6	Fluorescence Applications	167
	dsDNA Fluorescence or RNA Fluorescence	167
	Best practices for fluorescence measurements	167
	Measure dsDNA Fluorescence or RNA Fluorescence	172
	Reported Results	175
	Settings	176
	Calculations	177
	Fluorometer	181
	Measure Fluorophore-Labeled Samples	181
	Reported Results	184
	Settings	185
	Reagent Calculator	187
	Use the Reagent Calculator	187
	Reagent Calculator Results	188
Chapter 7	Acclaro Pro Applications	189
	Activating Acclaro Pro	189
	dsDNA, ssDNA or RNA Pro	189
	Best practices for nucleic acid measurements	190
	Measure dsDNA Pro, ssDNA Pro or RNA Pro	191
	Reported Results	193
	Settings	195
	Calculations	196
	Custom Factor Pro	199
	Measure Custom Factor Pro	199
	Reported Results	201
	Settings	202
	Detection Limits	203
	Oligo DNA Pro or Oligo RNA Pro	205
	Measure Oligo DNA Pro or Oligo RNA Pro	205
	Reported Results	208
	Settings	210
	Detection Limits	212
	Calculations	213
	Protein A280 Pro	217
	Best practices for protein measurements	217
	Measure Protein A280 Pro	218
	Reported Results	221
	Settings	223
	Detection Limits	226
	Calculations	226

Chapter 8	More Applications	231
	UV-Vis	232
	Best practices for UV-Vis measurements	232
	Measure UV-Vis	233
	Reported Results	235
	Settings	237
	Custom Methods	239
	Manage Custom Methods	239
	Measure using a Custom Method	242
	Reported Results	244
	Settings	245
	OD600	251
	Measure OD600	253
	Reported Results	255
	Settings	257
	Calculations	259
	Kinetics	261
	Manage Kinetics Methods	261
	Measure Kinetics	264
	Reported Results	267
	Settings	272
Chapter 9	Learning Center	275
	Micro-Volume Sampling—How it Works	276
	Measure a Micro-Volume Sample	279
	Best practices for micro-volume measurements	279
	Recommended sample volumes	280
	Measure a Sample Using a Cuvette	284
	Best practices for cuvette measurements	285
	Cuvette Settings	287
	Prepare Samples and Blanks	287
	Preparing Samples	287
	Run a Blanking Cycle	291
	Basic Instrument Operations	293
	NanoDrop Ultra Home Screen	293
	NanoDrop Ultra Measurement Screens	298
	Measurement Screen Display Options	299
	NanoDrop Ultra General Operations	320
	Acclaro Sample Intelligence	329
	Activate Detection	329
	View Acclaro Sample Intelligence Information	330
	Contaminant Analysis	331
	On-Demand Technical Support	334
	Alarm Notifications	335

Contents


Instrument Settings	336
System Settings.	336
Connectivity.	337
Protein Editor.	350
Kit Editor	350
Dye Editor	350
Preferences	350
Export Setup	351
SciVault 2 Activation	353
Update Software	353
Database Backup and Restore	355
About Instrument.	356
PC Control Software	356
PC Control Home Screen overview	356
PC Control Measurement Screen	357
Chapter 10 Maintenance	359
Maintenance Schedule	360
Daily Maintenance	360
Periodic Maintenance	360
Every 6 Months	360
Cleaning the Touchscreen.	361
Maintaining the Pedestals	361
Cleaning the Pedestals	361
Reconditioning the Pedestals.	363
Decontaminating the Instrument	365
Maintaining the Cuvette Sampling System.	366
Instrument Diagnostics	367
Intensity Check	368
Performance Verification	370
Pedestal Image Check.	376
Cuvette Check.	377
Fluorescence Performance Check.	379
Acclaro Pro Check.	383
Export Logs	386
Diagnostics Schedule	386
Chapter 11 Safety and Operating Precautions	389
Operating Precautions.	390

Safety Information391
 Safety and Special Notices391
 When the System Arrives393
 Lifting or Moving the Instrument393
 Electrical Requirements and Safety394
 Power Cords394
 Fire Safety and Burn Hazards396
 Optical Safety396
 Hazardous Materials396

Contents

About the Spectrophotometer

NanoDrop Ultra^C FL Spectrophotometer and Fluorometer



Note Locate the instrument away from air vents and exhaust fans to minimize evaporation

The Thermo Scientific™ NanoDrop™ Ultra instrument series is composed of compact, stand-alone UV-Visible spectrophotometers and fluorometers developed for micro-volume analysis of a wide variety of analytes. The pioneering [sample retention system](#) enables the measurement of highly concentrated samples without the need for dilutions.

The NanoDrop Ultra system comes with preloaded software and a touchscreen display. NanoDrop Ultra PC control software can be installed on a local PC and used to control the instrument and view data. The instrument can be connected to an optional printer with a USB cable or to a remote printer through an Ethernet connection or wireless network using the provided NanoDrop Ultra Wi-Fi and Bluetooth Dongle.

1 About the Spectrophotometer Features

Note Before operating a NanoDrop Ultra instrument, please read the [safety and operating precautions](#) and then follow their recommendations when using the instrument.

Features

The NanoDrop Ultra spectrophotometers and fluorometers feature the [micro-volume sample retention system](#). The NanoDrop Ultra^c and NanoDrop Ultra^c FL models additionally feature a cuvette holder for analyzing dilute samples using standard UV-Visible cuvettes. The NanoDrop Ultra FL and NanoDrop Ultra^c FL models are capable of both micro-volume absorbance and fluorescence measurements by including red and blue solid-state light emitting diodes (LED's), which are oriented 90° to the detector.

Touchscreen



The NanoDrop Ultra comes with a built-in, 10-inch, high-resolution touchscreen preloaded with easy-to-use instrument control software. The touchscreen can tilt forward or back for optimal viewing.

Cuvette Holder




The NanoDrop Ultra^c and NanoDrop Ultra^c FL include a cuvette holder for measuring dilute samples, colorimetric assays, cell cultures and kinetic studies. The cuvette system has these features:

- extended lower [detection limits](#)
- 37 °C heater option for temperature-sensitive samples and analyses
- micro-stirring option to ensure sample homogeneity and support kinetic studies

For details, see [Measure a Sample using a Cuvette](#).

Fluorescence LEDs



The NanoDrop Ultra FL and NanoDrop Ultra^c FL are enhanced with the inclusion of one red LED and one blue LED for fluorescence measurements. This addition allows users to perform fluorescence analysis, providing an extra dimension to the spectrophotometric capabilities of the instrument. The red and blue LEDs serve as excitation sources, enabling the detection of fluorescent emissions from samples that respond to these specific wavelengths. The operating software automatically limits the amount of time the LED source is on in order to prevent too much light from saturating the spectrometer detector for the emission wavelength. The fluorescence system has these features:

- **High Specificity** —provides greater specificity than absorbance-based quantification by using fluorescent dyes that bind to or interact with specific target molecules. These dyes emit light at characteristic wavelengths upon excitation, allowing detection even in complex or unpurified samples
- **Sample Prep Calculator**—integrated reagent calculator determines the amount of dye and buffer needed for the selected NanoDrop Ultra Fluorescence Assay
- **NanoDrop Ultra Fluorescence Assay Kits**—fluorescence assays to be used with built-in software applications for dsDNA and RNA detection

USB-A port

One USB-A port located on the instrument rear panel.

USB-B port

One USB-B port located on instrument rear panel.

USB-C ports

One USB-C port located on the front of the instrument and one USB-C port dedicated for power, located on the instrument rear panel.

Wi-Fi™ and Bluetooth® Dongle port

One recessed USB-A port located on the instrument rear panel dedicated for Wi-Fi and Bluetooth connectivity. This port also acts as a standard USB-A port in the event that the Wi-Fi and Bluetooth dongle is not used.

Instrument Status Indicator



Instrument Status indicator

The instrument status indicator located below the touchscreen will display specific colors and patterns to indicate current status of the instrument.

Indicator	Instrument Status
Off	Powered off/Energy Saver Mode
Breathing white light	Powering on or initializing
Constant white light	Ready/standby
Sweeping blue light	Measuring
Constant blue light	Diagnostic - pass result Measurement does not trigger Acclaro alert or Acclaro contaminant
Constant amber light	Diagnostic - conditional pass result Measurement triggers Acclaro alert or Acclaro contaminant
Blinking amber light	Diagnostic - fail result

Accessories

This section lists the accessories for use with the NanoDrop Ultra instruments.

Note Some accessories listed below are not included with the purchase of NanoDrop Ultra spectrophotometers and fluorometers.

NanoDrop Ultra USB Wi-Fi and Bluetooth Dongle

A USB Wi-Fi and Bluetooth dongle used to enable Wi-Fi and Bluetooth connectivity.



The NanoDrop Ultra USB Wi-Fi and Bluetooth Dongle is intended to be plugged into the recessed USB-A port in the back of the instrument.

DYMO™ LabelWriter™ 550 USB Label Printer



Prints two 5/16-in x 4-in self-adhesive labels for transferring sample data directly into laboratory notebooks or posting on bulletin boards. The software allows [printing of data](#) from each sample measurement or from a group of samples logged and measured together.

The printer connects to the instrument (back panel) via a USB cable (included).

PR-1 Pedestal Reconditioning Kit



Specially formulated conditioning compound that can be applied to the pedestals to restore them to a hydrophobic state (required to achieve adequate surface tension for accurate sample measurements). The kit includes conditioning compound and applicators. For more information, see [Reconditioning the Pedestals](#).

PV-1 Performance Verification Solution

Liquid photometric standard used to check instrument absorbance performance. For more information, see [Performance Verification](#).

FL-1 Fluorescence Verification Kit

Reagent based assay used to check instrument fluorescence performance. For more information, see [Fluorescence Performance Check](#).

NanoDrop Ultra dsDNA BR Fluorescence Assay

An assay kit that provides an accurate and selective method for the quantitation of sensitive double-stranded DNA samples. To be used with the dsDNA Fluorescence application.

NanoDrop Ultra RNA BR Fluorescence Assay

An assay kit that provides an accurate and selective method for the quantitation of sensitive RNA samples. To be used with the RNA Fluorescence application.

NanoDrop Ultra Instrument Dust Cover



Transparent plastic cover protects instrument against dust and any liquid exposure while not in use.

NanoDrop Ultra Microfiber Screen Wipe



Specialized cleaning cloth designed for safely and effectively cleaning the NanoDrop Ultra touch screen.

USB-C Battery Pack

The NanoDrop Ultra instruments are designed to be compatible with USB-C portable chargers (75 Wh provides approximately 8 hours of run time).

Instrument Detection Limits



Absorbance-Based Detection Limits

Measurement Location	Pathlength (mm)	Upper Detection Limit (10 mm Equivalent Absorbance)
Pedestal	1.0	12.5
	0.2	62.5
	0.1	150
	0.05	300
	0.03	550
Cuvette	10	1.5
	5	3
	2	7.5
	1	15

Fluorescence-Based Detection Limits

The fluorescence-based detection limits are dependent on the fluorescence assay used with the instrument. See the kit's protocol for more information.

The NanoDrop Ultra fluorescence kits have been optimized for microvolume measurements. Performance of other fluorescence dyes/reagents on the microvolume pedestal cannot be guaranteed.

This page is intentionally blank.

Instrument Set up


Register Your Instrument

Register your instrument to receive e-mail updates on software and accessories for the NanoDrop Ultra instrument. An Internet connection is required for registration.

To register your instrument

From any PC that is connected to the Internet, use any web browser to navigate to www.thermofisher.com/nanodropsv. On the website, locate the “Register your instrument” button, and follow the instructions to register the instrument.

Set Up Your Instrument



Connect Power



CAUTION Avoid shock hazard. Each wall outlet used must be equipped with a ground. The ground must be a noncurrent-carrying wire connected to earth ground at the main distribution box.

Connect the provided power cord to a grounded wall outlet and then to the designated USB-C port on the back of the instrument. See [“Power Cords”](#) on [page 394](#) for more information.

Connect an Accessory

To connect a compatible printer or other compatible accessory such as a USB keyboard and/or mouse to the instrument, use any USB port on the instrument (front or back). See [Accessories](#) for information about accessories compatible with the NanoDrop Ultra instruments.

Connect the USB Wi-Fi and Bluetooth Dongle




To connect the provided USB Wi-Fi and Bluetooth Dongle, use the dedicated recessed USB port on the back panel of the instrument. The Wi-Fi and Bluetooth features of this instrument will only function when a USB Wi-Fi and Bluetooth Dongle is connected.

Out-of-Box Experience

The Out-of-Box Experience (OOBE) is the initial setup process that you encounter when you first start a NanoDrop Ultra instrument. It is designed to help you configure essential settings, personalize your environment, and get your device ready for use.

1. Power on the instrument for the first time.
2. Read and **Accept** the End User License Agreement (EULA).
3. Select the desired language for the software from the provided drop-down menu and select **Next**.
4. If it is desired to setup a Wi-Fi connection at this time, select the toggle button next to Wi-Fi to move it to the right, it will turn blue and a checkmark will be displayed.
5. Select from the list of available Wi-Fi networks (input password if required, then select **OK**) then select **Next**.




See [Instrument Settings](#) in the [Learning Center](#) chapter for more information on how to connect to a Wi-Fi network.

2 Instrument Set up

Set Up Your Instrument

6. If it is not desired to connect to a Wi-Fi network at this time, select **Next** without enabling Wi-Fi.
7. Select the desired date and time settings from the appropriate menus then select **Next**.



See [Instrument Settings](#) in the [Learning Center](#) chapter for more information on how to configure the date and time settings.

8. Select **Finish** to complete the OOB setup. the instrument must restart for the changes to take effect.

Set Up Bluetooth Connections

Use Bluetooth to connect the instrument to one or more Bluetooth (wireless) input devices such as a Bluetooth keyboard, mouse or barcode scanner.

Note Make sure the device is labeled “Bluetooth” and not just “wireless.” All Bluetooth devices are wireless but not all wireless devices will run with Bluetooth.

Ensure the NanoDrop Ultra Wi-Fi and Bluetooth Dongle is connected to the instrument before proceeding with setup. See [“Connect the USB Wi-Fi and Bluetooth Dongle”](#) on [page 22](#).

See the section on [bluetooth connectivity](#) for additional details on connecting to a Bluetooth device

Set up Wireless Connection

Ensure the NanoDrop Ultra USB Wi-Fi and Bluetooth Dongle is connected to the instrument before proceeding with setup. See [“Connect the USB Wi-Fi and Bluetooth Dongle”](#) on [page 22](#).

See the section on [Wi-Fi connectivity](#) for additional details on connecting to a Wi-Fi network.

Set Up Cloud Connection

If the instrument is connected to a Cloud account, you can export data files to the following Cloud services: Thermo Fisher Connect, Google Drive, and Microsoft OneDrive. Multiple Cloud accounts can be added to the instrument allowing the operator the ability to select from the available list when exporting data.

Note An ethernet or Wi-Fi connection to a network with internet access is required for this feature.

See the section on [cloud connectivity](#) for additional details on connecting to a Cloud platform.

Set Up Ethernet Connection

The instrument Ethernet port can be used to set up a wired connection between the instrument and an active network wall jack.

If the instrument is connected to a network wall jack, you can export data files to a network location, for example, in order to transfer them to another computer. You can define multiple network paths that the operator can select from when exporting data. See [Export Settings](#) for details. If the connected network has internet access, the ability to setup Cloud exporting and use online software updates are enabled.

Parts needed:


- Shielded (straight through) Ethernet cable (CAT5e or newer is recommended)


Set up Ethernet connection

1. Connect one end of Ethernet cable to the Ethernet port on the instrument back panel.

2 Instrument Set up

Set Up Your Instrument



2. Connect the other end of the Ethernet cable to an active network wall jack.
3. When connected to a network via an Ethernet connection, the Ethernet Connectivity icon  will appear at the top of the screen.

Set Up USB Connection

The instrument USB-B port can be used to set up a wired connection between the instrument and a personal computer (PC) or to connect to a USB printer.


Parts needed:

- Standard USB-A to USB-B cable (not supplied)


Set up USB connection

1. Connect the USB-B end of USB cable to the USB-B port on the instrument back panel.



2. Connect the USB-A end of the USB cable to either the computer or printer USB-A port.
3. When connected to a printer via a USB connection, a printer icon  will appear at the top of the screen.

Assess Instrument Connectivity

Use the System Status icon  at the top right of the instrument home screen to quickly assess the instrument's connectivity status including Bluetooth and Wi-Fi. "System Status" on [page 294](#) for more details.

Operating Specifications

The instrument operates reliably when the room environment meets these specifications:

- operating temperatures: 5 °C - 35 °C (41 °F - 95 °F)
- relative humidity (non-condensing): 20-80%

Locate the instrument away from air vents and exhaust fans to minimize evaporation.

Note If operating the instrument at the low end of the recommended humidity range, use adequate sample volume to avoid evaporation.

Computer Requirements

When controlling the instrument through a connected computer as opposed to the local touch screen, use the information below to select a compatible computer to operate the software.

Required Windows Operating System

- Windows 10/11 Enterprise or Professional, build 1607 or greater

Minimum hardware configuration

- 2.0 GHz dual-core processor enabled
- 4 GB RAM with system managed memory enabled
- 5 GB available on drive C
- Display resolution 1366 x 768

Recommended hardware configuration

- 2.33 GHz 4-core processor enabled (or greater)
- 8 GB RAM with system managed memory enabled (or greater)
- 200 GB available on drive C (or greater)

Update Software

Quickly and easily download and install the latest NanoDrop Ultra software and release notes from our website. Follow the steps to update or upgrade the software on your local instrument and/or install or update the NanoDrop Ultra software on a personal computer (PC). A PC with an Internet connection is required to download software.

See [“Update Software”](#) on [page 353](#) for information on updating the NanoDrop Ultra software.

Technical Support

For U.S./Canada Support, please contact:

Thermo Fisher Scientific
3411 Silverside Road
Tatnall Building, Suite 100
Wilmington, DE 19810 U.S.A.

Telephone: 302 479 7707
Toll Free: 1 877 724 7690 (U.S. & Canada only)
Fax: 302 792 7155
E-mail: nanodrop@thermofisher.com
Website: www.thermofisher.com/nanodrop

For International Support, please contact:

Contact your local distributor. For contact information go to:

<http://www.thermofisher.com/NanoDropDistributors>

If you are experiencing an issue with your system, refer to the troubleshooting information. If the issue persists, contact us. If you are outside the U.S.A. and Canada, please contact your local distributor.

If your instrument requires maintenance or repair, contact us or your local distributor.

This page is intentionally blank.

Application Measurement Ranges

Detection Limits for All Applications



Note Absorbance-based detection limits provided in the tables below are approximate and apply to micro-volume measurements only; they are based on the instrument's photometric absorbance range (10 mm equivalent) of 0–550 A. For measurements with 10 mm pathlength cuvettes, the photometric absorbance range is 0–1.5 A.

Detection limits for absorbance-based applications

Sample Type	Lower Detection Limit	Upper Detection Limit	Typical Reproducibility ^a
dsDNA	1.0 ng/μL (pedestal)	27,500 ng/μL (pedestal)	±1.0 ng/μL for sample concentrations between 1.0 and 100 ng/μL samples; ±2% for samples >100 ng/μL
	0.20 ng/μL (cuvette)	75 ng/μL (cuvette)	
ssDNA	0.66 ng/μL (pedestal)	18,150 ng/μL (pedestal)	±0.66 ng/μL for sample concentrations between 0.66 and 100 ng/μL samples; ±2% for samples >100 ng/μL
	0.13 ng/μL (cuvette)	49.5 ng/μL (cuvette)	

3 Application Measurement Ranges
Detection Limits for All Applications

Sample Type	Lower Detection Limit	Upper Detection Limit	Typical Reproducibility^a
RNA	0.8 ng/μL (pedestal) 0.16 ng/μL (cuvette)	22,000 ng/μL (pedestal) 60 ng/μL (cuvette)	±0.8 ng/μL for sample concentrations between 0.8 and 100 ng/μL samples; ±2% for samples >100 ng/μL
DNA Microarray (ssDNA)	1.3 ng/μL (pedestal) 0.13 ng/μL (cuvette)	495 ng/μL (pedestal) 49.5 ng/μL (cuvette)	±1.3 ng/μL for sample concentrations between 1.3 and 100 ng/μL samples; ±2% for samples >100 ng/μL
Purified BSA by Protein A280	0.03 mg/mL (pedestal) 0.006 mg/mL (cuvette)	820 mg/mL (pedestal)	±0.03 mg/mL (for 0.03–10 mg/mL samples); ±2% for samples >10 mg/mL
IgG by Protein A280	0.02 mg/mL (pedestal) 0.003 mg/mL (cuvette)	402 mg/mL (pedestal)	
Purified BSA by Proteins & Labels	0.06 mg/mL (pedestal) 0.006 mg/mL (cuvette)	19 mg/mL (pedestal)	±0.06 mg/mL for 0.06–10 mg/mL samples
Protein BCA	0.125 mg/mL (20:1 reagent/sample volume) 0.02 mg/mL (1:1 reagent/sample volume)	2.0 mg/mL 0.20 mg/mL	2% over entire range 0.01 mg/mL over entire range
Protein Lowry	0.2 mg/mL	1.5 mg/mL	2% over entire range
Protein Bradford	100 μg/mL (30:1 reagent/sample volume) 15 μg/mL (1:1 reagent/sample volume)	1000 μg/mL 100 μg/mL	±25 μg/mL for 100–500 μg/mL samples ±5% for 500–1000 μg/mL samples ±4 μg/mL for 15–50 μg/mL samples ±5% for 50–100 μg/mL samples
Protein Pierce 660	50 μg/mL (15:1 reagent/sample volume) 25 μg/mL (7.5:1 reagent/sample volume)	2000 μg/mL 1000 μg/mL	±3 μg/mL for 50–125 μg/mL samples ±2% for samples > 125 μg/mL ±3 μg/mL for 25–125 μg/mL samples ±2% for samples >125 μg/mL

^a Based on five replicates (SD=ng/μL; CV=%)

Note To minimize instrument error with highly concentrated samples, make dilutions to ensure that measurements are made within these absorbance limits:

- For micro-volume measurements, maximum absorbance at 260 nm (for nucleic acids) or 280 nm (for proteins) should be less than 62.5 A.
- For measurements with 10 mm pathlength cuvettes, maximum absorbance at 260 nm (or 280 nm for proteins) should be less than 1.5 A, which is approximately 75 ng/μL dsDNA.

Detection limits (absorbance) for pre-defined fluorescent dyes

Sample Type	Lower Detection Limit	Upper Detection Limit ^a	Typical Reproducibility ^b
Cy3, Cy3.5, Alexa Fluor 555, Alexa Fluor 660	0.2 pmol/μL (pedestal)	100 pmol/μL (pedestal)	±0.20 pmol/μL for sample concentrations between 0.20 and 4.0 pmol/μL; ±2% for samples >4.0 pmol/μL
Cy5, Cy5.5, Alexa Fluor 647	0.12 pmol/μL (pedestal)	60 pmol/μL (pedestal)	±0.12 pmol/μL for sample concentrations between 0.12 and 2.4 pmol/μL; ±2% for samples >2.4 pmol/μL
Alexa Fluor 488, Alexa Fluor 594	0.4 pmol/μL (pedestal)	215 pmol/μL (pedestal)	±0.40 pmol/μL for sample concentrations between 0.40 and 8.0 pmol/μL; ±2% for samples >8.0 pmol/μL
Alexa Fluor 546	0.3 pmol/μL (pedestal)	145 pmol/μL (pedestal)	±0.30 pmol/μL for sample concentrations between 0.30 and 6.0 pmol/μL; ±2% for samples >6.0 pmol/μL

^a Values are approximate

^b Based on five replicates (SD=ng/μL; CV=%)

Detection Limits for Fluorescence-based Applications

The fluorescence-based detection limits are dependent on the fluorescence assay used with the instrument. See the kit's protocol for more information.

Note The NanoDrop Ultra Fluorescence Assays have been optimized for microvolume measurements. Performance of other fluorescence dyes / reagents on the microvolume pedestal cannot be guaranteed.

3 Application Measurement Ranges

This page is intentionally blank.

Nucleic Acid Applications

dsDNA, ssDNA or RNA

Measures the concentration of purified dsDNA, ssDNA or RNA samples that absorb at 260 nm.

[Measure dsDNA, ssDNA or RNA](#)


[Reported Results](#)

[Settings](#)

[Kit Editor](#)

[Detection Limits](#)

[Calculations](#)



Best practices for nucleic acid measurements

- Isolate and purify nucleic acid samples before measurement to remove impurities. Depending on the sample, impurities could include DNA, RNA, free nucleotides, proteins, some buffer components and dyes. See [Preparing Samples](#) for more information.

Note Extraction reagents such as guanidine, phenol, and EDTA contribute absorbance between 230 nm and 280 nm and will affect measurement results if present in samples (even residual amounts).

- Ensure the sample absorbance is within the instrument's [absorbance detection limits](#).
- Blank with the same buffer solution used to resuspend the analyte of interest. The blanking solution should be a similar pH and ionic strength as the analyte solution.

4 Nucleic Acid Applications

dsDNA, ssDNA or RNA

- Run a [blanking cycle](#) to assess the absorbance contribution of your buffer solution. If the buffer exhibits strong absorbance at or near the analysis wavelength (typically 260 nm), you may need to choose a different buffer or application. See [Choosing and Measuring a Blank](#) for more information.
- For micro-volume measurements:
 - Ensure pedestal surfaces are properly [cleaned](#) and [conditioned](#).
 - If possible, heat highly concentrated or large molecule samples, such as genomic or lambda DNA, to 63 °C (145 °F) and gently (but thoroughly) vortex before taking a measurement. Avoid introducing bubbles when mixing and pipetting.
 - Follow [best practices for micro-volume measurements](#).
 - Use a 1-2 µL sample volume. See [Recommended Sample Volumes](#) for more information.
- For cuvette measurements (NanoDrop Ultra^C and NanoDrop Ultra^C FL instruments only), use compatible cuvettes and follow [best practices for cuvette measurements](#).

Related Topics

- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Best Practices for Micro-Volume Measurements](#)
- [Best Practices for Cuvette Measurements](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Measure dsDNA, ssDNA or RNA

Use the dsDNA, ssDNA and RNA applications to quantify purified double-stranded (ds) or single-stranded (ss) DNA or RNA samples. These applications report nucleic acid concentration and two absorbance ratios (A260/A280 and A260/A230). A single-point baseline correction can also be used.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **Nucleic Acids** tab, then select **dsDNA, ssDNA or RNA**, depending on the samples to be measured.
2. Configure any of the [setup options](#) if desired and select **Save**.
3. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Pipette 1–2 µL blanking solution onto the lower pedestal and lower the arm, or insert the blanking cuvette into the cuvette holder.


Tip: If using a cuvette, make sure to [align the cuvette light path](#) with the instrument light path.
5. Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements.)
6. Lift the arm and clean both pedestals with a new laboratory wipe, or remove the blanking cuvette.
7. Pipette 1-2 µL sample solution onto the pedestal, or insert the sample cuvette into the cuvette holder.
8. Start the sample measurement:
 - Pedestal: If [Auto-Measure](#) is On, lower arm; if Auto-Measure is Off, lower arm and select **Measure**.
 - Cuvette: Select **Measure**.


When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

4 Nucleic Acid Applications
dsDNA, ssDNA or RNA

- When you are finished measuring samples, select **End Experiment**.
- Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.



Typical nucleic acid spectrum




Comparison of nucleic acid spectra with and without two common contaminants

Reported Results

dsDNA measurement screen (local control)

For each measured sample, the dsDNA, ssDNA and RNA applications show the UV absorbance spectrum and a summary of the results. The following is an example of the measurement screen of the NanoDrop Ultra local control software:



Note Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.


dsDNA measurement screen (PC control)

For each measured sample, the dsDNA, ssDNA and RNA applications show the UV absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See “Measurement Screen Display Options” on page 299 for an example.

4 Nucleic Acid Applications

dsDNA, ssDNA or RNA

qPCR Kit Data

The NanoDrop Ultra software includes an integrated qPCR calculator designed to swiftly determine the precise amount of your sample required for downstream qPCR reactions. When selecting the qPCR Kit Data icon  at the top of the measurement screen, a summary of the calculations will be displayed, these calculations are specific to the Kit that was selected during setup. The following is an example of the qPCR Kit Data screen on the NanoDrop Ultra local control software screen:



Sample Name	Dilution Required		Kit Reaction		Final Template Conc.
	Nucleic Acid Volume (µL)	Diluent Volume (µL)	Nucleic Acid Volume (µL)	Water Volume (µL)	
Sample 1	1.00	2999.00	1.03	3.47	5 ng/µL
Sample 2	3.00	97.00	1.08	3.42	50 ng/µL
Sample 3	3.00	97.00	1.08	3.42	50 ng/µL
Sample 4	1.00	299.00	1.47	3.03	50 ng/µL

Dilution Required

If it is necessary to dilute the nucleic acid before adding it to the reaction mix, follow the recommended dilution scheme outlined in the **Dilution Required** section.

If the **Dilution Required** section is empty, no dilution is needed before adding the nucleic acid to the reaction mix.

Kit Reaction

Follow recommended volumes in the **Kit Reaction** section to achieve the desired Final Template Concentration in the kit reaction.

Note Final Template Concentration may be modified.



If a dilution is required in the Dilution Required section, the Nucleic Acid Volume called for in the Kit Reaction section will indicate the required amount of the diluted nucleic acid product to add to the reaction.

If a dilution is not required in the Dilution Required section, the Nucleic Acid Volume called for in the Kit Reaction section will indicate the required amount of the original nucleic acid sample to add to the reaction.

Reported values

The initial screen that appears after each measurement (see previous image) shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:

Sample Details ✕			
#:	1	Contaminant:	-
Date:	11/29/2022 4:45:07 PM	Corrected:	-
Sample Name:	Sample 1	Species:	-
ng/μL:	5811.7	Kit name:	TaqMan Fast Advanced Master Mix
A260/A280:	2.19		(0.1 mL plate)
A260/A230:	2.42	Sample Dilution Ratio:	1:1200
A260:	116.23	Dilution Required	
A280:	53.15	Nucleic Acid Volume (μL):	1
Factor:	50	Diluent Volume (μL):	1199
Baseline Correction:	25.972A@340nm	Kit Reaction	
Location:	Pedestal	Nucleic Acid Volume (μL):	1.03
Monitored Wavelength:	0.000	Water Volume (μL):	3.47
Pathlength Used:	@260 0.05mm	Final Template Conc.:	5 ng/μL



Close


- date (date and time sample measurement was taken)
- [sample name](#)
- # (sample number)
- [nucleic acid concentration](#)
- [A260/A280](#)
- [A260/A230](#)
- [A260](#)
- [A280](#)
- [factor](#)
- [baseline correction](#)
- [location](#)
- [monitored wavelength](#)
- [pathlength used](#)
- [contaminant](#)
- [corrected nucleic acid concentration](#)


4 Nucleic Acid Applications

dsDNA, ssDNA or RNA

- species
- kit name
- sample dilution ratio
- dilution required: nucleic acid volume (μL)
- dilution required: diluent volume (μL)
- kit reaction: nucleic acid volume (μL)
- kit reaction: water volume (μL)
- kit reaction: final template conc.
- stirrer
- heater
- measured temperature
- target temperature

Settings

The dsDNA, ssDNA, or RNA Setup screen appears after you select the respective application from the Nucleic Acids tab on the home screen. To show the dsDNA, ssDNA or RNA settings, from the dsDNA, ssDNA or RNA measurement screen, select the respective Setup icon .




dsDNA Setup

Auto Naming

Baseline Correction 340 nm (Default value is 340nm)

RNA Contaminant Detection Mammal

 qPCR Setup

Kit: TaqMan Fast Advanced Master Mix (0.1 mL plate)

Tips: The qPCR kit list is located in the Settings menu.

Cancel Save

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Baseline Correction	On or off Enter baseline correction wavelength in nm or use default value (340 nm)	Optional user-defined baseline correction. When enabled, the software corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength.
RNA Contaminant Detection	On or off Use the drop-down menu to select between mammal, plant, or bacteria	When enabled, RNA Contaminant Detection/DNA Contaminant detection will apply mathematical models to predict the amount of RNA contaminant in dsDNA or dsDNA in RNA. These models are specific to the source of the nucleic acid. If you are measuring nucleic acid from a source for which we do not have a mathematical model, leave this feature off.
qPCR Setup	On or off Use the drop-down menu to select one of the available qPCR kits.	When enabled, the software automatically calculates sample volumes needed for a qPCR reaction based on sample concentration and the qPCR kit selected at setup.

Detection Limits

The lower detection limits and reproducibility specifications for nucleic acids are provided [here](#). The upper detection limits are dependent on the [upper absorbance limit](#) of the instrument and the user-defined extinction coefficients.

To calculate upper detection limits for nucleic acid samples

To calculate upper detection limits in ng/μL, use the following equation:

$$(\text{upper absorbance limit}_{\text{instrument}} * \text{extinction coefficient}_{\text{sample}})$$

For example, for a sample measurement using an extinction coefficient of 55, the equation looks like this:

$$(550 \text{ AU} * 55 \text{ ng-cm}/\mu\text{L}) = 30,250 \text{ ng}/\mu\text{L}$$

Note For measurements with 10 mm pathlength cuvettes, the upper absorbance limit is 1.5 AU, which is approximately 75 ng/μL for dsDNA.


Related Topics

- [Detection Limits for All Applications](#)

Kit editor

Use the **Kit Editor** to add a custom qPCR kit to the list of available kits in [Settings for Nucleic Acid Measurements](#).


To access the Kit Editor:

- From the software home screen, select  **Settings > Kit Editor**.

Select to exit Kit Editor

Select to edit custom kit

Custom kits (will appear in qPCR Setup list in dsDNA and RNA Setup)



Name	Total Reaction Volume (μL)	Primer/Probe and Master Mix Volume (μL)	Final Template Conc.
Power SYBR Green RNA-to-CT 1-Step Kit (0.1 mL plate)	10	11.15	50 ng/mL
Power SYBR Green RNA-to-CT 1-Step Kit (0.2 mL plate)	50	27.9	50 ng/mL
PowerTrack™ SYBR Green Master Mix (0.1 mL plate)	10	5.75	5 ng/μL
PowerTrack™ SYBR Green Master Mix (0.2 mL plate)	20	11.5	5 ng/mL
TaqMan Fast Advanced Master Mix (0.1 mL plate)	10	5.5	5 ng/mL
TaqMan Fast Advanced Master Mix (0.2 mL plate)	10	5.75	5 ng/mL
TaqMan Fast Virus 1-Step Master Mix (0.1 mL plate)	10	5.75	5 ng/mL
TaqMan Fast Virus 1-Step Master Mix (0.2 mL plate)	10	5.75	5 ng/mL

Select to select custom kit


Select to delete selected custom kit

Select to add custom kit

Kit Editor as seen on the NanoDrop Ultra local control software



These operations are available from the Kit Editor:

Add custom kit


1. In Kit Editor, select  to show the **Add Kit** box.
2. Enter a unique name for the new kit (from local control, tap field to display keyboard, tap **Done** key to close keyboard).
3. Enter the Total Reaction Volume (μL), Primer/Probe and Master Mix Volume (μL), and Final Template Concentration with it's respective units for the kit in their respective fields.
4. Select **Save** to close the **Add Kit** window.

The new custom kit appears in the Kit list in dsDNA and RNA Setup.

Edit custom kit

1. In Kit Editor, select  in the row of the desired kit that needs to be edited.
2. Select  to show the Edit Kit window.
3. Edit all available fields. For predefined kits, only the Final Template Concentration can be edited.
4. Select **Save**.

Delete custom kit

1. In Kit Editor, select one or more custom kits to delete by using the check boxes to the left of each desired kit.
2. Select .
3. Confirm deletion of the custom kit(s) by selecting **Yes**.

Calculations

The Nucleic Acid applications use a modification of the Beer-Lambert equation (shown at right) to calculate sample concentration where the extinction coefficient and pathlength are combined and referred to as a “factor.”

For the dsDNA, ssDNA and RNA applications, the generally accepted factors for nucleic acids are used in conjunction with Beer’s Law to calculate sample concentration. For the Custom Factor application, the user-specified factor is used.

Extinction Coefficients vs Factors

Using the terms in the Beer-Lambert equation, factor (f) is defined as:

$$\text{factor (f)} = 1/(\mathcal{E} * b)$$

where:

\mathcal{E} = wavelength-dependent molar extinction coefficient in ng-cm/ μ L

b = **sample pathlength** in cm

As a result, analyte concentration (c) is calculated as:

$$c = A * [1/(\mathcal{E} * b)]$$

or

$$c = A * f$$

where:

c = analyte concentration in ng/ μ L

A = absorbance in absorbance units (A)

f = factor in ng-cm/ μ L (see below)

Factors Used

- **dsDNA** (factor = 50 ng-cm/ μ L)
- **ssDNA** (factor = 33 ng-cm/ μ L)
- **RNA** (factor = 40 ng-cm/ μ L)
- **Custom Factor** (user entered factor between 15 ng-cm/ μ L and 150 ng-cm/ μ L)

Calculated nucleic acid concentrations are based on the absorbance value at 260 nm, the factor used and the sample pathlength. A single-point baseline correction (or analysis correction) may also be applied.

Concentration is reported in mass units. Calculators are available on the Internet to convert concentration from mass to molar units based on sample sequence.

Absorbance values at 260 nm, 280 nm and sometimes 230 nm are used to calculate purity ratios for the measured nucleic acid samples. Purity ratios are sensitive to the presence of contaminants in the sample, such as residual solvents and reagents typically used during sample purification.

Measured Values

Note: For micro-volume absorbance measurements and measurements taken with nonstandard (other than 10 mm) cuvettes, the spectra are normalized to a 10 mm pathlength equivalent.

A260 absorbance

- Nucleic acid absorbance values are measured at 260 nm using the normalized spectrum. This is the reported A260 value if Baseline Correction is not selected.
- If [Baseline Correction](#) is selected, the absorbance value at the correction wavelength is subtracted from the absorbance at 260 nm. The corrected absorbance at 260 nm is reported and used to calculate nucleic acid concentration.

A230 and A280 absorbance

- Normalized and baseline-corrected (if selected) absorbance values at 230 nm and 280 nm are used to calculate A260/A230 and A260/A280 ratios.

Sample Pathlength

- For micro-volume measurements, the software selects the optimal pathlength (between 1.0 mm and 0.03 mm) based on sample absorbance at the analysis wavelength.
- For cuvette measurements, the pathlength selected after switching to cuvette mode is utilized. (see [Cuvette Settings](#))
- Displayed spectra and absorbance values are normalized to a 10 mm pathlength equivalent.

Reported Values

- **Nucleic acid concentration.** Reported in selected unit (i.e., ng/ μ L, μ g/ μ L or μ g/mL). Calculations are based on modified Beer's Law equation using corrected nucleic acid absorbance value.
- **A260/A280 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 280 nm. An A260/A280 purity ratio of ~1.8 is generally accepted as "pure" for DNA (~2.0 for RNA). Acidic solutions may under represent the reported value by 0.2-0.3; the opposite is true for basic solutions.
- **A260/A230 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 230 nm. An A260/A230 purity ratio between 1.8 and 2.2 is generally accepted as "pure" for DNA and RNA.

Note: Although purity ratios are important indicators of sample quality, the best quality indicator quality is functionality in the downstream application of interest (e.g., real-time PCR).

- **Factor.** Used in conjunction with Beer's Law to calculate sample concentration
- **Contaminant** - If a contaminant was identified by the Acclaro software, the contaminant will be displayed in this column.
- **Factor.** Used in conjunction with Beer's Law to calculate sample concentration.
- **Baseline correction.** Wavelength selected for baseline correction and the absorbance detected at that wavelength.
- **Location.** Displays whether the measurement was taken from the pedestal or cuvette mode.
- **Monitored wavelength.** Enter an additional wavelength whose absorbance value you want included in the report.

- **Contaminant.** If a contaminant was identified by the Acclaro software, the contaminant will be displayed in this column.
- **Corrected.** Displays the corrected analyte concentration determined using the Acclaro software, if one is available.
- **Species.** The species of nucleic acid contaminant detected by the Acclaro software
- **Kit name.** Provides the name of the qPCR kit used to determine qPCR recipe calculations, if one was selected during setup.
- **Sample Dilution Ratio.** When dilution of the sample is recommended prior to qPCR, displays the volume of the measured sample relative to the final volume of the dilution.
- **Dilution required: diluent volume (µL).** When dilution of the sample is recommended prior to qPCR, displays the volume of the original nucleic acid sample to add to the dilution tube.
- **Dilution required: diluent volume (µL).** When dilution of the sample is recommended prior to qPCR, displays the volume of diluent to add to the dilution tube.
- **Kit reaction: nucleic acid volume (µL).** If no dilution is recommended in the Dilution Required fields, this indicates the volume of original nucleic acid sample to add to the qPCR reaction tube. If a dilution was recommended in the Dilution Required fields, this indicates the volume of the nucleic acid dilution tube to add to the qPCR reaction tube.
- **Kit reaction: water volume (µL).** Displays the volume of water to add to the qPCR reaction tube.
- **Kit reaction: final template conc.** Displays the concentration of the nucleic acid sample after it's been added to the qPCR reaction tube.

4 Nucleic Acid Applications

dsDNA, ssDNA or RNA

- **Stirrer.** "Off" will be displayed when the stirrer function on a cuvette model is not used. When the stirrer function is used, the stir speed is displayed.
- **Heater.** "Off" or "On" will be displayed to show if the cuvette port was heated during the measurement.
- **Measured Temperature.** Temperature of the cuvette port during the measurement will be displayed.
- **Target Temperature.** Desired temperature of the cuvette port will be displayed.

Microarray

Measures the concentration of purified nucleic acids that have been labeled with up to two fluorescent dyes for use in downstream microarray applications.

[Measure Microarray](#)

[Reported Results](#)

[Settings](#)

[Dye Chromophore Editor](#)

[Detection Limits](#)

[Calculations](#)



Measure Microarray

Use the Microarray application to quantify nucleic acids that have been labeled with up to two fluorescent dyes. The application reports nucleic acid concentration, an A260/A280 ratio and the concentrations and measured absorbance values of the dye(s), allowing detection of dye concentrations as low as 0.2 picomole per microliter.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **Nucleic Acids** tab, then select **Microarray**.

2. Configure any of the [setup options](#) if desired and select **Save**.

Tip: Select a dye from the pre-defined list or add a custom dye using the [Dye/Chromophore Editor](#).

3. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Pipette 1–2 μL blanking solution onto the lower pedestal and lower the arm, or insert the blanking cuvette into the cuvette holder.

Tip: If using a cuvette, make sure to [align the cuvette light path](#) with the instrument light path.


5. Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements.)

6. Lift the arm and clean both pedestals with a new laboratory wipe, or remove the blanking cuvette.
7. Pipette 1-2 μL sample solution onto the pedestal, or insert the sample cuvette into the cuvette holder.
8. Start the sample measurement:
 - Pedestal: If [Auto-Measure](#) is On, lower arm; if Auto-Measure is Off, lower arm and select **Measure**.
 - Cuvette: Select **Measure**.

When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

9. When you are finished measuring samples, select **End Experiment**.
10. Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.




Related Topics

- [Best Practices for Nucleic Acid Measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Best Practices for Micro-Volume Measurements](#)
- [Best Practices for Cuvette Measurements](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

Microarray measurement screen (local control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. Here is an example of the measurement screen of the NanoDrop Ultra local control software:



Note

- A baseline correction is performed at 850 nm (absorbance value at 850 nm is subtracted from absorbance values at all wavelengths in sample spectrum).
- Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

Microarray measurement screen (PC control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See “[Measurement Screen Display Options](#)” on [page 299](#) for an example.

Reported values

The initial screen that appears after each measurement (see previous image) shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:

#:	2	Nucleic Acid Factor:	50
Date:	10/28/2024 11:36:51 AM	Oligo Sequence:	-
Sample Name:	Sample 2	Analysis correction:	0.303A@340nm
ng/μL:	20.9	Location:	Pedestal
A260:	0.72	Pathlength Used:	1 mm
A280:	0.48	Monitor Wavelength:	-
A260/A280:	1.51	Stirrer:	OFF
Dye1:	Alexa Fluor 647	Heater:	OFF
Dye 1 Concentration:	0.996pmol/μL	Measured Temperature:	-
Dye 1 Absorbance:	0.27	Target Temperature:	-
Dye2:	Alexa Fluor 660		
Dye 2 Concentration:	0.596pmol/μL		
Dye 2 Absorbance:	0.08		


- # (sample number)
- date (date and time sample measurement was taken)
- [sample name](#)
- [nucleic acid concentration](#)
- [A260](#)
- [A260/A280](#)
- [A280](#)
- [dye 1/dye 2 absorbance](#)
- [dye 1/dye 2 concentration](#)
- dye 1/dye 2 name
- [sample type](#)
- [analysis correction](#)
- [factor](#)

4 Nucleic Acid Applications

Microarray

- oligo sequence
- location
- monitor wavelength
- pathlength used
- stirrer
- heater
- measured temperature
- target temperature

Settings

The Microarray Setup screen appears after you select the Microarray application from the Nucleic Acids tab on the home screen. To show the Microarray settings from the Microarray measurement screen, select Microarray Setup .

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Sample type and Factor	dsDNA (with non-editable factor of 50 ng-cm/ μ L)	Widely accepted value for double-stranded DNA
	ssDNA (with non-editable factor of 33 ng-cm/ μ L)	Widely accepted value for single-stranded DNA
	RNA (with non-editable factor of 40 ng-cm/ μ L)	Widely accepted value for RNA
	Oligo DNA with non-editable calculated factor in ng-cm/ μ L	Factor calculated from user-defined DNA base sequence. When selected, available DNA base units (i.e., G, A, T, C) appear as keys. Define sequence by tapping appropriate keys. Factor is calculated automatically based on widely accepted value for each base unit.

Setting	Available Options	Description
	Oligo RNA with non-editable calculated factor in ng-cm/ μ L	Factor calculated from user-defined RNA base sequence. When selected, available RNA base units (i.e., G, A, U, C) appear as keys. Define sequence by tapping appropriate keys. Factor is calculated automatically based on widely accepted value for each base unit.
	Custom (with user-specified factor in ng-cm/ μ L)	Enter factor between 15 ng-cm/ μ L and 150 ng-cm/ μ L
Dye 1/Dye 2 Type ^a	Cy3, 5, 3.5, or 5.5, Alexa Fluor 488, 546, 555, 594, 647, or 660	Select pre-defined dye(s) used to label sample material, or one that has been added using Dye Editor.
Dye 1/Dye 2 Unit	picomoles/microliter (pmol/ μ L), micromoles (μ M), or millimoles (mM)	Select unit for reporting dye concentrations
Analysis Correction ^b	On or off Enter analysis correction wavelength in nm or use default value (340 nm)	Corrects sample absorbance measurement for any offset caused by light scattering particulates by subtracting absorbance value at specified analysis correction wavelength from absorbance value at analysis wavelength. Corrected value is used to calculate sample concentration. Tip: If the sample has a modification that absorbs light at 340 nm, select a different correction wavelength or turn off Analysis Correction.

^a To add a custom dye or edit the list of available dyes, use the Dye/Chromophore Editor.

^b The Analysis Correction affects the calculation for nucleic acid concentration only.

Dye/chromophore editor

Use the Dye/Chromophore Editor to add a custom dye to the list of available dyes in [Microarray Setup](#) or [Proteins & Labels Setup](#). You can also specify which dyes are available in that list.

To access the Dye/Chromophore Editor:


- From the home screen, select **Settings** , then select **Dye Editor**

4 Nucleic Acid Applications
Microarray

Dye Editor
(PC control)

Select to exit Dye Editor


Selected dyes (will appear in Dye1 and Dye2 lists in Microarray Setup or Proteins & Labels Setup)



Select to delete selected custom dye

Select to add custom dye

Dye Editor
(local control)



Locked dye (pre-defined; cannot be edited or deleted)


Custom dye (user-defined; can be edited or deleted)

Edit custom dye

Select custom dye

These operations are available from the Dye/Chromophore Editor:

Add custom dye

1. In the Dye Editor, select  to show the Add Dye box.
2. Enter unique **Name** for new dye (tap field to display keyboard when using the local control software, select **Done** key to close keyboard).
3. Select default **Unit** that will be used to display dye concentration.
4. Enter dye's **Extinction Coefficient** (or molar absorptivity constant) in L/mole-cm (typically provided by dye manufacturer).
5. Specify **Wavelength** in nm (between 350 nm and 840 nm) that will be used to measure dye's absorbance.
6. Specify dye's correction values at 260 nm and 280 nm
7. Select **Save** to close the Add Dye box.

The new custom dye appears in the **Dye 1** and **Dye 2** lists in Microarray Setup and Proteins and Labels Setup.


Note To determine dye correction values (if not available from dye manufacturer):


- use instrument to measure pure dye and note absorbance at 260 nm, 280 nm and at analysis wavelength for dye (see above)
- calculate ratio of $A_{260}/A_{\text{dye wavelength}}$ and enter that value for 260 nm Correction
- calculate ratio of $A_{280}/A_{\text{dye wavelength}}$ and enter that value for 280 nm Correction

When a custom dye is selected before a measurement, the dye's absorbance and concentration values are reported and the corrections are applied to the measured sample absorbance values, and to the resulting sample concentrations and purity ratios.

Edit custom dye


Tip Dyes pre-defined in the software cannot be edited.

1. In Dye Editor, select  in the row of the desired custom dye that needs to be edited.

2. Select  **Edit**.
3. Edit any entries or settings.
4. Select **Save**.

Delete custom dye

Tip Dyes pre-defined in the software cannot be deleted.

1. In Dye Editor, select one or more custom dyes to delete by using the check boxes to the left of each desired custom dye.
2. Select .
3. Select **Yes** to confirm deletion.

NOTICE Deleting a custom dye permanently removes the dye and all associated information from the software.

Detection Limits

The lower detection limits and reproducibility specifications for nucleic acids are provided [here](#). The upper detection limits are dependent on the [upper absorbance limit](#) of the instrument and the user-defined extinction coefficients.

To calculate upper detection limits for nucleic acid samples

To calculate upper detection limits in ng/μL, use the following equation:

$$(\text{upper absorbance limit}_{\text{instrument}} * \text{extinction coefficient}_{\text{sample}})$$

For example, for a sample measurement using an extinction coefficient of 55, the equation looks like this:

$$(550 \text{ AU} * 55 \text{ ng-cm}/\mu\text{L}) = 30,250 \text{ ng}/\mu\text{L}$$

Note For measurements with 10 mm pathlength cuvettes, the upper absorbance limit is 1.5 AU, which is approximately 75 ng/μL for dsDNA.

Related Topics

- [Detection Limits for All Applications](#)

Calculations for Microarray Measurements

As with the other nucleic acid applications, the Microarray application uses a [modification of the Beer-Lambert equation](#) to calculate sample concentration where the extinction coefficient and pathlength are combined and referred to as a “factor.” The Microarray application offers six options (shown at right) for selecting an appropriate factor for each measured sample, to be used in conjunction with Beer’s Law to calculate sample concentration.

If the factor is known, choose the Custom Factor option and enter the factor in ng-cm/ μ L. Otherwise, choose the option that best matches the sample solution.

Tip: Ideally, the factor or extinction coefficient should be determined empirically using a solution of the study nucleic acid at a known concentration using the same buffer.

Available Options for Factors

- **dsDNA** (factor = 50 ng-cm/ μ L)
- **ssDNA** (factor = 33 ng-cm/ μ L)
- **RNA** (factor = 40 ng-cm/ μ L)
- **Oligo DNA** (calculated from user entered DNA nucleotide sequence)
- **Oligo RNA** (calculated from user entered RNA nucleotide sequence)
- **Custom Factor** (user entered factor between 15 ng-cm/ μ L and 150 ng-cm/ μ L)

Note: See [Sample Type](#) for more information.

4 Nucleic Acid Applications

Microarray

Calculated nucleic acid concentrations are based on the absorbance value at 260 nm, the factor used and the sample pathlength. A single-point baseline correction (or analysis correction) may also be applied.

Concentration is reported in mass units. Calculators are available on the Internet to convert concentration from mass to molar units based on sample sequence.

Absorbance values at 260 nm, 280 nm and sometimes 230 nm are used to calculate purity ratios for the measured nucleic acid samples. Purity ratios are sensitive to the presence of contaminants in the sample, such as residual solvents and reagents typically used during sample purification.

Measured Values

A260 absorbance

Note: The absorbance value at 850 nm is subtracted from all wavelengths in the spectrum. As a result, the absorbance at 850 nm is zero in the displayed spectra. Also, for micro-volume absorbance measurements and measurements taken with nonstandard (other than 10 mm) cuvettes, the spectra are normalized to a 10 mm pathlength equivalent.

- Nucleic acid absorbance values for all Microarray [sample types](#) are measured at 260 nm using the 850-corrected and normalized spectrum.
- If [Analysis Correction](#) is selected, the absorbance value at the correction wavelength is subtracted from the absorbance at 260 nm.
- If one or more dyes are selected, the [dye correction values](#) at 260 nm are also subtracted from the absorbance at 260 nm.
- The final corrected absorbance at 260 nm is reported and used to calculate sample concentration.

A280 absorbance

- 850-corrected and normalized absorbance value at 280 nm (minus the A280 dye correction) is used to calculate an A260/A280 ratio.

Dye concentrations are calculated from the absorbance value at the dye's analysis wavelength, the dye's extinction coefficient, and the sample pathlength.

Dye absorbance

- Dye absorbance values are measured at specific wavelengths. See [Dye/Chromophore Editor](#) for analysis wavelengths used.
- Baseline-corrected dye absorbance values are reported and used to calculate dye concentrations.

Dye correction

- Pre-defined dyes have known correction values for A260 and A280. See [Dye/Chromophore Editor](#) for correction values used.
- A260 dye corrections are subtracted from the [A260 absorbance value](#) used to calculate nucleic acid concentration, and from the A260 absorbance value used to calculate the [A260/A280 purity ratio](#).

Sample Pathlength

- For micro-volume measurements, the software selects the optimal pathlength (between 1.0 mm and 0.03 mm) based on sample absorbance at the analysis wavelength.
- For cuvette measurements, the pathlength selected after switching to cuvette mode is used. (see [Cuvette Settings](#)).
- Displayed spectra and absorbance values are normalized to a 10 mm pathlength equivalent.

Reported Values

- **Nucleic acid concentration.** Reported in selected unit (i.e., ng/ μ L, μ g/ μ L or μ g/mL). Calculations are based on modified Beer's Law equation using corrected nucleic acid absorbance value.
- **A260/A280 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 280 nm. An A260/A280 purity ratio of \sim 1.8 is generally accepted as "pure" for DNA (\sim 2.0 for RNA). Acidic solutions may under represent the reported value by 0.2-0.3; the opposite is true for basic solutions.
- **Dye1/Dye2 concentration.** Reported in pmol/ μ L. Calculations are based on Beer's Law equation using baseline-corrected dye absorbance value(s).

Note: Although purity ratios are important indicators of sample quality, the best indicator of DNA or RNA quality is functionality in the downstream application of interest (e.g., microarray).

Related Topics

- [Calculations for Nucleic Acid Measurements](#)

Custom Factor

Measures the concentration of purified nucleic acids using a custom factor for the calculations.

[Measure Custom Factor](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)

[Calculations](#)



Measure Custom Factor

Use the Custom Factor application to quantify purified DNA or RNA samples that absorb at 260 nm with a user-defined extinction coefficient or factor. The application reports nucleic acid concentration and two absorbance ratios (A_{260}/A_{280} and A_{260}/A_{230}). A single-point baseline correction can also be used.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **Nucleic Acids** tab, then select **Custom Factor**.


2. Configure any of the [setup options](#) if desired and select **Save**.
3. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Pipette 1–2 μL blanking solution onto the lower pedestal and lower the arm, or insert the blanking cuvette into the cuvette holder.

Tip: If using a cuvette, make sure to [align the cuvette light path](#) with the instrument light path.
5. Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements.)
6. Lift the arm and clean both pedestals with a new laboratory wipe, or remove the blanking cuvette.
7. Pipette 1-2 μL sample solution onto the pedestal or insert the sample cuvette into the cuvette holder.
8. Start the sample measurement:
 - Pedestal: If [Auto-Measure](#) is On, lower arm; if Auto-Measure is Off, lower arm and select **Measure**.
 - Cuvette: Select **Measure**.

When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

9. When you are finished measuring samples, select **End Experiment**.
10. Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.



Typical nucleic acid spectrum


Related Topics

- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Best Practices for Micro-Volume Measurements](#)
- [Best Practices for Cuvette Measurements](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

Custom Factor measurement screen (local control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. Here is an example of the measurement screen of the NanoDrop Ultra local control software:



Note The Custom Factor measurement screen is identical to the measurement screen for the other standard nucleic acid applications.

Custom Factor measurement screen (PC control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.

Related Topics

- [Basic Instrument Operations](#)
- [Nucleic Acid Reported Results](#)
- [Nucleic Acid Calculations](#)


Reported values

The initial screen that appears after each measurement (see previous image) shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:

#:	1	Stirrer:	OFF
Date:	10/28/2024 09:49:21 AM	Heater:	OFF
Sample Name:	Sample 1	Measured Temperature:	-
ng/ μ L:	1064.5	Target Temperature:	-
A260/A280:	2.05		
A260/A230:	2.37		
A260:	21.29		
A280:	10.41		
Factor:	50		
Baseline Correction:	4.797A@340nm		
Location:	Pedestal		
Monitor Wavelength:	-		
Pathlength Used:	@260 0.2 mm		

- # (sample number)
- date (date and time sample measurement was taken)
- sample name
- nucleic acid concentration
- A260/A280
- A260/A230
- A260
- A280
- factor
- baseline correction
- location
- monitor wavelength
- pathlength used
- stirrer
- heater
- measured temperature
- target temperature

Settings

The Custom Factor Setup screen appears after you select the Custom Factor application from the Nucleic Acids tab on the home screen. To show the Custom Factor settings, from the measurement screen, select Custom Factor Setup .

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Custom Factor	Enter an integer value between 15 ng-cm/μL and 150 ng-cm/μL	Constant used to calculate nucleic acid concentration in modified Beer’s Law equation . Based on extinction coefficient and pathlength: $f = 1/(\epsilon_{260} * b)$ where: f = factor ε = molar extinction coefficient at 260 nm in ng-cm/μL b = sample pathlength in cm (1 cm for nucleic acids measured with the NanoDrop Ultra instruments)
Baseline Correction	On or off Enter baseline correction wavelength in nm or use default value (340 nm)	Optional user-defined baseline correction. Can be used to correct for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength.

Related Topics

- [Instrument Settings](#)

Detection Limits

The lower detection limits and reproducibility specifications for nucleic acids are provided [here](#). The upper detection limits are dependent on the [upper absorbance limit](#) of the instrument and the user-defined extinction coefficients.

To calculate upper detection limits for nucleic acid samples

To calculate upper detection limits in ng/μL, use the following equation:

$$(\text{upper absorbance limit}_{\text{instrument}} * \text{extinction coefficient}_{\text{sample}})$$

For example, for a sample measurement using an extinction coefficient of 55, the equation looks like this:

$$(550 \text{ AU} * 55 \text{ ng-cm}/\mu\text{L}) = 30,250 \text{ ng}/\mu\text{L}$$

Note For measurements with 10 mm pathlength cuvettes, the upper absorbance limit is 1.5 AU, which is approximately 75 ng/μL for dsDNA.

Related Topics

- [Detection Limits for All Applications](#)

4 Nucleic Acid Applications

This page is intentionally blank.

Oligo DNA or Oligo RNA

Measures the concentration of purified ssDNA or RNA oligonucleotides that absorb at 260 nm.

[Measure Oligo DNA or RNA](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)

[Calculations](#)



Measure Oligo DNA or Oligo RNA

Use the Oligo DNA and Oligo RNA applications to quantify oligonucleotides that absorb at 260 nm. Molar extinction coefficients are calculated automatically based on the user-defined base sequence of the sample. These applications report nucleic acid concentration and two absorbance ratios (A_{260}/A_{280} and A_{260}/A_{230}). A single-point baseline correction can also be used.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

Note If the oligonucleotide has been modified, for example with a fluorophore dye, check with the oligo manufacturer to determine if the modification contributes absorbance at 260 nm. If it does, we recommend using the [Microarray](#) application to quantify nucleic acid concentration. The Microarray application includes a correction to remove any absorbance contribution due to the dye from the oligo quantification result.

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure


1. From the home screen, select the **Nucleic Acids** tab, then select either **Oligo DNA** or **Oligo RNA**, as needed.
2. Configure any of the [setup options](#) if desired and select **Save**.
3. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Pipette 1–2 µL blanking solution onto the lower pedestal and lower the arm, or insert the blanking cuvette into the cuvette holder.

Tip: If using a cuvette, make sure to [align the cuvette light path](#) with the instrument light path.
5. Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements.)
6. Lift the arm and clean both pedestals with a new laboratory wipe, or remove the blanking cuvette.
7. Pipette 1-2 µL sample solution onto the pedestal or insert the sample cuvette into the cuvette holder.
8. Start the sample measurement:
 - Pedestal: If [Auto-Measure](#) is On, lower arm; if Auto-Measure is Off, lower arm and select **Measure**.
 - Cuvette: Select **Measure**.

When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

9. When you are finished measuring samples, select **End Experiment**.
10. Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.




Related Topics

- [Best Practices for Nucleic Acid Measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Best Practices for Micro-Volume Measurements](#)
- [Best Practices for Cuvette Measurements](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

Oligo DNA measurement screen (local control)

For each measured sample, the Oligo DNA and Oligo RNA applications show the UV absorbance spectrum and a summary of the results. Here is an example of the measurement screen of the NanoDrop Ultra local control software:



¹Measured oligo: TTT TTT TTT TTT TTT TTT TTT TTT

Note Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

4 Nucleic Acid Applications

Oligo DNA or Oligo RNA


- pathlength used
- monitored wavelength
- stirrer
- heater
- measured temperature
- target temperature


Note The five nucleotides that comprise DNA and RNA exhibit widely varying A260/A280 ratios. See [Oligo Purity Ratios](#) for more information.

Related Topics

- [Basic Instrument Operations](#)
- [Oligo Calculations](#)

Settings

The Oligo Setup screen appears after you select the Oligo DNA or Oligo RNA application from the Nucleic Acids tab on the home screen. To show the Oligo settings from the Oligo measurement screen, select Oligo Setup .



The screenshot shows the Oligo Setup screen with the following elements:


- Auto Naming:** A toggle switch that is turned on (checked).
- Sequence Input:** A large text area with the placeholder text "Click to enter".
- Factor:** 0.00
- Molecular weight:** 0.00
- Number of bases:** 0
- Molar ext. coefficient (260):** 0
- %GC:** 0.00
- Sequence Input:** A row of five buttons labeled G, A, T, C, and a delete icon (X).
- Baseline Correction:** A toggle switch that is turned on (checked).
- Wavelength:** A numeric input field containing "340" followed by "nm" and the text "(Default value is 340nm)".

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.

Oligo Base Sequence
for DNA: Use the G, A, T and C keys to specify the DNA base sequence

for RNA: Use the G, A, U and C keys to specify the RNA base sequence

Specify your DNA or RNA base sequence.
Tap or click the corresponding keys:



From the PC control software, you can also enter base sequence using the keyboard, or by copy and pasting a sequence from another application.

Each time a base is added to the sequence, the software calculates the following:

- **Factor.** Constant used to calculate oligonucleotide concentration in [modified Beer’s Law equation](#). Based on extinction coefficient and pathlength:

$$f = 1/(\mathcal{E}_{260} * b)$$

where:

f= factor

ε = molar extinction coefficient at 260 nm in ng-cm/μL

b = [sample pathlength](#) in cm (0.1 cm for nucleic acids measured with the NanoDrop Ultra instrument)

Setting	Available Options	Description
		<ul style="list-style-type: none"> • Molecular Weight of oligo calculated from user-defined base sequence. • Number of Bases entered. • Molar Ext. Coefficient (260 nm). Molar extinction coefficient of oligo (in ng-cm/μL) at 260 nm calculated from entered base sequence. • %GC. Percentage of guanine and cytosine residues in total number of bases entered.
Baseline Correction	On or off Enter baseline correction wavelength in nm or use default value (340 nm)	Corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength. Tip: If the sample has a modification that absorbs light at 340 nm, select a different correction wavelength or turn off Baseline Correction.

Detection Limits

The lower detection limits and reproducibility specifications for the oligonucleotide sample types (ssDNA and RNA) are provided [here](#). The upper detection limits are dependent on the [upper absorbance limit](#) of the instrument and the extinction coefficients for the user-defined [base sequences](#).

To calculate upper detection limits for nucleic acid samples

To calculate upper detection limits in ng/ μ L, use the following equation:

$$(\text{upper absorbance limit}_{\text{instrument}} * \text{extinction coefficient}_{\text{sample}})$$

For example, for a sample measurement using an extinction coefficient of 55, the equation looks like this:

$$(550 \text{ AU} * 55 \text{ ng-cm}/\mu\text{L}) = 30,250 \text{ ng}/\mu\text{L}$$

Note For measurements with 10 mm pathlength cuvettes, the upper absorbance limit is 1.5 AU, which is approximately 75 ng/ μ L for dsDNA.

Calculations

As with the other nucleic acid applications, the Oligo applications use the [Beer-Lambert equation](#) to correlate absorbance with concentration based on the sample's extinction coefficient and pathlength. Because oligonucleotides are short, single-stranded molecules (or longer molecules of repeating sequences), their spectrum and extinction coefficient (ϵ) are closely dependent on base composition and sequence.

(The generally accepted extinction coefficients and factors for single-stranded DNA and RNA provide a reasonable estimate for natural, essentially randomized, sequences but not for short, synthetic oligo sequences.) To ensure the most accurate results, we use the exact value of ϵ_{260} to calculate oligonucleotide concentration.

The NanoDrop software allows you to specify the base sequence of an oligonucleotide before it is measured. For any entered base sequence, the software uses the equation at the right to calculate the extinction coefficient.

Tip: The extinction coefficient is wavelength specific for each oligonucleotide and can be affected by buffer type, ionic strength and pH.

Extinction Coefficients for Oligonucleotides

The software uses the nearest neighbor method and the following formula to calculate molar extinction coefficients for specific oligonucleotide base sequences:

$$\epsilon_{260} = \sum_{1}^{N-1} \epsilon_1 - \sum_{2}^{N-1} \epsilon_2 + \sum_{1}^N \epsilon_3$$

where:

ϵ = molar extinction coefficient in L/mole-cm

ϵ_1 = ϵ nearest neighbor

ϵ_2 = ϵ individual bases

ϵ_3 = ϵ modifications, such as fluorescent dyes

4 Nucleic Acid Applications

Oligo DNA or Oligo RNA

Calculated nucleic acid concentrations are based on the absorbance value at 260 nm, the factor used and the sample pathlength. A single-point baseline correction (or analysis correction) may also be applied.

Concentration is reported in mass units. Calculators are available on the Internet to convert concentration from mass to molar units based on sample sequence.

Absorbance values at 260 nm, 280 nm and sometimes 230 nm are used to calculate purity ratios for the measured nucleic acid samples. Purity ratios are sensitive to the presence of contaminants in the sample, such as residual solvents and reagents typically used during sample purification.

Measured Values

A260 absorbance

Note: For micro-volume absorbance measurements and measurements taken with nonstandard (other than 10 mm) cuvettes, the spectra are normalized to a 10 mm pathlength equivalent.

- Nucleic acid absorbance values are measured at 260 nm using the normalized spectrum. This is the reported A260 value if Baseline Correction is not selected.
- If [Baseline Correction](#) is selected, the absorbance value at the correction wavelength is subtracted from the sample absorbance at 260 nm. The corrected absorbance at 260 nm is reported and used to calculate nucleic acid concentration.

A230, A280 absorbance

- Normalized absorbance values at 230 nm, 260 nm and 280 nm are used to calculate A260/A230 and A260/A280 ratios.

Sample Pathlength

- For micro-volume measurements, the software selects the optimal pathlength (between 1.0 mm and 0.03 mm) based on sample absorbance at the analysis wavelength.
- For cuvette measurements, the pathlength selected after switching to cuvette mode is utilized (see [Cuvette Settings](#)).
- Displayed spectra and absorbance values are normalized to a 10 mm pathlength equivalent.

The five nucleotides that comprise DNA and RNA exhibit widely varying A260/A280 ratios. Estimated A260/A280 ratios for each independently measured nucleotide are provided below:

Guanine: 1.15
Adenine: 4.50
Cytosine: 1.51
Uracil: 4.00
Thymine: 1.47

The A260/A280 ratio for a specific nucleic acid sequence is approximately equal to the weighted average of the A260/A280 ratios for the four nucleotides present.

Note: RNA will typically have a higher 260/280 ratio due to the higher ratio of Uracil compared to that of Thymine.

Reported Values

- **Nucleic acid concentration.** Reported in selected unit (i.e., ng/ μ L, μ g/ μ L or μ g/mL). Calculations are based on modified Beer's Law equation using corrected nucleic acid absorbance value.
- **A260/A280 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 280 nm.
- **A260/A230 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 230 nm.

Note: The traditional purity ratios (A260/A280 and A260/A230), which are used as indicators of the presence of various contaminants in nucleic acid samples, do not apply for oligonucleotides because the shapes of their spectra are highly dependent on their base compositions. See side bar for more information.

Related Topics

- [Calculations for Nucleic Acid Measurements](#)

4 Nucleic Acid Applications

This page is intentionally blank.

Protein Applications

Protein A280

Measures the concentration of purified protein samples that absorb at 280 nm.

[Measure Protein A280](#)

[Reported Results](#)

[Settings](#)

[Protein Editor](#)

[Detection Limits](#)

[Calculations](#)



Best practices for protein measurements

- Isolate and purify protein samples before measurement to remove impurities. Depending on the sample, impurities could include DNA, RNA and some buffer components. See [Preparing Samples](#) for more information.

Note Extraction reagents that contribute absorbance between 200 nm and 280 nm will affect measurement results if present in samples (even residual amounts).

- Ensure the sample absorbance is within the instrument's [absorbance detection limits](#).
- Choosing a blank:
 - For the Protein A280, Protein A205, and Proteins & Labels applications, blank with the same buffer solution used to resuspend the analyte of interest. The blanking solution should be a similar pH and ionic strength as the analyte solution.

- For the Protein BCA, Protein Bradford, and Protein Lowry applications, blank with deionized water (DI H₂O).
- For the Protein Pierce 660 application, blank with the reference solution used to make the standard curve (reference solution should contain none of the standard protein stock). For more information, see [Working with standard curves](#).
- Run a [blanking cycle](#) to assess the absorbance contribution of your buffer solution. If the buffer exhibits strong absorbance at or near the analysis wavelength (typically 280 nm or 205 nm), you may need to choose a different buffer or application, such as a colorimetric assay (for example, BCA or Pierce 660). See [Choosing and Measuring a Blank](#) for more information.

Note Buffers such as Triton X, RIPA, and NDSB contribute significant absorbance and are not compatible with direct A280 or A205 measurements.

- For micro-volume measurements:
 - Ensure pedestal surfaces are properly [cleaned](#) and [conditioned](#). (Proteins tend to stick to pedestal surfaces.)
 - Gently (but thoroughly) vortex samples before taking a measurement. Avoid introducing bubbles when mixing and pipetting.
 - Follow [best practices for micro-volume measurements](#).
 - Use a 2 µL sample volume. See [Recommended Sample Volumes](#) for more information.
- For cuvette measurements (NanoDrop Ultra^C and NanoDrop Ultra^C FL instruments only), use compatible cuvettes and follow [best practices for cuvette measurements](#).

Related Topics

- [Best practices for protein measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Measure Protein A280

Use the Protein A280 application to quantify purified protein samples that contain amino acids such as tryptophan or tyrosine, or cys-cys disulfide bonds, which exhibit absorbance at 280 nm. This application reports protein concentration measured at 280 nm and one absorbance ratio (A260/A280). A single-point baseline correction can also be used. This application does not require a standard curve.

Before you begin...

Before taking pedestal measurements with one of the NanoDrop Ultra instruments, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

Note If your samples contain mainly peptide bonds and little or no amino acids, use the [Protein A205](#) application instead of Protein A280.

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **Proteins** tab, then select **Protein A280**.
2. Configure any of the [setup options](#) if desired and select **Save**.
3. If using a NanoDrop Ultra^C or NanoDrop Ultra^C FL model, select the correct measurement pathway.
 - When using a cuvette, from the drop-down menu at the top of the screen, select **Cuvette** to display the cuvette settings. Select the desired pathlength, stir speed, and heating, and then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Pipette 1–2 µL blanking solution onto the lower pedestal and lower the arm, or insert the blanking cuvette into the cuvette holder.

Tip: If using a cuvette, make sure to [align the cuvette light path](#) with the instrument light path.

5. Select **Blank** and wait for the measurement to complete.

Tip: If **Auto-Blank** is On, the blank measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements.)

6. Lift the arm and clean both pedestals with a new laboratory wipe, or remove the blanking cuvette.

7. Pipette 2 μL sample solution onto the pedestal, or insert the sample cuvette into the cuvette holder.

8. Start the sample measurement:


– Pedestal: If **Auto-Measure** is On, lower arm; if Auto-Measure is Off, lower arm and select **Measure**.

– Cuvette: Select **Measure**.


When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

9. When you are finished measuring samples, select **End Experiment**.

10. Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.



High concentration BSA sample




Low concentration BSA sample

Reported Results

Protein A280 measurement screen (local control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. The following is an example of the measurement screen of the NanoDrop Ultra local control software:



Note Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

Protein A280 measurement screen (PC control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.

Reported values

The initial screen that appears after each measurement (see previous image) shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:

#:	4	Pathlength Used:	@280 0.2 mm
Date:	10/28/2024 11:52:44 AM	Contaminant:	-
Sample Name:	Sample 4	Corrected:	-
mg/mL:	25.631	Stirrer:	OFF
A280:	17.17	Heater:	OFF
A260/A280:	0.60	Measured Temperature:	-
Sample Type:	BSA	Target Temperature:	-
Baseline Correction:	0.607A@340nm		
Location:	Pedestal		
Monitor Wavelength:	-		
Ext.Coeff.E1%/gm-cm:	6.7		
$\epsilon/1000$:	-		
MW(kDa):	-		


- # (sample number)
- date (date and time sample measurement was taken)
- [sample name](#)
- [protein concentration](#)
- [A280](#)
- [A260/A280](#)
- [sample type](#)
- [baseline correction](#)
- [location](#)
- [monitored wavelength](#)
- mass extinction coefficient (1% solution)
- molar extinction coefficient
- molecular weight (kDA)
- pathlength used
- [contaminant](#)
- [corrected protein concentration](#)
- [stirrer](#)

- heater
- measured temperature
- target temperature


Related Topics

- Basic Instrument Operations
- Protein A280 Calculations

Settings

The Protein A280 Setup screen appears after you select the application from the Proteins tab on the home screen. To show the Protein A280 settings, from the Protein A280 measurement screen, select the setup icon .

The Protein A280 application provides a variety of sample type options for purified protein analysis.



Each sample type applies a unique extinction coefficient to the protein calculations. If the extinction coefficient of the sample is known, choose the ϵ + MW (molar) or ϵ 1% (mass) option and enter the value. Otherwise, calculate the extinction coefficient or choose the option that best matches the sample solution. If you only need a rough estimate of protein concentration and the sample extinction coefficient is unknown, select the 1 Abs=1 mg/mL sample type option.

Tip Ideally, the extinction coefficient should be determined empirically using a solution of the study protein at a known concentration using the same buffer.

Setting	Available Options	Mass Ext. Coefficient (L/gm-cm)	Description
Auto Naming	On or Off	N/A	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Baseline Correction	On or Off Enter baseline correction wavelength in nm or use default value (340 nm)	N/A	Corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength. Tip: If the sample has a modification that absorbs light at 340 nm, select a different correction wavelength or turn off Baseline Correction.
Sample type ^a	1 Abs = 1 mg/mL	General reference	Recommended when extinction coefficient is unknown and rough estimate of protein concentration is acceptable for a solution with no other interfering substances. Assumes 0.1% (1 mg/mL) protein solution produces 1.0A at 280 nm (where pathlength is 10 mm), i.e., $\epsilon_{1\%} = 10$.
	BSA	6.7	Calculates BSA (Bovine Serum Albumin) protein concentration using mass extinction coefficient (ϵ) of 6.7 L/gm-cm at 280 nm for 1% (i.e., 10 mg/mL) BSA solution. Assuming MW is 66,400 daltons (Da), molar extinction coefficient at 280 nm for BSA is approximately $43,824 \text{ M}^{-1}\text{cm}^{-1}$.

5 Protein Applications
Protein A280

Setting	Available Options	Mass Ext. Coefficient (L/gm-cm)	Description
	IgG	13.7	Suitable for most mammalian antibodies (i.e., immunoglobulin G or IgG). Calculates protein concentration using mass extinction coefficient (ϵ) of 13.7 L/gm-cm at 280 nm for 1% (i.e., 10 mg/mL) IgG solution. Assuming MW is 150,000 Da, molar extinction coefficient at 280 nm for IgG is approximately 210,000 M ⁻¹ cm ⁻¹ .
	Lysozyme	26.4	Calculates lysozyme protein concentration using mass extinction coefficient (ϵ) of 26.4 L/gm-cm at 280 nm for 1% (i.e., 10 mg/mL) lysozyme solution. Assumes molar extinction coefficient for egg white lysozyme ranges between 36,000 M ⁻¹ cm ⁻¹ and 39,000 M ⁻¹ cm ⁻¹ .
	Other protein (ϵ + MW)	User entered molar extinction coefficient and molecular weight	Assumes protein has known molar extinction coefficient (ϵ) and molecular weight (MW), where: $(\epsilon_{\text{molar}}) * 10 = (\epsilon_{\text{percent}}) * (MW_{\text{protein}})$ Enter MW in kiloDaltons (kDa) and molar extinction coefficient (ϵ) in M ⁻¹ cm ⁻¹ divided by 1000 (i.e., $\epsilon/1000$). For example, for protein with molar extinction coefficient of 210,000 M ⁻¹ cm ⁻¹ , enter 210.
	Other protein (ϵ 1%)	User entered mass extinction coefficient	Assumes protein has known mass extinction coefficient (ϵ). Enter mass extinction coefficient in L/gm-cm for 10 mg/mL (ϵ 1%) protein solution.


^a To add or edit a custom protein, use Protein Editor.

Protein editor

Use the Protein Editor to add a custom protein to the list of available protein sample types in [Protein A280 Setup](#) and [Proteins & Labels Setup](#).

To access the Protein Editor:


- From the software home screen, select  **Settings > Protein Editor**.




Protein Editor as seen on the NanoDrop Ultra local control software

These operations are available from the Protein Editor:

Add custom protein

- In Protein Editor, select  to show the Add Protein box.
- Enter a unique **Name** for the new protein (from local control, tap field to display keyboard, tap **Done** key to close keyboard).
- Enter a **Description** for the new protein.
- Specify whether to enter **Molar Extinction** coefficient or **Mass Extinction** coefficient for the custom protein by selecting the appropriate radio button.
 - If **Mass Extinction** coefficient is selected, enter mass extinction coefficient in L/gm-cm for 10 mg/mL (E1%) protein solution.





- If **Molar Extinction** is selected,
 - Enter molar extinction coefficient (ϵ) in $M^{-1}cm^{-1}$ divided by 1000 (that is, $\epsilon/1000$). For example, for protein with molar extinction coefficient of $210,000 M^{-1}cm^{-1}$, enter 210.
 - Enter molecular weight (MW) in kiloDaltons (kDa).


5. Select **Save** to close the **Add Protein** box.

The new custom protein appears in the **Sample Type** list in Protein A280 Setup, Proteins & Labels Setup, and Protein A280 Pro Setup.

Edit custom protein

1. In Protein Editor, select  in the row of the desired Protein type that needs to be edited.
2. Select  **Edit** to show the **Edit Protein** box.
3. Edit any entries or settings.
4. Select **Save**.

Delete custom protein

1. In Protein Editor, select one or more custom proteins to delete by using the check boxes to the left of each desired Protein.
2. Select .

3. Confirm deletion of the custom protein(s) by selecting **Yes**.

Note Deleting a custom protein permanently removes the protein and all associated information from the software.

Detection Limits

Detection limits and reproducibility specifications for purified BSA proteins are provided [here](#). The BSA lower detection limit and reproducibility values apply to any protein sample type. The upper detection limits are dependent on the [upper absorbance limit](#) of the instrument and the sample's extinction coefficient.

To calculate upper detection limits for other (non-BSA) protein sample types

To calculate upper detection limits in ng/μL for proteins, use the following equation:

$$(\text{upper absorbance limit}_{\text{instrument}} / \text{mass extinction coefficient}_{\text{sample}}) * 10$$

For example, if the sample's mass extinction coefficient at 280 nm is 6.7 for a 1% (10 mg/mL) solution, the equation looks like this:

$$(550 / 6.7) * 10 = 824.6 \text{ (or } \sim 825)$$

Calculations

The Protein A280 application uses the [Beer-Lambert equation](#) to correlate absorbance with concentration. Solving Beer's law for concentration yields the equation at the right.

Beer-Lambert Equation (solved for concentration)

$$c = A / (\epsilon * b)$$

where:

A = UV absorbance in absorbance units (AU)

ε = wavelength-dependent molar absorptivity coefficient (or extinction coefficient) in liter/mol-cm

b = pathlength in cm

c = analyte concentration in moles/liter or molarity (M)

Note: Dividing the measured absorbance of a sample solution by its molar extinction coefficient yields the molar concentration of the sample. See [Published Extinction Coefficients](#) for more information regarding molar vs. mass concentration values.

5 Protein Applications

Protein A280

The extinction coefficient of a peptide or protein is related to its tryptophan (W), tyrosin (Y) and cysteine (C) amino acid composition.

Tip: The extinction coefficient is wavelength specific for each protein and can be affected by buffer type, ionic strength and pH.

This application offers six options (shown at right) for selecting an appropriate extinction coefficient for each measured sample, to be used in conjunction with Beer's Law to calculate sample concentration.

If the extinction coefficient of the sample is known, choose the ϵ + MW (molar) or ϵ 1% (mass) option and enter the value. Otherwise, calculate the extinction coefficient or choose the option that best matches the sample solution.

Tip: Ideally, the extinction coefficient should be determined empirically using a solution of the study protein at a known concentration using the same buffer.

Extinction Coefficients for Proteins

At 280 nm, the extinction coefficient is approximated by the weighted sum of the 280 nm molar extinction coefficients of the three constituent amino acids, as described in this equation:

$$\epsilon = (nW * 5500) + (nY * 1490) + (nC * 125)$$

where:

ϵ = molar extinction coefficient

n = number of each amino acid residue

5500, 1490 and 125 = amino acid molar absorptivities at 280 nm

Available Options for Extinction Coefficient

- **1 Abs = 1 mg/mL**, where sample type and/or ext. coefficient is unknown (produces rough estimate of protein concentration)
- **BSA** (Bovine Serum Albumin, 6.7 L/gm-cm)
- **IgG** (any mammalian antibody, 13.7 L/gm-cm)
- **Lysozyme** (egg white lysozyme, 26.4 L/gm-cm)
- **Other protein** (ϵ + MW), user-specified molar ext. coefficient
- **Other protein** (ϵ 1%), user-specified mass ext. coefficient

Note: See [Sample Type](#) for details.

Most sources report extinction coefficients for proteins measured at or near 280 nm in phosphate or other physiologic buffer. These values provide sufficient accuracy for routine assessments of protein concentration.

The equation at the right shows the relationship between molar extinction coefficient (ϵ_{molar}) and percent extinction coefficient ($\epsilon 1\%$).

To determine concentration (c) of a sample in mg/mL, use the equation at the right and a conversion factor of 10.

Tip: The NanoDrop Ultra software includes the conversion factor when reporting protein concentrations.

Published Extinction Coefficients

Published extinction coefficients for proteins may be reported as:

- wavelength-dependent molar absorptivity (or extinction) coefficient (ϵ) with units of $\text{M}^{-1}\text{cm}^{-1}$
- percent solution extinction coefficient ($\epsilon 1\%$) with units of $(\text{g}/100 \text{ mL})^{-1}\text{cm}^{-1}$ (i.e., 1% or 1 g/100 mL solution measured in a 1 cm cuvette)
- protein absorbance values for 0.1% (i.e., 1 mg/mL) solutions

Tip: Assess published values carefully to ensure unit of measure is applied correctly.

Conversions Between ϵ_{molar} and $\epsilon 1\%$

$$(\epsilon_{\text{molar}}) * 10 = (\epsilon 1\%) * (\text{MW}_{\text{protein}})$$

Example: To determine percent solution extinction coefficient ($\epsilon 1\%$) for a protein that has a molar extinction coefficient of $43,824 \text{ M}^{-1}\text{cm}^{-1}$ and a molecular weight (MW) of 66,400 daltons (Da), rearrange and solve the above equation as follows:

$$\epsilon 1\% = (\epsilon_{\text{molar}} * 10) / (\text{MW}_{\text{protein}})$$

$$\epsilon 1\% = (43,824 * 10) / 66,400 \text{ Da}$$

$$\epsilon 1\% = 6.6 \text{ g}/100 \text{ mL}$$

Conversions Between g/100 mL and mg/mL

$$C_{\text{protein}} \text{ in mg/mL} = (A / \epsilon 1\%) * 10$$

Example: If measured absorbance for a protein sample at 280 nm relative to the reference is 5.8 A, protein concentration can be calculated as:

$$C_{\text{protein}} = (A / \epsilon 1\%) * 10$$

$$C_{\text{protein}} = (5.8/6.6 \text{ g}/100 \text{ mL}) * 10$$

$$C_{\text{protein}} = 8.79 \text{ mg/mL}$$

5 Protein Applications

Protein A280

Calculated protein concentrations are based on the absorbance value at 280 nm, the selected (or entered) extinction coefficient and the sample pathlength. A single-point baseline correction (or analysis correction) may be applied.

Concentration is reported in mass units. Calculators are available on the Internet to convert concentration from mass to molar units based on sample sequence.

Absorbance values at 260 nm and 280 nm are used to calculate purity ratios for the measured protein samples.

Purity ratios are sensitive to the presence of contaminants in the sample, such as residual solvents and reagents typically used during sample purification.

Measured Values

A280 absorbance

Note: For micro-volume absorbance measurements and measurements taken with nonstandard (other than 10 mm) cuvettes, the spectra are normalized to a 10 mm pathlength equivalent.

- Protein absorbance values are measured at 280 nm using the normalized spectrum. If Baseline Correction is not selected, this is the reported A280 value and the value used to calculate protein concentration.
- If [Baseline Correction](#) is selected, the normalized and baseline-corrected absorbance value at 280 nm is reported and used to calculate protein concentration.

A260 absorbance

- Normalized and baseline-corrected (if selected) absorbance value at 260 nm is used to calculate A260/A280 ratios.

Sample Pathlength

- For micro-volume measurements, the software selects the optimal pathlength (between 1.0 mm and 0.03 mm) based on sample absorbance at the analysis wavelength.
- For cuvette measurements, the pathlength selected after switching to cuvette mode is utilized. (see [Cuvette Settings](#)).
- Displayed spectra and absorbance values are normalized to a 10 mm pathlength equivalent.

Reported Values

- **Protein concentration.** Reported in selected unit (mg/mL or $\mu\text{g/mL}$). Calculations are based on Beer-Lambert equation using corrected protein absorbance value.
- **A260/A280 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 280 nm. An A260/A280 purity ratio of ~ 0.57 is generally accepted as “pure” for proteins.

Note: Although purity ratios are important indicators of sample quality, the best indicator of protein quality is functionality in the downstream application of interest (e.g., real-time PCR).

- **Sample type.** Determines the extinction coefficient used in conjunction with Beer's Law to calculate sample concentration.
- **Baseline correction.** Wavelength selected for baseline correction and the absorbance detected at that wavelength.
- **Location.** Displays whether the measurement was taken from the pedestal or cuvette mode.
- **Monitored wavelength.** Enter an additional wavelength whose absorbance value you want included in the report.
- **Mass extinction coefficient (1% solution).**
- **Molar extinction coefficient.**
- **Molecular weight.**
- **Contaminant.** If a contaminant was identified by the Acclaro software, the contaminant will be displayed in this column.
- **Corrected.** Displays the corrected analyte concentration determined using the Acclaro software, if one is available.
- **Stirrer.** "Off" will be displayed when the stirrer function on a cuvette model is not used. When the stirrer function is used, the stir speed is displayed.

5 Protein Applications

Protein A280

- **Heater.** "Off" or "On" will be displayed to show if the cuvette port was heated during the measurement.
- **Measured Temperature.** Temperature of the cuvette port during the measurement will be displayed.
- **Target Temperature.** Desired temperature of the cuvette port will be displayed.

Protein A205

Measures the concentration of purified protein populations that absorb at 205 nm.

[Measure Protein A205](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)

[Calculations](#)



Measure Protein A205

Use the Protein A205 application to quantify purified peptides and other proteins that contain peptide bonds, which exhibit absorbance at 205 nm. This application reports protein concentration and two absorbance values (A205 and A280). A single-point baseline correction can also be used. This application does not require a standard curve.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

Note If your samples contain mainly amino acids such as tryptophan or tyrosine, or cys-cys disulfide bonds, use the [Protein A280](#) application instead of Protein A205.

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **Proteins** tab, then select **Protein A205**.

2. Configure any of the [setup options](#) if desired and select **Save**.
3. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Pipette 1–2 μL of the blanking solution onto the lower pedestal and lower the arm, or insert the blanking cuvette into the cuvette holder.

Tip: If using a cuvette, make sure to [align the cuvette light path](#) with the instrument light path.
5. Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements.)
6. Lift the arm and clean both pedestals with a new laboratory wipe, or remove the blanking cuvette.
7. Pipette 2 μL sample solution onto the pedestal and lower the arm, or insert the sample cuvette into the cuvette holder.
8. Start the sample measurement:
 - Pedestal: If [Auto-Measure](#) is On, lower arm; if Auto-Measure is Off, lower arm and select **Measure**.
 - Cuvette: Select **Measure**.

When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).
9. When you are finished measuring samples, tap **End Experiment**.
10. Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.


Related Topics

- [Best practices for protein measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Protein A205 Reported Results

Protein A205 measurement screen (local control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. The following is an example of the measurement screen of the NanoDrop Ultra local control software:



Note Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

Protein A205 measurement screen (PC control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See “[Measurement Screen Display Options](#)” on [page 299](#) for an example.

Protein A205 reported values

The initial screen that appears after each measurement (see previous image) shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:


#:	1	Heater:	OFF
Date:	10/28/2024 12:15:10 PM	Measured Temperature:	-
Sample Name:	Sample 1	Target Temperature:	-
mg/mL:	4.593		
A205:	142.37		
A280:	0.40		
Method:	31		
Baseline Correction:	0.340A@340nm		
Location:	Pedestal		
Monitor Wavelength:	-		
Extinction Coefficient:	31		
Pathlength Used:	@205 0.05 mm		
Stirrer:	OFF		

- # (sample number)
- date (date and time sample measurement was taken)
- sample name
- protein concentration
- A205
- A280
- method
- baseline correction
- location
- monitored wavelength
- mass extinction coefficient (1% solution)
- pathlength used
- stirrer
- heater
- measured temperature
- target temperature


Related Topics

- [Basic Instrument Operations](#)
- [Protein A205 Calculations](#)

Settings

The Protein A205 Setup screen appears after you select the application from the Proteins tab on the home screen. To show the Protein A205 settings, from the Protein A205 measurement screen, select the setup icon .

The Protein A205 application provides a variety of method options for protein analysis.



Setting	Available Options	Mass Ext. Coefficient (L/gm-cm)	Description
Method	31	31	Assumes ϵ 0.1% (1 mg/mL) at 205 nm = 31
	Scopes	$27 + 120 * (A280/A205)$	Assumes ϵ 0.1% (1 mg/mL) at 205 nm = $27 + 120 * (A280/A205)$
	Other protein (ϵ 1%)	User entered mass extinction coefficient	Assumes protein has known mass extinction coefficient (ϵ). Enter mass extinction coefficient in L/gm-cm for 1 mg/mL (ϵ 0.1%) protein solution.

Setting	Available Options	Mass Ext. Coefficient (L/gm-cm)	Description
Auto Naming	On or Off	N/A	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Baseline Correction	On or Off Enter baseline correction wavelength in nm or use default value (340 nm)	N/A	Corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength. Tip: If the sample has a modification that absorbs light at 340 nm, select a different correction wavelength or turn off Baseline Correction.

Detection Limits

See “[Detection Limits for All Applications](#)” on [page 31](#).

Calculations

As with the other protein applications, Protein A205 uses the [Beer-Lambert equation](#) to correlate absorbance with concentration based on the sample's extinction coefficient and pathlength.

This application offers three options (shown at right) for selecting an appropriate extinction coefficient for each measured sample, to be used in conjunction with Beer's Law to calculate sample concentration.

If the extinction coefficient of the sample is known, choose the **£1% (mass)** option and enter the value. Otherwise, calculate the extinction coefficient or choose the option that best matches the sample solution.

Tip: Ideally, the extinction coefficient should be determined empirically using a solution of the study protein at a known concentration using the same buffer.

Calculated protein concentrations are based on the absorbance value at 205 nm, the selected (or entered) extinction coefficient and the sample pathlength. A single-point baseline correction may also be applied.

Concentration is reported in mass units. Calculators are available on the Internet to convert concentration from mass to molar units based on the sample sequence.

Available Options for Extinction Coefficient

- **31**, assumes £0.1% (1 mg/mL) at 205 nm = 31
- **Scopes**, assumes £0.1% (1 mg/mL) at 205 nm = $27 + 120 * (A280/A205)$
- **Other protein**, enter mass extinction coefficient in L/gm-cm for 1 mg/mL (£0.1%) protein solution

Note: See [Sample Type](#) for details.

Measured Values

A205 absorbance

Note: For micro-volume absorbance measurements and measurements taken with nonstandard (other than 10 mm) cuvettes, the spectra are normalized to a 10 mm pathlength equivalent.

- Protein absorbance values are measured at 205 nm using the normalized spectrum. If Baseline Correction is not selected, this is the reported A205 value and the value used to calculate protein concentration.
- If [Baseline Correction](#) is selected, the normalized and baseline-corrected absorbance value at 205 nm is reported and used to calculate protein concentration.

A280 absorbance

- Normalized and baseline-corrected (if selected) absorbance value at 280 nm is also reported.

Sample Pathlength

- For micro-volume measurements, the software selects the optimal pathlength (between 1.0 mm and 0.03 mm) based on sample absorbance at the analysis wavelength.
- For cuvette measurements, the pathlength selected after switching to cuvette mode is utilized (see [Cuvette Settings](#)).
- Displayed spectra and absorbance values are normalized to a 10 mm pathlength equivalent.

Reported Values

- **Protein concentration.** Reported in selected unit (mg/mL or $\mu\text{g/mL}$). Calculations are based on Beer-Lambert equation using corrected protein absorbance value.
- **Method.** Determines the extinction coefficient used in conjunction with Beer's Law to calculate sample concentration.
- **Baseline correction.** Wavelength selected for baseline correction and the absorbance detected at that wavelength.
- **Location.** Displays whether the measurement was taken from the pedestal or cuvette mode.
- **Monitored wavelength.** Enter an additional wavelength whose absorbance value you want included in the report.
- **Mass extinction coefficient (1% solution).**
- **Stirrer.** "Off" will be displayed when the stirrer function on a cuvette model is not used. When the stirrer function is used, the stir speed is displayed.
- **Heater.** "Off" or "On" will be displayed to show if the cuvette port was heated during the measurement.
- **Measured Temperature.** Temperature of the cuvette port during the measurement will be displayed.
- **Target Temperature.** Desired temperature of the cuvette port will be displayed.

This page is intentionally blank.

Proteins and Labels

Measures the concentration of purified proteins that have been labeled with up to two fluorescent dyes.

[Measure Proteins and Labels](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)

[Calculations](#)



Measure Proteins and Labels

Use the Proteins and Labels application to quantify proteins and fluorescent dyes for protein array conjugates, as well as metalloproteins such as hemoglobin, using wavelength ratios. This application reports protein concentration measured at 280 nm, an A260/A280 absorbance ratio, and the concentrations and measured absorbance values of the dyes, allowing detection of dye concentrations as low as 0.2 picomole per microliter. This information is useful for evaluating protein/dye conjugation (degree of labeling) for use in downstream applications.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **Proteins** tab, then select **Protein & Labels**.

2. Configure any of the [setup options](#) if desired, then select **Save**.

Proteins & Labels Setup

Auto Naming

Sample type:

Sloping dye correction 400-850 nm

Analysis Correction nm (Default value is 340nm)

Dye 1:

Dye 2:

Tips: A general reference setting based on a 0.1% (1 mg/mL) protein solution producing an Absorbance at 280 nm of 1.0 A (where the pathlength is 10 mm or 1 cm)

Tip: Select a dye from the pre-defined list or add a custom dye using the [Dye/Chromophore Editor](#).


3. If using a NanoDrop Ultra^C or NanoDrop Ultra^C FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Pipette 1–2 μL of the blanking solution onto the lower pedestal and lower the arm, or insert the blanking cuvette into the cuvette holder.

Tip: If using a cuvette, make sure to [align the cuvette light path](#) with the instrument light path.
5. Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements.)
6. Lift the arm and clean both pedestals with a new laboratory wipe, or remove the blanking cuvette.
7. Pipette 2 μL sample solution onto the pedestal and lower the arm, or insert the sample cuvette into the cuvette holder.
8. Start the sample measurement:
 - Pedestal: If [Auto-Measure](#) is On, lower arm; if Auto-Measure is Off, lower arm and select **Measure**.
 - Cuvette: select **Measure**.

When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

9. When you are finished measuring samples, select **End Experiment**.
10. Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.




Related Topics

- [Best practices for protein measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

Proteins & Labels measurement screen (local control).

For each measured sample, this application shows the absorbance spectrum and a summary of the results. Below is an example of the measurement screen of the NanoDrop Ultra local control software:



Note

- A baseline correction is performed at 850 nm (absorbance value at 850 nm is subtracted from absorbance values at all wavelengths in sample spectrum).
- Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

Proteins & Labels measurement screen (PC control).

For each measured sample, this application shows the absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See “[Measurement Screen Display Options](#)” on [page 299](#) for an example.

Reported values

The initial screen that appears after each measurement (see previous image) shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:

#:	1	Dy2 DoL:	0.164
Date:	10/28/2024 12:29:53 PM	Sample Type:	BSA
Sample Name:	Sample 1	Sloping Dye Correction 400 - 850 nm:	ON
mg/mL:	1.657	Analysis correction:	0.681A@340nm
A280:	1.92	Location:	Pedestal
A260/A280:	1.69	MW(kDa):	-
Dye1:	Alexa Fluor 647	$\epsilon/1000$:	-
Dye 1 Concentration:	5.613pmol/ μ L	Ext.Coeff.E1%/gm-cm:	6.7
Dye 1 Absorbance:	1.52	Pathlength Used:	1 mm
Dy1 DoL:	0.217	Monitor Wavelength:	-
Dye2:	Alexa Fluor 660	Stirrer:	OFF
Dye 2 Concentration:	0.429pmol/ μ L	Heater:	OFF
Dye 2 Absorbance:	0.06	Measured Temperature:	-

Reported values for Proteins & Labels application

- # (sample number)
- date (date and time sample measurement was taken)
- [sample name](#)
- [protein concentration](#)
- [location](#)
- A280
- [A260/A280](#)
- Dye name, [concentration](#), and [absorbance](#)
- [degree of labeling](#)
- [Monitored wavelength](#)
- [Pathlength used](#)

5 Protein Applications


Proteins and Labels


- mass extinction coefficient (1% solution)
- molar extinction coefficient
- molecular weight (kDA)
- sample type
- Analysis Correction
- Sloping dye correction
- stirrer
- heater
- measured temperature
- target temperature

Related Topics

- Basic Instrument Operations
- Proteins & Labels calculations

Settings

The Protein & Labels Setup screen appears after you select the application from the Proteins tab on the home screen. To show the Proteins & Labels settings from the Proteins & Labels measurement screen, select .



The screenshot shows the Protein & Labels Setup screen with the following settings:

- Auto Naming
- Sample type: 1 Abs = 1mg/mL
- Sloping dye correction 400-850 nm
- Analysis Correction 340 nm (Default value is 340nm)
- Dye 1: None (unit: pmol/μL)
- Dye 2: None (unit: pmol/μL)

Tips: A general reference setting based on a 0.1% (1 mg/mL) protein solution producing an Absorbance at 280 nm of 1.0 A (where the pathlength is 10 mm or 1 cm)

Setting	Available Options	Mass Ext. Coefficient (L/gm-cm)	Description
Sample type ^a	1 Abs = 1 mg/mL	General reference	Select Sample type for detailed description of each available setting.
	BSA	6.7	Each sample type applies a unique extinction coefficient to the protein calculations. If the extinction coefficient of the sample is known, choose the ε + MW (molar) or ε1% (mass) option and enter the value. Otherwise, calculate the extinction coefficient or choose the option that best matches the sample solution. If you only need a rough estimate of protein concentration and the sample extinction coefficient is unknown, select the 1 Abs=1 mg/mL sample type option.
	IgG	13.7	
	Lysozyme	26.4	
	Other protein (ε + MW)	user-entered molar extinction coefficient/ molecular weight	
Other protein (ε1%)	User entered mass extinction coefficient		
			Tip: Ideally, the extinction coefficient should be determined empirically using a solution of the study protein at a known concentration using the same buffer.
Analysis Correction ^b	On or Off	N/A	Corrects sample absorbance measurement for any offset caused by light scattering particulates by subtracting absorbance value at specified analysis correction wavelength from absorbance value at analysis wavelength. Corrected value is used to calculate sample concentration.
	Enter analysis correction wavelength in nm or use default value (340 nm)		
			Tip: If the sample has a modification that absorbs light at 340 nm, select a different correction wavelength or turn off Analysis Correction.
Auto Naming	On or Off	N/A	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.

5 Protein Applications

Proteins and Labels

Setting	Available Options	Mass Ext. Coefficient (L/gm-cm)	Description
Dye 1/Dye 2 Type ^c	Cy3, 5, 3.5, or 5.5, Alexa Fluor 488, 546, 555, 594, 647, or 660	See Dye/Chromophore Editor for specific values for each dye	Select pre-defined dye used to label sample material, or one that has been added using Dye/Chrom. Editor.
Dye 1/Dye 2 Unit	picomoles/microliter (pmol/μL), micromoles (uM), or millimoles (mM)	not applicable	Select unit for reporting dye concentrations.
Sloping Dye Correction ^d	On or Off		Corrects dye absorbance measurements for any offset caused by light scattering particulates by subtracting absorbance value of a sloping baseline from 400 nm to 850 nm from absorbance value at dye's analysis wavelength.

^a To add or edit a custom protein, use [Protein Editor](#).

^b Analysis Correction affects calculation for protein concentration only.

^c To add custom dye or edit list of available dyes, use [Dye/Chromophore Editor](#).

^d Sloping Dye Correction affects calculations for dye concentration only.

Related Topics

- [Instrument Settings](#)
- [Protein Editor](#)
- [Dye/Chromophore Editor](#)

Detection Limits

Detection limits and reproducibility specifications for purified BSA proteins and dyes that are pre-defined in the software are provided [here](#). The BSA lower detection limit and reproducibility values apply to any protein sample type. The upper detection limits are dependent on the [upper absorbance limit](#) of the instrument and the sample's extinction coefficient.

To calculate upper detection limits for other (non-BSA) protein sample types

To calculate upper detection limits in mg/mL for proteins, use the following equation:

$$(\text{upper absorbance limit}_{\text{instrument}} / \text{mass extinction coefficient}_{\text{sample}}) * 10$$

For example, if the sample's mass extinction coefficient at 280 nm is 6.7 for a 1% (10 mg/mL) solution, the equation looks like this:

$$(550 / 6.7) * 10 = 820.9 \text{ (or } \sim 820)$$

Related Topics

- [Detection Limits for All Applications](#)

Calculations

As with the other protein applications, Proteins & Labels uses the [Beer-Lambert equation](#) to correlate absorbance with concentration based on the sample's extinction coefficient and pathlength.

This application offers six options (shown at right) for selecting an appropriate extinction coefficient for each measured sample, to be used in conjunction with Beer's Law to calculate sample concentration.

If the extinction coefficient of the sample is known, choose the ϵ + MW (molar) or ϵ 1% (mass) option and enter the value. Otherwise, calculate the extinction coefficient or choose the option that best matches the sample solution.

Tip: Ideally, the extinction coefficient should be determined empirically using a solution of the study protein at a known concentration using the same buffer.

Available Options for Extinction Coefficient

- **1 Abs = 1 mg/mL**, where sample type and/or ext. coefficient is unknown (produces rough estimate of protein concentration)
- **BSA** (Bovine Serum Albumin, 6.7 L/gm-cm)
- **IgG** (any mammalian antibody, 13.7 L/gm-cm)
- **Lysozyme** (egg white lysozyme, 26.4 L/gm-cm)
- **Other protein** (ϵ + MW), user-specified molar ext. coefficient
- **Other protein** (ϵ 1%), user-specified mass ext. coefficient

Note: See [Sample Type](#) for details.

5 Protein Applications

Proteins and Labels

Calculated protein concentrations are based on the absorbance value at 280 nm, the selected (or entered) extinction coefficient and the sample pathlength. A single-point baseline correction (or analysis correction) may be applied.

Concentration is reported in mass units. Calculators are available on the Internet to convert concentration from mass to molar units based on sample sequence.

Dye concentrations are calculated from the absorbance value at the dye's analysis wavelength, the dye's extinction coefficient, and the sample pathlength. A sloped-line dye correction may also be used.

Measured Values

A280 absorbance

Note: The absorbance value at 850 nm is subtracted from all wavelengths in the spectrum. As a result, the absorbance at 850 nm is zero in the displayed spectra. Also, for micro-volume absorbance measurements and measurements taken with nonstandard (other than 10 mm) cuvettes, the spectra are normalized to a 10 mm pathlength equivalent.

- Protein absorbance values are measured at 280 nm using the 850 nm-corrected and normalized spectrum. If Analysis Correction and Dye Correction are not selected, this is the reported A280 value and the value used to calculate protein concentration.
- If [Analysis Correction](#) is selected, the 850-corrected, normalized and analysis-corrected absorbance value at 280 nm is reported and used to calculate protein concentration.
- If a Dye is used, the 850-corrected, normalized, analysis-corrected and [dye-corrected](#) absorbance value at 280 nm is reported and used to calculate protein concentration.

Dye absorbance

- Dye absorbance values are measured at specific wavelengths. See [Dye/Chromophore Editor](#) for analysis wavelengths used.
- If Sloping Dye Correction is selected, a linear baseline is drawn between 400 nm and 850 nm and, for each dye, the absorbance value of the sloping baseline is subtracted from the absorbance value at each dye's analysis wavelength. Baseline-corrected dye absorbance values are reported and used to calculate dye concentrations.

Dye correction

- Pre-defined dyes have known correction values for A260 and A280. See [Dye/Chromophore Editor](#) for correction values used.
- A280 dye correction is subtracted from [A280 absorbance value](#) used to calculate protein concentration.

Sample Pathlength

- For micro-volume measurements, the software selects the optimal pathlength (between 1.0 mm and 0.03 mm) based on sample absorbance at the analysis wavelength.
- For cuvette measurements, the pathlength selected after switching to cuvette mode is utilized (see [Cuvette Settings](#)).
- Displayed spectra and absorbance values are normalized to a 10 mm pathlength equivalent.

Reported Values

- **Location.** Displays whether the measurement was taken from the pedestal.
- **Monitored wavelength.** Enter an additional wavelength whose absorbance value you want included in the report.
- **Protein concentration.** Reported in selected unit (mg/mL or $\mu\text{g/mL}$). Calculations are based on Beer-Lambert equation using corrected protein absorbance value.
- **Dye1/Dye2 concentration.** Reported in pmol/ μL . Calculations are based on Beer's Law equation using (sloping) baseline-corrected dye absorbance value(s).
- **Dye1/Dye 2 degree of labeling.** Specifies the average number of fluorophore molecules per molecule of conjugate.
- **Stirrer.** "Off" will be displayed when the stirrer function on a cuvette model is not used. When the stirrer function is used, the stir speed is displayed.
- **Heater.** "Off" or "On" will be displayed to show if the cuvette port was heated during the measurement.
- **Measured Temperature.** Temperature of the cuvette port during the measurement will be displayed.
- **Target Temperature.** Desired temperature of the cuvette port will be displayed.

Related Topics

- [Beer-Lambert Equation](#)
- [Protein A280 Calculations](#)

This page is intentionally blank.

Protein BCA

Measures total protein concentration of unpurified protein samples using a bicinchoninic acid colorimetric detection reagent.

[Measure Protein BCA](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)



Theory of Protein BCA assay

The Protein BCA assay uses bicinchoninic acid (BCA) as the detection reagent for Cu^{+1} , which is formed when Cu^{+2} is reduced by certain proteins in an alkaline environment. A purple reaction product is formed by the chelation of two molecules of BCA with one cuprous ion (Cu^{+1}). The resulting Cu-BCA chelate formed in the presence of protein is measured at 562 nm and baseline-corrected using the absorbance value at 750 nm. Pre-formulated kits of BCA reagent and CuSO_4 are available from us or a local distributor.

Protein assay kits and protocols

Please refer to the NanoDrop website for up-to-date kits and protocols for the NanoDrop Ultra instruments. Follow the assay kit manufacturer's recommendations for all standards and samples (unknowns). Ensure each is subjected to the same timing and temperature throughout the assay.

Protein standards for generating a standard curve may also be provided by the kit manufacturer. Since the NanoDrop Ultra pedestals can measure higher protein concentrations than traditional cuvette-based spectrophotometers, you may need to supply your own protein standards at higher concentrations than provided by the manufacturer. For example, additional standards may be required to ensure the standard curve covers the dynamic range of the assay and the expected range of the unknown samples.

Working with standard curves


A standard curve is required for colorimetric protein analysis.

- Each experiment requires a standard curve. You can run a new standard curve or import standard from a previously run experiment.
- To import a previously run standard curve, from the application setup screen, select the **Load Standards** option, select one of the available previously run standard curves and select **Save**.
- Prepare standards and unknown samples the same way. See the kit manufacturer's guidelines and recommendations.
 - **All reference and standards solutions** should be the same buffer used to resuspend the samples plus the same volume of reagent added to the samples.
 - **First standard** is a reference measurement. The reference solution should contain none of the analyte of interest. (The reference measurement is not the same as a blank measurement. This application requires both.)
 - **Concentration range of the standards** must cover the dynamic range of the assay and the expected range of the unknown samples. Sample analyte concentrations are not extrapolated beyond the concentration of the highest standard.
- Use the **Setup Standards** option in the application setup screen to enter concentration values for the standards and to specify how standards and samples will be measured (number of replicates, curve type, etc.).
 - Depending on the [Curve Type](#) setting, a standard curve can be generated using two or more standards.
 - The software **requires one reference measurement** and allows **up to 7 standards**.
 - **Concentration values for standards** can be entered in any order but the standards must be measured in the order in which they were entered; however, best practice dictates that standards be measured from the lowest concentration of the standard analyte stock to the highest.
- For all colorimetric assays except Protein Pierce 660, **blank the instrument** with DI H₂O (deionized water). For Protein Pierce 660, blank with the reference solution (see below).
- **Measure the reference and all standards** before you start analyzing samples. After the first sample has been measured, no additional changes are allowed to the standard curve.


As you measure the standards, a measurement screen appears, similar to the measurement screens for samples.

When using the PC control software, select the **Curve** tab to see the standard curve as you build it.

The R^2 value indicates how well the standard curve fits the standard data points (1.0 is a perfect fit; all points lie exactly on the curve).



After the minimum number of standards has been measured for the selected curve type, a message similar to the following appears:



Remeasure standards: returns to standard measurement screen, where previously run standards can be selected and rerun.


- From standards measurement screen, standard curve screen, or standards data table, press and hold the row to show **Sample Details** box.
- Select **Remeasure** then select **Yes** to confirm.

Load more standards: returns to the setup screen where you can add or edit the concentration value for any standard and then select **Save** to measure. This option will be unavailable if you have already measured the maximum allowable amount of standards.


Measure samples: continues to sample measurement screen, after which standards can no longer be edited.

Note You can add, edit or delete a standard any time before selecting **Measure Samples**.


Add standard:

- From standards measurement screen, select .
- Select the next empty Concentration field and enter the concentration value for the new standard.
- Select **Save**.

Edit standard:

- From standards measurement screen, select .
- Select the Concentration field and edit the concentration value.
- Select **Save**.

Delete standard:

- From standards measurement screen, standard curve screen, or standards data table, press and hold the row to show Standard Details box.
- Select .
- Select **Yes** to confirm.

The standard no longer appears in the table on the measurement screen and its concentration value no longer appears on the setup screen.

Note You can use this method to delete the reference measurement; however, a new reference must be measured immediately afterwards.

- After all standards have been measured for the selected curve type, the message "Standards Completed" will appear. If the message "Invalid Standards" appears after all entered standards have been measured, try:
 - selecting a different curve type
 - remeasuring standards using the correct standard material

Note This message is only an indicator that the required minimum number of points has been established for the selected curve type. It does not validate the integrity of the curve. For example, additional standards may be required to cover the expected assay concentration range.

Measure Protein BCA

The Protein BCA assay uses bicinchoninic acid as a colorimetric detection reagent to determine total protein concentration in unpurified protein samples. This application is useful for measuring dilute protein solutions or proteins in the presence of components that exhibit significant absorbance between 200 nm and 280 nm, which rules out direct protein measurements at 280 nm or 205 nm. This application measures absorbance at 562 nm and uses a standard curve to calculate protein concentration. A single-point baseline correction is applied.

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

Procedure

1. From the home screen, select the **Proteins** tab, then select **Protein BCA**.
2. Configure any of the [setup options](#) if desired and select **Save**. (Specify a [curve type](#) and number of [replicates for each standard](#) and enter the [concentration of each standard](#).)

Tip: For this assay, we recommend setting **Curve Type** to “Linear”.

3. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Measure blank:
 - Pipette 2 μL of DI H₂O onto lower pedestal and lower arm, or insert DI H₂O blanking cuvette into cuvette holder

Tip: If using a cuvette, make sure to [align cuvette light path](#) with instrument light path.

- Select **Blank** and wait for measurement to complete.
Tip: If **Auto-Blank** is On, the blank measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements).
 - Lift arm and clean both pedestals with new laboratory wipe, or remove cuvette.
5. Measure reference standard:
- Pipette 2 μ L of reference solution onto pedestal, or insert reference cuvette (reference solution should contain none of the standard protein stock, see [Working With Standard Curves](#) for details).
 - Select **Measure** and wait for measurement to complete.
Tip: If **Auto-Measure** is On, the standard measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements).
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
6. Measure remaining standards:
- Pipette 2 μ L of standard 1 onto pedestal, or insert standard 1 cuvette.
 - Select **Measure** and wait for measurement to complete.
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
 - Repeat substeps above for each additional standard (when specified number of standards and replicates have been measured, a message asks whether to remeasure standards, load more standards or begin measuring samples).
 - If finished measuring standards, select **Measure Samples** (Swipe left to view standard curve) when using the local control software or select the Curve tab when using the PC control software.
7. Measure samples:
- Pipette 2 μ L sample 1 onto pedestal, or insert sample 1 cuvette.
 - Select **Measure Samples** and wait for measurement to complete.
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.

8. When you are finished measuring samples, select **End Experiment**.
9. Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.

Related Topics

- [Working with standard curves](#)
- [Best practices for protein measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

Protein BCA measurement screen (local control software)

For each measured sample and standard, this application shows the visible absorbance spectrum and a summary of the results. When using the NanoDrop Ultra local control software, the standard curve is available by swiping left from the measurement screen (or in [History](#)).

Tap to end experiment

Settings

Protein concentration

Number of replicates

Absorbance at 562 nm

Tap row to select sample and update spectrum; tap more rows to overlay up to sixteen spectra. Press and hold sample row to view measurement details.

Drag tab down/up to see more/less sample data

Auto-Measure

Swipe screen left to view standard curve

#	mg/mL	# of Replicates	A562
Std 4	0.750	3	
Std 5	1.000	3	
Std 6	1.500	3	
Std 7	2.000	3	
1	0.255		0.318
2	0.544		0.65
3	0.816		0.962

Note

- A baseline correction is performed at 750 nm (absorbance value at 750 nm is subtracted from absorbance values at all wavelengths in sample spectrum).
- Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.


Protein BCA measurement screen (PC control software)

For each measured sample and standard, this application shows the visible absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See “Measurement Screen Display Options” on page 299 for an example.

Protein BCA standard curve screen

The standard curve screen shows graphically the relationship between the measured standards, the calculated standard curve, and the measured absorbance and calculated concentration for a selected sample. A horizontal line connects the sample absorbance value on the Y-axis to the standard curve. A vertical line connects that point to the sample concentration value on the X-axis.

The R^2 value indicates how well the standard curve fits the standard data points (1.0 is a perfect fit; that is, all points lie exactly on the curve).



When using the NanoDrop Ultra PC control software, select **Curve** to view the Standard curve.

5 Protein Applications

Protein BCA

Reported values

The initial screen that appears after each measurement and the standards screen (see previous image) show a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:





Sample Details			
#:	2	Equation:	-
Date:	10/30/2024 08:26:36 PM	Stirrer:	OFF
Sample Name:	Sample 2	Heater:	OFF
mg/mL:	0.544	Measured Temperature:	-
# of Replicates:	-	Target Temperature:	-
A562:	0.65		
#1 A562:	-		
#2 A562:	-		
#3 A562:	-		
Average A562:	-		
Baseline Correction:	0.242A@750nm		
Location:	Pedestal		
Monitor Wavelength:	-		

Related Topics

- [Example standard curve](#)
- [Basic Instrument Operations](#)
- [Protein A280 Calculations](#)

Settings

The Protein BCA Setup screen appears after you select the application from the Proteins tab on the home screen. To show the Protein BCA settings, from the Protein BCA measurement screen, select the setup icon .



Note You can edit the Curve Type setting when measuring standards by changing the list box at the top of the application measurement screen. You can edit the concentration value for a standard from the application setup screen. After selecting **Measure Samples**, these settings cannot be changed.

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Baseline Correction	"On or off Enter baseline correction wavelength in nm or use default value (340 nm)	Optional user-defined baseline correction. When enabled, the software corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength.

Setting	Available Options	Description
Curve Type		<p>Specify type of equation used to create standard curve from standard concentration values.</p> <p>Available options:</p> <ul style="list-style-type: none">– Linear: Draws the linear least squares line through all measured standards (requires reference measurement and at least one standard)– Interpolation: Draws a series of straight lines to connect all measured standards (requires reference measurement and at least one standard)– 2nd order polynomial: Draws the 2nd order least squares polynomial using all measured standards (requires reference measurement and at least two standards)– 3rd order polynomial: Draws the 3rd order least squares polynomial using all measured standards (requires reference measurement and at least three standards)
Replicates		<p>Enter number of measurements of the reference or the same standard or sample that are averaged together to produce its associated concentration value.</p> <p>Note: Replicates setting cannot be changed after the first standard has been measured.</p>
Standards		<p>Enter actual concentration value of each standard.</p> <p>Note: Concentration values can be entered in any order but the standards must be measured in the order they were entered.</p> <p>If you also want to use previously measured standards, select Load Standards.</p>

Protein Bradford

Measures total protein concentration of unpurified protein samples using a Coomassie Blue dye colorimetric detection reagent.

[Measure Protein Bradford](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)



Theory of Protein Bradford assay

The Protein Bradford assay uses the protein-induced absorbance shift of Coomassie Blue dye to determine total protein concentration. The bound protein-dye complex is measured at 595 nm and baseline-corrected using the absorbance value at 750 nm. Pre-formulated kits of stabilized reagent mixture containing Coomassie Blue dye, alcohol, and surfactant are available from us or a local distributor.

To maximize reliability with the Protein Bradford assay:

- **Work quickly and do not allow prepared standards or samples to sit longer than necessary.** Coomassie dye-dye and Coomassie dye-protein aggregates can form particulates with increasing development time, resulting in significant fluctuations in absorbance readings.
- **Measure standards and samples in triplicate** using a new aliquot for each measurement. For pedestal measurements, the total analyte (protein-dye) signal at 595 nm is limited to ~0.0150A due to the pedestal's 1.0 mm pathlength, the Coomassie dye concentration, and the acidic pH.

Note If you have a NanoDrop Ultra^C or a NanoDrop Ultra^C FL model instrument, using the cuvette option will result in a higher absorbance signal.

Protein assay kits and protocols

Please refer to the NanoDrop website for up-to-date kits and protocols for the NanoDrop Ultra instruments. Follow the assay kit manufacturer's recommendations for all standards and samples (unknowns). Ensure each is subjected to the same timing and temperature throughout the assay.

Protein standards for generating a standard curve may also be provided by the kit manufacturer. Since the NanoDrop Ultra pedestals can measure higher protein concentrations than traditional cuvette-based spectrophotometers, you may need to supply your own protein standards at higher concentrations than provided by the manufacturer. For example, additional standards may be required to ensure the standard curve covers the dynamic range of the assay and the expected range of the unknown samples.

Protein Bradford standard curves

A standard curve is required for colorimetric protein analysis.

- Each experiment requires a standard curve. You can run a new standard curve by selecting the **Setup standards** option or import standard from a previously run experiment by selecting **Load standards** on the application setup screen.

For more information, see [Working with Standard Curves](#).

Measure Protein Bradford

The Protein Bradford assay uses Coomassie Blue dye as a colorimetric detection reagent to determine total protein concentration in unpurified protein samples. This application is useful for measuring dilute protein solutions that require lower detection sensitivity or proteins in the presence of components that exhibit significant absorbance between 200 nm and 280 nm, which rules out direct protein measurements at 280 nm or 205 nm. This application measures absorbance at 595 nm and uses a standard curve to calculate protein concentration. See [Working with Standard Curves](#) for more information. A single-point baseline correction is applied.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **Proteins** tab, then select **Protein Bradford**.
2. Configure any of the [setup options](#) if desired and select **Save**. (Specify a [curve type](#) and number of [replicates for each standard](#) and enter the [concentration of each standard](#).)
Tip: For this assay, set **Curve Type** to “2nd Order Polynomial” and **Replicates** to 3.
3. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Measure blank:
 - Pipette 2 µL DI H₂O of onto lower pedestal and lower arm, or insert DI H₂O blanking cuvette into cuvette holder.
Tip: If using a cuvette, make sure to [align cuvette light path](#) with instrument light path.
 - Select **Blank** and wait for measurement to complete.
Tip: If [Auto-Measure](#) is On, the standard measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements).
 - Lift arm and clean both pedestals with new laboratory wipe, or remove cuvette.
5. Measure reference standard:
 - Pipette 2 µL of reference solution onto pedestal, or insert reference cuvette (reference solution should contain none of the standard protein stock, see [Working With Standard Curves](#) for details).
 - Select **Measure** and wait for measurement to complete.
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
6. Measure remaining standards:
 - Pipette 2 µL of standard 1 onto pedestal, or insert standard 1 cuvette.
 - Select **Measure** and wait for measurement to complete.

- Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
 - Repeat substeps above for each additional standard (when specified number of standards and replicates have been measured, a message asks whether to remeasure standards, load more standards, or begin measuring samples).
 - If finished measuring standards, select **Measure Samples** (Swipe left to view standard curve) when using the local control software or select the Curve tab when using the PC control software.
7. Measure samples:
- Pipette 2 μ L of sample 1 onto pedestal, or insert sample 1 cuvette.
 - Select **Measure samples** and wait for measurement to complete.
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
8. When you are finished measuring samples, select **End Experiment**.
9. Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.


Related Topics

- [Working with standard curves](#)
- [Best practices for protein measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

Protein Bradford measurement screen (local control software)

For each measured sample and standard, this application shows the visible absorbance spectrum and a summary of the results. When using the NanoDrop Ultra local control software, the standard curve is also available by swiping left from the measurement screen (or in [History](#)).



Note

- A baseline correction is performed at 750 nm (absorbance value at 750 nm is subtracted from absorbance values at all wavelengths in sample spectrum).
- Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.


Protein Bradford measurement screen (PC control)

For each measured sample and standard, this application shows the visible absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.

Protein Bradford standard curve screen

The standard curve screen shows graphically the relationship between the measured standards, the calculated standard curve, and the measured absorbance and calculated concentration for a selected sample. A horizontal line connects the sample absorbance value on the Y-axis to the standard curve. A vertical line connects that point to the sample concentration value on the X-axis.

The R^2 value indicates how well the standard curve fits the standard data points (1.0 is a perfect fit; that is, all points lie exactly on the curve).



When using the NanoDrop Ultra PC control software, select **Curve** to view the Standard curve.

Note

- A baseline correction is performed at 750 nm (absorbance value at 750 nm is subtracted from absorbance values at all wavelengths in sample spectrum).
- Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

Reported values


The initial screen that appears after each measurement and the standards screen (see previous image) show a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:


#:	3	Equation:	-
Date:	10/30/2024 03:15:02 PM	Stirrer:	OFF
Sample Name:	Sample 3	Heater:	OFF
mg/mL:	0.525	Measured Temperature:	-
# of Replicates:	-	Target Temperature:	-
A595:	0.864		
#1 A595:	-		
#2 A595:	-		
#3 A595:	-		
Average A595:	-		
Baseline Correction:	0.243A@750nm		
Location:	Pedestal		
Monitor Wavelength:	-		

Related Topics

- [Example standard curve](#)
- [Basic Instrument Operations](#)
- [Protein A280 Calculations](#)

Settings

The Protein Bradford Setup screen appears after you select the application from the Proteins tab on the home screen. To show the Protein Bradford settings from the Protein Bradford measurement screen, select the setup icon .



Note You can edit the Curve Type setting when measuring standards by changing the list box at the top of the application measurement screen. You can edit the concentration value for a standard from the application setup screen. After selecting **Measure Samples**, these settings cannot be changed.

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Baseline Correction	"On or off Enter baseline correction wavelength in nm or use default value (340 nm)	Optional user-defined baseline correction. When enabled, the software corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength.

Setting	Available Options	Description
Curve Type		<p>Specify type of equation used to create standard curve from standard concentration values. Available options:</p> <ul style="list-style-type: none"> – Linear: Draws the linear least squares line through all measured standards (requires reference measurement and at least one standard) – Interpolation: Draws a series of straight lines to connect all measured standards (requires reference measurement and at least one standard) – 2nd order polynomial: Draws the 2nd order least squares polynomial using all measured standards (requires reference measurement and at least two standards) – 3rd order polynomial: Draws the 3rd order least squares polynomial using all measured standards (requires reference measurement and at least three standards)
Replicates		<p>Enter number of measurements of the reference or the same standard or sample that are averaged together to produce its associated concentration value.</p> <p>Note: Replicates setting cannot be changed after the first standard has been measured.</p>
Standards		<p>Enter actual concentration value of each standard.</p> <p>Note: Concentration values can be entered in any order but the standards must be measured in the order they were entered.</p> <p>If you also want to use previously measured standards, select Load Standards.</p>

This page is intentionally blank.

Protein Lowry

Measures total protein concentration of unpurified protein samples using a Folin-Ciocalteu colorimetric detection reagent.

[Measure Protein Lowry](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)



Theory of Protein Lowry assay

The Protein Lowry assay involves the reaction of protein with cupric sulfate in alkaline solution, resulting in the formation of tetradentate copper-protein complexes. The Folin-Ciocalteu reagent is effectively reduced in proportion to the chelated copper-complexes. The water-soluble blue reaction product is measured at 650 nm and baseline-corrected using the absorbance value at 405 nm. Pre-formulated kits of Folin-Ciocalteu reagent and CuSO_4 are available from us or a local distributor.

Protein assay kits and protocols

Follow the assay kit manufacturer's recommendations for all standards and samples (unknowns). Ensure each is subjected to the same timing and temperature throughout the assay.

Protein Lowry standard curves

A standard curve is required for colorimetric protein analysis.

- Each experiment requires a standard curve. You can run a new standard curve by selecting the **Setup standards** option or import standard from a previously run experiment by selecting **Load Standards** on the application setup screen.

For more information, see [Working with Standard Curves](#).

Measure Protein Lowry

The Protein Lowry assay uses Folin-Ciocalteu as a colorimetric detection reagent to determine total protein concentration in unpurified protein samples. This application is an alternative to the other colorimetric applications for measuring dilute protein solutions or proteins in the presence of components that exhibit significant absorbance between 200 nm and 280 nm. This application measures absorbance at 650 nm and uses a standard curve to calculate protein concentration. See [Working with Standard Curves](#) for more information. A single-point baseline correction is applied.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **Proteins** tab, then select **Protein Lowry**.
2. Configure any of the [setup options](#) if desired and select **Save**.

Tip: For this assay, we recommend setting **Curve Type** to “2nd Order Polynomial.”

3. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Measure blank:
 - Pipette 2 μ L DI H₂O onto lower pedestal and lower arm, or insert DI H₂O blanking cuvette into cuvette holder.

Tip: If using a cuvette, make sure to [align cuvette light path](#) with instrument light path.

- Select **Blank** and wait for measurement to complete.
Tip: If [Auto-Measure](#) is On, the standard measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements).
 - Lift arm and clean both pedestals with new laboratory wipe, or remove cuvette.
5. Measure reference standard:
- Pipette 2 μ L of reference solution onto pedestal, or insert reference cuvette (reference solution should contain none of the standard protein stock, see [Working With Standard Curves](#) for details).
 - Select **Measure** and wait for measurement to complete.
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
6. Measure remaining standards:
- Pipette 2 μ L standard 1 onto pedestal, or insert standard 1 cuvette.
 - Select **Measure** and wait for measurement to complete.
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
 - Repeat substeps above for each additional standard (when specified number of standards and replicates have been measured, a message asks whether to remeasure standards, load more standards, or begin measuring samples).
 - If finished measuring standards, select **Measure Samples**. (Swipe left to view standard curve) when using the local control software or select the Curve tab when using the PC control software.
7. Measure samples:
- Pipette 2 μ L of sample 1 onto pedestal, or insert sample 1 cuvette.
 - Select **Measure Samples** and wait for measurement to complete.
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
8. When you are finished measuring samples, select **End Experiment**.
9. Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.

Related Topics

- [Working with standard curves](#)

5 Protein Applications


Protein Lowry

- Best practices for protein measurements
- Measure a Micro-Volume Sample
- Measure a Sample Using a Cuvette
- Prepare Samples and Blanks
- Basic Instrument Operations

Reported Results

Protein Lowry measurement screen (local control software)

For each measured sample and standard, this application shows the visible absorbance spectrum and a summary of the results. When using the NanoDrop Ultra local control software, the standard curve is available by swiping left from the measurement screen (or in [History](#)).



Note

- A baseline correction is performed at 405 nm (absorbance value at 405 nm is subtracted from absorbance values at all wavelengths in sample spectrum).
- Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.


Protein Lowry measurement screen (PC control)

For each measured sample and standard, this application shows the visible absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.

Protein Lowry standard curve screen

The standard curve screen shows graphically the relationship between the measured standards, the calculated standard curve, and the measured absorbance and calculated concentration for a selected sample. A horizontal line connects the sample absorbance value on the Y-axis to the standard curve. A vertical line connects that point to the sample concentration value on the X-axis.

The R^2 value indicates how well the standard curve fits the standard data points (1.0 is a perfect fit; that is, all points lie exactly on the curve).



When using the NanoDrop Ultra PC control software, select **Curve** to view the Standard curve.

Reported values


The initial screen that appears after each measurement and the standards screen (see previous image) show a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:


#:	2	Equation:	-
Date:	10/30/2024 05:32:09 PM	Stirrer:	OFF
Sample Name:	Sample 2	Heater:	OFF
mg/mL:	0.552	Measured Temperature:	-
# of Replicates:	-	Target Temperature:	-
A650:	0.694		
#1 A650:	-		
#2 A650:	-		
#3 A650:	-		
Average A650:	-		
Baseline Correction:	0.613A@405nm		
Location:	Pedestal		
Monitor Wavelength:	-		

Related Topics

- [Example standard curve](#)
- [Basic Instrument Operations](#)
- [Protein A280 Calculations](#)

Settings

The Protein Lowry setup screen appears after you select the application from the Proteins tab on the home screen. To show the Protein Lowry settings, from the Protein Lowry measurement screen, select the setup icon .



Note You can edit the Curve Type setting when measuring standards by changing the list box at the top of the application measurement screen. You can edit the concentration value for a standard from the application setup screen. After selecting **Measure Samples**, these settings cannot be changed.

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Baseline Correction	"On or off Enter baseline correction wavelength in nm or use default value (340 nm)	Optional user-defined baseline correction. When enabled, the software corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength.

Setting	Available Options	Description
Curve Type		<p>Specify type of equation used to create standard curve from standard concentration values. Available options:</p> <ul style="list-style-type: none"> – Linear: Draws the linear least squares line through all measured standards (requires reference measurement and at least one standard) – Interpolation: Draws a series of straight lines to connect all measured standards (requires reference measurement and at least one standard) – 2nd order polynomial: Draws the 2nd order least squares polynomial using all measured standards (requires reference measurement and at least two standards) – 3rd order polynomial: Draws the 3rd order least squares polynomial using all measured standards (requires reference measurement and at least three standards)
Replicates		<p>Enter number of measurements of the reference or the same standard or sample that are averaged together to produce its associated concentration value.</p> <p>Note: Replicates setting cannot be changed after the first standard has been measured.</p>
Standards		<p>Enter actual concentration value of each standard.</p> <p>Note: Concentration values can be entered in any order but the standards must be measured in the order they were entered.</p> <p>If you also want to use previously measured standards, select Load Standards.</p>

This page is intentionally blank.

Protein Pierce 660

Measures total protein concentration of unpurified protein samples using a proprietary colorimetric detection reagent.

[Measure Protein Pierce 660](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)



Theory of Protein Pierce 660 assay

The Protein Pierce 660 assay is based on the binding of a proprietary dye-metal complex to protein in acidic conditions that causes a shift in the dye's absorption maximum, which is measured at 660 nm. The dye-metal complex is reddish-brown and changes to green upon protein binding. The color change is produced by deprotonation of the dye at low pH facilitated by interactions with positively charged amino acid groups in proteins. The dye interacts mainly with basic residues in proteins such as histidine, arginine and lysine and to a lesser extent tyrosine, tryptophan and phenylalanine. The reaction product is measured at 660 nm and baseline-corrected using the absorbance value at 750 nm.

The color produced in the assay is stable and increases in proportion to a broad range of increasing protein concentrations. An optional Ionic Detergent Compatibility Reagent (IDCR) may be added to the assay reagent to increase compatibility with high amounts of ionic detergents, including Laemmli SDS sample buffer with bromophenol blue. The IDCR dissolves completely by thorough mixing and has no effect on the assay. Pre-formulated kits of the protein binding material are available from us or a local distributor. For information about IDCR, refer to the kit manufacturer.

Protein assay kits and protocols

Please refer to the NanoDrop website for up-to-date kits and protocols for the NanoDrop Ultra instruments. Follow the assay kit manufacturer's recommendations for all standards and samples (unknowns). Ensure each is subjected to the same timing and temperature throughout the assay.

Protein standards for generating a standard curve may also be provided by the kit manufacturer. Since the NanoDrop Ultra pedestals can measure higher protein concentrations than traditional cuvette-based spectrophotometers, you may need to supply your own protein standards at higher concentrations than provided by the manufacturer. For example, additional standards may be required to ensure the standard curve covers the dynamic range of the assay and the expected range of the unknown samples.

Protein Pierce standard curves

A standard curve is required for colorimetric protein analysis.

- Each experiment requires a standard curve. You can run a new standard curve by selecting the **Setup standards** option or import standard from a previously run experiment by selecting **Load Standards** on the application setup screen.

For more information, see [Working with Standard Curves](#).

Measure Protein Pierce 660

The Protein Pierce 660 assay uses a proprietary protein binding material as a colorimetric detection reagent to determine total protein concentration in unpurified protein samples. This application is suitable for protein solutions that contain high concentrations of detergents, reducing agents and other commonly used reagents. The Pierce 660 application measures absorbance at 660 nm and uses a standard curve to calculate protein concentration (see [Working with Standard Curves](#) for more information). A single-point baseline correction is applied.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **Proteins** tab, then select **Protein Pierce 660**.
2. Configure any of the [setup options](#) if desired and select **Save**. (Specify a [curve type](#) and number of [replicates for each standard](#) and enter the [concentration of each standard](#).)
Tip: For this assay, we recommend setting **Curve Type** to “Linear”.
3. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Measure blank:
 - Pipette 2 μ L of reference solution onto lower pedestal and lower arm, or insert reference solution blanking cuvette into cuvette holder (reference solution should contain none of the standard protein stock; see [Working With Standard Curves](#) for details).
Tip: If using a cuvette, make sure to [align cuvette light path](#) with instrument light path.
 - Select **Blank** and wait for measurement to complete.
Tip: If [Auto-Measure](#) is On, the standard measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements).
 - Lift arm and clean both pedestals with new laboratory wipe, or remove cuvette.
5. Measure reference standard:
 - Pipette 2 μ L of reference solution onto pedestal, or insert reference cuvette (reference solution should contain none of the standard protein stock, see [Working With Standard Curves](#) for details).
 - Select **Measure** and wait for measurement to complete.
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
6. Measure remaining standards:
 - Pipette 2 μ L of standard 1 onto pedestal, or insert standard 1 cuvette.

- Select **Measure** and wait for measurement to complete.
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
 - Repeat substeps above for each additional standard (when specified number of standards and replicates have been measured, a message asks whether to remeasure standards, load more standards, or begin measuring samples).
 - If finished measuring standards, select **Measure Samples** (swipe left to view standard curve) when using the local control software or select the Curve tab when using the PC control software.
7. Measure samples:
- Pipette 2 μL of sample 1 onto pedestal, or insert sample 1 cuvette.
 - Select **Measure Samples** and wait for measurement to complete.
 - Lift arm and clean both pedestals with new wipe, or remove cuvette.
 - If Replicates setting is greater than 1, repeat measurement.
8. When you are finished measuring samples, select **End Experiment**.
9. Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.


Related Topics

- [Working with standard curves](#)
- [Best practices for protein measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

Protein Pierce 660 measurement screen (local control software)

For each measured sample and standard, this application shows the visible absorbance spectrum and a summary of the results. In the NanoDrop Ultra local control software, the standard curve is also available by swiping left from the measurement screen (or in [History](#)).



Note

- A baseline correction is performed at 750 nm (absorbance value at 750 nm is subtracted from absorbance values at all wavelengths in sample spectrum).
- Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

When using the NanoDrop Ultra PC control software, select **Curve** to view the Standard curve.


Protein Pierce 660 measurement screen (PC control)

For each measured sample and standard, this application shows the visible absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.

Protein Pierce 660 standard curve screen

The standard curve screen shows graphically the relationship between the measured standards, the calculated standard curve, and the measured absorbance and calculated concentration for a selected sample. A horizontal line connects the sample absorbance value on the Y-axis to the standard curve. A vertical line connects that point to the sample concentration value on the X-axis.

The R^2 value indicates how well the standard curve fits the standard data points (1.0 is a perfect fit; that is, all points lie exactly on the curve).



When using the NanoDrop Ultra PC control software, select **Curve** to view the Standard curve.

Reported values


The initial screen that appears after each measurement and the standards screen (see previous image) show a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:


#:	2	Equation:	-
Date:	10/30/2024 06:45:25 PM	Stirrer:	OFF
Sample Name:	Sample 2	Heater:	OFF
mg/mL:	0.488	Measured Temperature:	-
# of Replicates:	-	Target Temperature:	-
A660:	0.45		
#1 A660:	-		
#2 A660:	-		
#3 A660:	-		
Average A660:	-		
Baseline Correction:	0.008A@750nm		
Location:	Pedestal		
Monitor Wavelength:	-		

Related Topics

- [Example standard curve](#)
- [Basic Instrument Operations](#)
- [Protein A280 Calculations](#)

Settings

The Protein Pierce 660 setup screen appears after you select the application from the Proteins tab on the home screen. To show the Protein Pierce 660 settings, from the Protein Pierce 660 measurement screen, select the setup icon .



Note You can edit the Curve Type setting when measuring standards by changing the list box at the top of the application measurement screen. You can edit the concentration value for a standard from the application setup screen. After selecting **Measure Samples**, these settings cannot be changed.

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Baseline Correction	"On or off Enter baseline correction wavelength in nm or use default value (340 nm)	Optional user-defined baseline correction. When enabled, the software corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength.

Setting	Available Options	Description
Curve Type		<p>Specify type of equation used to create standard curve from standard concentration values. Available options:</p> <ul style="list-style-type: none"> – Linear: Draws the linear least squares line through all measured standards (requires reference measurement and at least one standard) – Interpolation: Draws a series of straight lines to connect all measured standards (requires reference measurement and at least one standard) – 2nd order polynomial: Draws the 2nd order least squares polynomial using all measured standards (requires reference measurement and at least two standards) – 3rd order polynomial: Draws the 3rd order least squares polynomial using all measured standards (requires reference measurement and at least three standards)
Replicates		<p>Enter number of measurements of the reference or the same standard or sample that are averaged together to produce its associated concentration value.</p> <p>Note: Replicates setting cannot be changed after the first standard has been measured.</p>
Standards		<p>Enter actual concentration value of each standard.</p> <p>Note: Concentration values can be entered in any order but the standards must be measured in the order in which they were entered.</p> <p>If you also want to use previously measured standards, select Load Standards.</p> <p>If you also want to enter previously measured absorbance values for the standards, select this check box:</p> <div data-bbox="760 1577 1289 1675" style="border: 1px solid black; padding: 5px; background-color: #f0f0f0;"> <input type="checkbox"/> Absorbance data for standards can be either measured or entered manually. Check the box to manually enter the absorbance values. </div> <p>and the enter absorbance values for all the standards.</p>

This page is intentionally blank.

Fluorescence Applications

dsDNA Fluorescence or RNA Fluorescence

Measure the concentration of double-stranded DNA (dsDNA) or RNA tagged with a fluorophore using fluorescence. These applications are designed for use with a NanoDrop Ultra Fluorescence Assay. These applications are only available for use on the NanoDrop Ultra FL and NanoDrop Ultra^c FL and are not compatible with the standard cuvette mode.



[Measure dsDNA Fluorescence or RNA Fluorescence](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)

[Calculations](#)

Best practices for fluorescence measurements

- For untested contaminating substances, and for highest accuracy, the standards should be assayed under the same conditions as the experimental samples.
- Isolate and purify nucleic acid samples before measurement to remove impurities. Depending on the sample, impurities could include DNA, RNA, free nucleotides, proteins, some buffer components and dyes. See [Preparing Samples](#) for more information.
- Ensure the sample concentration is within the selected assay's [limits](#).

- The NanoDrop Ultra Fluorescence Assays deliver optimal performance when all solutions are at room temperature (18-28 °C). Temperature fluctuations can influence the accuracy of the assays.
- Ensure pedestal surfaces are properly cleaned and conditioned.
- Vortex and centrifuge before taking a measurement. Avoid introducing bubbles when mixing and pipetting.
- Follow best practices for micro-volume measurements.
- Use a 2 µL sample volume. See [Recommended Sample Volumes](#) for more information.
- No data are currently available that address the mutagenicity or toxicity of the **NanoDrop™ Ultra dsDNA BR Fluorescence Reagents** (the dye in Component A). These reagents are known to bind to nucleic acids. Treat buffers with the same safety precautions as all other potential mutagens and dispose of the dye in accordance with local regulations.
- The **NanoDrop™ Ultra RNA BR Fluorescence Assay** includes a high-quality rRNA standard. It is crucial to maintain the integrity and concentration of these standards for optimal performance. Therefore, we highly recommend treating the rRNA standards with the same care as any other RNA. This includes:
 - Using appropriate RNase-free handling techniques, such as RNase-free gloves, pipette tips, and tubes.
 - It is important to keep the tube lids closed whenever possible.
 - Avoid touching the pipette to the inside wall of the tube when withdrawing a sample.
 - After use, it is advisable to promptly return the rRNA standard to the refrigerator.

Related Topics

- [Measure a Micro-Volume Sample](#)
- [Best Practices for Micro-Volume Measurements](#)
- [Basic Instrument Operations](#)

Working with standard curves


A standard curve is required for fluorescence analysis.

- Each experiment requires a new standard curve.

Note NanoDrop Ultra BR Fluorescence Assays require 2 standards.

- **Measure all standards** before you start analyzing samples. After the first sample has been measured, no additional changes are allowed to the standard curve.


As you measure the standards, a dialog screen appears, prompting the user to measure the remaining standards.



6 Fluorescence Applications

dsDNA Fluorescence or RNA Fluorescence

Swipe left one screen to see the standard curve as you build it. Here is an example:



When using the Fluorometer application, an R^2 value will appear and indicates how well the standard curve fits the standard data points (1.0 is a perfect fit; all points lie exactly on the curve).

Swipe left one screen to see the data table for the standards. Here is an example:

Tap to select which details appear in data table

#	Date	Sample Name	ng/µL	Original Conc.	Dilution Factor	Sample RFU	Std Avg RFU	# of Replicates	Excitation L
Std 1	11/06/2024 05:13:48 PM	Standard 1	0.0	0.000	10.0		1008.32	3	Blue
Std 2	11/06/2024 05:14:23 PM	Standard 2	10.0	100.000	10.0		2872.90	3	Blue

Tap to view details

Press and hold any row to view details

Measure End Experiment

Press and hold a row in any of the previous screens to view details about an individual standard. Here is an example:

Sample Details


#:	Std 2
Date:	11/06/2024 05:14:23 PM
Sample Name:	Standard 2
ng/µL:	10.0
Original Conc.:	100.000
Dilution Factor:	10.0
Sample RFU:	-
Std Avg RFU:	2872.90
# of Replicates:	3
#1 RFU:	2874.69
#2 RFU:	2880.55
#3 RFU:	2863.47
Excitation LED:	Blue

Select to print sample details

Select to delete and remeasure standard

Close Remeasure

After the both standards have been measured for the selected curve type, a message similar to the following appears:



Remeasure standards: returns to standard measurement screen, where previously run standards can be selected and rerun.

- From the standard curve screen or standards data table, press and hold the row of the standard that is to be remeasured to show the **Sample Details** box.
- Select **Remeasure**.

Load more standards: returns to the setup screen where you can add additional standards to measure. A total of 1 reference and 7 standards can be used. This option will not appear when using the dsDNA Fluorescence or RNA Fluorescence applications as 2 standards must be used in those applications.

Measure samples: continues to sample measurement screen.

Note After both standards have been measured, the message "Standards Completed" should appear. If the "Invalid Standards" message appears after all standards have been measured, try remeasuring standards using the correct standard material.

Measure dsDNA Fluorescence or RNA Fluorescence

Use the dsDNA Fluorescence and RNA Fluorescence applications paired with NanoDrop Ultra Fluorescence Assays to quantify dsDNA or RNA samples. These applications report nucleic acid concentration. The NanoDrop Ultra dsDNA Broad Range (BR), and RNA Broad Range (BR) Fluorescence Assays contain selective dyes that fluoresce only when bound to the target molecule.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#). See [Working with standard curves](#) for more information on the options available when creating a standard curve.

NOTICE


- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.



Sample and standard preparation

For instructions on the preparation of the assay standards and samples, see the instructions that accompany the assay you are using.

Measure nucleic acid standards and samples using fluorescence



1. From the home screen, select the Fluorescence tab and select **dsDNA Fluorescence** or **RNA Fluorescence**, depending on the samples to be measured.
2. Configure any of the [setup options](#) if desired and select **Save**.
3. Measure standards.

Note After the standards are measured the Sample to Reaction Volume option can be changed for each sample by selecting  at the top of the screen if desired.

- a. Pipette 2 μ L of the prepared assay tube containing Standard #1 onto the lower pedestal and lower the arm.
 - If **Auto-Measure** is On , the measurement will begin automatically after the arm is lowered.
 - If **Auto-Measure** is Off , lower arm and select **Measure** to begin the measurement.

Note Always use a fresh 2 μ L aliquot for each measurement.

- b. After the measurement is complete, remove sample from both upper and lower pedestal using a dry laboratory wipe.
- c. If Replicates setting is greater than 1, repeat measurement until all replicates are complete.
- d. Repeat sub-steps above for Standard #2 including additional replicate measurements.
- e. When both standards and all replicates have been measured, a message will appear with options to remeasure standards or begin measuring samples.
- f. If finished measuring standards, select **Measure Samples**.

4. Measure samples.
 - a. Enter or edit the sample ID at the top of the screen by selecting the box and inputting a new sample name.
 - b. Pipette 2 μL of the assay tube containing the corresponding sample onto the lower pedestal and lower the arm.
 - If **Auto-Measure** is On , the measurement will begin automatically after the arm is lowered.
 - If Auto-Measure is Off , lower arm and select **Measure** to begin the measurement.


Note Always use a fresh 2 μL aliquot for each measurement.

- c. When the sample measurement is completed, the reported values are displayed (see the next section).
 - d. After the measurement is complete, remove sample from both upper and lower pedestal using a dry laboratory wipe.
 - e. Repeat sub-steps to measure all remaining samples (following the on-screen prompts).
 - f. When you are finished measuring samples, select **End Experiment**.
5. Lift the arm and clean both pedestals with a new wipe.

Reported Results

dsDNA Fluorescence and RNA Fluorescence measurement screen (local control)

For each measured sample, the dsDNA Fluorescence and RNA Fluorescence applications show the fluorescence standard curve and a summary of the results. The following is an example of the measurement screen of the NanoDrop™ Ultra local control software screen as seen from History:




dsDNA Fluorescence and RNA Fluorescence measurement screen (PC control)

For each measured sample, the dsDNA Fluorescence and RNA Fluorescence applications show the fluorescence standard curve and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See “Measurement Screen Display Options” on page 299 for an example.

dsDNA Fluorescence and RNA Fluorescence reported values


In the local control software, the initial screen that appears after each measurement shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:




- # (sample number)
- date (date and time sample measurement was taken)
- [sample name](#)
- [nucleic acid concentration](#)
- [original concentration](#)
- [dilution factor](#)
- [sample RFU](#)
- [excitation LED](#)

Settings

The dsDNA or RNA Fluorescence Setup screen appears after you select the respective application from the Fluorescence tab on the home screen.

To show the dsDNA Fluorescence or RNA Fluorescence settings from the respective measurement screen, select the Settings icon, .



Setting	Description
Auto Naming	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Standards	This setting cannot be adjusted, all current assays require the use of 2 standards to generate the standard curve.
Replicates	Adjust the number of measurements of the same standard that are averaged together to produce its associated RFU value. Up to 5 replicate measurements can be utilized. Note: Replicates setting cannot be changed after the first standard has been measured.
Sample to Reaction Volume	Select from the drop-down menu of available options that reflect the amount of sample in comparison to the total reaction volume. Each assay has 2 available options to select from. See the assay protocol for more details on which option to select.

Calculations

A curve-fitting algorithm which requires two standards is used in the calculation of concentration data. The NanoDrop Ultra dsDNA and RNA fluorescence assays use a modified Hill plot to generate concentration data based on the relationship between the two standards used.

Measured Values

Sample RFU. Sample RFU refers to the measurement of Relative Fluorescence Units from the given sample, which quantifies the fluorescence intensity emitted by the sample.

RFU #1. The Relative Fluorescence Units measured from the first standard replicate.

RFU #2. The Relative Fluorescence Units measured from the second standard replicate.

RFU #3. The Relative Fluorescence Units measured from the third standard replicate.

RFU #4. The Relative Fluorescence Units measured from the fourth standard replicate.

RFU #5. The Relative Fluorescence Units measured from the fifth standard replicate.

Reported Values

Nucleic acid concentration.

Reported in selected unit (i.e., ng/ μ L, μ g/ μ L or μ g/mL). The sample RFU is used to extrapolate the concentration of the sample from a standard curve (modified hill plot).

Original nucleic acid concentration.

The measured concentration and dilution factor are used to calculate the concentration of the original sample before it was added to the assay.

Dilution factor.

Ratio of the volume of the original sample to the total volume in the assay tube.

Standard average RFU.

Average of the measured RFU for all replicate measurements of a standard.

of replicates.

Number of replicate measurements of each standard.

Excitation LED.

Excitation light source used during the measurement.

This page is intentionally blank.

Fluorometer

Choose the excitation light source (blue LED or red LED) while reading fluorescence in both the green and far-red emission channels (for the blue LED) or in the far-red emission channel only (for the red LED). This application is only available for use on the NanoDrop Ultra FL and NanoDrop Ultra^c FL and is not compatible with the standard cuvette mode.



[Measure Fluorophore-Labeled Samples](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)

Measure Fluorophore-Labeled Samples

Use the Fluorometer application to measure the fluorescence emitted by a variety of samples that have been labeled with fluorophore. This application functions similarly to the dsDNA Fluorescence and RNA Fluorescence applications, using a generated standard curve to determine concentration of the measured samples. This application allows for the use of additional assays outside of the NanoDrop Ultra BR Fluorescence Assays.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.


Sample and standard preparation



For instructions on the preparation of the assay standards and samples, see the instructions that accompany the assay you are using.

See section [Working with standard curves](#) for more information on the options available when creating a standard curve.

Measure fluorophore-labeled standards and samples using fluorescence



1. From the home screen, select the **Fluorescence** tab, then select **Fluorometer**.
2. Select from one of the excitation LED options (Blue or Red).



3. Configure any of the [setup options](#) if desired and select **Save**.
4. Measure standards:
 - a. Pipette 2 μ L of the prepared assay tube containing the Reference onto the lower pedestal and lower the arm. The reference should contain none of the analyte of interest.
 - If the [Auto-Measure](#) feature is OFF , select **Measure** to begin the measurement.
 - If the [Auto-Measure](#) feature is ON , the measurement will begin automatically after the arm is lowered.

Note Always use a fresh 2 μ L aliquot for each measurement.

- b. After the measurement is complete, remove sample from both upper and lower pedestal using a dry laboratory wipe.
- c. If Replicates setting is greater than 1, repeat measurement until all replicates are complete.
 - Repeat sub-steps above for Standard #1 and all remaining standards including additional replicate measurements of each.
 - When all standards and all replicates have been measured, a message will appear with options to remeasure standards, load more standards or begin measuring samples.
 - If finished measuring standards, select **Measure Samples**.

5. Measure samples:
 - a. Enter or edit the sample ID at the top of the screen by selecting the box and inputting a new sample name.
 - b. Pipette 2 μL of the assay tube containing the corresponding sample onto the lower pedestal and lower the arm.
 - If the **Auto-Measure** feature is OFF , select **Measure** to begin the measurement.
 - If the **Auto-Measure** feature is ON , the measurement will begin automatically after the arm is lowered.

Note Always use a fresh 2 μL aliquot for each measurement.

- c. When the sample measurement is completed, the standard curve and reported values are displayed (see the next section).
 - d. After the measurement is complete, remove sample from both upper and lower pedestal using a dry laboratory wipe.
 - e. Repeat sub-steps to measure all remaining samples (following the on-screen prompts).
 - f. When you are finished measuring samples, select **End Experiment**.
6. Lift the arm and clean both pedestals with a new wipe.


Related Topics

- [Best Practices for Nucleic Acid Measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Best Practices for Micro-Volume Measurements](#)
- [Basic Instrument Operations](#)

Reported Results

Fluorometer measurement screen (local control)

For each measured sample, the Fluorometer application shows the fluorescence standard curve and a summary of the results. Here is an example of the NanoDrop Ultra local control software screen:




Fluorometer measurement screen (PC control)

For each measured sample, this application shows the fluorescence standard curve and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.


Reported values


In the local control software, the initial screen that appears after each measurement shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:



- date (date and time sample measurement was taken)
- sample name
- analyte concentration
- RFU
- Excitation LED

Settings

The Fluorometer Setup screen appears after you select the Fluorometer application from the Fluorescence tab on the home screen. To show the Fluorometer settings from the Fluorometer measurement screen, select the Settings icon, .



Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Curve Type		Specify type of equation used to create standard curve from standard concentration values. Available options: <ul style="list-style-type: none"> – Linear: Draws the linear least squares line through all measured standards (requires reference measurement and at least one standard) – 2nd order polynomial: Draws the 2nd order least squares polynomial using all measured standards (requires reference measurement and at least two standards) – 3rd order polynomial: Draws the 3rd order least squares polynomial using all measured standards (requires reference measurement and at least three standards)
Replicates		Enter number of duplicate measurements of the reference or the same standards that are then averaged together to produce its associated concentration value. Note: Replicates setting cannot be changed after the first standard has been measured.
Standards		Enter actual concentration value of each standard and select the concentration units. Note: Concentration values can be entered in any order but the standards must be measured in the order they were entered.

Reagent Calculator

Calculates the necessary volume of each NanoDrop Ultra Fluorescence Assay component (dye and buffer) required to make the appropriate amount of working solution needed to prepare your samples and standards.

[Use Reagent Calculator](#)

[Results](#)



Use the Reagent Calculator

1. From the home screen, select the **Fluorescence** tab, then select **Reagent Calculator**.
2. Select one of the available assays from the drop-down menu next to **Choose an assay**.
3. Select one of the available **Sample to Reaction Volume** options for the assay chosen above.
4. See the protocol provided with your assay for more information on which Sample to Reaction Volume is best for your samples.
5. Select the total number of samples and standards that will be measured by selecting each respective box and entering the desired amount.
6. Optional: Enable **Include overage** (1 extra tube) by selecting the empty check box to include an additional tube into the calculation to ensure enough working solution is created for all samples and standards.

Reagent Calculator Results

When all fields are populated the Results: window will display the necessary volume of each NanoDrop™ Ultra Fluorescence Assay component (dye and buffer) to create the necessary amount of working solution for the number of samples and standards selected. The total volume of the working solution created will also be displayed.

The calculator uses a 1:200 ratio of dye/reagent to buffer required by the NanoDrop™ Ultra Fluorescence Assays.

Results:

Add 5µl dye to 995µl buffer for a total volume of 1000µl.

Acclaro Pro Applications

New versions of the most popular measurement applications can be unlocked providing as little as $\pm 5\%$ error throughout the absorbance dynamic range. These applications are not compatible with the standard cuvette mode. Users may notice an increased measurement time required to achieve the higher level of accuracy compared to the standard applications.

Activating Acclaro Pro

Thermo Scientific™ Acclaro™ Pro is an optional software add-on. If you have not purchased this add-on and are interested in using these applications, please purchase it at thermofisher.com/nanodrop or reach out to your local sales representative. The NanoDrop Ultra Spectrophotometers and Fluorometers will need to be serviced to unlock the Acclaro Pro applications.

dsDNA, ssDNA or RNA Pro

More accurately measures the concentration of highly concentrated (> 62.5 A) purified dsDNA, ssDNA or RNA samples that absorb at 260 nm using the Acclaro Pro enhanced automated pathlength selection algorithm.

[Measure dsDNA Pro, ssDNA Pro, or RNA Pro](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)

[Calculations](#)



Best practices for nucleic acid measurements

- Isolate and purify nucleic acid samples before measurement to remove impurities. Depending on the sample, impurities could include DNA, RNA, free nucleotides, proteins, some buffer components and dyes. See [Preparing Samples](#) for more information.

Note Extraction reagents such as guanidine, phenol, and EDTA contribute absorbance between 230 nm and 280 nm and will affect measurement results if present in samples (even residual amounts).

- Ensure the sample absorbance is within the instrument's [absorbance detection limits](#).
- Blank with the same buffer solution used to resuspend the analyte of interest. The blanking solution should be a similar pH and ionic strength as the analyte solution.
- Run a [blanking cycle](#) to assess the absorbance contribution of your buffer solution. If the buffer exhibits strong absorbance at or near the analysis wavelength (typically 260 nm), you may need to choose a different buffer or application. See [Choosing and Measuring a Blank](#) for more information.
- For micro-volume measurements:
 - Ensure pedestal surfaces are properly [cleaned](#) and [conditioned](#).
 - If possible, heat highly concentrated or large molecule samples, such as genomic or lambda DNA, to 63 °C (145 °F) and gently (but thoroughly) vortex before taking a measurement. Avoid introducing bubbles when mixing and pipetting.
 - Follow [best practices for micro-volume measurements](#).
 - Use a 1-2 µL sample volume. See [Recommended Sample Volumes](#) for more information.

Related Topics

- [Measure a Micro-Volume Sample](#)
- [Best Practices for Micro-Volume Measurements](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Measure dsDNA Pro, ssDNA Pro or RNA Pro

Use the dsDNA Pro, ssDNA Pro and RNA Pro applications to quantify purified double-stranded (ds) or single-stranded (ss) DNA or RNA samples with increased accuracy at higher concentrations (> 62.5 A). These applications report nucleic acid concentration and two absorbance ratios (A260/A280 and A260/A230). A single-point baseline correction can also be used.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.


Procedure

1. From the home screen, select the **Acclaro Pro** tab, then select **dsDNA Pro**, **ssDNA Pro** or **RNA Pro**, depending on the samples to be measured.
2. Configure any of the [setup options](#) if desired and select **Save**.
3. Pipette 1-2 μ L blanking solution onto the lower pedestal and lower the arm.
4. Select **Blank** and wait for the measurement to complete.
Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm.
5. Lift the arm and clean both pedestals with a new laboratory wipe.
6. Pipette 1-2 μ L sample solution onto the pedestal.
7. Start the sample measurement:
 - If [Auto-Measure](#) is On, lower arm; if Auto-Measure is off, lower arm and select **Measure**.


When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

8. When you are finished measuring samples, select **End Experiment**.
9. Lift the arm and clean both pedestals with a new wipe.

7 Acclaro Pro Applications
dsDNA, ssDNA or RNA Pro



Typical nucleic acid spectrum




Comparison of nucleic acid spectra with and without two common contaminants

Reported Results

dsDNA Pro measurement screen (local control)

For each measured sample, the dsDNA Pro, ssDNA Pro and RNA Pro applications show the UV absorbance spectrum and a summary of the results. Below is an example of the measurement screen of the NanoDrop Ultra local control software:



Note Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

dsDNA Pro measurement screen (PC control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See “[Measurement Screen Display Options](#)” on [page 299](#) for an example.

Reported values


The initial screen that appears after each measurement (see previous image) shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:



#:	1
Date:	10/30/2024 07:43:36 PM
Sample Name:	Sample 1
ng/ μ L:	10381.1
A260/A280:	1.90
A260/A230:	2.31
A260:	207.62
A280:	109.08
R ² (260nm):	0.999953
Factor:	50
Baseline Correction:	0.311A@340nm
Location:	Pedestal
Monitor Wavelength:	-

- # (sample number)
- date (date and time sample measurement was taken)
- sample name
- nucleic acid concentration
- A260/A280
- A260/A230
- A260
- R² (260nm)
- A280
- factor
- baseline correction
- location
- monitor wavelength

Settings

The dsDNA Pro, ssDNA Pro, or RNA Pro Setup screen appears after you select the respective application from the Acclaro Pro tab on the home screen. To show the dsDNA Pro, ssDNA Pro or RNA Pro settings, from the dsDNA Pro, ssDNA Pro or RNA Pro measurement screen, select .

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Baseline Correction	On or off Enter baseline correction wavelength in nm or use default value (340 nm)	Optional user-defined baseline correction. When enabled, the software corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength.

Calculations

The Acclaro Pro Nucleic Acid applications use a modification of the Beer-Lambert equation (shown at right) to calculate sample concentration where the extinction coefficient and pathlength are combined and referred to as a “factor.”

For the dsDNA Pro, ssDNA Pro and RNA Pro applications, the generally accepted factors for nucleic acids are used in conjunction with Beer’s Law to calculate sample concentration. For the Custom Factor Pro application, the user-specified factor is used.

Extinction Coefficients vs Factors

Using the terms in the Beer-Lambert equation, factor (f) is defined as:

$$\text{factor (f)} = 1/(\epsilon * b)$$

where:

ϵ = wavelength-dependent molar extinction coefficient in ng-cm/ μ L

b = **sample pathlength** in cm

As a result, analyte concentration (c) is calculated as:

$$c = A * [1/(\epsilon * b)]$$

or

$$c = A * f$$

where:

c = analyte concentration in ng/ μ L

A = absorbance in absorbance units (A)

f = factor in ng-cm/ μ L (see below)

Factors Used

- **dsDNA** (factor = 50 ng-cm/ μ L)
- **ssDNA** (factor = 33 ng-cm/ μ L)
- **RNA** (factor = 40 ng-cm/ μ L)
- **Custom Factor** (user entered factor between 15 ng-cm/ μ L and 150 ng-cm/ μ L)

Calculated nucleic acid concentrations are based on the absorbance value at 260 nm, the factor used and the sample pathlength. A single-point baseline correction (or analysis correction) may also be applied.

Concentration is reported in mass units. Calculators are available on the Internet to convert concentration from mass to molar units based on sample sequence.

Absorbance values at 260 nm, 280 nm and sometimes 230 nm are used to calculate purity ratios for the measured nucleic acid samples. Purity ratios are sensitive to the presence of contaminants in the sample, such as residual solvents and reagents typically used during sample purification.

Measured Values

Note: For micro-volume absorbance measurements, the spectra are normalized to a 10 mm pathlength equivalent.

A260 absorbance

- Nucleic acid absorbance values are measured at 260 nm using the normalized spectrum with an improved algorithm. This is the reported A260 value if Baseline Correction is not selected.
- If [Baseline Correction](#) is selected, the absorbance value at the correction wavelength is subtracted from the absorbance at 260 nm. The corrected absorbance at 260 nm is reported and used to calculate nucleic acid concentration.

A230 and A280 absorbance

- Normalized and baseline-corrected (if selected) absorbance values at 230 nm and 280 nm are used to calculate A260/A230 and A260/A280 ratios.

Reported Values

- **Nucleic acid concentration.** Reported in selected unit (i.e., ng/ μ L, μ g/ μ L or μ g/mL). Calculations are based on modified Beer's Law equation using corrected nucleic acid absorbance value.
- **A260/A280 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 280 nm. An A260/A280 purity ratio of ~1.8 is generally accepted as "pure" for DNA (~2.0 for RNA). Acidic solutions may under represent the reported value by 0.2-0.3; the opposite is true for basic solutions.
- **A260/A230 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 230 nm. An A260/A230 purity ratio between 1.8 and 2.2 is generally accepted as "pure" for DNA and RNA.

Note: Although purity ratios are important indicators of sample quality, the best quality indicator quality is functionality in the downstream application of interest (e.g., real-time PCR).

- **A260 absorbance.**
- **A280 absorbance.**
- **R² (260nm).** Coefficient of determination for the line of best fit for 260 nm measurements.
- **Factor.** Used in conjunction with Beer's Law to calculate sample concentration.
- **Baseline correction.** Wavelength selected for baseline correction and the absorbance detected at that wavelength.
- **Location.** Displays that the measurement was taken from the pedestal mode.
- **Monitored wavelength.** Enter an additional wavelength whose absorbance value you want included in the report.

Custom Factor Pro

More accurately measures the concentration of highly concentrated (> 62.5 A) purified nucleic acids using a custom factor and the Acclaro Pro enhanced automated pathlength selection algorithm.

[Measure Custom Factor Pro](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)

[Calculations](#)



Measure Custom Factor Pro

Use the Custom Factor application to quantify purified DNA or RNA samples with increased accuracy at higher concentrations (> 62.5 A) that absorb at 260 nm with a user-defined extinction coefficient or factor. The application reports nucleic acid concentration and two absorbance ratios (A₂₆₀/A₂₈₀ and A₂₆₀/A₂₃₀). A single-point baseline correction can also be used.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **Acclaro Pro** tab, then select **Custom Factor Pro**.

7 Acclaro Pro Applications

Custom Factor Pro


2. Configure any of the [setup options](#) if desired and select **Save**.
3. Pipette 1–2 μL blanking solution onto the lower pedestal and lower the arm.
4. Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm.

5. Lift the arm and clean both pedestals with a new laboratory wipe.
6. Pipette 1-2 μL sample solution onto the pedestal.
7. Start the sample measurement:
 - If [Auto-Measure](#) is On, lower arm; if Auto-Measure is off, lower arm and select **Measure**.

When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

8. When you are finished measuring samples, select **End Experiment**.
9. Lift the arm and clean both pedestals with a new wipe.



Typical nucleic acid spectrum

Related Topics

- [Measure a Micro-Volume Sample](#)
- [Best Practices for Micro-Volume Measurements](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

Custom Factor Pro measurement screen (local control)

For each measured sample, the Custom Factor Pro application shows the absorbance spectrum and a summary of the results. Here is an example of the measurement screen of the NanoDrop Ultra local control software:


Note The Custom Factor Pro measurement screen is identical to the measurement screen for the other standard nucleic acid Pro applications.



Custom Factor Pro measurement screen (PC control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.

Settings

The Custom Factor Pro Setup screen appears after you select the application from the Acclaro Pro tab on the home screen. To show the Custom Factor Pro settings from the Custom Factor Pro measurement screen, select .

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Custom Factor	Enter an integer value between 15 ng-cm/μL and 150 ng-cm/μL	Constant used to calculate nucleic acid concentration in modified Beer’s Law equation . Based on extinction coefficient and pathlength: $f = 1/(\mathcal{E}_{260} * b)$ <p>where: f= factor ℰ = molar extinction coefficient at 260 nm in ng-cm/μL b = sample pathlength in cm (1 cm for nucleic acids measured with the NanoDrop Ultra instruments)</p>
Baseline Correction	On or off Enter baseline correction wavelength in nm or use default value (340 nm)	Optional user-defined baseline correction. Can be used to correct for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength.

Detection Limits

The lower detection limits and reproducibility specifications for nucleic acids are provided [here](#). The upper detection limits are dependent on the [upper absorbance limit](#) of the instrument and the user-defined extinction coefficients.

To calculate upper detection limits for nucleic acid samples

To calculate upper detection limits in ng/μL, use the following equation:

$$(\text{upper absorbance limit}_{\text{instrument}} * \text{extinction coefficient}_{\text{sample}})$$

For example, for a sample measurement using an extinction coefficient of 55, the equation looks like this:

$$(550 \text{ AU} * 55 \text{ ng-cm}/\mu\text{L}) = 30,250 \text{ ng}/\mu\text{L}$$

Related Topics

- [Detection Limits for All Applications](#)

This page is intentionally blank.

Oligo DNA Pro or Oligo RNA Pro

More accurately measures the concentration of highly concentrated (> 62.5 A) purified ssDNA or RNA oligonucleotides that absorb at 260 nm using the Acclaro Pro enhanced automated pathlength selection algorithm.

[Measure Oligo DNA Pro or RNA Pro](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)

[Calculations](#)



Measure Oligo DNA Pro or Oligo RNA Pro

Use the Oligo DNA Pro and Oligo RNA Pro applications to quantify oligonucleotides that absorb at 260 nm with increased accuracy at higher concentrations (> 62.5 A). Molar extinction coefficients are calculated automatically based on the user-defined base sequence of the sample. These applications report nucleic acid concentration and two absorbance ratios (A₂₆₀/A₂₈₀ and A₂₆₀/A₂₃₀). A single-point baseline correction can also be used.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

Note If the oligonucleotide has been modified, for example with a fluorophore dye, check with the oligo manufacturer to determine if the modification contributes absorbance at 260 nm. If it does, we recommend using the [Microarray](#) application to quantify nucleic acid concentration. The Microarray application includes a correction to remove any absorbance contribution due to the dye from the oligo quantification result.

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure


1. From the home screen, select the **Acclaro Pro** tab, then select either **Oligo DNA Pro** or **Oligo RNA Pro**, as needed.
2. Configure any of the [setup options](#) if desired and select **Save**.
3. Pipette 1–2 μL blanking solution onto the lower pedestal and lower the arm.
4. Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm.

5. Lift the arm and clean both pedestals with a new laboratory wipe.
6. Pipette 1-2 μL sample solution onto the pedestal.
7. Start the sample measurement:
 - If [Auto-Measure](#) is On, lower arm; if Auto-Measure is off, lower arm and select **Measure**.

When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

8. When you are finished measuring samples, select **End Experiment**.
9. Lift the arm and clean both pedestals with a new wipe.




Related Topics

- [Best Practices for Nucleic Acid Measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Best Practices for Micro-Volume Measurements](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

Oligo DNA Pro and Oligo RNA Pro measurement screen (local control)

For each measured sample, the Oligo DNA Pro and Oligo RNA Pro applications show the UV absorbance spectrum and a summary of the results. Here is an example of the measurement screen of the NanoDrop Ultra local control software:



¹Measured oligo: TTT TTT TTT TTT TTT TTT TTT TTT

Note Micro-volume absorbance measurements are normalized to a 10.0 mm pathlength equivalent.

7 Acclaro Pro Applications

Oligo DNA Pro or Oligo RNA Pro


- [baseline correction](#)
- [oligo sequence](#)
- [location](#)
- pathlength used
- [monitored wavelength](#)


Note The five nucleotides that comprise DNA and RNA exhibit widely varying A260/A280 ratios. See [Oligo Purity Ratios](#) for more information.

Related Topics

- [Basic Instrument Operations](#)
- [Oligo Calculations](#)

Settings

The Oligo DNA Pro or Oligo RNA Pro Setup screen appears after you select the respective application from the Acclaro Pro tab on the home screen. To show the Oligo Pro settings from the Oligo Pro measurement screen, select .



The screenshot displays the Oligo Pro Setup screen with the following elements:

- Auto Naming:** A toggle switch that is turned on (checked).
- Input Field:** A large dark grey box with the text "Click to enter" inside.
- Factor:** 0.00
- Molecular weight:** 0.00
- Number of bases:** 0
- Molar ext. coefficient (260):** 0
- %GC:** 0.00
- Nucleotide Selection:** A row of five buttons labeled G, A, T, C, and a delete icon (X).
- Baseline Correction:** A toggle switch that is turned on (checked).
- Wavelength:** A numeric input field containing "340" followed by "nm" and the text "(Default value is 340nm)".

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.

Oligo Base Sequence

for DNA: Use the G, A, T and C keys to specify the DNA base sequence

for RNA: Use the G, A, U and C keys to specify the RNA base sequence

Specify your DNA or RNA base sequence.
Tap or click the corresponding keys:



Add thymine (DNA) or uracil (RNA)

Add cytosine

From the PC control software, you can also enter base sequence using the keyboard, or by copy and pasting a sequence from another application.

Each time a base is added to the sequence, the software calculates the following:

- **Factor.** Constant used to calculate oligonucleotide concentration in [modified Beer’s Law equation](#). Based on extinction coefficient and pathlength:

$$f = 1/(\epsilon_{260} * b)$$

where:

f= factor

ε = molar extinction coefficient at 260 nm in ng-cm/μL

b = [sample pathlength](#) in cm (0.1 cm for nucleic acids measured with the NanoDrop Ultra instrument)

Setting	Available Options	Description
		<ul style="list-style-type: none"> • Molecular Weight of oligo calculated from user-defined base sequence. • Number of Bases entered. • Molar Ext. Coefficient (260 nm). Molar extinction coefficient of oligo (in ng-cm/μL) at 260 nm calculated from entered base sequence. • %GC. Percentage of guanine and cytosine residues in total number of bases entered.
Baseline Correction	On or off Enter baseline correction wavelength in nm or use default value (340 nm)	Corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength. Tip: If the sample has a modification that absorbs light at 340 nm, select a different correction wavelength or turn off Baseline Correction.

Related Topics

- [Instrument Settings](#)

Detection Limits

The lower detection limits and reproducibility specifications for the oligonucleotide sample types (ssDNA and RNA) are provided [here](#). The upper detection limits are dependent on the [upper absorbance limit](#) of the instrument and the extinction coefficients for the user-defined [base sequences](#).

To calculate upper detection limits for nucleic acid samples

To calculate upper detection limits in ng/μL, use the following equation:

$$(\text{upper absorbance limit}_{\text{instrument}} * \text{extinction coefficient}_{\text{sample}})$$

For example, for a sample measurement using an extinction coefficient of 55, the equation looks like this:

$$(550 \text{ AU} * 55 \text{ ng-cm}/\mu\text{L}) = 30,250 \text{ ng}/\mu\text{L}$$

Calculations

As with the other nucleic acid applications, the Oligo Pro applications use the [Beer-Lambert equation](#) to correlate absorbance with concentration based on the sample's extinction coefficient and pathlength. Because oligonucleotides are short, single-stranded molecules (or longer molecules of repeating sequences), their spectrum and extinction coefficient (ϵ) are closely dependent on base composition and sequence.

(The generally accepted extinction coefficients and factors for single-stranded DNA and RNA provide a reasonable estimate for natural, essentially randomized, sequences but not for short, synthetic oligo sequences.) To ensure the most accurate results, we use the exact value of ϵ_{260} to calculate oligonucleotide concentration.

The NanoDrop software allows you to specify the base sequence of an oligonucleotide before it is measured. For any entered base sequence, the software uses the equation at the right to calculate the extinction coefficient.

Tip: The extinction coefficient is wavelength specific for each oligonucleotide and can be affected by buffer type, ionic strength and pH.

Extinction Coefficients for Oligonucleotides

The software uses the nearest neighbor method and the following formula to calculate molar extinction coefficients for specific oligonucleotide base sequences:

$$\epsilon_{260} = \sum_{1}^{N-1} \epsilon_1 - \sum_{2}^{N-1} \epsilon_2 + \sum_{1}^N \epsilon_3$$

where:

ϵ = molar extinction coefficient in L/mole-cm

ϵ_1 = ϵ nearest neighbor

ϵ_2 = ϵ individual bases

ϵ_3 = ϵ modifications, such as fluorescent dyes

Calculated nucleic acid concentrations are based on the absorbance value at 260 nm, the factor used and the sample pathlength. A single-point baseline correction (or analysis correction) may also be applied.

Concentration is reported in mass units. Calculators are available on the Internet to convert concentration from mass to molar units based on sample sequence.

Absorbance values at 260 nm, 280 nm and sometimes 230 nm are used to calculate purity ratios for the measured nucleic acid samples. Purity ratios are sensitive to the presence of contaminants in the sample, such as residual solvents and reagents typically used during sample purification.

The five nucleotides that comprise DNA and RNA exhibit widely varying A260/A280 ratios. Estimated A260/A280 ratios for each independently measured nucleotide are provided below:

Guanine: 1.15
Adenine: 4.50
Cytosine: 1.51
Uracil: 4.00
Thymine: 1.47

The A260/A280 ratio for a specific nucleic acid sequence is approximately equal to the weighted average of the A260/A280 ratios for the four nucleotides present.

Note: RNA will typically have a higher 260/280 ratio due to the higher ratio of Uracil compared to that of Thymine.

Measured Values

A260 absorbance

Note: For micro-volume absorbance measurements, the spectra are normalized to a 10 mm pathlength equivalent.

- Nucleic acid absorbance values are measured at 260 nm using the normalized spectrum. This is the reported A260 value if Baseline Correction is not selected.
- If [Baseline Correction](#) is selected, the absorbance value at the correction wavelength is subtracted from the sample absorbance at 260 nm. The corrected absorbance at 260 nm is reported and used to calculate nucleic acid concentration.


A230, A280 absorbance

- Normalized absorbance values at 230 nm, 260 nm and 280 nm are used to calculate A260/A230 and A260/A280 ratios.

Reported Values

- **Nucleic acid concentration.** Reported in selected unit (i.e., ng/ μ L, μ g/ μ L or μ g/mL). Calculations are based on modified Beer's Law equation using corrected nucleic acid absorbance value.
- **A260/A280 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 280 nm.
- **A260/A230 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 230 nm.

Note: The traditional purity ratios (A260/A280 and A260/A230), which are used as indicators of the presence of various contaminants in nucleic acid samples, do not apply for oligonucleotides because the shapes of their spectra are highly dependent on their base compositions. See side bar for more information.



- **A260 absorbance.**
- **A280 absorbance.**
- **R² (260nm).** Coefficient of determination for the line of best fit for 260 nm measurements.
- **Factor.** Used in conjunction with Beer's Law to calculate sample concentration.
- **Baseline correction.** Wavelength selected for baseline correction and the absorbance detected at that wavelength.
- **Oligo Sequence.**
- **Location.** Displays that the measurement was taken from the pedestal.
- **Monitored wavelength.** Enter an additional wavelength whose absorbance value you want included in the report.

Related Topics

- [Calculations for Nucleic Acid Measurements](#)

This page is intentionally blank.

Protein A280 Pro

More accurately measures the concentration of highly concentrated (> 62.5 A) purified protein samples that absorb at 280 nm using the Acclaro Pro enhanced automated pathlength selection algorithm.

[Measure Protein A280 Pro](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)

[Calculations](#)



Best practices for protein measurements

- Isolate and purify protein samples before measurement to remove impurities. Depending on the sample, impurities could include DNA, RNA and some buffer components. See [Preparing Samples](#) for more information.

Note Extraction reagents that contribute absorbance between 200 nm and 280 nm will affect measurement results if present in samples (even residual amounts).

- Ensure the sample absorbance is within the instrument's [absorbance detection limits](#).
- Choosing a blank:
 - For the Protein A280 Pro application, blank with the same buffer solution used to resuspend the analyte of interest. The blanking solution should be a similar pH and ionic strength as the analyte solution.
- Run a [blanking cycle](#) to assess the absorbance contribution of your buffer solution. If the buffer exhibits strong absorbance at or near the analysis wavelength (typically 280 nm), you may need to choose a different buffer or application, such as a colorimetric assay (for example, BCA or Pierce 660). See [Choosing and Measuring a Blank](#) for more information.

Note Buffers such as Triton X, RIPA, and NDSB contribute significant absorbance and are not compatible with direct A280 measurements.

- For micro-volume measurements:
 - Ensure pedestal surfaces are properly [cleaned](#) and [conditioned](#). (Proteins tend to stick to pedestal surfaces.)

- Gently (but thoroughly) vortex samples before taking a measurement. Avoid introducing bubbles when mixing and pipetting.
- Follow [best practices for micro-volume measurements](#).
- Use a 2 μ L sample volume. See [Recommended Sample Volumes](#) for more information.

Related Topics

- [Best practices for protein measurements](#)
- [Measure a Micro-Volume Sample](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Measure Protein A280 Pro

Use the Protein A280 application to quantify purified protein samples that contain amino acids such as tryptophan or tyrosine, or cys-cys disulfide bonds, which exhibit absorbance at 280 nm with increased accuracy at higher concentrations (> 62.5 A). This application reports protein concentration measured at 280 nm and one absorbance ratio (A260/A280). A single-point baseline correction can also be used. This application does not require a standard curve.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

Note If your samples contain mainly peptide bonds and little or no amino acids, use the [Protein A205](#) application instead of Protein A280 Pro.

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure


1. From the home screen, select the **Acclaro Pro** tab, then select **Protein A280 Pro**.
2. Configure any of the [setup options](#) if desired and select **Save**.
3. Pipette 1–2 μL blanking solution onto the lower pedestal and lower the arm.
4. Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm.

5. Lift the arm and clean both pedestals with a new laboratory wipe.
6. Pipette 2 μL sample solution onto the pedestal.
7. Start the sample measurement:
 - If [Auto-Measure](#) is On, lower arm; if Auto-Measure is off, lower arm and select **Measure**.


When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

8. When you are finished measuring samples, select **End Experiment**.
9. Lift the arm and clean both pedestals with a new wipe.



High concentration BSA sample

7 Acclaro Pro Applications
Protein A280 Pro




Low concentration BSA sample

Reported Results

Protein A280 Pro measurement screen (local control)

For each measured sample, the Protein A280 Pro application shows the absorbance spectrum and a summary of the results. Here is an example of the measurement screen within the NanoDrop Ultra local control software:




Note Micro-volume absorbance measurements are normalized to a 10.0 mm pathlength equivalent.

Protein A280 Pro measurement screen (PC control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.

Reported values

The initial screen that appears after each measurement (see previous image) shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:




The screenshot shows a dark-themed dialog box titled "Sample Details" with a close button (X) in the top right corner. The dialog contains a list of parameters and their values. At the bottom, there are three buttons: a trash icon, a printer icon, and a "Close" button.

Parameter	Value
#:	2
Date:	10/30/2024 08:42:21 PM
Sample Name:	Sample 2
mg/mL:	131.183
A280:	87.89
A260/A280:	0.59
Sample Type:	BSA
R ² (280nm):	0.999933
Ext.Coeff.E1%/gm-cm:	6.7
ε/1000:	-
MW(kDa):	-
Baseline Correction:	0.982A@340nm
Location:	Pedestal

- # (sample number)
- date (date and time sample measurement was taken)
- sample name
- protein concentration
- A280
- A260/A280
- sample type
- R² (280nm)
- baseline correction
- location
- monitored wavelength
- mass extinction coefficient (1% solution)
- molar extinction coefficient
- molecular weight (kDA)


Settings

The Protein A280 Pro application provides a variety of sample type options for purified protein analysis.



Each sample type applies a unique extinction coefficient to the protein calculations. If the extinction coefficient of the sample is known, choose the ϵ + MW (molar) or ϵ 1% (mass) option and enter the value. Otherwise, calculate the extinction coefficient or choose the option that best matches the sample solution. If you only need a rough estimate of protein concentration and the sample extinction coefficient is unknown, select the 1 Abs=1 mg/mL sample type option.

Tip Ideally, the extinction coefficient should be determined empirically using a solution of the study protein at a known concentration using the same buffer.

The Protein A280 Pro Setup screen appears after you select the application from the Acclaro Pro tab on the home screen. To show the Protein A280 Pro settings from the Protein A280 Pro measurement screen, select .

7 Acclaro Pro Applications
Protein A280 Pro

Setting	Available Options	Mass Ext. Coefficient (L/gm-cm)	Description
Auto Naming	On or off	N/A	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Baseline Correction	On or off Enter baseline correction wavelength in nm or use default value (340 nm)	N/A	Corrects for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength. Tip: If the sample has a modification that absorbs light at 340 nm, select a different correction wavelength or turn off Baseline Correction.
Sample type ^a	1 Abs = 1 mg/mL	General reference	Recommended when extinction coefficient is unknown and rough estimate of protein concentration is acceptable for a solution with no other interfering substances. Assumes 0.1% (1 mg/mL) protein solution produces 1.0A at 280 nm (where pathlength is 10 mm), i.e., $\epsilon 1\% = 10$.
	BSA	6.7	Calculates BSA (Bovine Serum Albumin) protein concentration using mass extinction coefficient (ϵ) of 6.7 L/gm-cm at 280 nm for 1% (i.e., 10 mg/mL) BSA solution. Assuming MW is 66,400 daltons (Da), molar extinction coefficient at 280 nm for BSA is approximately $43,824 \text{ M}^{-1}\text{cm}^{-1}$.

Setting	Available Options	Mass Ext. Coefficient (L/gm-cm)	Description
	IgG	13.7	Suitable for most mammalian antibodies (i.e., immunoglobulin G or IgG). Calculates protein concentration using mass extinction coefficient (ϵ) of 13.7 L/gm-cm at 280 nm for 1% (i.e., 10 mg/mL) IgG solution. Assuming MW is 150,000 Da, molar extinction coefficient at 280 nm for IgG is approximately $210,000 \text{ M}^{-1}\text{cm}^{-1}$.
	Lysozyme	26.4	Calculates lysozyme protein concentration using mass extinction coefficient (ϵ) of 26.4 L/gm-cm at 280 nm for 1% (i.e., 10 mg/mL) lysozyme solution. Assumes molar extinction coefficient for egg white lysozyme ranges between $36,000 \text{ M}^{-1}\text{cm}^{-1}$ and $39,000 \text{ M}^{-1}\text{cm}^{-1}$.
	Other protein (ϵ + MW)	User entered molar extinction coefficient and molecular weight	Assumes protein has known molar extinction coefficient (ϵ) and molecular weight (MW), where: $(\epsilon_{\text{molar}}) * 10 = (\epsilon_{\text{percent}}) * (\text{MW}_{\text{protein}})$ <p>Enter MW in kiloDaltons (kDa) and molar extinction coefficient (ϵ) in $\text{M}^{-1}\text{cm}^{-1}$ divided by 1000 (i.e., $\epsilon/1000$). For example, for protein with molar extinction coefficient of $210,000 \text{ M}^{-1}\text{cm}^{-1}$, enter 210.</p>
	Other protein ($\epsilon 1\%$)	User entered mass extinction coefficient	Assumes protein has known mass extinction coefficient (ϵ). Enter mass extinction coefficient in L/gm-cm for 10 mg/mL ($\epsilon 1\%$) protein solution.

^a To add or edit a custom protein, use Protein Editor.

Detection Limits

Detection limits and reproducibility specifications for purified BSA proteins are provided [here](#). The BSA lower detection limit and reproducibility values apply to any protein sample type. The upper detection limits are dependent on the [upper absorbance limit](#) of the instrument and the sample's extinction coefficient.

To calculate upper detection limits for other (non-BSA) protein sample types

To calculate upper detection limits in ng/μL for proteins, use the following equation:

$$(\text{upper absorbance limit}_{\text{instrument}} / \text{mass extinction coefficient}_{\text{sample}}) * 10$$

For example, if the sample's mass extinction coefficient at 280 nm is 6.7 for a 1% (10 mg/mL) solution, the equation looks like this:

$$(550 / 6.7) * 10 = 824.6 \text{ (or } \sim 825)$$

Calculations

The Protein A280 Pro application uses the [Beer-Lambert equation](#) to correlate absorbance with concentration. Solving Beer's law for concentration yields the equation at the right.

Beer-Lambert Equation (solved for concentration)

$$c = A / (\epsilon * b)$$

where:

A = UV absorbance in absorbance units (AU)

ϵ = wavelength-dependent molar absorptivity coefficient (or extinction coefficient) in liter/mol-cm

b = pathlength in cm

c = analyte concentration in moles/liter or molarity (M)

Note: Dividing the measured absorbance of a sample solution by its molar extinction coefficient yields the molar concentration of the sample. See [Published Extinction Coefficients](#) for more information regarding molar vs. mass concentration values.

The extinction coefficient of a peptide or protein is related to its tryptophan (W), tyrosin (Y) and cysteine (C) amino acid composition.

Tip: The extinction coefficient is wavelength specific for each protein and can be affected by buffer type, ionic strength and pH.

This application offers six options (shown at right) for selecting an appropriate extinction coefficient for each measured sample, to be used in conjunction with Beer's Law to calculate sample concentration.

If the extinction coefficient of the sample is known, choose the ϵ + MW (molar) or ϵ 1% (mass) option and enter the value. Otherwise, calculate the extinction coefficient or choose the option that best matches the sample solution.

Tip: Ideally, the extinction coefficient should be determined empirically using a solution of the study protein at a known concentration using the same buffer.

Extinction Coefficients for Proteins

At 280 nm, the extinction coefficient is approximated by the weighted sum of the 280 nm molar extinction coefficients of the three constituent amino acids, as described in this equation:

$$\epsilon = (nW * 5500) + (nY * 1490) + (nC * 125)$$

where:

ϵ = molar extinction coefficient

n = number of each amino acid residue

5500, 1490 and 125 = amino acid molar absorptivities at 280 nm

Available Options for Extinction Coefficient

- **1 Abs = 1 mg/mL**, where sample type and/or ext. coefficient is unknown (produces rough estimate of protein concentration)
- **BSA** (Bovine Serum Albumin, 6.7 L/gm-cm)
- **IgG** (any mammalian antibody, 13.7 L/gm-cm)
- **Lysozyme** (egg white lysozyme, 26.4 L/gm-cm)
- **Other protein** (ϵ + MW), user-specified molar ext. coefficient
- **Other protein** (ϵ 1%), user-specified mass ext. coefficient

Note: See [Sample Type](#) for details.

Most sources report extinction coefficients for proteins measured at or near 280 nm in phosphate or other physiologic buffer. These values provide sufficient accuracy for routine assessments of protein concentration.

The equation at the right shows the relationship between molar extinction coefficient (ϵ_{molar}) and percent extinction coefficient ($\epsilon 1\%$).

To determine concentration (c) of a sample in mg/mL, use the equation at the right and a conversion factor of 10.

Tip: The NanoDrop Ultra software includes the conversion factor when reporting protein concentrations.

Published Extinction Coefficients

Published extinction coefficients for proteins may be reported as:

- wavelength-dependent molar absorptivity (or extinction) coefficient (ϵ) with units of $\text{M}^{-1}\text{cm}^{-1}$
- percent solution extinction coefficient ($\epsilon 1\%$) with units of $(\text{g}/100 \text{ mL})^{-1}\text{cm}^{-1}$ (i.e., 1% or 1 g/100 mL solution measured in a 1 cm cuvette)
- protein absorbance values for 0.1% (i.e., 1 mg/mL) solutions

Tip: Assess published values carefully to ensure unit of measure is applied correctly.

Conversions Between ϵ_{molar} and $\epsilon 1\%$

$$(\epsilon_{\text{molar}}) * 10 = (\epsilon 1\%) * (\text{MW}_{\text{protein}})$$

Example: To determine percent solution extinction coefficient ($\epsilon 1\%$) for a protein that has a molar extinction coefficient of $43,824 \text{ M}^{-1}\text{cm}^{-1}$ and a molecular weight (MW) of 66,400 daltons (Da), rearrange and solve the above equation as follows:

$$\epsilon 1\% = (\epsilon_{\text{molar}} * 10) / (\text{MW}_{\text{protein}})$$

$$\epsilon 1\% = (43,824 * 10) / 66,400 \text{ Da}$$

$$\epsilon 1\% = 6.6 \text{ g}/100 \text{ mL}$$

Conversions Between g/100 mL and mg/mL

$$C_{\text{protein}} \text{ in mg/mL} = (A / \epsilon 1\%) * 10$$

Example: If measured absorbance for a protein sample at 280 nm relative to the reference is 5.8 A, protein concentration can be calculated as:

$$C_{\text{protein}} = (A / \epsilon 1\%) * 10$$

$$C_{\text{protein}} = (5.8/6.6 \text{ g}/100 \text{ mL}) * 10$$

$$C_{\text{protein}} = 8.79 \text{ mg/mL}$$

Calculated protein concentrations are based on the absorbance value at 280 nm, the selected (or entered) extinction coefficient and the sample pathlength. A single-point baseline correction (or analysis correction) may be applied.

Concentration is reported in mass units. Calculators are available on the Internet to convert concentration from mass to molar units based on sample sequence.

Absorbance values at 260 nm and 280 nm are used to calculate purity ratios for the measured protein samples.

Purity ratios are sensitive to the presence of contaminants in the sample, such as residual solvents and reagents typically used during sample purification.

Measured Values

A280 absorbance

Note: For micro-volume absorbance measurements the spectra are normalized to a 10 mm pathlength equivalent.

- Protein absorbance values are measured at 280 nm using the normalized spectrum. If Baseline Correction is not selected, this is the reported A280 value and the value used to calculate protein concentration.
- If [Baseline Correction](#) is selected, the normalized and baseline-corrected absorbance value at 280 nm is reported and used to calculate protein concentration.

A260 absorbance

- Normalized and baseline-corrected (if selected) absorbance value at 260 nm is used to calculate A260/A280 ratios.

Reported Values

- **Protein concentration.** Reported in selected unit (mg/mL or $\mu\text{g/mL}$). Calculations are based on Beer-Lambert equation using corrected protein absorbance value.
- **A260/A280 purity ratio.** Ratio of corrected absorbance at 260 nm to corrected absorbance at 280 nm. An A260/A280 purity ratio of ~ 0.57 is generally accepted as “pure” for proteins.

Note: Although purity ratios are important indicators of sample quality, the best indicator of protein quality is functionality in the downstream application of interest (e.g., real-time PCR).

- **Sample type.** Determines the extinction coefficient used in conjunction with Beer's Law to calculate sample concentration.
- **R² (280nm).** Coefficient of determination for the line of best fit for 280 nm measurements.
- **Baseline correction.** Wavelength selected for baseline correction and the absorbance detected at that wavelength.
- **Location.** Displays that the measurement was taken from the pedestal.
- **Monitored wavelength.** Enter an additional wavelength whose absorbance value you want included in the report.
- **Mass extinction coefficient (1% solution).**
- **Molar extinction coefficient.**
- **Molecular weight.**

More Applications

Use the NanoDrop Ultra Spectrophotometers and Fluorometers to perform UV-Vis, OD600, Kinetics, or your own Custom measurements.

The UV-Vis application can be set up directly from the touchscreen and allows the instrument to function as a conventional spectrophotometer. Up to 40 wavelengths from 190 nm to 850 nm can be monitored and reported.

The Custom Methods application provides additional flexibility for users looking for unique information about their samples.

- [UV-Vis 232](#)
- [Custom Methods 239](#)
- [OD600 251](#)
- [Kinetics 261](#)

UV-Vis

Measures the absorbance of any sample at up to 40 wavelengths across the ultra-violet (UV) and visible regions of the spectrum.

[Measure UV-Vis](#)

[Reported Results](#)

[Settings](#)



Best practices for UV-Vis measurements

- Ensure the sample absorbance is within the instrument's [absorbance detection limits](#).
- Blank with the same buffer solution used to re-suspend the analyte of interest. The blanking solution should be a similar pH and ionic strength as the analyte solution.
- Run a [blanking cycle](#) to assess the absorbance contribution of your buffer solution. If the buffer exhibits strong absorbance at or near an analysis wavelength, you may need to choose a different buffer or application. See [Choosing and Measuring a Blank](#) for more information.
- For micro-volume measurements:
 - Ensure pedestal surfaces are properly [cleaned](#) and [conditioned](#).
 - Ensure samples are homogeneous before taking a measurement. Avoid introducing bubbles when mixing and pipetting.
 - Follow [best practices for micro-volume measurements](#).
 - Use a 1-2 μL sample volume. See [Recommended Sample Volumes](#) for more information.
- For cuvette measurements (NanoDrop Ultra^C and NanoDrop Ultra^C FL instruments only), use compatible cuvettes and follow [best practices for cuvette measurements](#).

Measure UV-Vis

The UV-Vis application allows the instrument to function as a conventional spectrophotometer. Sample absorbance is displayed on the screen from 190 nm to 850 nm. Up to 40 wavelengths can be designated for absorbance monitoring and inclusion in the report. Automatic pathlength adjustment and a single-point baseline correction can also be used.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **More Apps** tab, then select **UV-Vis**.
2. Configure any of the [setup options](#) if desired and select **Save**.
3. If using a NanoDrop Ultra^C or NanoDrop Ultra^C FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.
4. Pipette 1–2 μL blanking solution onto the lower pedestal and lower the arm, or insert the blanking cuvette into the cuvette holder.

Tip: If using a cuvette, make sure to [align the cuvette light path](#) with the instrument light path.
5. Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements.)

6. Lift the arm and clean both pedestals with a new laboratory wipe or remove the blanking cuvette.
7. Pipette 1-2 μL sample solution onto the pedestal, or insert the sample cuvette into the cuvette holder.
8. Start the sample measurement:
 - Pedestal: If **Auto-Measure** is On, lower arm; if Auto-Measure is off, lower arm and select **Measure**.
 - Cuvette: Select **Measure**.


When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

9. When you are finished measuring samples, select **End Experiment**.
10. Lift the arm and clean both pedestals with a new wipe or remove the sample cuvette.

Reported Results

UV-Vis measurement screen (local control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. Here is an example as it appears on the NanoDrop Ultra local control instrument display:



Note Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

UV-Vis measurement screen (PC control)

For each measured sample and standard, this application shows the visible absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.

8 More Applications

UV-Vis


UV-Vis reported values

The initial screen that appears after each measurement (see previous image) shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:

A screenshot of a UV-Vis measurement summary screen. The background is black with white text. The text is organized into a list of key-value pairs. The values are: #: 4, Date: 11/04/2024 06:11:33 PM, Sample Name: Sample 4, Location: Pedestal, Baseline Correction: -, Pathlength Used: @260 1 mm, Stirrer: OFF, Heater: OFF, Measured Temperature: -, 620: 12.758, and three instances of wavelength (λ): -.

#:	4	λ:	-
Date:	11/04/2024 06:11:33 PM		
Sample Name:	Sample 4		
Location:	Pedestal		
Baseline Correction:	-		
Pathlength Used:	@260 1 mm		
Stirrer:	OFF		
Heater:	OFF		
Measured Temperature:	-		
620:	12.758		
λ:	-		
λ:	-		
λ:	-		

Settings

The UV-Vis setup screen appears after you select the UV-Vis application from the More Apps tab on the home screen. To show the UV-Vis settings from the UV-Vis measurement screen, select .

Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Monitored wavelengths	Enter up to 40 wavelengths between 190 nm and 850 nm	<p>User-defined wavelengths to be measured and reported at run time. Absorbance values for the first three entered wavelengths are displayed in the measurement screen. To see absorbance values for 6 monitored wavelengths, swipe left in the measurement screen to show the Data table. To see all monitored wavelengths, press and hold a sample row to show the Sample Details screen (scroll up to display absorbance values for any additional user-defined wavelengths).</p> <p>Note: If Baseline Correction is selected, all displayed absorbance values are the corrected values.</p>
Analytical Wavelength	On or Off Enter analytical wavelength in nm or use default value (220 nm)	This is the wavelength the software will use to determine the pathlength selection. Automated Pathlength cannot be used when Analytical Wavelength is enabled.

Setting	Available Options	Description
Automated Pathlength	On or Off (affects pedestal measurements only)	<p>Optional automated pathlength selection. Allows the software to use the optimal (shorter) pedestal pathlength for high concentration samples to help prevent detector saturation (see Detection Limits for details).</p> <ul style="list-style-type: none">• When selected, the shorter pathlength is used when any wavelength between 220 nm and 850 nm has 10 mm equivalent absorbance value of 12.5 or higher. For wavelengths between 190 nm and 219 nm the change to the shorter pathlength occurs when any wavelength in this range has a 10 mm equivalent absorbance value of 10 or higher.• When deselected, the pedestal pathlength is restricted to 10 mm across all wavelengths. <p>Note: In either case, displayed absorbance values have been normalized to a 10 mm pathlength equivalent.</p>
Baseline Correction	On or off Enter baseline correction wavelength in nm or use default value (750 nm)	<p>Optional user-defined baseline correction. Can be used to correct for any offset caused by light scattering particulates by subtracting measured absorbance at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength.</p>

Custom Methods

Create and run a custom measurement method to display unique information for your samples.

[Manage Custom Methods](#)

[Measure Custom Method](#)

[Delete Custom Method](#)

[Reported Results](#)




Manage Custom Methods

Use the Custom Methods application to run a user-defined method created using the NanoDrop Ultra local or PC control software. Custom methods can be made with or without standards.

Create custom method

Create method to be used for sample measurements with user-defined settings.

1. From the home screen, select the **More Apps** tab, then select **Custom Methods**



2. Select **+ NEW METHOD** and choose one of the following:
 - **Formula** (if your method will not have standards)
 - **Standard Curve** (if your method will have standards)

8 More Applications

Custom Methods

3. In the **Create Formula Method** window, enter **Method Name**.
4. Enter detailed **Description** of method, if desired.
5. Specify how to calculate and report the method results:
 - If method does not have standards, specify [factor or extinction coefficient of analyte](#) (enter “1” to report absorbance measurements only) within the **Inputs** tab.
 - If method has standards, enter [name and concentration of each standard and select the curve fit type](#) within the **Standards** tab.

The screenshot shows the 'Create Standard Curve Method' window with the 'Standards' tab selected. The 'Curve fit Type' is set to 'Linear'. There is a '+ Standard' button. Below, a table lists five standards with their IDs and concentrations. At the bottom are 'Cancel', 'Save as', and 'Save' buttons.


Standard ID	Concentration (ng/μL)
Standard 1	10
Standard 2	20
Standard 3	30
Standard 4	40
Standard 5	50

Alternatively, [load a standard curve](#).

6. Enter or choose remaining [custom settings](#) as needed.
7. Choose **Save**

Note Any errors in the method will be listed in orange text at the bottom of the method editor screen. Errors must be addressed before the method can be saved.

View or edit custom method


1. From the Custom Methods screen, select  next to the desired method to edit.

2. Select **Edit**
3. View and adjust the method settings as desired
4. Select **Save**

Copy custom method


To create a custom method that is similar to an existing one, open the existing method, make your changes, then select **Save As** and enter a new name.

Copy custom method

1. From the Custom Methods screen, select  next to the desired method to copy.
2. Select **Edit**
3. View and adjust the method settings as desired.
4. Enter new **Method name** and **Description**.
5. Select **Save As**.
6. Enter a new filename for the method and click **Save**.

Export custom method

Export a custom method in order to run it and store the measurement results on another NanoDrop Ultra instrument or another computer running the NanoDrop Ultra PC control software.

1. From the Custom Methods screen, select  next to the desired custom method.
2. Select **Export** (if method is invalid, an error message is displayed; errors must be fixed before method can be exported).
3. Select an available location to export to from the dropdown menu and then choose **Export** (method is exported to method file (*.method filename extension in proprietary format).

To transfer the method to the NanoDrop Ultra instrument or NanoDrop PC control software, copy the method file to a USB memory device and then load the method (see [Import a Custom Method](#) for details)

Import custom method

Import a custom method back to a computer running the NanoDrop Ultra software in order to edit the method settings.

1. From the Custom Methods screen, select **Import**


8 More Applications

Custom Methods

2. Locate and select “.method” file
3. Choose **Import** (imported method is added to end of Select Method list)

Note Custom methods downloaded from the NanoDrop website have a .zip file-name extension and must be extracted using a third-party file decompression program before the software will recognize them as custom methods.


Delete custom method

1. From the Custom Methods screen, select  next to the desired custom method.
2. Select **Delete**.
3. After the confirmation message, choose **Yes**.

Load standard curve

1. From the Custom Methods screen, select **Load Standard**.
2. Select the desired previously run standard curve, then select **Load**.

The measurement screen opens with the standard curve loaded. You can begin by blanking and then measuring samples.



Measure using a Custom Method

Use the Custom application to create and run a user-defined method created. For more information, see [“Create custom method”](#) on [page 239](#).

Before you begin...


Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

NOTICE


- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **More Apps** tab, then select **Custom Methods**.
2. In the method selection pane, select the method to run.



Information about the selected method appears in the Method Details pane.

3. Select .
4. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.

8 More Applications

Custom Methods

- When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.

5. Follow the on-screen instructions to measure a sample.

Reported Results

Custom method measurement screen (local control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. Here is an example of the NanoDrop Ultra local control software screen:



Note Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

Custom method measurement screen (PC control)

For each measured sample and standard, this application shows the absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.

Reported values

The initial screen that appears after each measurement (see previous image) shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:

#:	1	Stirrer:	OFF
Date:	11/04/2024 06:59:30 PM	Heater:	OFF
Sample Name:	Sample 1	Measured Temperature:	-
Concentration(ng/μL):	22.5	Target Temperature:	-
Analysis Wavelength Absorbance:	0.450		
Alexa Fluor 647:	0.781		
Ext.Coeff.E1%/gm-cm:	0		
ε/1000:	0		
MW(kDa):	0		
Factor:	50		
Location:	Cuvette		
Baseline Correction:	-		
Pathlength Used:	10 mm		

Settings

These settings are available for creating custom methods.

Setting	Available Options
Result	Enter descriptive name for calculated concentration result (for example, “MTT Assay”) and use adjacent drop down list to select appropriate unit. Result name appears as column heading for reported concentration value.
Measurement range	Select spectral range in which method will acquire data. Available options: <ul style="list-style-type: none"> • Ultra-violet only (190 nm - 350 nm) • Visible only (350 nm - 850 nm) • Ultra-violet and visible (190 nm - 850 nm) • Custom (specify starting and ending point in nanometers) <p>Notes:</p> <ul style="list-style-type: none"> • If a Baseline correction and/or Analysis wavelength correction are used, make sure your selected spectral range includes your specified baseline correction and/or analysis correction wavelength. • For micro-volume absorbance measurements and measurements taken with nonstandard (other than 10 mm) cuvettes, the spectra are normalized to a 10 mm pathlength equivalent.

Setting	Available Options
Analysis wavelength correction	<p>Use this option to specify absorbance correction at analysis wavelength only. Available options:</p> <ul style="list-style-type: none">• None. No correction at analysis wavelength.• Single point. Enter wavelength for analysis correction. (Absorbance value at specified analysis correction wavelength is subtracted from absorbance value at analysis wavelength. Corrected value is used to calculate sample concentration.)• Sloping baseline. Enter two wavelengths that define sloping baseline for analysis correction. (Absorbance value of sloping baseline at analysis wavelength is subtracted from absorbance value at analysis wavelength. Corrected value is used to calculate sample concentration.)
Factor or Extinction coefficient at 1 cm pathlength (Formula methods only)	<p>Specify whether to use factor or extinction coefficient to calculate concentration result:</p> <ul style="list-style-type: none">• User-defined factor. Enter factor for 1 cm pathlength and use adjacent drop down list to select appropriate unit. Equation below shows how factor is used to calculate sample concentration: $c = (A * f) / b$<p>where: c = analyte concentration A = absorbance in absorbance units (A) f = factor (typically 1/ε, where ε = wavelength-dependent molar absorptivity coefficient, or extinction coefficient) b = pathlength in cm (determined at measurement time, then normalized to 10 mm (1 cm) pathlength equivalent)</p>• Extinction coefficient and molecular weight. Enter extinction coefficient for 1 cm pathlength and use adjacent drop down list to select appropriate unit. Equation below shows how extinction coefficient is used to calculate sample concentration: $c = A / (\epsilon * b)$<p>where: c = analyte concentration A = absorbance in absorbance units (A) ε = wavelength-dependent molar absorptivity coefficient (or extinction coefficient) b = pathlength in cm (determined at measurement time, then normalized to 10 mm (1 cm) pathlength equivalent)</p>

Setting**Available Options****Notes:**

- Refer to product literature for information about factors and extinction coefficients for specific materials.
- To set up a method that reports absorbance measurements only, select Factor or Extinction Coefficient with the factor or extinction coefficient set to “1”.
- If specified unit for factor or extinction coefficient is based on mass (such as mg/mL) and specified unit for calculated result is based on molarity (such as pmol/μL) or vice versa, enter **molecular weight** and use adjacent drop down list to select appropriate **unit**.


Standards (Standard curve methods only)

Define the standards:

- Enter name and analyte concentration of each standard and a reference, if desired:
 - Depending on the Curve Type setting, a standard curve can be generated using two or more standards. (The software **allows a reference and up to 15 standards**.)
 - **All reference and standards solutions** should be in the same buffer used to resuspend the samples plus the same volume of reagent added to the samples.
 - **First standard** can be a reference measurement. The reference solution should contain none of the analyte of interest. (The reference measurement is not the same as a blank measurement.)
 - **Concentration values for standards** can be entered in any order but the standards must be measured in the order in which they were entered; however, best practice dictates that standards be measured from the lowest concentration of the standard analyte stock to the highest.
 - **Concentration range of the standards** must cover the dynamic range of the assay and the expected range of the unknown samples. Sample analyte concentrations are not extrapolated beyond the concentration of the highest standard.

Setting	Available Options
	<ul style="list-style-type: none">• Select curve fit type. <p>Specify type of equation used to create standard curve from standard concentration values. Available options:</p> <ul style="list-style-type: none">– Linear: Draws the linear least squares line through all measured standards (requires reference measurement and at least one standard). Linear through zero is similar but with the constraint that the line passes through the origin (0,0)– Interpolation: Draws a series of straight lines to connect all measured standards (requires reference measurement and at least one standard)– 2nd order polynomial: Draws the 2nd order least squares polynomial using all measured standards (requires reference measurement and at least standards). 2nd order polynomial through zero is similar but with the constraint that the line passes through the origin (0,0)– 3rd order polynomial: Draws the 3rd order least squares polynomial using all measured standards (requires reference measurement and at least three standards). 3rd order polynomial through zero is similar but with the constraint that the line passes through the origin (0,0)
Analysis wavelength (Standard curve methods only)	<p>Monitor absorbance at specified wavelength (enter the wavelength in nanometers).</p> <p>Note: The specified wavelength must fall within the selected measurement range.</p> <p>The measurement results or the concentration will be calculated automatically using the absorbance value at the specified wavelength and applying the selected method type (factor or standard curve).</p>
Baseline correction	<p>Select this option to correct offset caused by light scattering particulates by subtracting the absorbance at a specified baseline point. Then specify wavelength for baseline correction.</p> <p>Note: Software subtracts absorbance value at specified baseline correction wavelength from absorbance values at all wavelengths in sample spectrum. As a result, absorbance of sample spectrum is zero at specified baseline correction wavelength.</p>

Setting	Available Options
Automated pathlength	<p data-bbox="505 289 1029 321">Affects micro-volume measurements only.</p> <ul data-bbox="505 348 1464 701" style="list-style-type: none"> <li data-bbox="505 348 1464 701">• When Automated Pathlength is selected, software selects the optimal pathlength (between 1.0 mm and 0.03 mm) based on sample absorbance at the analysis wavelength. For example, when sample absorbance at the analysis wavelength is less than or equal to 12.5 (10 mm pathlength equivalent), the optimal longer pathlength is used. When sample absorbance is greater than 12.5, the optimal shorter pathlength is used. Recommended for samples that are highly absorbing at the analysis wavelength. (This option may cause reduced sensitivity when the sample spectra have a large absorbance peak that is not at the analysis wavelength.) <p data-bbox="548 726 1446 863">Note: When the analysis wavelength is between 190 nm and 219 nm, the optimal longer pathlength is used when sample absorbance is less than or equal to 10 (10 mm pathlength equivalent), and the optimal shorter pathlength is used when sample absorbance is greater than 10.</p> <ul data-bbox="505 888 1464 1024" style="list-style-type: none"> <li data-bbox="505 888 1464 1024">• When Automated Pathlength is deselected, the software uses a 1 mm pathlength regardless of the sample absorbance. This can cause detector saturation (resulting in jagged peaks) for highly absorbing samples (e.g., ~15 A at 10 mm pathlength equivalent).

Setting	Available Options
Formula table (optional)	<p>Use the Formula table to specify additional reported results, such as a purity ratio, for each sample.</p> <p>Available options:</p> <ul style="list-style-type: none">• Predefined. Select from a list of predefined formulas, which can be used as is or edited, and choose Save. The predefined formula is listed in the Formula Table.• +Formula. Create formula for current method. Available options:<ul style="list-style-type: none">• Formula Name. Enter a name for the formula. After a measurement, the name is reported in Data Table and Sample Details screens.• Equation. Enter valid formula (see below for rules and examples). After a measurement, the measured or calculated value is reported in Data Table and Sample Details screens.• Unit. Select from a drop-down of available units for reported result. After a measurement, the unit is reported in Data Table and Sample Details screens.• Delete. Delete selected formula from current method by selecting .
Formula rules	<p>Custom formulas can include the following operators and functions:</p> <ul style="list-style-type: none">• Path(). Returns sample pathlength in cm.• A(nm). Returns sample absorbance at specified wavelength (for example, enter A(650) to add the measured absorbance at 650 nm to your equation).• Operators: + (add), - (subtract), * (multiply), / (divide).• Functions: Log(x), Pow(x,y). <p>Notes: Follow these additional rules for all languages:</p> <ul style="list-style-type: none">• Use period “.” decimal separators for floating point and double-floating point numbers.• Use comma “,” list separators (for example, “POW(2,8)”).• Do not use comma “,” group separators for large numbers (for example, enter 1000 rather than 1,000).

OD600

Measures the concentration of microbial cell cultures in solution by measuring scattered light at 600 nm.

[Measure OD600](#)

[Reported Results](#)

[Settings](#)

[Calculations](#)



Theory of OD600 application

The OD600 application measures light transmission and uses that value to calculate absorbance. In spectroscopy, transmitted light is defined as any light that is not absorbed by, reflected from and scattered off a sample.

In the case of living cells, most of the incident light is transmitted through the sample rather than scattered, reflected or absorbed. The amount of scattered light is low and can vary from instrument to instrument. As a result, calculated absorbance readings are typically very low.

The calculated absorbance values are used to determine the density of cells in solution in cells/mL. The physical concepts and formulas that relate optical properties of living cells to concentration include:

- Cells, which have a different index of refraction from the surrounding medium, randomly reflect and scatter light out of the incident light path. The amount of scattering is proportional to the density of cells in the sample.
- The Beer's Law equation is used to relate absorbance to concentration. See [Calculations for OD600 Measurements](#) for details.
- For cuvette reading with the NanoDrop Ultra^C and NanoDrop Ultra^C FL instruments, accurate absorbance readings are typically in the range between 0.04 A and 1.5 A. Serial dilutions of the sample are usually needed to bring the absorbance readings within this range.
- All measurements should be made on the same type of spectrophotometer and method (i.e., pedestal vs. cuvette) as the amount of scattered light captured varies based on the optical configuration. When using a different spectrophotometer or method, calculate and apply a conversion factor to the reported results. For example, to compare OD readings using the pedestal vs. a cuvette, a conversion factor can be calculated as follows:

$$\text{Conversion factor} = \text{Cuvette OD} / \text{Pedestal OD}$$

Best practices for OD600 measurements

- Ensure the sample is within the instrument's [absorbance detection limits](#).
- Blank with the growth or culture media the cells of interest are suspended in.
- Run a [blanking cycle](#) to assess the absorbance contribution of your media solution. If the media solution exhibits strong absorbance at or near the analysis wavelength (600 nm), you may need to choose a different media solution or application. See [Choosing and Measuring a Blank](#) for more information.
- Make dilutions as necessary to ensure sample cultures do not exceed the linear dynamic range of the assay before the culture reaches the stationary phase. The linear range depends largely on optical configuration and, therefore, differs for pedestal and cuvette measurements. To determine the linear range:
 - Measure a series of dilutions using a young overnight culture (~16 hrs) of the microbial strain
 - Graph the OD600 measurements against the dilution factor

The upper detection limit is the measured OD600 value at which there ceases to be a linear correlation between dilution factors and OD600 readings.

- Mix samples gently but thoroughly immediately before taking an aliquot for measurement.
- For micro-volume measurements:
 - Ensure pedestal surfaces are properly [cleaned](#) and [conditioned](#).
 - Avoid introducing bubbles when mixing and pipetting.
 - Start the measurement promptly to avoid settling or evaporation.
 - Follow [best practices for micro-volume measurements](#).
 - Use 2 μ L sample volume. See [Recommended Sample Volumes](#) for more information.
 - For dilute samples that exhibit low absorbance at 600 nm, use an alternative wavelength such as 400 nm to measure absorbance, or use cuvettes instead of micro-volume measurements.
- For cuvette measurements (NanoDrop Ultra^C and NanoDrop Ultra^C FL instruments only):
 - Use clean plastic, glass or quartz cuvettes.
 - Follow [best practices for cuvette measurements](#).
 - Do not use the automatic [stirring](#) feature for this assay.

Measure OD600

Use the OD600 application to monitor the growth rate of bacterial or other microbial cell cultures by measuring the optical density (absorbance) of the culture in growth media at 600 nm. The Beer-Lambert equation and a user-entered conversion factor are used to correlate absorbance with concentration. Reported concentration values can be used to identify the phase of cultured cell populations, e.g., log or exponential and stationary.

The OD600 application reports cell concentration in cells/mL. A single-point absorbance correction can be used. This application does not require a standard curve.

Before you begin...

Before taking pedestal measurements with the NanoDrop Ultra instrument, lift the instrument arm and clean the upper and lower pedestals. At a minimum, wipe the pedestals with a new laboratory wipe. For more information, see [Cleaning the Pedestals](#).

Note Due to the amount of scattered light present in this assay, absorbance readings are typically very low.

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

Procedure

1. From the home screen, select the **More Apps** tab, then select **OD600**.
2. Configure any of the [setup options](#) if desired and select **Save**.
3. If using a NanoDrop Ultra^c or NanoDrop Ultra^c FL model, select the correct measurement pathway.
 - When using a cuvette, select **Cuvette** from the drop-down menu at the top of the screen, this will present the cuvette settings. Select desired pathlength, stir speed, and heating then close the drop-down menu.
 - When using the pedestal for measurement, leave **Pedestal** as the selected setting at the top of the screen.

- Pipette 2 μL blanking solution (i.e., the media solution the cells of interest are suspended in) onto the lower pedestal and lower the arm, or insert the blanking cuvette into the cuvette holder.

Tip: If using a cuvette, make sure to [align the cuvette light path](#) with the instrument light path.

- Select **Blank** and wait for the measurement to complete.

Tip: If [Auto-Blank](#) is On, the blank measurement starts automatically after you lower the arm. (This option is not available for cuvette measurements.)

- Lift the arm and clean both pedestals with a new laboratory wipe, or remove the blanking cuvette.
- Pipette 2 μL sample solution onto the pedestal and lower the arm, or insert the sample cuvette into the cuvette holder.

- Start the sample measurement:

- Pedestal: If [Auto-Measure](#) is On, lower arm; if Auto-Measure is off, lower arm and select **Measure**.
- Cuvette: select **Measure**

When the sample measurement is completed, the spectrum and reported values are displayed (see the next section).

- When you are finished measuring samples, select **End Experiment**.
- Lift the arm and clean both pedestals with a new wipe, or remove the sample cuvette.


Related Topics

- [Measure a Micro-Volume Sample](#)
- [Measure a Sample Using a Cuvette](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

OD600 measurement screen (local control)

For each measured sample, this application shows the absorbance spectrum and a summary of the results. Here is an example of the NanoDrop Ultra local control software:



Note Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent.

OD600 measurement screen (PC control)

For each measured sample and standard, this application shows the absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See ["Measurement Screen Display Options"](#) on page 299 for an example.

Reported values


The initial screen that appears after each measurement shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:


#:	3	Heater:	OFF
Date:	11/04/2024 07:07:42 PM	Measured Temperature:	-
Sample Name:	Sample 3	Target Temperature:	-
A600:	0.71		
Correction:	0		
λ :	500		
A(λ):	1.17		
Factor (10 ⁸):	1.00		
Cells/mL (10 ⁸):	0.71		
Location:	Pedestal		
Monitor Wavelength:	-		
Pathlength Used:	1 mm		
Stirrer:	OFF		

Related Topics

- [Basic Instrument Operations](#)
- [OD600 Calculations](#)

Settings

The OD600 setup screen appears after you select the OD600 application from the More Apps tab on the home screen. To show the OD600 settings from the OD600 measurement screen, select .



Setting	Available Options	Description
Auto Naming	On or off	When enabled, each sample is given a default base name “sample” followed by the number sample in the sequence. For example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.
Absorbance correction	Absorbance value between 0 and 300 A	User-defined absorbance correction. Enter absorbance correction for displayed spectrum. This can be useful, for example, to correct baseline offset caused by any difference between the media solution used to blank the instrument and media used to suspend the cell culture sample, and because scattered light generally produces an offset. Absorbance correction value is subtracted from absorbance values at all wavelengths in sample spectrum. (All displayed absorbance values are corrected values.)

Setting	Available Options	Description
Additional monitored wavelength (λ)	Any wavelength between 250 nm and 700 nm	<p>User-defined wavelength. Enter an additional wavelength to measure if desired (useful for dilute samples that exhibit low absorbance at 600 nm).</p> <p>If an alternative wavelength is specified, use this equation to calculate cell concentration:</p> $c = A(\lambda) * \text{factor}(\lambda)$ <p>where:</p> <p>c = analyte concentration in cells/mL</p> <p>$A(\lambda)$ = UV-visible absorbance at specified wavelength in absorbance units (A)</p> <p>$\text{factor}(\lambda) = 1/(\epsilon(\lambda) * b)$ in mL/cell-cm</p> <p>where:</p> <p>$\epsilon(\lambda)$ = molar absorption coefficient (or extinction coefficient) at specified wavelength</p> <p>b = pathlength in cm (1.0 cm for the NanoDrop Ultra instruments)</p>
Cell number conversion factor (10^8)	Any number	<p>User-defined factor. Generally accepted factor for measured cell type, or one derived empirically using a solution of study cells at known concentration using the same media.</p> <p>Default value is 1×10^8 which is the generally accepted factor for most bacterial cell suspensions such as E. coli.</p> <p>Tip: The factor is wavelength specific for each cell type and can be affected by the type of media used for the measurements. Ideally, the factor should be determined empirically using a solution of the study cells at a known concentration using the same media.</p>

Related Topics

- [Instrument Settings](#)

Calculations

Similar to the nucleic acid applications, the OD600 application uses a [modification of the Beer-Lambert equation](#) to calculate sample concentration where the extinction coefficient and pathlength are combined and referred to as a “factor.”

The OD600 application offers a user-specified factor, to be used in conjunction with Beer's Law to calculate sample concentration. If the factor is known, enter the factor. Otherwise, use 1×10^8 , which is the generally accepted factor for most bacterial cell suspensions such as *E. coli*.

Calculated cell concentrations are based on the absorbance value at 600 nm, the entered factor and the sample pathlength. A single-point absorbance correction may be applied.

Measured Values

A600 absorbance

Note: For micro-volume absorbance measurements and measurements taken with nonstandard (other than 10 mm) cuvettes, the spectra are normalized to a 10 mm pathlength equivalent.

- Cell culture absorbance values are measured at 600 nm using the normalized spectrum. If no Absorbance Correction is specified, this is the reported A600 value and the value used to calculate cell concentration.
- If an [Absorbance Correction](#) is specified, the normalized and (absorbance) corrected absorbance value at 600 nm is reported and used to calculate cell concentration.

$A(\lambda)$ absorbance

- Normalized and (absorbance) corrected (if used) absorbance value at any specified [Additional Monitored Wavelength](#) (λ) is also reported.

Sample Pathlength

- For micro-volume measurements, the software selects the optimal pathlength (between 1.0 mm and 0.03 mm) based on sample absorbance at the analysis wavelength.
- For cuvette measurements, the pathlength selected after switching to cuvette mode is utilized. (see [Cuvette Settings](#)).
- Displayed spectra and absorbance values are normalized to a 10 mm pathlength equivalent.

Reported Values

Cell concentration. Reported in cells/mL. Calculations are based on Beer-Lambert equation using corrected A600 absorbance value.

This page is intentionally blank.

Kinetics

Make time-based kinetic measurements using the cuvette holder (NanoDrop Ultra^C and NanoDrop Ultra^C FL model instruments only).

[Manage Kinetics Methods](#)

[Measure Kinetics Method](#)

[Reported Results](#)

[Settings](#)

[Detection Limits](#)



Manage Kinetics Methods


Create kinetics method

1. From the home screen, select the **More Apps** tab, then select **Kinetics**.
2. Select **+ New Method**.
3. Enter **Method Name** and **Description** (if desired), select **Measurement range** (use the drop down to select ranges, shown below) and specify up to three **Wavelengths to monitor**.

Select all applicable cuvette specific settings (pathlength, stir speed, and heating)

The screenshot shows the 'Kinetics Setup' screen with two tabs: 'Name & Settings' (active) and 'Stages & Intervals'. The 'Method Information' section includes a 'Method Name' field with 'Method 1' and a 'Description' field with a placeholder 'Tap to enter description'. The 'Measurement Range' section features a dropdown menu set to 'UV range' and two input fields for '190 nm' and '350 nm'. The 'Wavelengths (nm) to monitor' section has three input fields for '250', '260', and '325'. The 'Default Cuvette Settings' section includes 'Pathlength' (1mm), 'Stir Speed' (OFF), and a toggle for 'Heat cuvette to 37°C'. At the bottom are 'Cancel', 'Save as', and 'Save' buttons. Two orange arrows point to the 'Measurement Range' dropdown and the 'Wavelengths' input fields.

4. Select the [Stages and Intervals](#) tab. The stages and intervals settings are displayed.




5. Select **Number of stages** and **Time unit** (minutes or seconds)
6. For each stage, specify **# intervals**, **Interval times** and any **Delays** between stages

The colored rows and boxes at the right visually represent the specified stages. The colored **rows** show the start and end times for each stage; the colored **boxes** correspond with the specified delay and number of intervals for each stage.

7. To save the method and return to the Kinetics menu, select **Save**.


Note The **Method Name**, **Interval Time**, and **# Intervals** must be defined before you are able to save the method.

View or edit kinetics method

1. From the Kinetics screen, select  next to the desired method to edit.
2. Select **Edit**.
3. View and adjust the [method settings](#) as desired.
4. Select **Save**.


Copy kinetics method

To create a kinetics method that is similar to an existing one, open the existing method, make your changes, then select Save As and enter a new name.

1. From the Kinetics screen, select  next to the desired method to copy.
2. Select **Edit**.
3. View and adjust the [method settings](#) as desired.
4. Enter new **Method Name** and **Description**.
5. Select **Save As**.
6. Enter a new name for the method and click **Save**.

Export kinetics method

Export a kinetics method in order to run it and store the measurement results on another NanoDrop Ultra instrument or another computer running the NanoDrop Ultra PC control software.

1. From the Kinetics screen, select  next to the desired method to export.
2. Select **Export** (if method is invalid, an error message is displayed; errors must be fixed before method can be exported).
3. Select an available export location from the drop-down menu and choose **Export** (method is exported to method file (*.method filename extension) in proprietary format).

To transfer the method to another NanoDrop Ultra instrument or NanoDrop Ultra PC control software, copy the method file to a USB memory device and then load the method (see [Import kinetics method](#) for details).

Import kinetics method

Import a kinetics method back to a NanoDrop Ultra instrument or a computer running the NanoDrop Ultra PC control software to run or edit the method.

1. From the **Kinetics** screen, select **Import**.
2. Locate and select ".method" file.
3. Select **Import** (imported method is added to end of the method selection pane).

Delete kinetics method

1. From the **Kinetics** screen, select  next to the desired method to delete.
2. Select **Delete**.
3. After the confirmation message, select **Yes**.

Measure Kinetics

The NanoDrop Ultra^C and NanoDrop Ultra^C FL model instruments can be used to make time-based kinetic measurements on samples in cuvettes. Up to 3 wavelengths between 190 nm and 850 nm can be designated for continuous absorbance monitoring at user-defined intervals in up to 5 stages. Cuvette measurements offer an extended lower [detection limit](#) and an optional 37 °C heater and micro-stirrer.

Note The instrument arm can be up during cuvette measurements, which allows you to add reagents to the sample solution if desired.


To make kinetic measurements

NOTICE

- To prevent damage from spills, keep containers of liquids away from the instrument.
- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.

Procedure

1. From the home screen, select the **More Apps** tab, then select **Kinetics**.
2. In the method selection pane, select the method to run.



Information about the selected method appears in the method details pane.


3. Select .

Note: If your cuvette pathlength is not 10 mm, specify the correct pathlength in [Kinetics Settings](#).

4. Measure a blank:

- Fill clean, dry cuvette with enough blanking solution to cover [instrument optical path](#)
- Lift instrument arm and insert blanking cuvette into cuvette holder, making sure to align light path of cuvette with light path of instrument
- Select **Blank**

If **Heat Cuvette to 37 °C** is selected in see [Kinetics Settings](#), a message tells you the current temperature and waits for the heater to reach the target temperature before starting the measurements:



To override the wait and start the blank measurement immediately, tap **Blank Now.**

- Wait for blank measurement to complete and then remove cuvette

Note: The heater target temperature is not adjustable.

5. Measure a sample:

- Fill clean, dry cuvette with enough sample solution to cover [optical path](#)
- Insert sample cuvette into cuvette holder, making sure to align light paths
- Select **Measure**

If **Heat Cuvette to 37 °C** is selected in [Kinetics Settings](#), a message tells you the current temperature and waits for the heater to reach the target temperature before starting the measurements:

Note: You may add reagents to the sample solution at any time during the measurement

Use the **Pause and Resume** button at the bottom of the measurement screen to pause and resume the experiment (if you need to end the experiment early, select **Stop**)



8 More Applications

Kinetics

- Wait for all measurement stages to complete
- Remove cuvette and clean it according to manufacturer specifications

Results for each measurement in each interval are displayed in real time. When all stages are completed, the [spectra and reported values](#) for the entire experiment are displayed.

6. When you are finished reviewing the data, select **End Experiment**. Each saved experiment contains one complete set of kinetic measurements based on the selected method.


Related Topics

- [Measure a Sample Using a Cuvette](#)
- [Best Practices for Cuvette Measurements](#)
- [Prepare Samples and Blanks](#)
- [Basic Instrument Operations](#)

Reported Results

Absorbance measurement screen (local control)

The absorbance measurement screen appears immediately after you select **Measure** in the kinetics experiment. This screen shows the absorbance spectrum for each measurement, with wavelength on the X-axis and absorbance on the Y-axis. Vertical lines indicate the specified wavelengths to monitor. The list at the right shows the time each measurement was taken in each specified stage (drag the tab down and up to see more entries). Each item in the list at the right has a corresponding absorbance spectrum at the left. The following image highlights the available features on the NanoDrop Ultra local control software.



Note For measurements taken with nonstandard (other than 10 mm) cuvettes, the spectra are normalized to a 10 mm pathlength equivalent.

Rate measurement screen (local control)

To see the Rate measurement screen, swipe the absorbance measurement screen (see above) to the left. The Rate measurement screen shows a sample's absorbance measured at each user-defined wavelength over time, with time on the X-axis and absorbance on the Y-axis. Measurements taken at each specified wavelength are presented in a unique color. A key showing the monitored wavelengths and their assigned colors appears in the upper left corner of the screen.

User-defined wavelengths

Method 1

Measure Stage 1 / 3 | Measurement 68/101

● 250 nm ● 260 nm ● 325 nm

Show delta absorbance vs time

#	Stage	Time(seconds)
62	1	125.00
63	1	127.00
64	1	129.00
65	1	131.00
66	1	133.00
67	1	135.00
68	1	137.00

10 mm Absorbance

Time (s)

Pause Stop End Experiment

Tap to update plot to show change in measured absorbance over time (see example below)

Swipe screen left to view data table; swipe right to return to absorbance measurement screen

Tap to end experiment and export data

Use the slider button next to **Show Delta Absorbance Vs Time** to show the change in measured absorbance over time, where each data point is the difference in absorbance from the previous measurement.

User-defined wavelengths

Method 1

Measure Stage 1 / 3 | Measurement 36/101

250 nm 260 nm 325 nm

Delta 10 mm Absorbance

Time (s)

Show delta absorbance vs time

#	Stage	Time(seconds)
30	1	61.00
31	1	63.00
32	1	65.00
33	1	67.00
34	1	69.00
35	1	71.00
36	1	73.00

Pause Stop End Experiment

Select to update plot to show measured absorbance over time (see example above)

Swipe screen left to view data table; swipe right to return to absorbance measurement screen

Select to end experiment and export data

Data Table (local control)

To see the data table, in local control, swipe the rate measurement screen to the left. Each row in the table shows the absorbance values at all user-defined wavelengths at a given stage and time. Scroll down to see measurement information that is out of view. The image below highlights the available features.

Tap to end experiment

Stage

Measurement time (click to specify unit)

Absorbance values for each user-defined wavelength

Method 1

Measure Stage 1 / 3 | Measurement 84/101

#	Date	Stage	Time(seconds)	A250	Δ A250	A260	Δ A260	A325	Δ A325	Path
78	11/04/2024 07:42:20 PM	1	157.00	0.127	-0.001	0.137	-0.002	0.027	-0.002	10 n
79	11/04/2024 07:42:21 PM	1	159.00	0.131	0.004	0.141	0.004	0.031	0.004	10 n
80	11/04/2024 07:42:23 PM	1	161.00	0.130	-0.001	0.139	-0.002	0.027	-0.004	10 n
81	11/04/2024 07:42:26 PM	1	163.00	0.128	-0.002	0.138	-0.001	0.026	-0.001	10 n
82	11/04/2024 07:42:28 PM	1	165.00	0.130	0.002	0.140	0.002	0.027	0.001	10 n
83	11/04/2024 07:42:29 PM	1	167.00	0.131	0.001	0.141	0.001	0.030	0.003	10 n
84	11/04/2024 07:42:32 PM	1	169.00	0.130	-0.001	0.140	-0.001	0.029	-0.002	10 n

Press and hold row to view measurement details

Pause Stop End Experiment

Swipe screen right to return to Rate measurement screen

Tap to end experiment and export data

Absorbance measurement screen (PC control)

For each measured sample and standard, this application shows the visible absorbance spectrum and a summary of the results. The layout of the measurement screen of the PC control software differs slightly from the local control. See [“Measurement Screen Display Options”](#) on [page 299](#) for an example.

Reported values

The initial screen that appears after each measurement shows a summary of the reported values. To view all reported values, press and hold the sample row. Here is an example:

Sample Details
✕

#:	15	Measured Temperature:	-
Date:	11/04/2024 07:22:09 PM	Target Temperature:	-
Stage:	3	Method Details	
Time(seconds):	43.20	Method Name:	Method 1
A250:	0.179	Measurement Range:	190 - 350(nm)
Δ A250:	0.000	Wavelength to Monitor:	250, 260, 325 (nm)
A260:	0.205	Number of stages:	3
Δ A260:	0.000	Time Unit:	Time(Minutes)
A325:	0.033		
Δ A325:	0.000		
Pathlength Used:	10 mm		
Stirrer:	OFF		
Heater:	OFF		

	Delay	Interval Time	# Intervals	Duration
Stage1	3	10.00	1	13.00
Stage2	0	5.00	2	10.00
Stage3	0	2.00	10	20.00

Close

Related Topics

- [Basic Instrument Operations](#)

Settings

These settings are available for creating kinetics methods.

Tab	Setting	Description
Name and Settings	Method name	Enter a name for this method (this name appears in the Kinetics Setup box after the method has been saved).
	Description	Enter a detailed description for this method, if desired, such as the type of samples, added reagents, etc.
	Measurement range	<p>Select the spectral range in which this method will acquire data. Available options:</p> <ul style="list-style-type: none"> • Ultra-violet only (190 nm - 350 nm) • Visible only (350 nm - 850 nm) • Ultra-violet and visible (190 nm - 850 nm) • Custom range (specify starting and ending point in nanometers) <p>Note: For measurements taken with nonstandard (other than 10 mm) cuvettes, the spectra are normalized to a 10 mm pathlength equivalent.</p>
	Monitored wavelengths	<p>Enter up to 3 wavelengths to be measured and reported at run time.</p> <p>Note: All specified wavelengths must fall within the selected measurement range.</p>
	Pathlength:	Enter cuvette pathlength (width) before taking blank or sample measurements with cuvettes (see cuvette manufacturer for cuvette specifications)
	Stir Speed:	If using automatic stirring, drop micro-stir bead into sample cuvette and set Stir Speed (levels 1 through 9 correspond with range from 10 RPM to 850 RPM with controlled ramping from zero)
	Heat cuvette to 37 °C:	<p>Select this option if sample cuvettes require heating.</p> <p>Cuvette heater increases from room temperature to 37 °C at rate of 5 °C/minute</p>

Tab	Setting	Description
Stages and Intervals	Number of Stages	<p>Specify up to 5 stages for kinetic measurements. Each stage can have unique Delay, Interval Time and # Intervals settings.</p> <p>Note: Many kinetic measurements include only one stage. Additional stages are necessary only when a variation in stage interval or duration is needed.</p>
	Time Unit	Select the unit for time-based measurements (seconds or minutes).
	Stage 1, 2, etc.	<p>Specify the available settings for each stage:</p> <ul style="list-style-type: none"> • Delay. Specify a delay before a stage starts. • Interval Time. Specify the length of time between measurements taken during this stage (minimum is 2 seconds). The first measurement occurs when the stage starts (or after the delay is completed if a Delay is specified). <p>Note: If two or more stages are specified with Delay set to zero, two measurements occur at the same time (the measurement at the beginning of the new stage directly overlaps the one at the end of the previous stage).</p> <ul style="list-style-type: none"> • # Intervals. Specify the number of absorbance measurements to take in this stage. <p>Note: Since the first measurement is taken when the stage starts, the number of measurements reported for each stage will be the # Intervals setting plus 1.</p> <ul style="list-style-type: none"> • Duration. Readout shows the total time required for this stage, including any delay and all specified intervals.

Tab	Setting	Description
-----	---------	-------------

The colored **rows** at the right (see image below) show the start and end times for each stage; the colored **boxes** at the right correspond with the specified delay and number of intervals for each stage.



If no delay is specified, absorbance measurements are taken at the start and end of each stage and after each specified interval. If a delay is specified, as in stage 3 above, the first measurement occurs at the start of the first interval. If the unit is seconds in the example above, a total of 11 measurements are taken at the following times over a period of 32 seconds:

- Stage 1: 0, 3, 6 and 9 seconds
- Stage 2: 9, 11, 13, 15 and 17 seconds
- Stage 3: 22, 27 and 32 seconds

Note: Kinetic experiments are limited to 1000 measurements. This means the total number of measurements from all intervals in all stages must be less than 1000. Consider available instrument or computer disc space for lengthy experiments.

Related Topics

- [Instrument Settings](#)

Learning Center

Contents

- [Micro-Volume Sampling—How it Works](#) 276
- [Measure a Micro-Volume Sample](#) 279
- [Measure a Sample Using a Cuvette](#) 284
- [Prepare Samples and Blanks](#) 287
- [Basic Instrument Operations](#) 293
- [Acclaro Sample Intelligence](#) 329
- [Instrument Settings](#) 336
- [About Instrument](#) 356
- [PC Control Software](#) 356

Micro-Volume Sampling—How it Works

Surface Tension

Absorbance Spectrum



Sample Absorbance

Sample Concentration

Baseline Correction

Sample Fluorescence


Fluorescence Sample
Concentration



Surface Tension

The NanoDrop Ultra Spectrophotometers and Fluorometers use surface tension to hold a small volume of sample between two pedestals. The patented sample retention system enables the measurement of highly concentrated samples without the need for dilutions.

A fiber optic cable embedded in the upper pedestal leads to a xenon light source. A second cable embedded in the lower pedestal leads to a detector. When the instrument arm is down, the sample forms a liquid column, essentially bridging the gap between the two fiber optic cables.



Absorbance Spectrum

The light passes through the liquid column to the detector, which generates a spectrum of absorbance versus wavelength. The spectrum shows the amount of light absorbed by the molecules of the sample at each measured wavelength.

Note: To prevent evaporation, which affects measurement accuracy, close the arm quickly after you finish loading a sample or blank.

The example at the left shows a typical absorbance spectrum taken of a nucleic acid sample. The spectrum is measured from 190 nm to 850 nm. The displayed range may vary for each application.

Sample Absorbance

When the instrument is blanked, a reference spectrum is taken of the blanking solution and stored in memory. For each sample measurement, the sample intensities along with the blank intensities are used to calculate the total absorbance of the sample according to the equation at the left.

$$\text{Absorbance} = -\log \left[\frac{\text{intensity}_{\text{sample}}}{\text{intensity}_{\text{blank}}} \right]$$

Beer-Lambert equation

$$A = \epsilon \cdot b \cdot c$$

where:

A = absorbance in absorbance units (A)

ε = wavelength-dependent molar absorptivity coefficient (or extinction coefficient) in liter/mol-cm

b = pathlength in cm

c = analyte concentration in moles/liter or molarity (M)

Sample Concentration

The Beer-Lambert equation (Beer's law) shown at the left is used to correlate sample absorbance with concentration.

The pathlength is the distance between the two pedestals, which varies in real time during each measurement. This auto-ranging pathlength technique produces accurate concentration results over a wide dynamic range.

Baseline Correction

For some applications, the instrument can be set up to apply a baseline correction to each measurement to minimize any offset caused by light scattering particulates in the sample spectra. The correction subtracts the absorbance value at a reference wavelength that is close to zero from the absorbance value at each wavelength across the spectrum, essentially “anchoring” the spectrum to zero absorbance units at the reference wavelength.

Sample Relative Fluorescence


The Red or Blue LED is used as the light source, the light passes through an excitation filter, through the liquid column, to the detector, which records the intensity of the fluorescence signal. This is then converted to relative fluorescence units.

Fluorescence Sample Concentration

A curve-fitting algorithm which requires a minimum of two standards is used in the calculation of concentration data. The NanoDrop Ultra dsDNA BR and RNA BR Fluorescence Assays both use a modified Hill plot.

Measure a Micro-Volume Sample

The NanoDrop Ultra spectrophotometers and fluorometers use surface tension to hold a small volume of sample between two pedestals. The patented sample retention system enables the measurement of highly concentrated samples without the need for dilutions. [Tap here](#) for details.




Supplies needed

- NanoDrop Ultra Spectrophotometer (Ultra, Ultra^C, Ultra FL, Ultra^C FL)
- lint-free laboratory wipes
- calibrated precision pipettor (0–2 μ L)
- sample material resuspended in appropriate buffer solution (see [Preparing Samples](#))
- pure buffer solution for blanking instrument (see [Choosing and Measuring a Blank](#)).

Best practices for micro-volume measurements

Cleaning pedestals for daily operation

- Before first measurement, clean both pedestals with a new laboratory wipe.
- [Run a blanking cycle](#) to verify pedestals are clean.
- After each measurement, clean both pedestals with new wipe to prevent carryover.
- After each set of measurements, clean pedestals with DI H₂O (see [Clean pedestals between users](#))
- [Recondition pedestals](#) periodically to maintain their hydrophobic property.



Pipetting Samples

- Use [recommended sample volumes](#) to ensure proper liquid column formation.
- Use calibrated precision pipettor (0–2 μL volume range) with well-fitting, low-retention precision tips to apply sample material to instrument for measurement.

If using low accuracy (0-10 μL) pipettor, use 2 μL sample volumes.

- Filter tips are not recommended as filter particulates can impact absorbance measurements at 230 nm
- Use new tip for each blank and sample aliquot.
- Use new aliquot of sample for each measurement.
- When the measurement is complete, open the sampling arm and wipe the samples from both the upper and lower pedestals using a soft laboratory wipe.
- If solvents are used, make sure they are compatible with the pedestals. (see “Compatible Solvents” in [Hazardous Materials](#)).



Recommended sample volumes

Application	Sample Volume
Nucleic acid (aqueous solution)	1 μL ^a
Purified protein	2 μL
Other protein applications such as Bradford or BCA	2 μL
Microbial cell suspensions	2 μL
Fluorescence	2 μL

^a Use 2 μL for samples that contain materials that may reduce surface tension such as a surfactant.

To measure a micro-volume sample

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.



1. From the instrument home screen, select an application from one of the application categories, such as **dsDNA**.



2. Lift the instrument arm and clean the upper and lower pedestals with new laboratory wipe.




3. Measure a blank:
 - Pipette 1–2 μL blanking solution onto the lower pedestal and quickly lower the arm
 - Select **Blank** and wait for the measurement to complete

Tip: If **Auto-Blank** is On, blank measurement starts automatically after you lower the arm.

 - Lift the arm and clean both pedestals with a new laboratory wipe

9 Learning Center

Measure a Micro-Volume Sample




4. Measure the first sample:
 - Pipette 1-2 μL sample solution onto the pedestal and quickly lower the arm (see [Recommended Sample Volumes](#) for more information)
 - Start the sample measurement:
 - if [Auto-Measure](#) is On, lower arm
 - if Auto-Measure is Off, lower arm and select **Measure**
 - When the sample measurement is completed, the spectra and reported values are displayed.



Tap to end experiment

5. To measure another sample:
 - Lift the arm
 - Clean both pedestals with new wipe
 - Load the next sample and quickly lower the arm
 - Start the sample measurement
 - Wait for the measurement to complete

The new spectrum replaces the previous one on the spectral display and the new reported values appear under the previous ones in the table. (When measuring many samples, press and hold the chart area while moving up or down to show information for samples that may be offscreen.)



6. When you are finished measuring samples:

- Select End Experiment (see previous image)
- Enter an experiment name (tap **Experiment Name** box to display keyboard), or leave the default experiment name
- Select **Next**.
- Lift the arm and clean both pedestals with a new wipe

If finished with the instrument for the day, clean the pedestals with DI H₂O (see [Clean pedestals between users](#))

Acquired data are automatically saved in an experiment with the entered name. In the default configuration, experiments are stored in a database on the local instrument according to acquisition date, experiment name, [application used](#) and any assigned labels (see [Manage identifiers on the instrument](#)).

Measure a Sample Using a Cuvette

The NanoDrop Ultra^C or NanoDrop Ultra^C FL spectrophotometer models include a cuvette holder for measuring dilute samples, colorimetric assays, cell cultures and kinetic studies. The cuvette system offers an extended lower [detection limit](#) and an optional 37 °C heater and micro-stirrer.

Supplies needed

- NanoDrop Ultra^C or NanoDrop Ultra^C FL
- lint-free laboratory wipes
- two [compatible cuvettes](#)
- sample material resuspended in appropriate buffer solution (see [Preparing Samples](#))
- pure buffer solution for blanking instrument (see [Choosing and Measuring a Blank.](#))



Best practices for cuvette measurements

- The instrument arm can be up or down for cuvette measurements.
- Use 10 mm, 5 mm, 2 mm or 1 mm cuvettes.
- Clean and dry cuvette after each measurement.
- Use cuvettes that are free of scratches and avoid fingerprints which may affect results.
- Use quartz cuvettes or UV-grade plastic cuvettes to measure samples with analysis wavelengths in the UV range (<340 nm).
- Micro, semi-micro, and ultra-micro cuvettes should be masked.
- Fill cuvettes with enough blanking or sample solution to cover instrument optical path (2 mm sample beam is 8.5 mm above cuvette bottom).
- Lift instrument arm and make sure cuvette holder is free of debris.
- When inserting quartz or masked plastic cuvettes, align cuvette light path with instrument light path.




To measure a sample using a cuvette

NOTICE


- To prevent damage from spills, keep containers of liquids away from the instrument.
- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.

9 Learning Center


Measure a Sample Using a Cuvette




1. Open one of the applications and select **Save**.
2. The dropdown will default with **Pedestal** selected, open the dropdown and select **Cuvette** to switch to the Cuvette mode. The dropdown must then be closed for the change to take effect.



3. Specify the cuvette options:
 - Set Pathlength to pathlength (width) of cuvette (see cuvette manufacturer for specifications)
 - Set Stir Speed and heater if desired
 - Select anywhere outside of the box, to close the window



4. Measure a blank:
 - Fill clean, dry cuvette with enough blanking solution to cover **instrument optical path**
 - Lift instrument arm and insert blanking cuvette into cuvette holder, making sure to align light path of cuvette with light path of instrument
 - Select **Blank** and wait for the measurement to complete



5. Measure a sample:
 - Fill clean cuvette to same height with sample solution
 - Replace blanking cuvette with sample cuvette, making sure to align light paths
 - Select **Measure**
 - Wait for measurement to complete
 - Remove cuvette
 - Clean cuvette according to manufacturer specifications

Cuvette Settings

Select cuvette sampling mode. When selected, these additional options are available:

Pathlength: Enter cuvette pathlength (width) before taking blank or sample measurements with cuvettes (see cuvette manufacturer for cuvette specifications)

Stir Speed: If using automatic stirring, drop micro-stir bead into sample cuvette and set Stir Speed (levels 1 through 9 correspond with range from 10 RPM to 850 RPM with controlled ramping from zero)

Heat cuvette to 37 °C: Select this option if sample cuvettes require heating. Cuvette heater increases from room temperature to 37 °C at rate of 5 °C/minute.

Prepare Samples and Blanks

Preparing Samples

- Isolate and purify samples before measuring them with the instrument. Commercial sample isolation kits are available for these purposes, or use an in-house protocol. After purification, analyte of interest is typically dissolved in aqueous buffer solution before it is measured.


Tip: Any molecule that absorbs light at analysis wavelength will contribute to total absorbance value used to calculate sample concentration.
- Ensure final analyte concentration is within instrument's [absorbance detection limits](#).
- For micro-volume measurements, gently (but thoroughly) vortex each sample before taking a measurement.

Avoid introducing bubbles when mixing and pipetting.

Note Samples dissolved in extremely volatile solvent such as hexane may work best with [cuvette sampling option](#) (NanoDrop Ultra and NanoDrop Ultra^C FL instruments only).

Choosing and Measuring a Blank


The buffer used to resuspend a sample analyte can contribute absorbance. Blanking minimizes any absorbance contribution due to the buffer components from the sample measurement. The resulting sample spectrum represents the absorbance of only the analyte of interest.



For best results:

- For most applications, blank with the same buffer solution used to resuspend the analyte of interest. The blanking solution should be a similar pH and ionic strength as the analyte solution.
- Measure new blank before each set of samples. It is not necessary to blank the instrument before each sample measurement unless the samples are dissolved in different buffer solutions.
- Measure a new blank every 30 minutes.
- Run a [blanking cycle](#) to assess the suitability of your blanking solution before using it to perform sample measurements.

The resulting spectrum should vary no more than 0.02 A or 0.4 A (10 mm equivalent) depending on the application across the spectrum, especially at the analysis wavelength as in the example below.




Good blanking buffer (measured abs < 0.02)

If the resulting spectrum is greater than 0.02 A or 0.4 A (10 mm equivalent) depending on the application around the analysis wavelength, that buffer solution may interfere with the sample analyses, especially for low concentration samples. See below for details.

Problems associated with blanking

- Residual sample was left on pedestal or in cuvette before blank measurement was performed. (Resulting sample spectra may exhibit negative absorbance values, indicating blank had more absorbance than sample in that region of spectrum.)
- Blank measurement exhibits higher absorbance than unknown sample at analysis wavelength. (If buffer used as blank differs in composition from that used to resuspend sample, measurement results will be incorrect.)
- Sample was inadvertently used to blank instrument. (Resulting sample spectra may exhibit negative absorbance values or, in some cases, resemble a mirror image of a typical pure nucleic acid or protein spectrum as in example below.)



Nucleic acid sample solution used to blank instrument results in “mirror image” spectrum

Solutions for blanking problems

- Thoroughly **clean** and/or **recondition both pedestals** and then:
 - rerun blanking cycle, or
 - measure new blank using new aliquot of appropriate buffer solution, then measure new aliquot of unknown sample

For most applications, blank with the same buffer solution used to resuspend the analyte of interest. The blanking solution should be a similar pH and ionic strength as the analyte solution.

Run a Blanking Cycle

Run a blanking cycle to verify the following:

- instrument is operating normally (with flat baseline)
- pedestals are clean (i.e., no dried-down sample material on pedestals)
- absorbance contribution of buffer solution you plan to use for sample analyses

Supplies needed

- lint-free laboratory wipes
- 0–2 μ L calibrated precision pipettor (low retention tips)
- buffer solution for evaluation

To run a blanking cycle

NOTICE

- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.

9 Learning Center

Prepare Samples and Blanks

1. From the home screen, select an application and save the settings.
2. Lift the instrument arm and clean the upper and lower pedestals with new laboratory wipe.
3. Measure a water blank:
 - Pipette exactly 1 μL deionized water (DI H_2O) onto the lower pedestal and lower the arm.
 - Select **Blank** and wait for the measurement to complete.
 - Lift the arm and clean both pedestals with new laboratory wipe.


4. Measure the buffer solution:
 - Pipette 1-2 μL buffer solution onto the pedestal and lower the arm.
 - Start the sample measurement:
 - if **Auto-Measure** is On, lower arm
 - if Auto-Measure is off, lower arm and select **Measure**
 - Wait for measurement to complete.

The resulting spectrum should vary no more than 0.02 A (or 0.04 depending on the application used) from the baseline at the analysis wavelength.


If your spectrum does not meet these criteria, repeat steps 2–4.

If spectrum is still outside specifications, see [Solutions for Blanking Problems](#).

5. When you are finished with the blanking cycle, select **End Experiment**.
6. Lift the arm and clean both pedestals with a new wipe.



Example spectrum of buffer suitable for Protein A280 protein quantification




Example spectrum of buffer unsuitable for Protein A280 protein quantification

Basic Instrument Operations

- [NanoDrop Ultra Home Screen](#)
- [NanoDrop Ultra Measurement Screens](#)
- [Measurement Screen Display Options](#)
- [NanoDrop Ultra General Operations](#)

NanoDrop Ultra Home Screen


These operations are available from the NanoDrop Ultra home screen.




Applications

The NanoDrop Ultra software offers several configurable applications, which gives users full control of the measurement.

System Status

Select  on the instrument home screen to open the system status box.








The available information is described below.

Instrument type	Instrument model (NanoDrop Ultra)
Serial number	Instrument serial number
Instrument status	Current status of the instrument
Data storage location	Indicates location of database set where instrument is currently storing data.
Ethernet status	Status of Ethernet connections for the instrument (“Connected to...”, or “Not connected”)
Wi-Fi status	Status of WiFi connections for the instrument (“Connected to...”, “Enabled and not connected”, or “Disabled”)
Bluetooth status	Status of Bluetooth connections for the instrument (“Connected to...”, “Enabled-[list of any paired devices]”, or “Disabled”)
Package version	Version of entire package that is installed on the system. Helps manage dependencies and ensures compatibility between firmware, OS, and software versions
Software version	Version of instrument operating software installed
Firmware version	Version of instrument firmware installed



OS version	Version of customized Linux operating system software installed
Hostname	Unique identifier of instrument including serial number and last 4 digits of MAC address

Connectivity Status

Ethernet Connectivity		Shows that the instrument is successfully connected to an Ethernet network.
Wi-Fi Connectivity		Shows that the instrument is successfully connected to a Wi-Fi network.
Bluetooth Connectivity		Shows that the instrument is successfully connected to a Bluetooth device.
Printer Connectivity		Shows that the instrument is successfully connected to a printer.
USB Connectivity		Shows that the instrument is successfully connected to a USB device.


System Alarms and Alerts




The NanoDrop Ultra software will alert users of any potential issues, available software updates, or if any Acclaro Sample Intelligence information is available during an experiment. Depending on the severity of the notification, either the alert icon  or the alarm icon  will be displayed at the top of the screen. Both icons

contain a blue circle with a number indicating the amount of alert or alarm messages that are available. A detailed explanation will appear of all status alerts or alarms by selecting their respective icon. Selecting some individual alerts or alarms may prompt additional information.

The software will also display a banner that provides additional information about the alert or alarm at the time it was triggered. Selecting the banner will provide you additional information.




The alert and alarm banners can be minimized by selecting  to the right of the message.

The alert for software updates will only occur if the instrument is connected to a network with internet access.


Navigation Bar

The following options make up the Navigation Bar. The Navigation Bar can be found on the left side of the screen at various times in the software including on the home screen.


Home Screen

Select  to return to the home screen.


History

Select  to view any data acquired earlier today, last week, last month, last six months, last year or in a specific date range. See [“View History”](#) on [page 313](#) for more information about the History feature on the instrument.


Instrument Settings

Select  to access instrument settings for software updates, cuvette sampling, networking and more. See [“Instrument Settings”](#) on [page 336](#) for detailed information about all available instrument settings.


Instrument Diagnostics

Select  to verify instrument operation. Instrument diagnostics should be run periodically according to the recommended [maintenance schedule](#). See “[Instrument Diagnostics](#)” on [page 367](#) for information about how to run the available instrument diagnostics.

SciVault 2

Select  to open the optional add-on Thermo Scientific™ SciVault™ 2 software. This companion software allows users to operate their NanoDrop Ultra instrument in a manner that complies with US FDA 21 CFR Part 11. When purchased, the SciVault 2 software and activation key is shipped to you on a USB stick and integrates directly into the NanoDrop Ultra software user interface. For more information, visit thermofisher.com/nanodrop.

Help

Select  to launch the interactive help guide.

NanoDrop Ultra Measurement Screens

These operations are available from any measurement screen within an [Application](#).


The screenshot shows the NanoDrop Ultra measurement interface. At the top, there are labels for 'Application' (dsDNA), 'Sample name; tap to edit' (Sample 5), 'Settings' (gear icon), 'qPCR Kit Data' (wavy line icon), and 'Measurement results; see Applications for details'. A 'Sample alert' icon (warning triangle) is in the top right. The main area features a 'UV absorbance spectrum for selected sample' graph showing 10mm Absorbance vs Wavelength (nm) for five samples. Below the graph is a table of measurement results:

#	ng/μL	A260/A280	A260/A230
1	399.3	1.93	2.41
2	327.9	1.86	2.38
3	228.6	1.89	2.45
4	170.6	1.88	2.35

At the bottom, there are buttons for 'Blank', 'Measure', 'Auto-Measure' (person icon), and 'End Experiment'. A page control indicator (two dots) is also present. An orange arrow points from the table to the 'End Experiment' button.

Annotations:

- Application: dsDNA
- Sample name; tap to edit: Sample 5
- Settings: Gear icon
- qPCR Kit Data: Wavy line icon
- Measurement results; see Applications for details
- Sample alert: Warning triangle icon
- UV absorbance spectrum for selected sample: Graph area
- Tap row to select sample and update spectrum; tap more rows to overlay up to five spectra. Press and hold sample row to view sample details.
- Acclaro Contaminant alert; Tap to learn more
- Drag tab down/up to see more/less sample data
- Tap to measure blank solution: Blank button
- Tap to measure sample solution: Measure button
- Auto-Measure: Person icon
- Tap to end experiment and export data: End Experiment button
- Page control; swipe screen left to view table with more measurement results




Measurement Screen Display Options

When performing measurements, right-click the graph on the measurement screen or select the Spectra settings icon **Spectrum Operation** to bring up the following display options:


- Overlay Mode Display multiple samples in overlay.
- Show CrossHairs Hover spectrum to see plot data
- Show Legend Display graph legend
- Autoscale Scale axes to fit spectra measurement
- Find Peaks Calculate peaks for specified range
- Format X-axis Select to manually enter x-axis range
- Format Y-axis Select to manually enter y-axis range

Find Peaks

Select **Find Peaks** to view calculated peaks for specified range. You can enter the range by dragging the color-coded limit lines, or enter values into the fields at the top of the spectrum. Found peaks for the defined range are listed in the table below the spectrum.



Setup

Select  in any measurement screen to see the available setup options for that application.


The types of setup options that appear in the setup screen depend on the selected application. For details, see the settings section of that application in this guide:

Applications > [application group] > Settings

Sample Name

Select the **Sample Name** field in any measurement screen to edit the sample name.

When Auto-Naming is enabled in the application setup screen, each sample is automatically assigned a sample name using the default base name followed by a unique number starting with “1.” The first time this appears is after the first blank measurement and before the first sample measurement in each experiment as shown below.



In this example, the first sample would be named “Sample 1” followed by “Sample 2,” etc. You can edit the default base name and overwrite any sample name.


Note If you edit the sample base name during an experiment when Auto-Naming is selected, the assigned sample ID numbers restart.

Edit default sample base name


After you measure a blank and before the first sample is measured:

1. Select **Sample Name** field to display keyboard
2. Enter new base name
3. Select **Done** key

Edit sample name

1. From home screen, select  to open History

2. Select experiment
3. **Swipe left** to show data table
4. Select the **Sample Name** field to display keyboard



The screenshot shows a table with the following columns: #, Sample Name, $\mu\text{g/mL}$, A260/A280, A260/A230, A260, A280, Factor, Baseline Correction, and Species. The first row contains the sample name 'Hela bracca1 dsDNA 20231018'. The second row is highlighted, and an orange arrow points to the 'Sample Name' field, which is currently empty and has a keyboard cursor over it.

#	Sample Name	$\mu\text{g/mL}$	A260/A280	A260/A230	A260	A280	Factor	Baseline Correction	Species
1	Hela bracca1 dsDNA 20231018	0.2	0.61	0.61	0.61	0.61	50	340	Mammal
2	Sample 2	0.2	0.61	0.61	0.61	0.61	50	340	Mammal

5. Enter new sample name
6. Select **Done** key to close keyboard

Measurement Results

The types of results that appear in the measurement screens depend on the selected application. For details, see the reported results section of that application in this guide:

Applications > [application group] > Measure [application name] > Reported Results

Absorbance Spectrum


For each measured sample taken in one of the applications (Fluorescence applications not included), the UV or UV-Visible absorbance spectrum and a summary of the results is shown. The vertical axis shows absorbance in absorbance units (A). The horizontal axis shows wavelength in nm. Here is an example for a UV-Vis method.

All absorbance-based applications display the sample pathlength along the spectrum's vertical axis. Micro-volume absorbance measurements and measurements taken with nonstandard cuvettes are normalized to a 10.0 mm pathlength equivalent. Here is an example.



Fluorescence Standard Curve

For each measured sample taken in one of the Fluorescence applications, a standard curve and a summary of the results is shown. The vertical axis shows fluorescence in relative fluorescence units (RFU). The horizontal axis shows concentration in ng/μL. Here is an example from the dsDNA Fluorescence application.



Acclaro Measurement Alerts

#	ng/µL	A260/A280	A260/A230
36	257.8	1.82	1.49
37	198.0	1.72	1.22
38	198.5	1.72	1.20
9	190.7	1.80	2.30
10	190.2	1.80	2.29
11	191.5	1.79	2.29

The [Acclaro Sample Intelligence technology](#) built into the NanoDrop Ultra instruments provides important features to help you assess sample integrity. Tap or click a Sample Intelligence icon in the software to view its associated information.



[contaminant analysis](#) is available to help qualify a sample before use in downstream applications



[on-demand technical support](#) is available for measurements that are atypical or very low concentrations

Blank Button

Select **Blank** to measure a blank for the selected experiment.

A blank must be measured before each group of similar samples. The blank solution is typically the pure buffer that was used to resuspend the sample. For more information, see [Choosing and Measuring a Blank](#).

Measure Button

Select **Measure** to measure a sample for the selected experiment.

Samples must be properly isolated and prepared before they can be measured with the instrument and the concentration must be within the instrument's absorbance detection limits. For more information, see [Preparing Samples](#), and [Measure a Micro-Volume Sample](#) or [Measure a Cuvette Sample](#) and [Detection Limits](#).

Note The **Measure** button is enabled after a valid blank measurement is completed.


Auto-Measure and Auto-Blank Options

Speed up sample analysis with the NanoDrop Ultra Auto-Measure and Auto-Blank features, which cause the instrument to start the measurement immediately after you lower the instrument arm. These options eliminate the need for repetitive Measure or Blank operations for large batches of samples.

Note Auto-Measure and Auto-Blank are available for micro-volume measurements only.

Auto-Measure

To select or deselect Auto-Measure, from any sample measurement screen, select the **On** or **Off** button at the right of the Measure button.



Auto-Blank

To select or deselect Auto-Blank, from any blank measurement screen, select the **On** or **Off** button at the right of the Blank button.



End Experiment Button

Select **End Experiment** when you are ready to name and save your experiment, add a label to help you locate the experiment later, print, or export the data. Depending on the administrative settings, you may be prompted to sign the experiment upon ending the experiment.



Note The **End Experiment** button is enabled after the first sample measurement is completed.

After you select End Experiment, the information tab within the End Experiment box is displayed:

The screenshot shows the 'End Experiment' dialog box. It has two tabs: 'Information' (selected) and 'Data Export & Report'. Under the 'Information' tab, there are two input fields: 'Experiment Name:' with the value 'dsDNA 10/18/2024 01:11:19 PM' and 'Tags:' with the placeholder '+ Tap to enter tags'. At the bottom, there are 'Cancel' and 'Next' buttons.

Available options:

Experiment Name Enter a name for this group of measurements. The measurement results are saved in the selected database location using the entered experiment name.


Tags Enter a descriptive label to help you find this experiment later or to associate it with another experiment (see [Manage identifiers on the instrument](#) for details).

When using the local control software, Tap the **Tags** box to display a keyboard to enter the label text.

Press the enter key (PC control) or **Done** (local control) to add the label.

Export Data Select **Next** to move to the Data Export & Report tab within the End Experiment window.

Select **Continue** to open the Data Export & Report window



**Thermo Fisher
User Account**

**Thermo
Fisher PIN**

**Available Cloud
accounts**

Available options (Data Export):

Export File Type The first section allows you to select a file format for exporting the measurements in this experiment.

Available export file formats:

- Customized report
 - comma-separated values spreadsheet (.csv) file
 - tab-separated values spreadsheet (.tsv) file (spectral data only)
- NanoDrop Ultra experiment data (.ndu) file
 - tab-separated values spreadsheet "TQ Analyst" (TQ) file used for troubleshooting purposes

The filename is the entered experiment name (see above). The file is stored in a folder named "NanodropUltra" followed by the instrument serial number. (Use **System Status** to view your instrument serial number.)

Customized Report Options The second section is dedicated to what information will appear within one of the customized report options (.csv or .tsv).

Available customizable options:

- Application name - when enabled will include the name of the application used.
 - Serial number - when enabled will include the serial number of the instrument that was used during the experiment.
 - Report data - when enabled will include information from all columns currently displayed in the report table.
 - Spectral data - when enabled will include absorbance data for displayed spectrum in 0.5 nm increments. For fluorescence-based measurements, this option will display intensity values for the measurement spectrum in 0.3-.0.4 nm increments.
-

Export Location The last section is dedicated to where the export file will be saved.

USB/Network Path-

Select the USB/Network path radio button to export to a USB device connected to any USB port on the local instrument (front, back-left or back-right), or to a [network location](#). When using the NanoDrop Ultra PC control software, use the Browse option to select a File Path to export to.

Cloud-

Select the Cloud radio button to export to one of the available Cloud options: Thermo Fisher Cloud Connect™, Google Drive, or Microsoft OneDrive (see [Cloud](#) for more details). The Thermo Fisher Scientific Cloud Connect™ username and PIN will need to be selected/entered before finally selecting the available Cloud location to export to.

Export Preview Select the **Preview** tab to see what the exported file will look like including any customization that was selected. This feature will not work for the .ndu export file type.


Date	Sample Name	mg/mL	A280	A260/A280	Sample Type	Baseline Correction	Location	333nm	Ext. Coeff. E1%/L/gm-cm	c/1000	M/W(kDa)	Pathlength Used	Contaminant	Corrected	Stirrer	Heater	Me
11/6/2024 3:24:14 PM	Sample 1	2504.3	2504.32	2.23	1 Abs = 1 mg / mL	1180.117A@340nm	Pedestal	-74.729	10.000			@280 0.03 mm	Nucleic Acid	0.000	OFF	OFF	

Cancel (Return to Experiment)

Close the **End Experiment** box and display the results for the most recent measurement. From there you can add measurements to the current experiment and save it later.

Export Button

Export measurement results for current experiment.



Available options (Report)

Customized Report Options Section dedicated to what information will appear within the printable report.

Available customizable options:

- Sample Spectra Display - option to overlay all sample spectra or display up to 4 sample spectra per page.
- Header/Footer Display - option to input a title and sub-title on the report to appear on all pages as well as enter any notes to appear at the end of the report.
- Report data - option to include information from specific columns currently displayed in the report table

Print to PDF Location


Section dedicated to where the export file will be saved.

Select the USB/Network path radio button to export to a USB device connected to any USB port on the local instrument (front, back-left or back-right), or to a [network location](#). When using the NanoDrop Ultra PC control software, use the Browse option to select a File Path to export to.

Select the Cloud radio button to export to one of the available Cloud options: Thermo Fisher Cloud Connect™, Google Drive, or Microsoft OneDrive (see [Cloud](#) for more details). The Thermo Fisher Scientific Cloud Connect™ username and PIN will need to be selected/entered before finally selecting the available Cloud location to export to.

Print Preview

Select the **Preview** tab to see what the printable report file will look like including any customization that was selected.



Cancel (Return to Experiment)	Close the End Experiment box and display the results for the most recent measurement. From there you can add measurements to the current experiment and save it later.
Print button	Select printer to print from and print measurement results for current experiment.
Save to .pdf button	Export the .pdf report for measurement results for current experiment.

Sample Details

Press and hold a **sample row** in any [measurement screen](#) or [data table](#) to show the sample details, which include all available measurement results and associated details for the selected sample. Here is an example:

Sample Details

#:	1	Contaminant:	-
Date:	11/29/2022 4:45:07 PM	Corrected:	-
Sample Name:	Sample 1	Species:	-
ng/μL:	5811.7	Kit name:	TaqMan Fast Advanced Master Mix
A260/A280:	2.19		(0.1 mL plate)
A260/A230:	2.42	Sample Dilution Ratio:	1:1200
A260:	116.23	Dilution Required	
A280:	53.15	Nucleic Acid Volume (μL):	1
Factor:	50	Diluent Volume (μL):	1199
Baseline Correction:	25.972A@340nm	Kit Reaction	
Location:	Pedestal	Nucleic Acid Volume (μL):	1.03
Monitored Wavelength:	0.000	Water Volume (μL):	3.47
Pathlength Used:	@260 0.05mm	Final Template Conc.:	5 ng/μL

Close

Information about the measured values displayed in Sample Details is provided in this Help system, under the [application](#) used to acquire the data.

Note You can also [edit the sample name](#) from the Sample Details box.

Data Table

Swipe left in any measurement screen to see the data table for the current experiment. The data table contains the measurement results for all samples in the experiment. The image below highlights the available features.

Application used

Sample name; tap to edit

Measurement results; see Applications for details

#	Sample Name	µg/mL	A260/A280	A260/A230	A260	A280	Factor	Baseline Correction	Species
1	Hela bracca1 dsDNA 20231018	0.2	0.61	0.61	0.61	0.61	50	340	Mammal
2	Sample 2	0.2	0.61	0.61	0.61	0.61	50	340	Mammal
3	Sample 3	0.2	0.61	0.61	0.61	0.61	50	340	Mammal
4	Sample 4	0.2	0.61	0.61	0.61	0.61	50	340	Mammal
5	Sample 5	0.2	0.61	0.61	0.61	0.61	50	340	Mammal
6	Sample 6	0.2	0.61	0.61	0.61	0.61	50	340	Mammal
7	Sample 7	0.2	0.61	0.61	0.61	0.61	50	340	Mammal
8	Sample 8	0.2	0.61	0.61	0.61	0.61	50	340	Mammal

Tap row to select sample; Press and hold row for sample details

Acclaro Contaminant alert; tap to learn more

Blank Measure End Experiment


Page control; swipe screen right to return to [measurement screen](#)

View History

Whether you collect one sample or many in a row, after you choose End Experiment, the acquired data are automatically saved in an experiment with an experiment name. In the default configuration, experiments are stored in the NanoDrop Ultra database on the local instrument according to acquisition date, experiment name, [application used](#) and any assigned labels.

Use the History feature to open the database in order to view acquired spectra and associated data from any experiment at any time.

Open instrument database of measurement results

To open NanoDrop Ultra database on instrument, tap  (History) on instrument home screen.

History Menu

Available options:

Delete	Delete the experiment
Import	Import data from a USB flash drive or folder on your computer or network hard drive (NanoDrop Ultra PC control software only)
Export	Export the experiment

Search Experiment Database

Use the **Search** feature in History to search the selected database for an experiment or to change the time range or other search filters. The database is filtered using the current settings in the Search box. Filters include time range, application type and any user-defined labels (see [Manage Identifiers](#) for information about adding and deleting labels). Here is an example:

The screenshot shows the 'NanoDrop Ultra History' interface. At the top, there is a search bar with a 'Search' button and a 'Clear' button. To the right of the search bar are two dropdown menus: 'All Applications' and 'All time'. Below the search bar is a table with the following columns: Date, Name, Application, Created User, and Samples. The table contains several rows of data. At the bottom of the screen, there are 'Export' and 'Import' buttons, and a 'Rows' indicator showing '10' and navigation arrows. Two callout boxes with arrows point to specific elements: one points to the 'All Applications' dropdown with the text 'Tap to select or deselect application filters', and another points to the 'All time' dropdown with the text 'Tap to change time range filter'. A third callout box points to the three-dot menu icon in the first column of the table with the text 'Tap to select or deselect user-defined labels'.

Date	Name	Application	Created User	Samples
9/13/2020 07:38:32 PM GMT -08:00	UV-Vis 9/13/2022 07:38:03 PM tags1	UV-Vis	Thermo	1
9/13/2020 07:38:32 PM GMT -08:00	UV-Vis 9/13/2022 07:38:03 PM tags1	UV-Vis	Thermo	1
9/13/2020 07:38:32 PM GMT -08:00	UV-Vis 9/13/2022 07:38:03 PM tags1	UV-Vis	Thermo	1
9/13/2020 07:38:32 PM GMT -08:00	UV-Vis 9/13/2022 07:38:03 PM tags1	UV-Vis	Thermo	1
9/13/2020 07:38:32 PM GMT -08:00	UV-Vis 9/13/2022 07:38:03 PM tags1	UV-Vis	Thermo	1
9/13/2020 07:38:32 PM GMT -08:00	UV-Vis 9/13/2022 07:38:03 PM tags1	UV-Vis	Thermo	1
9/13/2020 07:38:32 PM GMT -08:00	UV-Vis 9/13/2022 07:38:03 PM tags1	UV-Vis	Thermo	1


Tap to select or deselect user-defined labels

Export Selected Experiments

To the left of each experiment an empty checkbox will appear. Select one or more checkboxes in History to select experiments to be exported.


Export selected experiments

- Within History, filter using the **Search** feature to find the desired experiment




Change a filter to display updated list of experiments


- Click or tap the checkbox to select the row or rows containing the desired experiment(s) to be exported (click or tap again to deselect an experiment)



- To select or deselect all experiments in database, click or tap on the checkbox directly below the Search window



- Click or tap **Export**
- Select one or more formats to export to (see “Export Selected Experiments” on page 322)




- Select the applicable details to appear in the exported file
- Designate the export location and select **Export**.
- After "Export Completed" message, select **OK**

Delete Selected Experiments


Select experiments to be deleted.

Delete selected experiments

- Within History, filter using the **Search** feature to find the desired experiment
- Click or tap the checkbox to select the row or rows containing the desired experiment(s) to be deleted (click or tap again to deselect an experiment)
- Select  and **Yes**

Note Deleted data cannot be recovered.

Open Experiment and View Associated Data

Use  (History) to locate and open any experiment to see the measurement data it contains.

Open an experiment


- Within History, if you don't see the experiment you want to open, you can use the Search feature to find the desired experiment
- Double click or tap the experiment name to open the experiment

The History provides measurement data as [spectral data](#), [standard curve data](#), and [data tables](#), similar to what you see after you complete a measurement.

Note The data shown are dependent upon the application used to measure the samples (nucleic acids in these examples). For more information, see the [application](#) details.

Spectral data—

After you open an absorbance-based experiment, the software shows the UV or UV-visible absorbance spectrum and a summary of the associated data for the first sample measurement, much like it appears during a measurement. The image below describes the available features.



Standard curve data—

After you open a fluorescence-based experiment, the software shows the fluorescence standard curve and a summary of the associated data for the first sample measurement, much like it appears during a measurement.

Data Table—

Swipe left in any Spectral or Standard Curve Data screen to see the data table for the current experiment. The data table contains the measurement results for all samples in the experiment. The image below describes the available features.

Return to History

Selected experiment

Measurement results; see Applications for details

#	Date	Sample Name	ng/µL	A260/A280	A260/A230	A260	A280
1	10/23/2024 04:17:29 PM	Sample 1	399.3	1.93	2.41	7.99	4.13
2	10/23/2024 04:18:05 PM	Sample 2	327.9	1.86	2.38	6.56	3.52
3	10/23/2024 04:18:50 PM	Sample 3	228.6	1.89	2.45	4.57	2.42
4	10/23/2024 04:20:45 PM	Sample 5	170.6	1.88	2.35	3.41	1.81


Tap to select unit

Tap row to select sample; Press and hold row for sample details

Acclaro Contaminant alert; tap to learn more

Page control; swipe screen right to view previous screen

Menu

Select  at the top of any Spectral Data, Standard Curve, or Data Table screen to see the available menu options.

Data Export & Report	Export or print opened experiment
Tags	Add or delete labels for selected experiment to make it easier to find (see Manage identifiers on the instrument)
Details	Open experiment details box which contains the experiment name, application type, date the experiment was created, number of measurements contains within the experiment, serial number of the instrument used, and information on the software used during the experiment

NanoDrop Ultra General Operations


These operations are available from any measurement screen or from the [History](#).


Manage Identifiers

You can add one or more tags (i.e., labels or metadata tags) to an experiment to make the experiment easier to find. Labels can be added from the NanoDrop Ultra local control software running on the instrument, or from the NanoDrop Ultra PC control software installed on a personal computer.

Use the History to add labels to experiments, assign existing labels, view assigned labels and remove or delete labels on the instrument. You can filter the list of experiments in the History based on one or more user-defined labels.



Label new experiment when you save it

- After the last sample has been measured, tap 
- In End Experiment window, tap on the field that says "Tap to enter tags." A keyboard will appear
- Enter user defined labels and tap **Done** to close the keyboard and add the tag




- Select **Next** to Export or Print
- Select **Finish** to close the experiment

Label experiment in History


- From the home screen, select  to open History
- Select  to open the menu options of the experiment row and select **Tags**

- in Manage Tags window, tap on the field that says "Tap to enter tags." A keyboard will appear
- Enter user defined labels and tap **Done** to close the keyboard
- Select **Save** to confirm the changes and close the Manage Tags window. Add additional tags as needed




View assigned labels for an experiment

- From the home screen, select  to open History
- The tag(s) will appear below the experiment name

Find labeled experiments

- From the home screen, select  to open History
- Enter the tag name into the **Search** bar


Remove a label


- From home screen, select  to open History
- Select  to open the menu options of the experiment row and select **Tags**
- In Manage Tags window, select  for any tag you wish to delete
- Select **Save** to confirm the changes and close the Manage Tags window

Edit Experiment Name

You can edit the experiment name when you save the experiment or afterwards from the [History](#).


Edit experiment name at end of experiment

- When finished measuring samples, select 
- Enter a name for this group of measurements in the **Experiment Name** field



- Select **Next**
- Select **Continue** to export or print the experiment
- Select **Finish** to close the experiment

Edit experiment name from History

- From home screen, tap  to open History
- Use [Search](#) feature to find experiment
- Double click or tap the experiment to open experiment select the **Experiment Name** field to display keyboard
- Enter new experiment name
- Tap **Done** key to close keyboard and save the new name

Export Selected Experiments

You can export measurement data when you save the experiment or afterwards from the [History](#).

Note Data exported during a save are still saved to a database (local or remote, depending on the Data Storage setting).

Measurement data can be exported in three formats:

- As comma-separated values (.csv) files containing one or more of the following:
 - whole spectrum data in columns
 - report data in columns
 - application name
 - serial number
- As tab-separated values (.tsv) files containing one or more of the following:
 - whole spectrum data in columns
 - report data in columns
 - application name
 - serial number
- As NanoDrop Ultra software (.ndu) files containing spectra and measurement results for each exported experiment

Use any spreadsheet or word processing application to open a CSV or TSV file. Here is an example of several sample measurement results in CSV format:


#	Date	Sample Name	mg/mL	Original Conc.	Dilution Factor	RFU	Average RFU	# of Replicates	Excitation LED
1	10/16/2024 9:01	Reference	0	0	1		105.603	1	Blue
2	10/16/2024 9:02	Standard 1	5	5	1		1.742	1	Blue
3	10/16/2024 9:02	Sample 1	4.122	4.122	1	19.978			Blue
4	10/16/2024 9:02	Sample 2	3.992	3.992	1	22.678			Blue
5	10/16/2024 9:03	Sample 3	3.87	3.87	1	25.206			Blue

Note The types of data exported are dependent upon the application used to measure the samples (nucleic acids in this example). For more information, see the [application](#) details.

Data can be exported to a USB device connected to any USB port on the local instrument (front, back-left or back-right), a [connected Cloud account](#), or to a [network location](#). If you select multiple experiments for export, each exported experiment has a corresponding file. The filenames are the same as the [experiment names](#). The files are stored in a folder named “NanodropOne” followed by the instrument serial number. (Use [System Status](#) to view your instrument serial number.)


Export data at end of experiment


- When finished measuring samples, select End Experiment
- From End Experiment window, enter Tags and edit experiment name if desired then select **Next** to move to the Data Export & Report tab within the End Experiment window
- Select **Continue** to open the Data Export & Report window
- Select one or more [export file types](#) and one or more customized report options



- Designate the export location
- Select the **Preview** tab to see what the exported file will look like including any customization that was selected. This feature will not work for the .ndu export file type
- Select **Export**
- After "Export Completed" message, select **OK**

Export data from History

- From home screen, tap  to open History
- Select the checkbox for the experiment(s) you want to export. You can use the [Search](#) feature to find experiment
- Select **Export**
- Select one or more [export file types](#) and one or more customized report options
- Designate the export location




- Select **Export**
- After "Export Completed" message, select **OK**

Delete Selected Measurements



You can delete selected sample measurements from any experiment, or all the measurements in the database.

Note Deleted data can only be recovered if a database back up file containing that data was saved prior to the deletion.

Delete data from any measurement screen

- Press and hold sample row to open Sample Details window
- Select 
- Select **Yes** to confirm the deletion of the sample measurement

Delete data from History

- From home screen, tap  to open History
- Select the checkbox for the experiment(s) you want to delete. You can use the [Search](#) feature to find experiment
- Select 
- Select **Yes** to confirm the deletion of the experiment and all sample measurements within that experiment


Print Selected Measurements


Connect a [compatible printer](#) to the instrument to quickly print measurement results, including spectral data, standard curves, data tables, sample details and diagnostic results. You can print to a USB printer (label or full service) or to a remote printer through an Ethernet connection or wireless network.

Note

- A wireless printer or the device it is connected to must be available on the same wireless network as the instrument. The wireless printer must also have its wireless function enabled.
- Full service printer options are not available if you have a label printer connected. Disconnect the label printer to access the full service printer options.

Print data at end of experiment



- When finished measuring samples, select 
- From End Experiment window, enter Tags and edit experiment name if desired then select **Next** to move to the Data Export & Report tab within the End Experiment window
- Select **Continue** to open the Data Export & Report window
- Select the **Report** tab
- Select one or more [customized report options](#)




- Designate the print to PDF location if necessary
- Select the **Preview** tab to see what the printed report will look like including any customization that was selected
- Select **Print** to select an available printer to print to or select **Save to .pdf** to export a .PDF file to the specified location
- After "Export Completed" message, select **OK**

If a label printer is connected to the instrument, the software prints one label for each selected measurement. If a full service printer is connected, the selected measurement screen is printed for each selected measurement.

Print data from History

- From home screen, select  to open History
- Select the checkbox for the experiment(s) you want to print. You can use the [Search](#) feature to find experiment
- Double click or tap the experiment name to open the experiment
- Select  at the top of any Spectral Data, Standard Curve, or Data Table screen to see the available menu options and select **Data Export & Report**
- Select the **Report** tab
- Select one or more customized report options



- Designate the print to PDF location if necessary
- Select the **Preview** tab to see what the printed report will look like including any customization that was selected
- Select **Print** to print to a connected printer or select **Save to .pdf** to export a .PDF file to the specified location
- After "Export Completed" message, select **OK**

If a label printer is connected to the instrument, the software prints one label for each selected measurement. If a full service printer is connected, the selected measurement screen is printed for each selected measurement.

Print sample details

- From the [spectral data](#), [standard curve](#), or [data table](#) in any measurement screen or from [History](#), press and hold sample row to open Sample Details box

Sample Details			
#:	1	Contaminant:	-
Date:	11/29/2022 4:45:07 PM	Corrected:	-
Sample Name:	Sample 1	Species:	-
ng/μL:	5811.7	Kit name:	TaqMan Fast Advanced Master Mix (0.1 mL plate)
A260/A280:	2.19	Sample Dilution Ratio:	1:1200
A260/A230:	2.42	Dilution Required	
A260:	116.23	Nucleic Acid Volume (μL):	1
A280:	53.15	Diluent Volume (μL):	1199
Factor:	50	Kit Reaction	
Baseline Correction:	25.972A@340nm	Nucleic Acid Volume (μL):	1.03
Location:	Pedestal	Water Volume (μL):	3.47
Monitored Wavelength:	0.000	Final Template Conc.:	5 ng/μL
Pathlength Used:	@260 0.05mm		

  Close

- Select 
- Select an available printer from the down-down menu and select Print

If a label printer is connected to the instrument, the software prints a label for the selected measurement. If a full service printer is connected, the selected sample details screen is printed

LIMS API Integration

The REST API allows for efficient data exchange, including the submission, retrieval, and updating of data between the instrument and your LIMS. For more information on how to utilize the REST API, please refer to the [NanoDrop Ultra API Reference Manual](#).

Acclaro Sample Intelligence

The Thermo Scientific™ Acclaro™ Sample Intelligence technology built into the NanoDrop Ultra instruments provides these exclusive features to help you assess sample integrity:



contaminant analysis to help qualify a sample before use in downstream applications




on-demand technical support for measurements that are atypical or very low concentration



Use these embedded resources to quickly troubleshoot possible problem measurements and make informed decisions on whether to use, re-purify or take other actions with an atypical sample result. The Sample Intelligence feature also serves as a resource for further study and a learning tool for new or novice users.

Activate Detection


From the measurement setup screen, select the activate contaminant detection option and choose between Mammal, Plant, and Bacteria.



When enabled, RNA Contaminant Detection/DNA Contaminant detection will apply mathematical models to predict the amount of RNA contaminant in dsDNA or dsDNA in RNA for the selected species. These models are specific to the source of the nucleic acid. For example, if mammal is selected, the mathematical model will be representative of all mammalian sources of nucleic acid. If you are measuring nucleic acid from a source for which we do not have mathematical model, leave the selections unchecked.

View Acclaro Sample Intelligence Information

Measurements that include a contaminant analysis or technical information are flagged automatically (see examples below). Select the icon to review the associated data or information. The alert icon and banner will also appear at the top of the screen.



The icons appear next to the [measurement results](#) (see above), and in the [data table](#), as well as in [History](#) (see below).


#	Date	Sample Name	ng/μL	A260/A280	A260/A230	A260	A280
2	10/23/2024 08:19:35 PM	Sample 2	11.4	0.66	0.08	0.23	0.34
3	10/23/2024 08:20:00 PM	Sample 3	106.1	1.65	0.73	2.12	1.29
4	10/23/2024 08:20:23 PM	Sample 4	10.4	0.58	0.07	0.21	0.36


The icons are active in all three places; the information remains with the data indefinitely, even after it has been exported.

Contaminant Analysis

For the dsDNA, RNA and Protein A280 applications, the NanoDrop Ultra software automatically initiates a spectral analysis for several known contaminants during the measurement. Examples of known contaminants include:


- For dsDNA and RNA measurements:
 - in the analysis region: protein and phenol
 - detects presence of guanidine HCl and guanidinium isothiocyanate
 - detects species-specific dsDNA contamination in the RNA application and detects species-specific RNA contamination in the dsDNA application
- For protein measurements:
 - in the analysis region: nucleic acids and phenol

If contaminants are identified in a sample, the “Contaminant Analysis” icon  appears to the left of the measurement results.



Tap the icon to view the contaminant analysis and associated information.

The following is an example of results from a nucleic acid contaminant analysis that contains enough protein contaminant to affect the measurement results.



Original nucleic acid concentration¹

Purity ratio

% coefficient of variation

Corrected nucleic acid concentration²

Protein absorbance value at 260 nm

¹Based on total sample absorbance (sample plus contaminant)

²Based on corrected sample absorbance (sample minus contaminant)

Since proteins absorb light near the analysis wavelengths for nucleic acid (230 nm, 260 nm, and 280 nm), the presence of protein in the nucleic acid sample shown above has pushed the A260/A280 and A260/A230 ratios out of range and caused the reported nucleic acid concentration to be higher than the real value. The software identifies the impurity (protein), and reports the following:

- baseline-corrected absorbance due to protein (0.18) at the analysis wavelength (260 nm)

- % coefficient of variation for the measurement result (uncertainty x 100/measurement result = 14.03%; a high %CV indicates the measurement result is close to the instrument detection limit or there is an interfering component)
- original nucleic acid concentration (59.5 ng/μL), which is based on the total baseline-corrected absorbance (sample plus contaminant) at the analysis wavelength
- corrected nucleic acid concentration (49.5 ng/μL), which is based on the corrected absorbance (sample minus contaminant) at the analysis wavelength

Theory behind contaminant analysis

UV and UV-visible absorbance measurements are used to quantify nucleic acid and protein samples at 260 nm and 280 nm, respectively. The analysis is based on the fact that the total absorbance of a mixture solution at a given wavelength is the sum of the absorbance values of each component in the mixture.

An ongoing challenge of this method is that a number of materials used in the extraction process can absorb in various regions across the spectrum. When these contaminants are present in a sample, they can interfere with the analysis by artificially inflating the absorbance at the wavelength of interest, which causes the analyte concentration to be overestimated.

Traditionally, purity ratios are used to detect the presence of contaminants that could affect downstream applications. However, purity ratios do not always provide a complete picture of possible contamination. When a purity ratio falls outside the expected range, the spectral profile is often examined qualitatively.

Our Acclaro technology applies a quantitative approach to contaminant analysis. Through sophisticated mathematical algorithms, Acclaro analyzes the spectral data to identify probable contaminants in a sample and removes any contribution due to the contaminant from the sample result. This results in a more accurate concentration value of the analyte of interest and a more quantitative analysis of the level of contamination.


Since the spectrum of a pure compound is unique to that compound, a mixture spectrum of mostly known materials that have few interactions can be mathematically broken down into its component spectra and the components identified. The contaminant analysis algorithm uses a narrow spectral region (220-285 nm) around the analysis wavelength (260 nm for nucleic acids, 280 nm for proteins) to determine any absorbance contribution from possible known contaminants (protein or nucleic acid, and phenol) that absorb in that region. The entire spectrum is analyzed to determine the presence of other possible contaminants such as guanidine HCl and/or guanidinium isothiocyanate, which are common reagents used for nucleic acid purification.


Note Achieving consistent, high quality contaminant analysis results is dependent on the quality of the measured sample spectra, which is dependent on the maintenance status of the instrument. For more information, see [Maintenance Schedule](#).

On-Demand Technical Support

For the dsDNA and Protein A280 applications, the NanoDrop Ultra software monitors all sample measurements for the presence of contaminants or other anomalies that may affect the measurement. Examples of monitored characteristics include:


- Absorbance ratios, which indicate the presence of compounds that may interfere with sample measurements (also referred to as “purity ratios”). For more information, see [What is a Purity Ratio?](#).
- Bubble check, which looks for bubbles or other reflective materials in a sample or blank.

If technical information is available, the “information” icon  appears to the left of the measurement results. The alert icon and banner will also appear at the top of the screen.




Tap the icon to view the information.

Here are results from a nucleic acid analysis for which two measured purity ratios are either above or below the expected value and the sample contained enough bubbles to possibly affect the measurement results.



Tap an information button for more information. Here is the information provided for the bubble error:

Bubbles Detected

 [Learn more](#)

These measurement results may be invalid.

Possible causes:


- Bubbles in sample or blank, which can cause poor liquid column formation and diffract UV light. Both issues will impact reproducibility
- Particulates in sample or blank, which can diffract UV light and affect reproducibility


Possible solutions:

- To remove bubbles, briefly centrifuge sample and/or blank at low speed
- To remove reflective particles such as magnetic beads, try extracting the sample again taking care to avoid particle contamination
- For DNA samples, heat to 63 °C for 10 minutes, then centrifuge at low speed. For other sample types, centrifuge at high speed to condense particulates before taking an aliquot

Tap **Learn More** for the next level of information.

Bubbles Detected - Learn More

 [Spectrum of Sample with Bubbles](#)


 [Effects of Bubbles in Sample](#)


Alarm Notifications

The NanoDrop Ultra software uses an embedded image sensor to monitor all measurements for conditions (such as a broken liquid column) that are likely to invalidate the measurement results.

After an invalid-result, the Alarm icon  is displayed and the measurement is stopped. See [Troubleshooting](#) for more information.

Instrument Settings

From the home screen, select . Available options include System Settings, Connectivity Settings, Protein Editor, Kit Editor, Dye Editor, Preferences, Export Settings, SciVault 2 Mode, Update Software, Database Settings, and About Instrument.



System Settings



These options are available:

- General** Increase or decrease screen brightness and volume on instrument by using the respective sliders
- Language** Select language for displaying NanoDrop Ultra software and for any connected input device such as a keyboard, mouse or barcode scanner
- Notice:** Changing the language requires a software restart.
- Date and Time** **Automatic Time Zone & Time:** synchronize instrument time zone, time, and date with available network
- Select Time Zone:** manually select instrument time zone (this option is disabled when Automatic Time Zone & Time is selected)
- Set Time & Date:** manually set instrument time and date (this option is disabled when Automatic Time & Date is selected)
- Choose Date Format:** choose an available date format
- Choose Time Format:** choose an available time format
- Energy Saver** The touchscreen will turn off and a screen saver will be displayed when the instrument has been idle for the selected time. To reactivate the instrument, tap the touchscreen.

Connectivity


These options are available:

- Wi-Fi** Use this tab to enable and specify a Wi-Fi connection for the instrument. The NanoDrop Ultra Wi-Fi and Bluetooth Dongle must be inserted into the instrument, or this feature will be unavailable.
- Bluetooth** Use this tab to enable and specify a Bluetooth connection for the instrument. The NanoDrop Ultra Wi-Fi and Bluetooth Dongle must be inserted into the instrument, or this feature will be unavailable.
- Cloud** Use this tab to connect to the Thermo Fisher Connect Platform which enables direct data export to Thermo Fisher Cloud Connect, Google Drive, or Microsoft OneDrive. Some Cloud platforms may only be available in select global locations.

Wi-Fi


Use this tab to enable and specify a Wi-Fi connection for the instrument. The NanoDrop Ultra Wi-Fi and Bluetooth Dongle must be inserted into the instrument, or this feature will be unavailable.

Setup Wi-Fi network on the instrument

1. From the instrument home screen, select  (**Settings**).
2. Select **Connectivity**, then select the **Wi-Fi** tab at the top.
3. Use the slide toggle to enable Wi-Fi connectivity (turns blue with a checkmark and software automatically searches for any available Wi-Fi network)


If no Wi-Fi networks are found, after a few seconds the message "No nearby Wi-Fi networks were found" is displayed

4. The NanoDrop Ultra instrument will continuously refresh the list of available networks.
5. The Wi-Fi name should appear in the provided list.






6. To connect to a Wi-Fi network, tap its name in Available Network list.

- If the Wi-Fi network is password protected, a window will pop-up requesting entry of the Wi-Fi password. Enter the password and select **OK**.



- A green checkmark will appear next to the name if the pairing was successful.



- When Wi-Fi is enabled and connected, a Wi-Fi icon  will appear at the top of the screen.
- Select back  to return to the Settings screen.

Disconnect Wi-Fi network


To disconnect from a connected Wi-Fi network, select the disconnect icon  and then select **Forget** to confirm.

Bluetooth

Use this tab to enable and specify a Bluetooth connection for the instrument. The NanoDrop Ultra Wi-Fi and Bluetooth Dongle must be inserted into the instrument, or this feature will be unavailable.


Note Make sure the device is labeled "Bluetooth" and not just "wireless." All Bluetooth devices are wireless but not all wireless devices will run with Bluetooth.


Set up Bluetooth connections on the instrument

1. From the instrument home screen, select  (**Settings**).
2. Select **Connectivity**, then select the **Bluetooth** tab at the top.
3. Use the slide toggle to enable Bluetooth connectivity (turns blue with a checkmark and software automatically searches for any available Bluetooth input devices)

If no Bluetooth devices are found, after a few seconds the message "No nearby Bluetooth devices were found" is displayed

4. To add a Bluetooth device, follow the manufacturer instructions to pair the device (for example, you may need to hold down a button). The NanoDrop Ultra instrument will continuously refresh the list of available devices.
5. The device name should then appear in the provided list.



6. To pair a device, tap its name in Available Devices list, a green checkmark will appear next to the name if the pairing was successful.
7. When Bluetooth is enabled and connected, a Bluetooth icon  will appear at the top of the screen.

- Repeat steps above to add another Bluetooth device or select back  to return to the Settings screen.

Note If your Bluetooth device does not pair, restart the device and then repeat the steps above to pair it with the instrument (you may also try turning Bluetooth off and back on). After a device is paired, it remains paired even after the instrument is restarted.

Disconnect Bluetooth device

To disconnect from a connected Bluetooth device, select the disconnect icon  and then select **Forget** to confirm.

Note

- If no Bluetooth device is selected for input, the instrument relies on the integrated touchscreen keyboard for input.
- To select the device again, follow the steps above and select the device's Use for Input checkbox.

Cloud

Use this tab to connect to the Thermo Fisher Connect Platform which enables direct data export to Thermo Fisher Connect, Google Drive, or Microsoft OneDrive.

Create a Thermo Fisher Scientific Cloud Connect Account


During the initial Cloud setup, the instrument needs to be connected to the Thermo Fisher Cloud Connect Platform. A Thermo Fisher account must be used during this process and requires a one-time setup.


- From a PC browse to www.thermofisher.com/cloud (if you are located outside of China) or www.thermofisher.cn/cloud (if you are located in China)
- Select **Sign in** which will display a dropdown menu
- Select either **Sign in** again if you already own a Thermo Fisher account or **Create Account** if you do not already have one and complete the additional steps to create a free account.


Create a Thermo Fisher Cloud Connect PIN

- After signing into your Thermo Fisher account in your preferred browser, select **My Account**

2. Select **Connect: Lab, Data, Apps**.




3. Select the  icon on the left panel then select **Instruments** followed by **Connected Instruments**
4. Select **Change PIN** towards the upper right of the Instrument Connect page




5. Enter a unique PIN number and save this information as you will need it later

Connect to a Thermo Fisher Cloud Account

1. From the instrument home screen, select  (Settings).


2. Select **Connectivity**, then select the **Cloud** tab at the top.





3. Ensure the appropriate Region is selected, the options are either Others or China.

Note When using China as a region, users will not be able to export to a Google Drive due to country limitations.


4. Select Dashboard on Internet Browser to display a unique linking code for the instrument.





5. From a PC with internet access browse to <https://apps.thermofisher.com/>
6. Sign into your Thermo Fisher account in your preferred web browser on a PC with internet access.
7. Select **Connect: Lab, Data, Apps**.
8. Select the  icon on the left panel then select **Instruments** followed by **Connected Instruments**.




9. Select **Link instrument** and enter the unique linking code provided by the instrument.
10. Select **Link Instrument**.
11. The connected Thermo Fisher Scientific account will now appear in the Settings > Connectivity > Cloud tab.



12. Use  to display any Cloud accounts that have been associated with the Thermo Fisher Scientific account.




13. The Thermo Fisher Cloud Connect PIN will be requested, enter and select **OK**.




14. Select the **Connect Account** option and follow steps 5-10 to add additional Thermo Cloud Connect accounts.

Connect to a Google Drive or Microsoft OneDrive Account

The connection to the Thermo Fisher Connect platform must be established before connecting to other cloud platforms, see the above section for more information.


1. From the Thermo Fisher Connect home page, open the settings by selecting  at the top left of the screen.
2. Open the **Instruments** dropdown, then select **Connected Instruments**.

- Use the instrument settings  button below the now connected NanoDrop Ultra and select **Storage Settings**.




- Select **Add account**.


5. Select **Add account** under the desired cloud platform.



6. Grant thermofisher.com access to the cloud drive.
 - Ensure the option **See, edit, create, and delete** all of your Google Drive files is checked off for Google Drive connection; if it is unchecked the files will not be saved to the drive.





7. In the **Cloud** tab within the instrument settings, select the down arrow followed by the PIN to view the newly added cloud drive(s). The newly added cloud drive should be listed as active and will now appear as an export option.




Note As a security measure, access to the connected Google Drive or Microsoft OneDrive account will expire after 90 days. At which point the user will need to perform these steps again to re-setup the account.

Disconnect a Thermo Fisher Cloud Account

1. After signing into your Thermo Fisher account in your preferred browser, select **My Account**.
2. Select **Connect: Lab, Data, Apps** which will take you to the Thermo Fisher Connect home page.
3. Open the settings by selecting  at the top left of the screen.
4. Open the **Instruments** dropdown, then select **Connected Instruments**.

5. Select  at the lower right of the connected instrument profile on the instrument dashboard.



6. Select **Disconnect** followed by **Confirm**, the profile will then disappear from the Cloud tab on the once connected instrument.

Protein Editor

See [“Protein editor”](#) on [page 95](#) for information on using the protein editor


Kit Editor

See [Kit editor](#) on [page 44](#) for information on using the kit editor.

Dye Editor

See [Dye/chromophore editor](#) on [page 57](#) for information on using the dye editor.


Preferences



Set your preferences for remembering certain settings when restarting the application.

Export Setup

These options are available:



Network Paths

Use this tab to specify one or more network paths for exporting acquired data when the instrument is connected to a network (connection can be wired or wireless). Network paths defined here will appear in the USB/Network path drop-down menu when exporting data, from both the [History](#) and the [End Experiment](#) box after you complete a measurement.


Add Network path

1. Select **+ New Path** to add a new entry box for a network path each entry will have a unique ID.
2. Enter the Network Path in the following format: \\IP Address of Server or Server name\Name of the share drive\Name of folder on share drive. (from local control, tap field to display keyboard, tap **Done** key to close keyboard).
3. Enter a unique Path Name which will appear in the USB/Network path drop-down menu when exporting data.


Test Network path

1. After all information has been input, select **Test** to confirm connection and access to the network location.

2. Enter a User name and Password for the network and are trying to connect to and select **OK**.




Edit or Delete Network path

1. The Network Path and Path Name can be edited at any time by selecting each entry box and making any necessary changes.
2. Select  to the right of the row containing the Network Path to delete it.

Auto Export

Select a file path for exported files and enable/disable Auto-export at the end of experiments. A [Network Path](#) must be setup prior to enabling Auto Export. Select the desired network path location, file type(s) to export, and customizable settings. When enabled the software will save a copy of the experiment in the provided file type(s) to the location specified.



SciVault 2 Activation

Setup the optional add-on Thermo Scientific™ SciVault™ 2 software. This companion software allows users to operate their NanoDrop Ultra instrument in a manner that complies with US FDA 21 CFR Part 11. When purchased, the SciVault 2 software and activation key is shipped to you on a USB stick and integrates directly into the NanoDrop Ultra software user interface.

See the SciVault 2 User Guide for information on activating and installing SciVault 2 software.

Update Software


Quickly and easily download and install the latest NanoDrop Ultra software and release notes from our website. Follow the steps to update or upgrade the software on your local instrument and/or install or update the NanoDrop Ultra software on a personal computer (PC). An Internet connection is required to download software.

To install or update NanoDrop Ultra software on a PC

1. Insert the USB flash drive containing the installer software into an available USB port on your PC, or open the installation folder downloaded from the internet.

2. Launch **Start.exe** and click **Install**. The software installer will run.

To install or update NanoDrop Ultra software on the instrument



Through USB:

1. Copy the .zip file with the new software from your computer to a USB storage device. **Do not attempt to unzip the folder.**
2. Insert the USB device into any USB port on the NanoDrop Ultra instrument.
3. From the instrument home screen, tap the Settings icon, then **Software Update**.
4. Select **Update** in the row with the latest version of software.


Through Online Update

Note Instrument must be connected to a network with internet access. When connected to a network with internet access, any time a software update is released, a system alert will be provided to the user to inform them they are able to update.

1. From the instrument home screen, tap the Settings icon > **Software Update**.
2. Select **Download** in the row with "Online update" as the source.
 - Once the download is complete, select **Update** to install the software package.

Database Backup and Restore

You can Backup & Restore the internal measurement database to a specified location of your choosing. From the **Settings** screen, select **Database**.



Use the Backup location drop-down menu to select a backup location.

You can configure a database backup as a one-time event, or as a periodic scheduled backup. Specify the date and time of a one-time backup, or toggle **Backup Schedule** to then select the period of the backup.

Select **Save** to save the database to the backup location.

Use the Restore location drop-down menu to select the location of a previously saved Backup file, then select **Restore**.

Note Restoring the database will overwrite any data that is not in the Backup file and that data will be lost.

About Instrument

Provides all the same information provided in the [System Status](#) information screen with a few exceptions.

Memory Left Displays the remaining data storage space on the instrument.


See “[System Status](#)” on [page 294](#) for details on remaining information provided on this screen.

EULA Select **EULA** to display Thermo Fisher Scientific's End User License Agreement. Once displayed, the EULA can be exported by selecting **Export**.

PC Control Software

Control your NanoDrop Ultra from a PC through a USB A/B connection with the PC control software. You can store or view data acquired with a NanoDrop Ultra instrument on the PC, as well as change instrument settings, and create or edit custom methods. The NanoDrop Ultra local control and PC control software applications are almost identical, a few of the more significant differences will be highlighted below.

PC Control Home Screen overview



PC Control Home Screen

Select your application from the categorized icons just as you would with the NanoDrop Ultra instrument Home Screen.

PC Control Measurement Screen

The screenshot shows the PC Control Measurement Screen with the following labeled components:

- Application:** dsDNA
- Measurement pathway:** Pedestal
- Next sample name; select to edit:** Sample 4
- Menu of options; select to open:** Gear icon
- qPCR Kit Data:** Waveform icon
- Status alert:** Warning icon with '3' notification
- Auto-measure:** Blank, Measure, End Experiment buttons
- Select to end experiment and export data:** End Experiment button

The graph displays **UV spectrum** with 10mm Absorbance on the y-axis (0 to 2.100) and Wavelength (nm) on the x-axis (220 to 350). Three samples are overlaid: Sample 1 (green), Sample 2 (purple), and Sample 3 (yellow).

#	Date	Sample Name	ng/μL	A260/A280	A260/A230	A260	A280
	11/06/2024 12:37:39 PM	Blank					
1	11/06/2024 12:37:58 PM	Sample 1	30.1	2.28	0.65	0.60	0.26
2	11/06/2024 12:38:18 PM	Sample 2	47.6	2.09	0.88	0.95	0.46
3	11/06/2024 12:38:39 PM	Sample 3	68.3	1.96	1.08	1.37	0.70

Additional labels for the table:

- Select row to select sample and update spectrum; select more rows to overlay up to sixteen spectra. Click and hold sample row to view measurement details.** (points to row 1)
- Nucleic acid concentration** (points to ng/μL column)
- Purity ratios** (points to A260/A280, A260/A230, A260, A280 columns)
- select to assign unit** (points to the unit dropdown arrow)
- Scroll down/up to see more/less sample data** (points to the table scrollbar)

PC Control Settings

The NanoDrop Ultra PC control settings are very similar to the settings seen on the local instrument. Some settings like [Export](#) and [About Instrument](#) are not available on the PC control. Additionally, some settings like [System Settings](#) have limited options in comparison to the local control software as some of the settings are pulled from the PC.

To access Thermo Fisher Scientific's End User License Agreement, open **System Status** and select **EULA**. Select and confirm a location to save the EULA.

This page is intentionally blank.

Maintenance

- [Maintenance Schedule 360](#)
- [Cleaning the Touchscreen 361](#)
- [Maintaining the Pedestals 361](#)
- [Decontaminating the Instrument 365](#)
- [Maintaining the Cuvette Sampling System 366](#)
- [Instrument Diagnostics 367](#)

Maintenance Schedule

Daily Maintenance

- [Clean pedestals with deionized water](#)

Periodic Maintenance

- [Clean touchscreen](#)
- [Clean and recondition pedestals](#)



Every 6 Months

- [Clean and recondition pedestals](#)
- Run [Intensity Check](#)
- Run [Performance Verification](#)
- Run [Pedestal Image Check](#)
- Run [Cuvette Check](#)
- Run [Fluorescence Performance Check](#)
- Run [Acclaro Pro Check](#)
- [Export Logs](#)

If you are experiencing an issue with your system, refer to the troubleshooting information. If the issue persists, contact us. If you are outside the U.S.A. and Canada, please contact your local distributor.

If your instrument requires maintenance or repair, [contact us](#) or your local distributor.

Cleaning the Touchscreen

Note To avoid causing permanent damage to the touchscreen, **do not:**

- clean the touchscreen with abrasive material such as paper towel
- apply excessive pressure
- spray liquid directly onto the touchscreen
- apply lubricant to the touchscreen slide mechanism

To clean the touchscreen

Gently wipe the touchscreen with a soft, lint-free cloth such as microfiber.

If necessary, use a cleaner intended for glass LCD displays and follow the manufacturer's recommendations.

Maintaining the Pedestals

The pedestals require periodic maintenance to maintain measurement integrity. Time lines and procedures for cleaning and reconditioning the pedestals are provided below.

Cleaning the Pedestals

To avoid carryover and cross contamination, clean the pedestals before the first blank or sample measurement and at the end of each measurement. Additional cleaning (see below) or [reconditioning](#) may be required for periodic maintenance.

Note

- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.
- To prevent damage from spills, keep containers of liquids away from the instrument.
- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not attempt to remove the diaphragm around the lower pedestal as it is permanently affixed to the instrument.
- Do not allow HCl, alcohol, bleach, acetone or any other solvent to remain on the diaphragm for more than one minute or it may loosen the seals. If the diaphragm becomes loose, [contact us](#).

10 Maintenance

Maintaining the Pedestals

Note Solutions containing detergent or isopropyl alcohol may uncondition the pedestals. If these are required for sample analyses, follow immediately with 3–5 μL DI H_2O .

Supplies needed

- lint-free laboratory wipes
- deionized water (DI H_2O)
- for thorough cleaning: [PR-1 kit](#) or 0.5M HCl

To clean the pedestals between measurements

Lift the instrument arm and clean the upper and lower pedestal with a new laboratory wipe.

To clean the pedestals between users

1. Lift the arm and clean both pedestals with a new laboratory wipe.
2. Pipette 3–5 μL DI H_2O onto the lower pedestal.
3. Lower the arm and wait 2–3 minutes.
4. Lift the arm and clean both pedestals with a new wipe.


Tip: When thorough cleaning is required (for example, to remove dried sample left on the pedestals), [clean and recondition the pedestals](#) using PR-1 compound. If you do not have PR-1, you can also substitute 0.5M HCl for the DI H_2O in the procedure above and follow with 3 μL DI H_2O .




Reconditioning the Pedestals

The pedestal surfaces may lose their “conditioned” properties over time, especially after measurements with isopropyl alcohol or solutions that contain surfactants or detergents such as the Bradford reagent. An unconditioned pedestal causes droplets on the lower pedestal to “flatten out,” preventing proper formation of the liquid column when the arm is lowered. The resulting spectrum may look “rough” or “jagged.”

If samples flatten out on the pedestal (rather than “beading up” or forming a rounded droplet) or the liquid column breaks during a measurement, recondition the pedestals.



Unconditioned pedestal
(droplet flattens out)



Properly conditioned pedestal
(droplet beads up)

Supplies needed

- lint-free laboratory wipes
- [PR-1 pedestal reconditioning kit](#) (available from us or a local distributor)
- calibrated precision pipettor (0-2 μ L)
- canned air

10 Maintenance

Maintaining the Pedestals

To recondition the pedestals



1. Open the container of PR-1 compound and use the provided applicator to remove a pin-head sized amount of the compound.
2. Apply a thin, even layer of reconditioning compound to the surface of the upper and lower pedestal.

Wait 30 seconds for the PR-1 compound to dry.



3. Fold a clean laboratory wipe into quarters and use it to vigorously buff the surface of each pedestal.

Notice: Support the instrument arm with one hand while you buff the upper pedestal to avoid damaging the arm.


Tip: Black residue on the wipe is normal.

4. Repeat step 3 with a new folded wipe until all residue is removed and the pedestals buff clean.
5. Use canned air to remove any paper residue from the pedestals.



6. Pipette 1 μ L DI H₂O onto the lower pedestal.

The DI H₂O should “bead up” or form a rounded droplet.



Tip The PR-1 pedestal reconditioning compound is the easiest way to recondition the pedestals. If you don't have a PR-1 kit, follow these steps:

1. Lift the instrument arm and pipette 3 μ L 0.5M HCl onto the lower pedestal.
2. Lower the arm and wait 2–3 minutes.
3. Lift the arm and clean both pedestals with a new laboratory wipe.
4. Pipette 3 μ L DI H₂O onto the lower pedestal.
5. Lower the arm and wait 2–3 minutes.
6. Lift the arm and clean both pedestals with a new wipe.

NOTICE: Support the instrument arm with one hand while you buff the upper pedestal to avoid damaging the arm.

7. Fold a clean laboratory wipe into quarters and use it to vigorously buff the surface of each pedestal at least 50 times.
8. Use canned air to remove any paper residue from the pedestals.

Decontaminating the Instrument

Decontaminate the instrument after measurements with samples that contain [hazardous materials](#) and before returning the instrument to us for maintenance or repair.

Note If your instrument requires maintenance or repair, [contact us](#) or your local distributor.

Note

- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.
- To prevent damage from spills, keep containers of liquids away from the instrument.
- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not allow HCl, alcohol, bleach, acetone or any other solvent to remain on the diaphragm for more than one minute or it may loosen the seals. If the diaphragm becomes loose, [contact us](#).

10 Maintenance

Maintaining the Cuvette Sampling System

Supplies needed

- lint-free laboratory wipes
- deionized water (DI H₂O)
- 0.5% sodium hypochlorite solution (1:10 dilution of commercial bleach, freshly prepared)
- pipettor

To decontaminate the pedestals

1. Lift the instrument arm and clean the upper and lower pedestal with a new laboratory wipe.
2. Pipette 2–3 μ L diluted bleach solution (see [Supplies needed](#)) onto the lower pedestal.
3. Lower the arm and wait 2–3 minutes.
4. Lift the arm and clean both pedestals with a new wipe.
5. Pipette 3–5 μ L DI H₂O onto the lower pedestal.
6. Lower the arm and wait 2–3 minutes.
7. Lift the arm and clean both pedestals with a new wipe.



To decontaminate the instrument surfaces

1. Dampen a clean, soft cloth or laboratory wipe with the diluted bleach solution (see [Supplies needed](#)) and use it to gently wipe the outside surfaces of the instrument.
2. Use a clean cloth or wipe dampened with DI H₂O to remove the bleach solution.

Maintaining the Cuvette Sampling System

The cuvette sampling system is included only with the NanoDrop Ultra^C and NanoDrop Ultra^C FL model instruments. For information about compatible cuvettes, see [Measuring a Sample using a Cuvette](#).

Note Clean and dry cuvettes after each measurement. Use cuvettes that are free of scratches and avoid fingerprints which may affect results.

Note Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.

To maintain the cuvette sampling system

- Keep the instrument arm closed when the instrument is not in use.
- Use canned air to remove any dust from the cuvette holder.
- Clean up any spills inside the cuvette holder with a new laboratory wipe.

To clean and maintain cuvettes, follow the recommendations of the cuvette manufacturer.



Instrument Diagnostics

Every 6 months, run the following performance and quality checks to verify instrument operation.

[Intensity Check 368](#)

[Performance Verification 370](#)

[Pedestal Image Check 376](#)


[Cuvette Check 377](#)


[Fluorescence Performance Check 379](#)

[Acclaro Pro Check 383](#)

[Export Logs 386](#)

[Diagnostics Schedule 386](#)

Diagnostics can be performed using the NanoDrop Ultra local instrument software or the PC control software. **Performance Verification, Intensity Check, Pedestal Image Check, Cuvette Check, Fluorescence Performance Check, and Acclaro Pro Check** are all accessible from **Diagnostics**  on the software home screen.



Diagnostic options

Performance Verification:	Run a pedestal performance verification test using PV-1
Intensity Check:	Run an intensity check for the cuvette or pedestal
Pedestal Image Check:	Run a pedestal image check
Cuvette Check:	Run a cuvette transmittance test
Fluorescence Performance Check:	Run a fluorescence performance test using FL-1
Acclaro Pro Check:	Run an Acclaro Pro performance check using PV-1
Export Logs:	Export instrument related log files for troubleshooting
Diagnostics Schedule	Establish a diagnostic schedule that provides timely alerts, reminding you when the next diagnostic checks are due


Intensity Check

It is recommended to run an Intensity Check every 6 months to verify operation of the instrument's internal components. The test measures the intensity of light from the xenon source through the instrument to verify that throughput, wavelength accuracy, and bias are within specifications. The test is automatically performed using the pedestal and the cuvette optical paths.

Supplies needed

- lint-free laboratory wipes

To run Intensity Check

1. Lift the instrument arm and clean the upper and lower pedestal with a new laboratory wipe.
2. Remove any cuvette from the cuvette holder.
3. Lower the arm.
4. From the home screen, select the **Diagnostics** icon , followed by **Intensity Check**.
5. When using the NanoDrop Ultra^c or NanoDrop Ultra^c FL, select which pathway to test.
 - Use the dropdown menu at the top of the screen to select **Pedestal** or **Cuvette**.
6. Tap or select **Measure** and wait for the measurement to complete.


Here is an example of a typical Intensity Check result screen from the NanoDrop Ultra local control software.



Swipe screen left to view detailed results

7. To rerun the Intensity Check, select **Measure**.
8. When finished, select **End Experiment**.

After the test is completed, the results are available from the History (see example below). See [Manage identifiers on the instrument](#) for details.



The screenshot shows the 'NanoDrop Ultra History' interface. At the top, there is a search bar with a 'Clear' button and a dropdown menu set to 'Diagnostics / Inte...'. Below this is a table with the following columns: Date, Name, Application, Created User, and Samples. A single row is visible with the following data: Date: 03/25/2024 11:53:27 AM GMT+00:00, Name: Intensity Check 03/25/2024 11:53:29 AM, Application: Intensity Check, Created User: Instrument, Samples: 1.

Date	Name	Application	Created User	Samples
03/25/2024 11:53:27 AM GMT+00:00	Intensity Check 03/25/2024 11:53:29 AM	Intensity Check	Instrument	1

To interpret Intensity Check results

If one of these indicators:

- UV
- Visible
- Bias

has an adjacent orange exclamation mark instead of the green check marks shown above, [clean the pedestals with deionized water](#) and then repeat the Intensity Check. When using the PC control software, the status column in the results table will display "Pass" or "Fail" for each criterion and the column beside it will display any applicable details.

If an orange exclamation mark appears next to the Bias indicator, or it's status is marked as fail, make sure the room is within the temperature specifications for the instrument.

If the Intensity Check fails again, [contact us](#).

Performance Verification

It is recommended to run a Performance Verification every 6 months to confirm pathlength accuracy is within specifications.


Supplies needed

- lint-free laboratory wipes
- deionized water (DI H₂O)
- calibrated precision pipettor (0–2 µL)
- [PV-1 performance verification solution](#) (liquid photometric standard available only from us or a local distributor)
- laboratory gloves

Note The PV-1 solution comes in a single-use ampoule. Before you open the ampoule, ensure that the solution is thoroughly mixed by vigorously shaking/inverting the ampoule. Allow the solution to collect in the bottom portion of the ampoule. If needed, gently tap or flick the ampoule. After the ampoule is opened, its contents must be used within one hour. Pipette directly from the ampoule; do not transfer the solution.


Before you begin

First make sure the pedestals are properly conditioned. To test pedestal conditioning, clean the pedestals with a new laboratory wipe, then pipette 1 μ L DI H₂O onto the lower pedestal. The droplet should “bead up” as shown below. If it does not, [recondition both pedestals](#).




Droplet “beads up” on properly conditioned pedestal

To run the performance verification


1. From the home screen, select the Diagnostics icon , followed by **Performance Verification**.

A message asks for the PV-1 lot number and target absorbance values.



2. Enter the lot number of the PV-1 ampoule (found on the PV-1 ampoule label) that is going to be used for this test. Select **Done** when using the local instrument control to close the keyboard.
3. Enter each lot-specific target absorbance value from the label on the PV-1 ampoule in its associated entry box and then select **Done**.
4. Once all values have been entered, select **Run**.
5. Lift the instrument arm and clean the upper and lower pedestal with a new laboratory wipe.
6. Pipette 1 μ L DI H₂O onto the lower pedestal, lower the arm and select **Blank**.
7. Lift the arm and clean both pedestals with a new wipe.
8. Ensure PV-1 solution is thoroughly mixed by vigorously shaking/ inverting the ampoule. Allow the solution to collect in the bottom portion of the ampoule. If needed, tap or flick the ampoule.
9. Carefully snap off the top portion of the ampoule using a plastic ampoule cracker. Discard top along with ampoule cracker (use proper safety precautions for disposal).
10. Pipette 1 μ L PV-1 solution onto the lower pedestal and start the sample measurement:
 - If **Auto-Measure** is On, lower arm
 - If **Auto-Measure** is Off, lower arm and select **Measure**.


After the measurement, the software displays the results. Here is an example of the performance verification result screen.




11. Repeat [step 10](#) to measure the PV-1 solution nine more times using a new 1 μ L aliquot for each measurement and cleaning both pedestals after each measurement. After each measurement, a new sample result is added to the display.

After the tenth measurement, a message containing the overall result of the performance verification will be displayed. Below is an example of what would be seen after passing a test. Select **OK** to see a more detailed summary of the results.


10 Maintenance Instrument Diagnostics



Swipe the screen left to see a summary of the 10 sample results.




Swipe left again to see additional measurement details, along with the overall test result.



- When finished, select **End Experiment** and [clean the pedestals with 3–5 \$\mu\text{L}\$ DI \$\text{H}_2\text{O}\$](#) .

After the test is completed, the results are available from the History (see example below). See [Manage identifiers on the instrument](#) for details.



To interpret performance verification results

The overall result of the test will be displayed at the top of the screen as a Green "Pass"/ Orange "Conditional Pass"/ Red "Fail".

- If results are not within specifications, repeat procedure using 2 μL aliquots of PV-1.
- If results fail to meet specifications using 2 μL aliquots, [contact us](#).
- A conditional pass means that the instrument is working slightly outside of our factory specifications, but not yet failing. With a conditional pass, it is left up to the end user to determine if those results are acceptable for their intended purposes.


Pedestal Image Check

Run the Pedestal Image Check periodically to verify the instrument's column sensor which monitors for possible errors such as an empty column or bubbles in a sample. The Pedestal Image Check can be used for routine quality control purposes. It also provides important diagnostic information if a detection system component fails.

Supplies needed

- lint-free laboratory wipes


To run the Pedestal Image Check

1. Lift the instrument arm and clean the upper and lower pedestal with a new laboratory wipe.
2. Lower the arm.
3. From the instrument home screen, select , (Diagnostics) and then select **Pedestal Image Check**.
4. Select **Measure**.

The instrument runs a series of tests to check pedestal position and image quality. After the measurements are completed, the results are displayed. The display of a Pass result at the top of the screen indicates the instrument passed the Pedestal Image Check.

5. When finished, select **End Experiment**.

To interpret Pedestal Image Check results



Item	Value	Acceptance limits
Empty column detection	Pass	Pass
Sharpness	16.14	>5
Background exposure	142.00	80 - 215
Background quality	4.40	<9.00
Edge detection	2	2

If the Pedestal Image Check displays a Fail result, follow the on-screen instructions to fix any possible problems. Then rerun the Pedestal Image Check. If the instrument fails again, [contact us](#).

Cuvette Check

Run the Cuvette Check test every 6 months to confirm photometric accuracy is within specification when using the cuvette mode. This module is only available on the NanoDrop Ultra^C and NanoDrop Ultra^C FL instruments.


Supplies needed

- Cuvette Qualification Filter Kit for NanoDrop Instruments (PN 840-257300)
- NIST-Traceable Certified 10% Transmittance (at 590 nm) Neutral Density Filter
- NIST-Traceable Certified 50% Transmittance (at 590 nm) Neutral Density Filter

To run the Cuvette Check

1. From the home screen, select , (Diagnostics) and then select **Cuvette Check**.

A message asks for the serial number of the cuvette qualification filter kit and the target absorbance values of each filter at 590 nm.



The screenshot shows the 'Cuvette Test Kit Values' screen on a NanoDrop Ultra instrument. The interface is dark-themed with a sidebar on the left containing navigation icons. The main content area has the following elements:

- Header: NanoDrop Ultra^C Cuvette Check
- Title: Cuvette Test Kit Values
- Serial Number: A text input field with a 'Tap to enter' prompt and a small note below it: 'Enter serial number format as SA1111'.
- Filter 50% 590.0nm: A text input field with a 'Tap to enter' prompt.
- Filter 10% 590.0nm: A text input field with a 'Tap to enter' prompt.
- Run: A large blue button at the bottom center.


2. Enter the serial number of the qualification filter kit and the lot-specific target absorbance at 590 nm for each filter into their respective box. When using the local instrument software, select **Done** to close the number pad window for each entry.

This information can be found on the certificate of calibration provided with the cuvette qualification filter kit.

3. After all values have been input, select **Run**.
4. Ensure the cuvette holder is empty and clear from obstruction.


5. With no cuvette inside the cuvette holder, select **Blank** to record the blank measurement. A blank filter is not needed.
6. Load the 50%T filter into the cuvette holder ensuring that the filter is oriented correctly, select **Measure**.
 - The arrow etched into the assembly indicates the direction of the light path.
7. Load the 10%T filter into the cuvette holder ensuring that the filter is oriented correctly, select **Measure**.
8. After each measurement, the software will display the result for the measured filter in the results table with either a green checkmark Pass or a red X Fail. If both filters provide passing results, an overall Pass result will be displayed.

Here is an example of the **Cuvette Check** result screen from the NanoDrop Ultra local control software.



9. If one or both filters fail, the cuvette check will also fail. In the case of a failed cuvette check, immediately repeat the test.
10. When finished, select **End Experiment**. Remove any remaining filters from the cuvette holder.

After the test is completed, the results are available from the **History** (see example below). See **Manage identifiers** on the instrument for details.



To interpret Cuvette Check results

If your instrument failed the cuvette check more than once, [contact us](#).

Fluorescence Performance Check

Run the Fluorescence Performance Check every 6 months to confirm fluorescence performance and accuracy are within specifications. This module is only available on the NanoDrop Ultra FL and NanoDrop Ultra^c FL instruments.

Supplies needed

- FL-1 Fluorescence Verification Assay Kit (CHEM-FL1-KIT)
 - Blank reagent
 - Green fluorescence reagent
 - Far-red fluorescence reagent
 - Diluent
- plastic tubes, 0.2 mL, flat cap or similar
- lint-free laboratory wipes
- deionized water (DI H₂O)
- Single-channel, low-volume (0.5-10 µL) Pipettor (with appropriate tips), calibrated
- Single-channel, 10-100 µL Pipettor (with appropriate tips), calibrated
- laboratory gloves
- vortex mixer
- mini centrifuge

Before you begin

Set up three sterile plastic amber tubes for three of the four reagents provided in the FL-1 Fluorescence Verification Kit (Blank reagent, Green Fluorescence Reagent, and Far Red Fluorescence Reagent) and label them **Blank**, **Green Standard 2**, and **Far Red Standard 2**.

Set up two additional sterile plastic amber tubes for dilutions of both Green Fluorescence Reagent and Far Red Fluorescence Reagent and label them **Green Standard 1** and **Far Red Standard 1**.

Blank Tube

- Add 50 µL of Blank Reagent to the Blank tube.

Green Standard 2 Tube

- Add 33 μL of Green Fluorescence Reagent.

Far Red Standard 2 Tube

- Add 45 μL of Far Red Fluorescence Reagent.




Green Standard 1 Tube

- Add 27 μL of Diluent.
- Transfer 3 μL of Green Standard 2 to this tube and vortex for 3-5 seconds.
- Centrifuge briefly, a 10-fold diluted Green Fluorescence Reagent is ready.

Far Red Standard 1 Tube

- Add 15 μL of Diluent.
- Transfer 15 μL of Far Red Standard 2 to this tube and vortex for 3-5 seconds.
- Centrifuge briefly, a 2-fold diluted Far-red Fluorescence Reagent is ready.

To run the Fluorescence Performance Check

1. From the home screen, select the Diagnostics icon , followed by **Fluorescence Performance Check**.
2. A message will appear stating the test requires the use of the FL-1 Fluorescence Verification Kit, select **Run**.
3. Ensure nothing is on the pedestal and select **Blank**.
4. Pipette 2 μL of the Blank Reagent from the Blank tube onto the lower pedestal.
 - a. If Auto-Measure is ON , lower arm.
 - b. If Auto-Measure is OFF , lower arm and select **Measure**.
5. Lift the arm and clean both pedestals with a new wipe.
6. Repeat steps 4 & 5 for a total of five measurements.
7. Pipette 2 μL of the diluted Green Fluorescence Reagent from the Green Standard 1 tube onto the lower pedestal.
 - a. If Auto-Measure is ON, lower arm.
 - b. If Auto-Measure is OFF, lower arm and select **Measure**.
8. Lift the arm and clean both pedestals with a new wipe.
9. Repeat steps 7 & 8 for a total of five measurements.

10. Pipette 2 μ L of the Green Fluorescence Reagent from the Green Standard 2 tube onto the lower pedestal.
 - a. If Auto-Measure is ON, lower arm.
 - b. If Auto-Measure is OFF, lower arm and select **Measure**.
11. Lift the arm and clean both pedestals with a new wipe.
12. Repeat steps 10 & 11 for a total of five measurements.
13. Pipette 2 μ L of the Blank Reagent from the Blank tube onto the lower pedestal.
 - a. If Auto-Measure is ON, lower arm.
 - b. If Auto-Measure is OFF, lower arm and select **Measure**.
14. Lift the arm and clean both pedestals with a new wipe.
15. Repeat steps 13 & 14 for a total of five measurements.
16. Pipette 2 μ L of the diluted Far-Red Fluorescence Reagent from the Far Red Standard 1 tube onto the lower pedestal.
 - a. If Auto-Measure is ON, lower arm.
 - b. If Auto-Measure is OFF, lower arm and select **Measure**.
17. Lift the arm and clean both pedestals with a new wipe.
18. Repeat steps 16 & 17 for a total of five measurements.
19. Pipette 2 μ L of the Far-Red Fluorescence Reagent from the Far Red Standard 2 tube onto the lower pedestal.
 - a. If Auto-Measure is ON, lower arm.
 - b. If Auto-Measure is OFF, lower arm and select **Measure**.

20. Lift the arm and clean both pedestals with a new wipe. Repeat steps 19 & 20 for a total of five measurements.

The screenshot displays the 'Fluorescence Performance Check' results for 07/08/2025 at 09:23:36 AM. The overall status is 'Pass' (green button). The instrument ID is NUC2400034. The results are presented in two tables, each with an R-squared value.

Sample Name	# of Replicates	#1 RFU	#2 RFU	#3 RFU	#4 RFU	#5 RFU	Average	Std Dev	%CV
Blank reagent	5	3759.0	3770.8	4093.4	4126.0	3778.4	3905.528	186.852	4.784
Green Standard 1	5	11608.9	11156.0	11757.5	10784.5	10404.1	11142.206	563.33	5.056
Green Standard 2	5	98295.4	97156.9	102635.8	104182.4	94567.2	99367.523	3966.837	3.992

R²:0.9995

Sample Name	# of Replicates	#1 RFU	#2 RFU	#3 RFU	#4 RFU	#5 RFU	Average	Std Dev	%CV
Blank reagent	5	985.7	1016.8	983.8	1004.6	1008.0	999.788	14.462	1.446
Far Red Standar...	5	6528.1	7557.9	7274.3	7272.5	6531.2	7032.816	473.787	6.737
Far Red Standar...	5	20376.1	21030.6	21845.4	22493.7	20703.4	21289.856	866.606	4.071

R²:0.9481

The bottom screenshot shows the 'NanoDrop Ultra^C FL History' screen. It includes a search bar, a 'Clear' button, and dropdown menus for 'Diagnostics / Fluor...' and 'All time'. A table lists the history of checks:

Date (GMT +00:00)	Name	Application	Created User	Samples
11/04/2024 09:17:02 PM	Fluorescence Performance Check 11/04/2...	Fluorescence Perfor...	Instrument	7

To interpret Fluorescence Performance Check results

The overall result of the test will be displayed at the top of the screen as a Green "Pass" or Red "Fail".

If results are not within specifications, repeat procedure with new dilutions of the FL-1 reagents.

If results fail to meet specifications after the second attempt, [contact us](#).

Acclaro Pro Check

It is recommended to run the Acclaro Pro Check every 6 months to confirm pathlength accuracy is within specifications for the Acclaro Pro applications.


Supplies needed

- lint-free laboratory wipes
- deionized water (DI H₂O)
- calibrated precision pipettor (0-2 μ L)
- PV-1 performance verification solution (liquid photometric standard available only from us or a local distributor)
- laboratory gloves

Note The PV-1 solution comes in a single-use ampoule. Before you open the ampoule, ensure that the solution is thoroughly mixed by vigorously shaking/inverting the ampoule. Allow the solution to collect in the bottom portion of the ampoule. If needed, gently tap or flick the ampoule. After the ampoule is opened, its contents must be used within one hour. Pipette directly from the ampoule; do not transfer the solution.


Before you begin


First make sure the pedestals are properly conditioned. To test pedestal conditioning, clean the pedestals with a new laboratory wipe, then pipette 1 μ L DI H₂O onto the lower pedestal. The droplet should “bead up” as shown below. If it does not, [recondition both pedestals](#).



Droplet “beads up” on properly conditioned pedestal

To run the Acclaro Pro Check

1. From the home screen, select the Diagnostics icon , followed by **Acclaro Pro Check**. A message asks for the PV-1 lot number and target absorbance values.



Tap an entry box to display the alphanumeric keyboard

2. Enter the lot number of the PV-1 ampoule (found on the PV-1 ampoule label) that is going to be used for this test. Select **Done** when using the local instrument control to close the keyboard.
3. Enter each lot-specific target absorbance value from the label on the PV-1 ampoule in its associated entry box. Select **Done** when using the local instrument control to close the keyboard.
4. Once all values have been entered, select **Run**.
5. Lift the instrument arm and clean the upper and lower pedestal with a new laboratory wipe.
6. Ensure PV-1 solution is thoroughly mixed by vigorously shaking/ inverting the ampoule. Allow the solution to collect in the bottom portion of the ampoule. If needed, tap or flick the ampoule.
7. Carefully snap off the top portion of the ampoule using a plastic ampoule cracker. Discard top along with ampoule cracker (use proper safety precautions for disposal).

8. Pipette 1 μL PV-1 solution onto the lower pedestal and start the sample measurement:
 - If **Auto-Measure** is On, lower arm
 - If Auto-Measure is Off, lower arm and select **Measure**.
9. Repeat step 8 to measure the PV-1 solution four more times using a new 1 μL aliquot for each measurement and cleaning both pedestals after each measurement.

After the fifth measurement, the software displays the results. Here is an example of the Acclaro Pro Check result screen from the NanoDrop Ultra local control software.

Name	Target Absorbance	Measure Absorbance	R ²	% Error
Sample 1	10.115	10.019	1.000	0.951%
Sample 2	10.115	10.062	1.000	0.529%
Sample 3	10.115	10.067	1.000	0.474%
Sample 4	10.115	10.053	1.000	0.617%
Sample 5	10.115	10.040	1.000	0.740%

% CV: 0.192%

10. When finished, select End Experiment and clean the pedestals with 3-5 μL DI H₂O.

After the test is completed, the results are available from the **History** (see example below). See **Manage identifiers** on the instrument for details.

Date (GMT +00:00)	Name	Application	Created User	Samples
11/04/2024 09:35:34 PM	Acclaro Pro Check 11/04/2024 09:35:34 PM	Acclaro Pro Check	Instrument	5


To interpret Acclaro Pro Check results

1. The overall result of the test will be displayed at the top of the screen as a Green "Pass" or a Red "Fail".
 - If results are not within specifications, repeat procedure using 2 μ L aliquots of PV-1.
 - If results fail to meet specifications using 2 μ L aliquots, [contact us](#).

Export Logs

The Export Logs feature is a functionality that allows users to retrieve logs generated by the instrument software for diagnostic purposes.


To export instrument log file

1. From the home screen, select , (diagnostics) and then select **Export Logs**.
2. Select preferred export method and then select **Export**.


Diagnostics Schedule

The Diagnostics Schedule feature enables users to set specific intervals for receiving alerts, ensuring timely execution of diagnostic checks on the instrument. The NanoDrop Ultra Spectrophotometers and Fluorometers track the last run date and time for all available diagnostics, providing an automated, reliable system to maintain optimal instrument performance.


To enable the diagnostics schedule

1. From the home screen, select , (diagnostics) and then select Diagnostics Schedule.
2. Use the toggle switch next to Diagnostics Schedule to enable or disable the feature.

3. Once enabled, use the entry box besides Expiration interval to define the number of days (90-999) that will pass before a diagnostic test is deemed expired.



Once a diagnostic test has expired, an alert will then trigger to alert message to the user. An example of this alert has been provided below.



This page is intentionally blank.

Safety and Operating Precautions

Contents

- [Operating Precautions 390](#)
- [Safety Information 391](#)



Note Be sure that all persons operating this system read the safety manual first.

Operating Precautions



CAUTION Do not remove the instrument cover. Removing the cover exposes the operator to sharp edges and delicate fiber optic cables. The instrument warranty is void if the cover has been removed.

NanoDrop Ultra spectrophotometers and fluorometers are designed to operate indoors in an environment that meets our specifications. For details, see the site preparation guide for your instrument.

Follow these precautions to avoid damaging your NanoDrop spectrophotometer and fluorometer during use:

- Use a grounded power cord appropriate for your electrical service. If the supplied power cord is incompatible or if it becomes damaged, [contact us](#).
- Do not remove the instrument cover.
- The LCD display uses heat treated, chemical tempered glass and is rugged and difficult to break. However, should the display become cracked or broken, contact us for replacement.
- Use solvents that are compatible with the instrument (see [Hazardous Materials](#))
- Do not use hydrofluoric acid (HF) on the pedestals. Fluoride ions will permanently damage the quartz fiber optic cables.
- To prevent damage from spills, keep containers of liquids away from the instrument.
- Do not use a squirt or spray bottle on or near the instrument as liquids will flow into the instrument and may cause permanent damage.
- Do not attempt to remove the diaphragm around the lower pedestal as it is permanently affixed to the instrument.
- Do not allow HCl, alcohol, bleach, acetone or any other solvent to remain on the diaphragm for more than one minute or it may loosen the seals. If the diaphragm becomes loose, [contact us](#).

Safety Information

Before operating a NanoDrop Ultra instrument, please read the safety information and follow its recommendations for the system.

Safety and Special Notices

In many cases, safety information is displayed on the instrument itself. The symbol indicates that there is additional safety information in the documentation and failure to heed the safety precautions could result in injury.



WARNING Indicates a hazardous situation which, if not avoided, could result in death or serious injury.



CAUTION Indicates a hazardous situation which, if not avoided, could result in minor or moderate injury.

















Note Follow instructions with this label to avoid damaging the system hardware or losing data.

Note Contains helpful supplementary information.

11 Safety and Operating Precautions

Safety Information

The following table lists some of the safety symbols and their indications that may appear in the user documentation.

Symbols	Indication
	This is a mandatory action symbol. It is used to indicate that an action shall be taken to avoid a hazard.
	This is a prohibition symbol. The graphic in this symbol is used to alert the user to actions that shall not be taken or shall be stopped.
	This is the general warning sign. Failure to heed the safety precautions could result in personal injury.
	Avoid shock hazard. If you see either of these symbols, there is a risk of electrical shock in the vicinity. Only qualified persons shall perform the related procedures.
	Avoid fire hazard. Do not test flammable or explosive samples. Read and follow the associated instructions carefully.
	Avoid eye injury. If you see these symbols, there is a risk of exposure to ultraviolet light, which can harm your eyes if safety glasses are not worn.
	Avoid Biohazard. This icon informs of a biological hazard in the area. Read and follow the associated instructions carefully.
	Avoid chemical burns. This symbol alerts you to possible skin irritation. Wear gloves when handling toxic, carcinogenic, mutagenic, or corrosive or irritant chemicals. Use approved containers and proper procedures to dispose of waste.
Symbol	Description
	Alternating current
	Earth terminal or ground
	Direct current
	Protective conductor terminal
	Frame or chassis terminal
	Fuse
	Power on
	Power off

When the System Arrives



WARNING Avoid personal injury. If this equipment is used in a manner not specified in the accompanying documentation, the protection provided by the equipment may be impaired.



CAUTION Avoid personal injury. Perform *only* those procedures described in the documentation. If there are other problems, contact us. Any other service must be performed by trained personnel.



CAUTION Avoid shock hazard. Do not remove the cover of the instrument. All service to the instrument must be performed by trained personnel.

When the instrument arrives, check the exterior of the shipping box for signs of damage. If damage is apparent, contact us or your local distributor for instructions.

- Move the shipping box to the installation location at least 24 hours before installation.

Note

- Inside the shipping box, the instrument is sealed in a plastic bag to keep the unit dry.
- Allow 24 hours for the instrument to reach room temperature before opening the bag. If the bag is opened before the instrument reaches room temperature, moisture could condense on the optical components and cause permanent damage.
- Keep the instrument upright at all times.

The warranty will not cover:

- Damage due to improper moving techniques.
- Damage due to removing the sealed plastic bag before the instrument has come to room temperature.

Note It is important to have all system utilities installed before the instrument arrives. Utility installations must comply with all local building and safety codes.

Lifting or Moving the Instrument

To avoid risk of injury, use proper lifting techniques when lifting or moving the instrument or other system components.

Electrical Requirements and Safety

Power supplied to the system must be from dedicated, uninterrupted sources. Power must be free of voltage dropouts, transient spikes, frequency shifts, and other line disturbances that impair reliable performance.

If you suspect power quality problems at your site, or if your system will be installed in a heavy industrial environment, we recommend a power quality audit before installation. Contact us or your local electrical authority for more information.



CAUTION Avoid shock hazard.

- Only a qualified person using the appropriate measuring device shall check the line voltage, current and frequency.
- Only our trained and certified service representatives shall attempt to service a component that carries this symbol.
- If a protective cover on a system component appears damaged, turn off the system and secure it against any unintended operation. Always examine the protective cover for transport stresses after shipping.
- Even after this instrument has been disconnected from all voltage sources, capacitors may remain charged for up to 30 seconds and can cause an electrical shock.
- Do not allow liquid to run over or into any surface where it may gain entry into the instrument.
- Do not attempt to remove the cover of the instrument.

Grounding



CAUTION Avoid shock hazard. Each wall outlet used must be equipped with a ground. The ground must be a noncurrent-carrying wire connected to earth ground at the main distribution box.

Power Cords

Be sure to use an appropriate grounded power cord for your electrical service. If the power cord received is not appropriate for the electrical system in your location, or if the power cord becomes damaged, [contact us](#).

Power Line Conditioning Accessories

A UPS reduces the probability of a system shutdown if power is lost elsewhere in the building. Power line conditioners (which ensure that your service is free from sags, surges or other line disturbances) also are available in the U.S.A. from us for 120 volt operation. Line conditioners for 220 volt operation can be purchased locally. Contact technical support for information about power conditioners and UPS.

Electrical Service Specifications

The following table lists the specifications for electrical service. Contact our service representative in your area if you have questions about the requirements.

Requirements	Specifications
Input current	5.0 A (max.)
Input voltage	100-240 VAC
Line frequency	50/60 Hz
Line disturbances	Sags, surges or other line disturbances must not exceed 10% of input voltage (even for a half cycle).
Noise	< 2 V (common mode) < 20 V (normal mode)

Power Consumption

Generally, 50% more power should be available than the entire system (including accessories) typically uses. Maximum power consumption and heat dissipation specifications for the spectrometer and fluorometer and accessories are shown below. The values are approximate.

Item	Power Consumption	Max. Heat Dissipation
instrument	60 W	205 Btu/hr

FCC/IC Notices

This system complies with Part 15 of the FCC Rules and Industry Canada ICES-003 for a Class A system. Operation is subject to the following two conditions:

1. This system may not cause harmful interference, and
2. This system must accept any interference received, including interference that may cause undesired operation.

Note This is a Class A product. In a domestic environment, this product may cause radio interference, in which case the user may be required to take adequate measures.

Fire Safety and Burn Hazards

Note Do not position the instrument so that it is difficult to operate the power switch or access the power supply and power cord.

To avoid a burn injury and the risk of fire or explosion:

- Use caution when testing flammable or explosive samples (see the “[Hazardous Materials](#)” section)
- Never block any of the vents on the instrument or its power supply
- Only use exact replacement power supplies from us

Optical Safety

This instrument was designed with a protective housing to prevent user exposure to ultraviolet light.

Hazardous Materials

Many standard spectroscopy methods are based on the use of solvents. Others involve corrosive samples or pressurized samples in a gaseous state.

Volatile Solvents and Flammable Samples



CAUTION Avoid personal injury. Do not leave solvents or flammable samples near the instrument. Be sure that the workspace is properly ventilated.

Compatible Solvents

Most solvents typically used in life science laboratories are compatible with the fiber optic pedestals of all NanoDrop spectrophotometers and fluorometers. However, the high vapor pressure properties of some solvents may not be conducive to small volume measurements when using the pedestal for measurements on any of the NanoDrop instruments. If you are measuring samples with high vapor pressures, use an instrument with provision for measuring samples in cuvettes.

The following solvents are compatible for use on the pedestals of all NanoDrop instruments.

Note Spillage of these solvents on surfaces other than the pedestals may damage the instrument.

- methanol
- ethanol
- n-propanol
- isopropanol
- butanol
- acetone
- ether
- chloroform
- carbon tetrachloride
- DMSO
- DMF
- acetonitrile
- THF
- toluene
- hexane
- benzene
- sodium hydroxide
- sodium hypochlorite (bleach)
- dilute HCl
- dilute HNO₃
- dilute acetic acid

It is recommended that all corrosive solvents be wiped from the pedestal immediately upon completion of a measurement. It is also recommended that the user end a series of measurements with a diH₂O sample to ensure that solvents are not inadvertently left on the pedestal.

The diaphragm around the pedestal of the NanoDrop is permanently affixed to the instrument. Do not attempt to remove the diaphragm or break the seal. Avoid prolonged exposure of the diaphragm to HCl, alcohol, bleach, acetone or other solvents as the adhesive securing the seal may be affected. If the seal comes loose please contact us.

Note All forms of Hydrofluoric Acid (HF) are incompatible as the fluoride ion will etch the fiber optic cable.

Biohazard or Radioactive Materials and Infectious Agents

Biological samples such as tissues, body fluids, infectious agents, and blood of humans and other animals have the potential to transmit infectious diseases. Wear appropriate protective equipment. Individuals should be trained according to applicable regulatory and organization requirements before working with potentially infectious materials. Follow your organization's Biosafety Program protocols for working with and/or handling potentially infectious materials.

Instruments, accessories, components or other associated materials should not be disposed of and may not be returned to us or other accessory manufacturers if they are contaminated with biohazard or radioactive materials, infectious agents, or any other materials and/or conditions that could constitute a health or injury hazard to employees. Contact us if you have questions about decontamination requirements.

This page is intentionally blank.