	Document Owner Department: QC	BT-SPEC-0659
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
GENTAMICIN CN30 CT0072B		

Product Details

Product Name	GENTAMICIN
Disc Code	CN30
Disc Concentration	30µg
CT Code	CT0072B
Shelf Life (years)	3
CLSI Recommended Concentration	No
EUCAST Recommended Concentration	Yes

Microbiological Performance Criteria

EUCAST		
Organism	ATCC/NCTC Number	Range (mm)
<i>Enterococcus faecalis</i>	ATCC®29212	12 – 18
<i>Enterococcus faecalis</i>	ATCC®51299	6 – 6


References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- EUCAST Routine and extended internal quality control for MIC determination and disk diffusion as recommended by EUCAST

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0659
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
GENTAMICIN CN30 CT0072B		

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Entire document/ References	Update to current details/ Removal of all BSAC reference / Update to current performance standards	Change control	BT-CC-2848