

Certificate of Analysis

Certificate of Analysis

European Pharmacopoeia 0.1 M Ferrous sulphate

Lot N: CXXXXXX
Description: Ph Eur N 3001400
Ref N: 3001400
Measured Concentration: 0.1000 M +/- 0.2%
Method: Titration with 0.02 M potassium permanganate (Ph Eur N 3005300)
Titer*: 1.000
Expiry Date: XXXXXX

**The Titer (correction factor) is defined as the ratio from the actual molar concentration [c(x)] of a standard solution (actual value) and the intended molar concentration [c1(x)] of the same solution (required value)*

The reagent is specially intended for use in testing Ph Eur and monographs. This solution was produced according 4.2.2 of Ph Eur specifications using Ph Eur reagent quality components and water meeting the Ph Eur monograph requirement for Purified Water.

Instruction for use

The solution bottle should be open for the minimum time required to dispense the solution. After use, the bottle should be tightly recapped and stored under normal laboratory conditions.

Standardise immediately before use.

This certificate relates solely to the lot number given above.

Certification date: XXXXXXXXXXXX

*Responsible for quality control: B. COULANGE**

**This certificate has been computer generated and does not signed*

