



CERTIFICATE OF COMPLIANCE

CERTIFICATE NUMBER: 060203 - E215059

ISSUE DATE: February 06, 2003

Page 1 of 1

Issued to: Eppendorf A G - Dept Mp
Barkhausenweg 1
D-22339 Hamburg Germany

Report Reference: E215059, February 19, 1999

This is to Certify that representative samples of: Laboratory Centrifuges, Models 5804, 5804R, 5810, 5810R.

Have been investigated by Underwriters Laboratories Inc.® in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL 3101-1 - Electrical Equipment for Laboratory Use; Part 1: General Requirements
UL 3101-2-20 - Electrical Equipment for Laboratory Use, Part 2: Laboratory Centrifuges
CSA C22.2 No. 1010-1 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part I: General Requirements


Additional Information: Electrical Ratings:


Ratings	Voltage	Frequency(Hz)	Power(W)
5804	120	60	850
5804R	120	60	1400
5810	120	60	950
5810R	120	60	1650

Only those products bearing the UL Listing Mark for the US and Canada should be considered as being covered by UL's Listing and Follow-Up Service meeting the appropriate requirements for US and Canada.

The UL Listing Mark for the US and Canada generally includes: the UL in a circle symbol with "C" and "US" identifiers:  the word "LISTED"; a control number (may be alphanumeric) assigned by UL; and the product category name (product identifier) as indicated in the appropriate UL Directory.

LOOK FOR THE UL LISTING MARK ON THE PRODUCT

Engineer: 
Bogdan Maliszewski - Senior Project Engineer
UL International Germany GmbH

Review Engineer: 
Daniel Bejnarowicz - Engineering Group Leader
UL International Germany GmbH

Pursuant to the Corporate Services Agreement between UL International Germany GmbH and Underwriters Laboratories Inc. ("UL"), UL hereby accepts and issues this Certificate of Compliance. For questions in Germany, you may call 49 6102 3690.