	Document Owner Department: QC	BT-SPEC-0732
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
AZITHROMYCIN AZM15 CT0906B		

Product Details

Product Name	AZITHROMYCIN
Disc Code	AZM15
Disc Concentration	15µg
CT Code	CT0906B
Shelf Life (years)	3
CLSI Recommended Concentration	Yes
EUCAST Recommended Concentration	No

Microbiological Performance Criteria

CLSI		
Organism	ATCC/NCTC Number	Range (mm)
<i>Haemophilus influenzae</i>	ATCC®49247	13 – 21
<i>Staphylococcus aureus</i>	ATCC®25923	21 – 26
<i>Streptococcus pneumoniae</i>	ATCC®49619	19 – 25


References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- CLSI: M100 Performance standards for antimicrobial susceptibility testing

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0732
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
AZITHROMYCIN AZM15 CT0906B		

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Microbiological Performance Criteria	Remove BSAC	Change control	BT-CC-2142
Entire document/ References	Update to current format/ Update to current performance standards	Change control	BT-CC-2820