

	Document Owner Department: QC	MBD-BT-SPEC-0836
		Page 1 of 2
		Revision 01
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
CEFIDEROCOL FDC30 CT1964B		

Product Details

Product Name	CEFIDEROCOL
Disc Code	FDC30
Disc Concentration	30µg
CT Code	CT1964B
Shelf Life (years)	1
CLSI Recommended Concentration	Yes
EUCAST Recommended Concentration	Yes

Microbiological Performance Criteria

CLSI		
Organism	ATCC/NCTC Number	Range (mm)
<i>Escherichia coli</i>	ATCC® 25922	25 – 31
<i>Pseudomonas aeruginosa</i>	ATCC® 27853	22 – 31

EUCAST		
Organism	ATCC/NCTC Number	Range (mm)
<i>Escherichia coli</i>	ATCC® 25922	24 – 30
<i>Pseudomonas aeruginosa</i>	ATCC® 27853	23 – 29

References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- CLSI: M100 Performance standards for antimicrobial susceptibility testing
- EUCAST: Routine and extended internal quality control for MIC determination and disc diffusion as recommended by EUCAST

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

<div><div>ThermoFisher</div><div>SCIENTIFIC</div></div>	Document Owner Department: QC	MBD-BT-SPEC-0836
		Page 2 of 2
		Revision 01
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
CEFIDEROCOL FDC30 CT1964B		

Revision History

Section / Step	Description of Change	Reason for Change	Reference
N/A	Creation of new document	No previous product specification	MOC-2024-0056