BT-SPEC-0727

Page 1 of 2

OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION FOSFOMYCIN TROMETAMOL FOT200 CT0758B

Product Details

Product Name	FOSFOMYCIN TROMETAMOL
Disc Code	FOT200
Disc Concentration	200μg
CT Code	CT0758B
Shelf Life (years)	2
CLSI Recommended Concentration	Yes
EUCAST Recommended Concentration	Yes

Microbiological Performance Criteria

CLSI				
Organism	ATCC/NCTC Number	Range (mm)		
Staphylococcus aureus	ATCC®25923	25 – 33		

EUCAST				
Organism	ATCC/NCTC Number	Range (mm)		
Escherichia coli	ATCC®25922	26 – 34		

References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- CLSI: M100 Performance standards for antimicrobial susceptibility testing
- EUCAST Routine and extended internal quality control for MIC determination and disc diffusion as recommended by EUCAST

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.



Document Owner Department: QC

BT-SPEC-0727

Page 2 of 2

OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION FOSFOMYCIN TROMETAMOL FOT200 CT0758B

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Microbiological criteria/ References	Addition of EUCSAT recommended concentration testing / Addition of EUCAST current performance standards	Change control	MOC-2022- 0230