CD199 (CCR9) Monoclonal Antibody (BBC3M4), eFluor™ 660, eBioscience™

Product Details		
Size	100 Tests	
Species Reactivity	Human	
Host/Isotype	Mouse / IgG1, kappa	
Recommended Isotype Control	Mouse IgG1 kappa Isotype Control (P3.6.2.8.1), eFluor™ 660, eBioscience™	
Class	Monoclonal	
Туре	Antibody	
Clone	BBC3M4	
Conjugate	eFluor™ 660	
Excitation/Emission Max	651/668 nm	
Form	Liquid	
Concentration	5 µL/Test	
Purification	Affinity chromatography	
Storage buffer	PBS, pH 7.2, with 0.2% BSA	
Contains	0.09% sodium azide	
Storage conditions	4° C, store in dark, DO NOT FREEZE!	
RRID	AB_2574193	

Applications	Tested Dilution	Publications
Flow Cytometry (Flow)	5 μL (0.25 μg)/test	-

Product Specific Information

Description: This BBC3M4 monoclonal antibody reacts with human CD199 (CCR9), a 7 transmembrane spanning G-proteincoupled receptor. CCR9 expression is regulated during thymic development and CCR9 deficiency results in mild impairment in thymocyte development. In the periphery, CCR9 and alpha 4 beta 7 integrin are upregulated in response to dendritic cellderived retinoic acid, conferring preferential trafficking to gut-associated tissues. CCR9 is the receptor for CCL25 (TECK).

Applications Reported: This BBC3M4 antibody has been reported for use in flow cytometric analysis.

Applications Tested: This BBC3M4 antibody has been pre-titrated and tested by flow cytometric analysis of stimulated normal human peripheral blood cells. This can be used at 5 μ L (0.25 μ g) per test. A test is defined as the amount (μ g) of antibody that will stain a cell sample in a final volume of 100 μ L. Cell number should be determined empirically but can range from 10^5 to 10^8 cells/test.


eFluor® 660 is a replacement for Alexa Fluor® 647. eFluor® 660 emits at 659 nm and is excited with the red laser (633 nm). Please make sure that your instrument is capable of detecting this fluorochome.

Excitation: 633-647 nm; Emission: 668 nm; Laser: Red Laser.

Filtration: 0.2 µm post-manufacturing filtered.

1

Product Images For CD199 (CCR9) Monoclonal Antibody (BBC3M4), eFluor™ 660, eBioscience™



CD199 (CCR9) Antibody (50-1999-42) in Flow

CD4+ T cells isolated from normal human perpheral blood cells were stimulated for 3 days with all-trans retinoic acid, Anti-Human CD3, and Anti-Human CD28 Functional Grade Purifieds (Product # 16-0037-81 and Product # 16-0289-81). The cells were cultured for 2 additional days with all-trans retinoic acid and Human IL-2 Recombinant Protein (Product # 14-8029-81). These cells were surface stained with Anti-Human/Mouse Integrin beta 7 FITC (Product # 11-5867-42) and Mouse IgG1 K Isotype Control eFluor® 660 (Product # 50-4714-82) (left) or Anti-Human CD199 (CCR9) eFluor® 660 (right). Total viable cells were used for analysis.

For Research Use Only. Not for use in diagnostic procedures. Not for resale without express authorization. Products are warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Production documentation, specifications and/or accompanying package inserts ("Documentation"). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is ilmited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than the Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample. No OTHER WARRANTIES, EXERCISS OR IMPLED, ARRANTIES, DERGENT, BUYER'S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS UNING THE WARRANTIES OF MENY PARTICULAR PUPRESS OR ANY PARTICULAR PUPRES OR ANY PARTICULAR PUPRES OR ANGL