

**OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION**
**RAPPAPORT-VASSILIADIS (RV) ENRICHMENT BROTH CM0669**
**RAPPAPORT-VASSILIADIS (RV) ENRICHMENT BROTH**
**CM0669**
**Typical Formula\***

grams per litre

Soya peptone	5.0
Sodium chloride	8.0
Potassium dihydrogen phosphate	1.6
Magnesium chloride .6H <sub>2</sub> O	40.0
Malachite green	0.04

\* adjusted as required to meet performance standards

**Directions**

Suspend 30g (the equivalent weight of dehydrated medium per litre) in 1 litre of distilled water. Heat gently until dissolved completely. Mix well and distribute into final containers. Sterilize by autoclaving at 115°C for 15 minutes. The medium is very hygroscopic and must be protected from moisture.

**Physical Characteristics**

Straw/green, free-flowing coarse powder  
 Colour on reconstitution - blue  
 Moisture level - less than or equal to 7%  
 pH - 5.2 ± 0.2 at 25°C  
 Clarity - clear

**Microbiological Tests Using Optimum Inoculum Dilution**

Control Medium: Tryptone Soya Agar

Inoculate 10ml quantities of medium to achieve 10-100 colony-forming units/ml (cfu/ml). Incubate broths at 41°C for 24 hours. Subculture onto Tryptone Soya Agar (CM0131) and incubate plates at 37°C for 24 hours.

<i>Salmonella enteritidis</i>	ATCC®13076	1-3mm straw colonies
<i>Salmonella typhimurium</i>	ATCC®14028	1-3mm straw colonies
<i>Salmonella nottingham</i>	NCTC7832	1-3mm straw colonies


A satisfactory result is represented by recovery of *Salmonella* strains equal to or greater than a 2 log (10) increase.

<b>ThermoFisher</b> SCIENTIFIC	Document Owner Department: QC	BT-SPEC-0190
		Page 2 of 3
<b>OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION</b>		
<b>RAPPAPORT-VASSILIADIS (RV) ENRICHMENT BROTH CM0669</b>		

Inoculate 10ml quantities of medium to achieve 1E+02 to 1E+04 cfu/ml of the test organism. Incubate broths at 41°C for 24 hours. Subculture onto Tryptone Soya Agar (CM0131) and incubate plates at 37°C for 24 hours.

*Escherichia coli*                      ATCC®8739  
*Enterococcus faecalis*            ATCC®29212

Negative strains are inhibited or shall produce at least a 2 log (10) reduction.

	Document Owner Department: QC	BT-SPEC-0190
		Page 3 of 3
<b>OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION</b>		
<b>RAPPAPORT-VASSILIADIS (RV) ENRICHMENT BROTH CM0669</b>		

### Revision History

Section / Step	Description of Change	Reason for Change	Reference
Entire Document	Removal of Statement (ISO/CEN 11133-2 control strains are included in the test panel). Update to new format and the correction of typographical/minor errors. Addition of Control Media and Result Criteria.	Change control	BT-CC-2384