

Certificate of Conformity

VWR Catalogue Number	See List
Description	Pipettes, mechanical, variable volume, Ergonomic High Performance (EHP), VWR®
Country of Origin	Manufactured in Poland
Date of Issue (yyyy-mm-dd)	2023-06-27

Multi-channel pipettes, 8-channel

Catalogue Number	Volume [µl]	Accuracy	Imprecision
613-5250	0.5 - 10	± 10.0 - ± 2.0 %	< 8.0 - < 1.2 %
613-5252	20 - 200	± 2.0 - ± 1.0 %	< 1.5 - < 0.8 %
613-5253	50 - 300	± 1.6 - ± 1.0 %	< 1.5 - < 0.8 %
613-5251	5 - 50	± 4.0 - ± 1.6 %	< 4.0 - < 1.6 %

Multi-channel pipettes, 12-channel

Catalogue Number	Volume [µl]	Accuracy	Imprecision
613-5254	0.5 - 10	± 10.0 - ± 2.0 %	< 8.0 - < 1.2 %
613-5256	20 - 200	± 2.0 - ± 1.0 %	< 1.5 - < 0.8 %
613-5257	50 - 300	± 1.6 - ± 1.0 %	< 1.5 - < 0.8 %
613-5255	5 - 50	± 4.0 - ± 1.6 %	< 4.0 - < 1.6 %

Single-channel pipettes

Catalogue Number	Volume [µl]	Accuracy	Imprecision
613-5258	0.2 - 2	± 12.0 - ± 1.5 %	< 6.0 - < 0.7 %
613-5259	0.5 - 10	± 1.2 - ± 0.6 %	< 2.8 - < 0.4 %
613-5260	2 - 20	± 1.0 - ± 0.6 %	< 1.5 - < 0.3 %
613-5261	5 - 50	± 0.9 - ± 0.6 %	< 2.0 - < 0.4 %
613-5262	10 - 100	± 0.6 - ± 0.5 %	< 0.8 - < 0.2 %
613-5263	20 - 200	± 2.5 - ± 0.5 %	< 0.6 - < 0.2 %
613-5264	50 - 250	± 1.0 - ± 0.6 %	< 0.4 - < 0.3 %
613-5265	100 - 1000	± 0.9 - ± 0.6 %	< 0.4 - < 0.15 %
613-5266	1000 - 5000	± 0.6 - ± 0.5 %	< 0.25 - < 0.15 %
613-7137	500 - 5000	± 0.6 - ± 0.5 %	< 0.25 - < 0.15 %
613-5267	1000 - 10000	± 2.5 - ± 0.5 %	< 0.6 - < 0.2 %

Starter Kits

Catalogue Number	Description
613-6963	3 Pack, incl. 3 pipettes (0,5 - 10 / 10 - 100 / 100 - 1000 µl) and 4 pipette linear stand
613-7141	4 Pack incl. 4 pipettes (0,5 - 10 / 2 - 20 / 20 - 200 / 100 - 1000 µl), 3 racks of tips (10, 200 and 1000 µl) and 4 pipette plexi stand
VWRI613-5465	3 Pack, incl. 3 pipettes (2 - 20 / 20 - 200 / 100 - 1000 µl), 2 racks of tips (200 and 1000 µl) and 4 pipette linear stand
613-6962	3 Pack, incl. 3 pipettes (2 - 20 / 20 - 200 / 100 - 1000 µl) and 4 pipette linear stand
613-6964	4 Pack, incl. 4 pipettes (0,5 - 10 / 2 - 20 / 20 - 200 / 100 - 1000 µl) and 4 pipette linear stand

Quality System Compliance

Products are manufactured under the **ISO 9001:2015 and ISO 14001:2015** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	Plastic, mainly Polypropylene and PVDF.
Sterilization	Not Applicable.
ATP Assay	Not Applicable.
DNase & RNase Free	Not Applicable.
BSE/TSE	Not Applicable.
Cytotoxicity	Not Applicable.
Non-Pyrogenic Statement	Not Applicable.
Latex Statement	Not Applicable.
BPA Statement	Not Applicable.
DEHP Statement	Not Applicable.
RoHS	Not Applicable.
REACH Statement	No substances as listed on >Substance of Very High Concern< candidate list published 2023-01-17 are used in manufacturing of the raw materials used to produce this product. No routinely analyse is performed.
Storage Conditions	-25°C to 55°C.
Shelf Life	Not Applicable.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.