




Contents

Catalog No.
10488023

Size
50 µg

 Kit contents




Storage

- Product is shipped at [ambient temperature](#).
- Store at room temperature or at 4°C for up to 6 months, or at -20°C for long term storage.




Product description

- The Invitrogen™ TrackIt™ Ultra Low Range DNA Ladder is designed for sizing and quantification of double stranded DNA on 4% to 5% agarose gels.
- The TrackIt™ Ultra Low Range DNA Ladder consists of 9 individual chromatography-purified DNA fragments ranging in size from 10 bp to 300 bp.
- A reference band at 100 bp is included for easy orientation.
- The ladder is supplied with 6X TrackIt™ Cyan/ Yellow Loading Buffer for sample DNA.



Required materials

 List of materials



Online resources

- Visit our [product pages](#) for additional information and protocols.
- Go online to view related [DNA ladders and markers](#).
- For support, visit thermofisher.com/support.



Important guidelines

- Do not heat the TrackIt™ Ultra Low Range DNA Ladder before loading.
- Load the same volume of DNA sample and DNA ladder.
- For quantification, adjust the concentration of the sample to equalize it approximately with the amount of DNA in the nearest band of the ladder.
- For DNA bands visualization with GelRed™ use gel staining after electrophoresis to avoid aberrant DNA migration.



Guidelines for agarose gel preparation

- Determine the required agarose concentration for your gel based on the size of DNA fragments to be separated.

Fragment size	Recommended agarose gel %	
	1X TAE	1X TBE
800–10,000	0.8	0.7
400–8,000	1.0	0.85
300–7,000	1.2	1.0

- Prepare agarose in a flask with 2-4 times the volume of the agarose solution.
- Exercise caution when handling microwaved agarose. The solution may become superheated and foam over when agitated.
- Refer to the product insert for [UltraPure™ Agarose](#) for detailed instructions on agarose preparation.



Guidelines for staining gels








Troubleshooting



Limited product warranty and disclaimer details

Prepare DNA ladders and samples for electrophoresis

Step		Action												
1 	Cast agarose gel	a. Prepare agarose solution (w/v) for the gel percentage appropriate for separating your DNA fragments. b. Microwave agarose solution. c. Cast agarose gel.												
2 	Prepare DNA ladder	a. Thaw, mix and briefly centrifuge DNA ladder before use. b. Mix gently. c. Load the gel with 1 µL of DNA ladder per 1 mm of well width.												
3 	Prepare samples	a. Dilute your sample with 6X TrackIt™ Cyan/Yellow Loading Buffer (Cat. no. 10482035): mix 1 volume of loading dye with 5 volumes of the DNA sample. b. Mix gently. c. Load DNA ladder on gel.												
4 	Perform electrophoresis	a. Add appropriate amount of UltraPure TAE or UltraPure TBE buffer to chamber. b. Set appropriate voltage and perform electrophoresis of samples. <table border="1" data-bbox="978 857 1978 1023"> <thead> <tr> <th>DNA size</th><th>Voltage</th><th>Buffer</th></tr> </thead> <tbody> <tr> <td><1 kb</td><td>5–10 V/cm</td><td>TBE</td></tr> <tr> <td>1–5 kb</td><td>4–10 V/cm</td><td>TAE or TBE</td></tr> <tr> <td>>5 kb</td><td>1–3 V/cm</td><td>TAE</td></tr> </tbody> </table>	DNA size	Voltage	Buffer	<1 kb	5–10 V/cm	TBE	1–5 kb	4–10 V/cm	TAE or TBE	>5 kb	1–3 V/cm	TAE
DNA size	Voltage	Buffer												
<1 kb	5–10 V/cm	TBE												
1–5 kb	4–10 V/cm	TAE or TBE												
>5 kb	1–3 V/cm	TAE												
5 	Stain agarose gel	a. Incubate gel in staining buffer for 30 minutes. b. Visualize DNA ladder and samples. <ul style="list-style-type: none"> ▪ Use UV transilluminator to detect DNA bands stained with ethidium bromide. ▪ Use blue light transilluminator to detect DNA bands stained with SYBR™ stains. 												