

QuantStudio™ 12K Flex OpenArray® FAM™ Dye (-20°C)

Pub. Part no. 4469966 Rev. B Rev. Date May 2012

Part Number	Part	Storage Conditions
4469963	QuantStudio™ 12K Flex OpenArray® FAM™ Dye (-20°C)	Store at -15 to -25°C

Note: For safety and biohazard guidelines, refer to the “Safety” section in the *Applied Biosystems QuantStudio™ 12K Flex Real-Time PCR System Maintenance and Administration Guide* (Part no. 4470689). For every chemical, read the Safety Data Sheets (SDSs) and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves.

Description

The QuantStudio™ 12K Flex OpenArray® FAM™ Dye (-20°C) is used for the FAM™ Dye calibration of the the Applied Biosystems QuantStudio™ 12K Flex Real-Time PCR Instrument.

It is used with the QuantStudio™ 12K Flex Calibration Starter Kit (Part no. 4478650) for instrument calibration during installation and as part of routine instrument maintenance.

For information on instrument setup, the calibration process, and data analysis, please refer to the *Applied Biosystems QuantStudio™ 12K Flex Real-Time PCR System Maintenance and Administration Guide* (Part no. 4470689).

For more information, visit our web site: www.lifetechnologies.com.



For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.

NOTICE TO PURCHASER: PLEASE REFER TO THE APPLIED BIOSYSTEMS QUANTSTUDIO™ 12K FLEX REAL-TIME PCR SYSTEM USER GUIDE PRODUCT INSERT AND PROTOCOL FOR LIMITED LABEL LICENSE OR DISCLAIMER INFORMATION.

© 2012 Life Technologies Corporation. All rights reserved. The trademarks mentioned herein are the property of Life Technologies Corporation or their respective owners.

For support visit www.lifetechnologies.com/support
www.lifetechnologies.com

