

# MagMAX™ mirVana™ Total RNA Isolation Kit

High-throughput isolation of RNA (including small RNA) from cells

Catalog Number A27828

Pub. No. MAN0011138 Rev. D.0

**WARNING!** Read the Safety Data Sheets (SDSs) and follow the handling instructions. Wear appropriate protective eyewear, clothing, and gloves. Safety Data Sheets (SDSs) are available from [thermofisher.com/support](http://thermofisher.com/support).

## Product description

The MagMAX™ mirVana™ Total RNA Isolation Kit is designed for isolation of total RNA, including microRNA, from a wide variety of sample matrices. The kit uses MagMAX™ magnetic-bead technology, ensuring reproducible recovery of high-quality RNA that is suitable for a broad range of applications, including TaqMan™ miRNA Detection Assays.

This protocol describes isolation of RNA from cells, optimized for use with the MagMAX™ Express-96 Deep Well Magnetic Particle Processor, the KingFisher™ Flex Magnetic Particle Processor 96DW (96-well deep well setting), the KingFisher™ Apex with 96 Deep-Well head, and the KingFisher™ Duo Prime Magnetic Particle Processor (12-well deep well setting).

## Kit contents and storage

**Table 1** MagMAX™ mirVana™ Total RNA Isolation Kit (Cat. no. A27828, 96 reactions)

Contents	Amount	Storage
<b>Box 1 of 2</b>		
Proteinase K <sup>[1]</sup> , 50 mg/mL	0.48 mL	-25°C to -15°C
Lysis/Binding Enhancer	0.96 mL	
TURBO DNase™, 20 U/μL	0.2 mL	
<b>Box 2 of 2</b>		
Lysis Buffer	115 mL	15°C to 25°C
PK Digestion Buffer <sup>[1]</sup>	4.4 mL	
RNA Binding Beads <sup>[2]</sup>	2 mL	
Wash Solution 1 Concentrate <sup>[3]</sup>	20 mL	
Wash Solution 2 Concentrate <sup>[3]</sup>	60 mL	
Rebinding Buffer	4.8 mL	
MagMAX™ TURBO DNase™ Buffer	4.6 mL	
Elution Buffer	9.6 mL	
Processing Plate <sup>[1]</sup>	1	
Elution Plates	2	
Plate Covers	4	

<sup>[1]</sup> Not used for RNA isolation from cells.

<sup>[2]</sup> Do not freeze the RNA Binding Beads.

<sup>[3]</sup> Final volume; see "Before first use: prepare Wash Solutions" on page 2.

## Materials required but not supplied

Unless otherwise indicated, all materials are available through [thermofisher.com](http://thermofisher.com). MLS: Fisher Scientific ([fisherscientific.com](http://fisherscientific.com)) or other major laboratory supplier.

Catalog numbers that appear as links open the web pages for those products.

Item	Source
<b>Magnetic particle processor, one of the following:</b>	
MagMAX™ Express-96 Deep Well Magnetic Particle Processor	— <sup>[1]</sup>
KingFisher™ Flex Magnetic Particle Processor 96DW <sup>[2]</sup>	<a href="#">5400630</a>
KingFisher™ Apex with 96 Deep-Well head <sup>[2]</sup>	<a href="#">5400930</a>
KingFisher™ Duo Prime Magnetic Particle Processor <sup>[2]</sup>	<a href="#">5400110</a>
<b>Other equipment</b>	
Thermo Scientific™ Compact Digital Microplate Shaker	Fisher Scientific <a href="#">11-676-337</a>
Fisher Scientific™ Analog Vortex Mixer	Fisher Scientific <a href="#">02-215-365</a>
One of the following incubators, or an equivalent incubator with slatted shelves and thermometer and able to reach 65°C:	
Economy Standard Incubator, 19.8 L, aluminum	Fisher Scientific <a href="#">S50441A</a>
Andwin Scientific Digital Mini Incubator	Fisher Scientific <a href="#">50-112-6011</a>
Adjustable micropipettors	<a href="#">MLS</a>
Multi-channel micropipettors	<a href="#">MLS</a>
Ice bucket	<a href="#">MLS</a>
<b>Plates and combs<sup>[3]</sup></b>	
Deep-well plates, one of the following:	
KingFisher™ Flex Microtiter Deep-Well 96 plate, sterile	<a href="#">95040460</a>
KingFisher™ 96 Deep-Well Plate, v-bottom, polypropylene	<a href="#">95040450</a>
Standard well plate:	
KingFisher™ 96 KF microplate	<a href="#">97002540</a>
One of the following tip combs, depending on the magnetic particle processor used:	
KingFisher™ 96 tip comb for deep-well magnets	<a href="#">97002534</a>
KingFisher™ 12-tip comb, for 96 deep-well plate <sup>[4]</sup>	<a href="#">97003500</a>
<b>Other consumables</b>	
MicroAmp™ Clear Adhesive Film	<a href="#">4306311</a>
Nonstick, RNase-free Microfuge Tubes, 1.5 mL	<a href="#">AM12450</a>
Nonstick, RNase-Free Microfuge Tubes, 2.0 mL	<a href="#">AM12475</a>
Conical Tubes (15 mL)	<a href="#">AM12500</a>
Aerosol-resistant pipette tips	<a href="#">MLS</a>
Reagent reservoirs	<a href="#">MLS</a>
<b>Reagents</b>	
Isopropanol, 100% (molecular grade or higher)	<a href="#">MLS</a>
Ethanol, 200 proof (absolute)	<a href="#">MLS</a>
2-Mercaptoethanol	<a href="#">MLS</a>

<sup>[1]</sup> Not available for sale.

<sup>[2]</sup> See "If needed, download the KingFisher™ Apex, Flex, or Duo program" on page 2

<sup>[3]</sup> KingFisher™ Duo Combi Pack (Cat. no. [97003530](#)) includes plates and combs for the KingFisher™ Duo Prime Magnetic Particle Processor.

<sup>[4]</sup> For use with the KingFisher™ Duo Prime instrument only.

## Sample collection and storage

We recommend using up to  $1 \times 10^6$  cells grown in 96-well or 24-well cell culture plates. Cells can be adherent or in suspension.

We recommend the methods described in this User Guide to prepare cells from adherent and liquid cultures.

- See “Lyse the cells and bind the RNA to the RNA Binding Beads” on page 2 if you are using the MagMAX™ Express-96 Deep Well Magnetic Particle Processor or the KingFisher™ Flex Magnetic Particle Processor 96DW.
- See “Lyse the cells and bind the RNA to the RNA Binding Beads” on page 5 if you are using the KingFisher™ Duo Prime Magnetic Particle Processor.

## Important procedural guidelines

- Perform all steps at room temperature (20–25°C) unless otherwise noted.
- When mixing samples by pipetting up and down, avoid creating bubbles.
- Cover the plate during the shaking steps to prevent spill-over and cross-contamination. The same Plate Cover can be used throughout the procedure, unless it becomes contaminated.
- If you use a titer plate shaker other than the Thermo Scientific™ Compact Digital Microplate Shaker, verify that:
  - The plate fits securely on your titer plate shaker.
  - The recommended speeds are compatible with your titer plate shaker. Ideal speeds should allow for thorough mixing without splashing.
- Volumes for reagent mixes are given per well. We recommend that you prepare master mixes for larger sample numbers. To calculate volumes for master mixes, refer to the per-well volume and add 5% overage.
- Lysed samples can be stored in Lysis Binding Mix at –20°C for up to 4 days before adding the Binding Beads Mix. Thaw frozen samples to room temperatures before use.

## Perform RNA extraction from cells

Isolate RNA using the MagMAX™ Express-96 Deep Well Magnetic Particle Processor or the KingFisher™ Flex Magnetic Particle Processor 96DW

- 1** Lyse the cells and bind the RNA to the RNA Binding Beads

- 1.1. Prepare sufficient Lysis Binding Mix, according to the following table.

Component	Volume per well
Lysis Buffer	99 µL
Isopropanol	100 µL
2-Mercaptoethanol	1 µL
<b>Total Lysis Binding Mix</b>	<b>200 µL</b>

- 1.2. Collect the cells according to the following methods:

- **Adherent cells:** remove the media from the wells of a 96-well or 24-well culture plate containing up to  $1 \times 10^6$  cells and add 200 µL of Lysis Binding Mix to each sample.
- **Suspension cells:** pellet cells (up to  $1 \times 10^6$ ) in a 96-well or 24-well culture plate by spinning the culture plate at  $1000 \times g$  for 4 minutes at 4°C, remove the media from the wells, and add 200 µL of Lysis Binding Mix to each sample.

**IMPORTANT!** Add Lysis Binding Mix to the cells immediately after they have been harvested.

- 1.3. Lyse the samples by pipetting up and down 5 times.  
1.4. Incubate for 5 minutes, then transfer the cell lysates from the cell culture plate to a KingFisher™ 96 Deep-Well Plate.

## If needed, download the KingFisher™ Apex, Flex, or Duo program

The program required for this protocol is not pre-installed on the KingFisher™ instrument.

1. On the MagMAX™ *mirVana*™ Total RNA Isolation Kit web page, scroll down to the **Product Literature** section.
2. Right-click on the appropriate program for your instrument:
  - **A27828\_FLEX\_Tissue\_Cells** for KingFisher™ Flex Magnetic Particle Processor 96DW.
  - **MagMAX\_mirVana\_TissueCells** for KingFisher™ Apex with 96 Deep-Well head.
  - **A27828\_DUO\_Tissue\_cells** for KingFisher™ Duo Prime Magnetic Particle Processor.
3. Select **Save as Target** to download to your computer.
4. Refer to the manufacturer's documentation for instructions for installing the program on the instrument.

## Before first use: prepare Wash Solutions

Prepare the Wash Solutions from the concentrates:

- Add 10 mL of isopropanol to Wash Solution 1 Concentrate, mix, and store at room temperature.
- Add 48 mL of ethanol to Wash Solution 2 Concentrate, mix, and store at room temperature.

## Before each use: prepare TURBO DNase™ Solution and Binding Beads Mix

- Prepare the TURBO DNase™ Solution as indicated in the following table, mix, and store on ice until use.

Component	Volume per well
MagMAX™ TURBO DNase™ Buffer	48 µL
TURBO DNase™	2 µL
<b>Total TURBO DNase™ Solution</b>	<b>50 µL</b>

- Prepare the Binding Beads Mix as indicated in the following table, mix, and store on ice until use.

Component	Volume per well
RNA Binding Beads	10 µL
Lysis/Binding Enhancer	10 µL
<b>Total Binding Beads Mix</b>	<b>20 µL</b>

1 Lyse the cells and bind the RNA to the RNA Binding Beads (continued)

1.5. Cover and shake the plate as indicated.

Time	Speed
5 minutes	1050 rpm (Speed 8) <sup>[1]</sup>

<sup>[1]</sup> Setting for Lab-Line™ shaker.

During the incubation, set up the processing plates (next section).

1.6. Remove the plate from the shaker and add 20 µL of Binding Beads Mix to each sample.

1.7. Proceed directly to “Wash, rebind, and elute the RNA” on page 3.

2 Set up the processing plates

While the samples are incubating, set up the Wash, DNase, Elution, and Tip Comb Plates outside the instrument as described in the following table.

Table 2 Processing plates

Plate ID	Plate position <sup>[1]</sup>	Plate type	Reagent	Volume per well
Wash Plate 1	2	Standard	Wash Solution 1	150 µL
Wash Plate 2	3	Standard	Wash Solution 2	150 µL
DNase Plate <sup>[2]</sup>	4	Standard	TURBO DNase™ Solution	50 µL
Wash Plate 3	5	Standard	Wash Solution 2	150 µL
Wash Plate 4	6	Standard	Wash Solution 2	150 µL
Elution Plate	7	Standard	Elution Buffer	50–100 µL <sup>[3]</sup>
Tip Comb	8	Deep Well or standard	Place a KingFisher™ 96 tip comb for deep-well magnets in a KingFisher™ 96 Deep-Well Plate or in a KingFisher™ 96 KF microplate.	

<sup>[1]</sup> Position on the instrument

<sup>[2]</sup> The instrument prompts the user to add 50 µL of Rebinding Buffer and 100 µL of isopropanol to the DNase Plate after the DNase treatment step.

<sup>[3]</sup> Use 50 µL for low input or 100 µL for high input.

3 Wash, rebind, and elute the RNA

3.1. Ensure that the instrument is set up for processing with the deep well magnetic head and select the program on the instrument.

- **AM1830DW** on MagMAX™ Express-96 Deep Well Magnetic Particle Processor
- **A27828\_FLEX\_Tissue\_Cells** on KingFisher™ Flex Magnetic Particle Processor

3.2. Start the run and load the prepared processing plates in their positions when prompted by the instrument (see Table 2).

3.3. Load the sample plate (containing lysate, isopropanol, and Binding Beads Mix) at position 1 when prompted by the instrument.

3.4. When prompted by the instrument (30–35 minutes after the initial start):

- Remove the DNase Plate from the instrument.
- Add 50 µL of Rebinding Buffer and 100 µL of isopropanol to each sample well.

Add Rebinding Buffer and isopropanol immediately after the prompt, to prevent excessive drying of any beads that are still captured on the Tip Comb.

**IMPORTANT!** Do not pre-mix the Rebinding Buffer and isopropanol. Add them separately to the samples.

c. Load the DNase Plate back onto the instrument, and press **Start**.

3.5. At the end of the run (approximately 45 minutes after the initial start), remove the Elution Plate from the instrument and seal immediately with a new MicroAmp™ Clear Adhesive Film.

- (Optional) Eluates can be transferred to a storage plate after collection.
- If excess bead residue is seen in the wells, place the Elution Plate on the Magnetic Stand-96 to capture any residue prior to downstream use of the RNA.

**IMPORTANT!** Do not allow the purified samples to sit uncovered at room temperature for more than 10 minutes, to prevent evaporation and contamination.

The purified samples are ready for immediate use. Alternatively, store the covered Elution Plate:

- On ice for up to 8 hours.
- At –20°C or –80°C for long-term storage.

## Isolate RNA using the KingFisher™ Apex with 96 Deep-Well head

1 Lyse the cells and bind the RNA to the RNA Binding Beads

1.1. Prepare sufficient Lysis Binding Mix, according to the following table.

Component	Volume per well
Lysis Buffer	99 µL
Isopropanol	100 µL
2-Mercaptoethanol	1 µL
<b>Total Lysis Binding Mix</b>	<b>200 µL</b>

**1** Lyse the cells and bind the RNA to the RNA Binding Beads (continued)

- 1.2. Collect the cells according to the following methods:
  - **Adherent cells:** remove the media from the wells of a 96-well or 24-well culture plate containing up to  $1 \times 10^6$  cells and add 200  $\mu\text{L}$  of Lysis Binding Mix to each sample.
  - **Suspension cells:** pellet cells (up to  $1 \times 10^6$ ) in a 96-well or 24-well culture plate by spinning the culture plate at  $1000 \times g$  for 4 minutes at  $4^\circ\text{C}$ , remove the media from the wells, and add 200  $\mu\text{L}$  of Lysis Binding Mix to each sample.

**IMPORTANT!** Add Lysis Binding Mix to the cells immediately after they have been harvested.

- 1.3. Lyse the samples by pipetting up and down 5 times.
- 1.4. Incubate for 5 minutes, then transfer the cell lysates from the cell culture plate to a KingFisher™ 96 Deep-Well Plate.
- 1.5. Cover and shake the plate as indicated.

Time	Speed
5 minutes	1050 rpm (Speed 8) <sup>[1]</sup>

<sup>[1]</sup> Setting for Lab-Line™ shaker.

During the incubation, set up the processing plates (next section).

- 1.6. Remove the plate from the shaker and add 20  $\mu\text{L}$  of Binding Beads Mix to each sample.
- 1.7. Proceed directly to “Wash, rebind, and elute the RNA” on page 4.

**2** Set up the processing plates

While the samples are incubating, set up the Wash, DNase, Elution, and Tip Comb Plates outside the instrument as described in the following table.

**Table 3 Processing plates**

Plate ID	Plate position <sup>[1]</sup>	Plate type	Reagent	Volume per well
Wash Plate 1	3	Standard	Wash Solution 1	150 $\mu\text{L}$
Wash Plate 2	4	Standard	Wash Solution 2	150 $\mu\text{L}$
DNase Plate <sup>[2]</sup>	5	Deep Well	TURBO DNase™ Solution	50 $\mu\text{L}$
Wash Plate 3	6	Standard	Wash Solution 2	150 $\mu\text{L}$
Wash Plate 4	7	Standard	Wash Solution 2	150 $\mu\text{L}$
Elution Plate	8	Standard	Elution Buffer	50–100 $\mu\text{L}$ <sup>[3]</sup>
Tip Comb	1	Deep Well	Place a KingFisher™ 96 tip comb for deep-well magnets into a KingFisher™ 96 Deep-Well Plate.	

<sup>[1]</sup> Position on the instrument

<sup>[2]</sup> The instrument prompts the user to add 50  $\mu\text{L}$  of Rebinding Buffer and 100  $\mu\text{L}$  of isopropanol to the DNase Plate after the DNase treatment step.

<sup>[3]</sup> Use 50  $\mu\text{L}$  for low input or 100  $\mu\text{L}$  for high input.

**3** Wash, rebind, and elute the RNA

- 3.1. Ensure that the instrument is set up for processing with the deep well magnetic head and select the program on the instrument.
- 3.2. Start the run and load the prepared processing plates in their positions when prompted by the instrument (see “Set up the processing plates” on page 4).
- 3.3. Load the sample plate (containing lysate, isopropanol, and Binding Beads Mix) when prompted by the instrument.

3.4. When prompted by the instrument (30–35 minutes after the initial start):

- a. Remove the DNase Plate from the instrument.
- b. Add 50  $\mu\text{L}$  of Rebinding Buffer and 100  $\mu\text{L}$  of isopropanol to each sample well.  
Add Rebinding Buffer and isopropanol immediately after the prompt, to prevent excessive drying of any beads that are still captured on the Tip Comb.

**IMPORTANT!** Do not pre-mix the Rebinding Buffer and isopropanol. Add them separately to the samples.

c. Load the DNase Plate back onto the instrument, and press **Start**.

- 3.5. At the end of the run (approximately 45 minutes after the initial start), remove the Elution Plate from the instrument and seal immediately with a new MicroAmp™ Clear Adhesive Film.
  - (Optional) Eluates can be transferred to a storage plate after collection.
  - If excess bead residue is seen in the wells, before using the RNA in downstream applications, place the Elution Plate on the Magnetic Stand-96, then transfer eluates to a fresh Elution Plate.

**IMPORTANT!** Do not allow the purified samples to sit uncovered at room temperature for more than 10 minutes, to prevent evaporation and contamination.

The purified samples are ready for immediate use. Alternatively, store the covered Elution Plate:

- On ice for up to 8 hours.
- At  $-20^\circ\text{C}$  or  $-80^\circ\text{C}$  for long-term storage.

**1** Lyse the cells and bind the RNA to the RNA Binding Beads

1.1. Prepare sufficient Lysis Binding Mix, according to the following table.

Component	Volume per well
Lysis Buffer	99 µL
Isopropanol	100 µL
2-Mercaptoethanol	1 µL
<b>Total Lysis Binding Mix</b>	<b>200 µL</b>

1.2. Collect the cells according to the following methods:

- **Adherent cells:** remove the media from the wells of a 96-well or 24-well culture plate containing up to  $1 \times 10^6$  cells and add 200 µL of Lysis Binding Mix to each sample.
- **Suspension cells:** pellet cells (up to  $1 \times 10^6$ ) in a 96-well or 24-well culture plate by spinning the culture plate at  $1000 \times g$  for 4 minutes at 4°C, remove the media from the wells, and add 200 µL of Lysis Binding Mix to each sample.

**IMPORTANT!** Add Lysis Binding Mix to the cells immediately after they have been harvested.

1.3. Lyse the samples by pipetting up and down 5 times.

1.4. Incubate for 5 minutes, then transfer the cell lysates from the cell culture plate to Row H of a KingFisher™ 96 Deep-Well Plate.

1.5. Cover and shake the plate as indicated.

Time	Speed
5 minutes	1050 rpm (Speed 8) <sup>[1]</sup>

<sup>[1]</sup> Setting for Lab-Line™ shaker.

1.6. Remove the plate from the shaker and add 20 µL of Binding Beads Mix to each sample.

**2** Set up the processing plate

Add processing reagents as indicated in the following table.

**Table 4 Volume of processing reagents and plate location**

Row ID	Plate row <sup>[1]</sup>	Reagent	Volume per well
Elution	A	Elution Buffer	50–100 µL <sup>[2]</sup>
Tip Comb	B	Place a KingFisher™ Duo 12-Tip Comb in Row B.	
Wash 4	C	Wash Solution 2	150 µL
Wash 3	D	Wash Solution 2	150 µL
DNase <sup>[3]</sup>	E	TURBO DNase™ Solution	50 µL
Wash 2	F	Wash Solution 2	150 µL
Wash 1	G	Wash Solution 1	150 µL

<sup>[1]</sup> Row on the MagMAX™ Express-96 Deep Well Plate.

<sup>[2]</sup> Use 50 µL for low input or 100 µL for high input.

<sup>[3]</sup> The instrument prompts the user to add 50 µL of Rebinding Buffer and 100 µL of isopropanol to the DNase Plate after the DNase treatment step.

**3** Wash, rebind, and elute the RNA

3.1. Ensure that the instrument is set up for processing with the deep well 96-well plates and select the program **A27828\_DUO\_Tissue\_cells** on the instrument.

3.2. Start the run and load the prepared processing plate when prompted by the instrument (see “Wash, rebind, and elute the RNA” on page 5).

3.3. When prompted by the instrument (30–35 minutes after the initial start):

- Remove the plate from the instrument.
- Add 50 µL of Rebinding Buffer and 100 µL of isopropanol to each sample well in Row E. Add Rebinding Buffer and isopropanol immediately after the prompt, to prevent excessive drying of any beads that are still captured on the Tip Comb.

**IMPORTANT!** Do not pre-mix the Rebinding Buffer and isopropanol. Add them separately to the samples.

c. Load the plate back onto the instrument, and press **Start**.

3.4. At the end of the run (approximately 45 minutes after the initial start), remove the Elution Plate from the instrument and transfer the eluted RNA (Row A) to an Elution Plate.

3.5. Seal immediately with a new MicroAmp™ Clear Adhesive Film.

**IMPORTANT!** Do not allow the purified samples to sit uncovered at room temperature for more than 10 minutes, to prevent evaporation and contamination.

The purified samples are ready for immediate use. Alternatively, store the covered Elution Plate:

- On ice for up to 8 hours.
- At –20°C or –80°C for long-term storage.

## Limited product warranty

Life Technologies Corporation and/or its affiliate(s) warrant their products as set forth in the Life Technologies' General Terms and Conditions of Sale at [www.thermofisher.com/us/en/home/global/terms-and-conditions.html](http://www.thermofisher.com/us/en/home/global/terms-and-conditions.html). If you have any questions, please contact Life Technologies at [www.thermofisher.com/support](http://www.thermofisher.com/support).



Thermo Fisher Scientific Baltics UAB | V.A. Graiciuno 8, LT-02241 | Vilnius, Lithuania

For descriptions of symbols on product labels or product documents, go to [thermofisher.com/symbols-definition](http://thermofisher.com/symbols-definition).

Revision history: Pub. No. MAN0011138

Revision	Date	Description
D.0	9 August 2022	The automated protocol for the KingFisher™ Apex Purification System was updated with minor corrections.
C.0	19 April 2021	Support added for KingFisher™ Apex Purification System.
B.0	14 November 2018	Updated centrifugation speeds.
A.0	14 May 2015	Initial release.

The information in this guide is subject to change without notice.

**DISCLAIMER:** TO THE EXTENT ALLOWED BY LAW, THERMO FISHER SCIENTIFIC INC. AND/OR ITS AFFILIATE(S) WILL NOT BE LIABLE FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE, OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING YOUR USE OF IT.

**Important Licensing Information:** This product may be covered by one or more Limited Use Label Licenses. By use of this product, you accept the terms and conditions of all applicable Limited Use Label Licenses.

©2022 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified. TaqMan is a trademark of Roche Molecular Systems, Inc., used under permission and license.