	Document Owner Department: QC	BT-SPEC-0630
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
NALIDIXIC ACID NA30 CT0031B		

Product Details

Product Name	NALIDIXIC ACID
Disc Code	NA30
Disc Concentration	30µg
CT Code	CT0031B
Shelf Life (years)	3
CLSI Recommended Concentration	Yes
EUCAST Recommended Concentration	Yes

Microbiological Performance Criteria

CLSI:		
Organism	ATCC/NCTC Number	Range (mm)
<i>Escherichia coli</i>	ATCC®25922	22 – 28

EUCAST		
Organism	ATCC/NCTC Number	Range (mm)
<i>Escherichia coli</i>	ATCC®25922	22 – 28
<i>Haemophilus influenzae</i>	ATCC®49766	26 – 32


References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- CLSI: M100 Performance standards for antimicrobial susceptibility testing
- EUCAST Routine and extended internal quality control for MIC determination and disk diffusion as recommended by EUCAST

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0630
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
NALIDIXIC ACID NA30 CT0031B		

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Microbiological Performance Criteria	Update to organisms tested / zone size ranges	Change control	BT-CC-1420
Microbiological Performance Criteria	Removal of BSAC	Change control	BT-CC-2142
Entire document/ Microbiological Performance Criteria/ References	Update to current format. Update EUCAST <i>H. influenzae</i> ATCC®49766 limits from 27-33mm to 26-32mm/Update to current performance standards	Change control	BT-CC-2807