	Document Owner Department: QC	BT-SPEC-0728
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
CEFEPIME FEP30 CT0771B		

Product Details

Product Name	CEFEPIME
Disc Code	FEP30
Disc Concentration	30µg
CT Code	CT0771B
Shelf Life (years)	3
CLSI Recommended Concentration	Yes
EUCAST Recommended Concentration	Yes

Microbiological Performance Criteria

CLSI		
Organism	ATCC/NCTC Number	Range (mm)
<i>Escherichia coli</i>	ATCC® 25922	31 – 37
<i>Staphylococcus aureus</i>	ATCC® 25923	23 – 29
<i>Pseudomonas aeruginosa</i>	ATCC® 27853	25 – 31
<i>Haemophilus influenzae</i>	ATCC® 49247	25 – 31
<i>Neisseria gonorrhoeae</i>	ATCC® 49226	37 – 46
<i>Streptococcus pneumoniae</i>	ATCC® 49619	28 – 35

EUCAST		
Organism	ATCC/NCTC Number	Range (mm)
<i>Escherichia coli</i>	ATCC® 25922	31 – 37
<i>Pseudomonas aeruginosa</i>	ATCC® 27853	25 – 31
<i>Streptococcus pneumoniae</i>	ATCC® 49619	31 – 37


References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- CLSI: M100 Performance standards for antimicrobial susceptibility testing
- EUCAST Routine and extended internal quality control for MIC determination and disc diffusion as recommended by EUCAST

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0728
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
CEFEPIME FEP30 CT0771B		

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Microbiological Performance Criteria	Removal of BSAC	Change control	BT-CC-2142
Entire document/ References	Update to current format/ Update to current performance standards	Change control	BT-CC-2820