

	Document Owner Department: QC	BT-SPEC-0650
		Page 1 of 1
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
AMOXYCILLIN AML25 CT0061B		

Product Details

Product Name	AMOXYCILLIN
Disc Code	AML25
Disc Concentration	25µg
CT Code	CT0061B
Shelf Life (years)	3
CLSI Recommended Concentration (inc Vet Standard)	No
EUCAST Recommended Concentration	No
BSAC Recommended Concentration	No

Microbiological Performance Criteria

Internal Test Strains (based on CLSI methodology)		
Organism	ATCC/NCTC	Range (mm)
<i>Staphylococcus aureus</i>	ATCC 25923	29 – 35
<i>Haemophilus influenzae</i>	ATCC 49247	20 – 26

References

The zone ranges given are internally generated Quality Control parameters using CLSI methodology. These ranges have been produced from historic testing.

Other additional challenging strains may be employed

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Microbiological Performance Criteria	Updated to include internal organisms and removal of SFM organism	Change control	BT-CC-1309