	Document Owner Department: QC	BT-SPEC-0713
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
AMOXYCILLIN / CLAVULANIC ACID AMC3 CT0538B		

## Product Details

<b>Product Name</b>	AMOXYCILLIN / CLAVULANIC ACID
<b>Disc Code</b>	AMC3
<b>Disc Concentration</b>	2/1µg
<b>CT Code</b>	CT0538B
<b>Shelf Life (years)</b>	3
<b>CLSI Recommended Concentration</b>	No
<b>EUCAST Recommended Concentration</b>	Yes

## Microbiological Performance Criteria

EUCAST:		
Organism	ATCC/NCTC Number	Range (mm)
<i>Haemophilus influenzae</i>	ATCC® 49766	17 – 23
<i>Staphylococcus aureus</i>	ATCC® 29213	19 – 25


## References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- EUCAST Routine and extended internal quality control for MIC determination and disc diffusion as recommended by EUCAST

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0713
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
AMOXYCILLIN / CLAVULANIC ACID AMC3 CT0538B		

## Revision History

Section / Step	Description of Change	Reason for Change	Reference
Microbiological Performance Criteria	Update to organisms tested / zone size ranges	Change control	BT-CC-1420
Microbiological Performance Criteria	Removal of BSAC	Change Control	BT-CC-2142
Microbiological Performance Criteria	Update to current format. Addition of <i>Staphylococcus aureus</i> ATCC® 29213	Change control	BT-CC-2795