

	Document Owner Department: QC	BT-SPEC-0652
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
CLINDAMYCIN DA2 CT0064B		

Product Details

Product Name	CLINDAMYCIN
Disc Code	DA2
Disc Concentration	2µg
CT Code	CT0064B
Shelf Life (years)	3
CLSI Recommended Concentration	Yes
EUCAST Recommended Concentration	Yes

Microbiological Performance Criteria

CLSI		
Organism	ATCC/NCTC Number	Range (mm)
<i>Staphylococcus aureus</i>	ATCC® 25923	24 – 30
<i>Streptococcus pneumoniae</i>	ATCC® 49619	19 – 25

EUCAST		
Organism	ATCC/NCTC Number	Range (mm)
<i>Staphylococcus aureus</i>	ATCC® 29213	23 – 29
<i>Streptococcus pneumoniae</i>	ATCC® 49619	22 – 28

References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- CLSI: M100 Performance standards for antimicrobial susceptibility testing
- CLSI: VET01 Performance standards for antimicrobial disc and dilution susceptibility tests for bacteria isolated from animals
- EUCAST Routine and extended internal quality control for MIC determination and disk diffusion as recommended by EUCAST

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

ThermoFisher S C I E N T I F I C	Document Owner Department: QC	BT-SPEC-0652 Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
CLINDAMYCIN DA2 CT0064B		

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Entire document/ References	Update to current format. Update of current performance standards	Change control	BT-CC-2802