

	Document Owner Department: QC	MBD-BT-SPEC-0874
		Rev 01
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
COLISTIN SULPHATE CT50 CT0664B		

Figure 1: *Product Details*

Product Name	COLISTIN SULPHATE
Disc Code	CT50
Disc Concentration	50µg
CT Code	CT0664B
Shelf Life (years)	3
CLSI Recommended Concentration	No
EUCAST Recommended Concentration	No
Legacy SFM	Yes

Microbiological Performance Criteria

Internal Test Strains (Legacy SFM Methodology)		
Organism	ATCC/NCTC Number	Range (mm)
<i>Pseudomonas aeruginosa</i>	ATCC®27853	17 – 22

References

Tested in accordance with the legacy SFM methodology, the control ranges are internally derived based on historical data.

Other additional challenging strains may be employed

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	MBD-BT-SPEC-0874
		Rev 01
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
COLISTIN SULPHATE CT50 CT0664B		

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Entire document	New document	No current Product specification	MOC-2025-0322