	Document Owner Department: QC	BT-SPEC-0730
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
GENTAMICIN CN120 CT0794B		

## Product Details

<b>Product Name</b>	GENTAMICIN
<b>Disc Code</b>	CN120
<b>Disc Concentration</b>	120µg
<b>CT Code</b>	CT0794B
<b>Shelf Life (years)</b>	3
<b>CLSI Recommended Concentration</b>	Yes
<b>EUCAST Recommended Concentration</b>	No

## Microbiological Performance Criteria

<b>CLSI:</b>		
<b>Organism</b>	<b>ATCC/NCTC Number</b>	<b>Range (mm)</b>
<i>Enterococcus faecalis</i>	ATCC® 29212	16 – 23


## References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- CLSI: M100 Performance standards for antimicrobial susceptibility testing
- CLSI: VET01 Performance standards for antimicrobial disk and dilution susceptibility tests for bacteria isolated from animals

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0730
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
GENTAMICIN CN120 CT0794B		

## Revision History

Section / Step	Description of Change	Reason for Change	Reference
Microbiological Performance Criteria	Update to organisms tested / zone size ranges	Change control	BT-CC-1420
Entire document/ References	Update to current format. Update of current performance standards	Change control	BT-CC-2790