

<b>ThermoFisher</b> SCIENTIFIC	Document Owner Department: QC	BT-SPEC-0684
		Page 1 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
VANCOMYCIN VA5 CT0188B		

## Product Details

<b>Product Name</b>	VANCOMYCIN
<b>Disc Code</b>	VA5
<b>Disc Concentration</b>	5µg
<b>CT Code</b>	CT0188B
<b>Shelf Life (years)</b>	3
<b>CLSI Recommended Concentration (inc Vet Standard)</b>	No
<b>EUCAST Recommended Concentration</b>	Yes

## Microbiological Performance Criteria

EUCAST:		
Organism	ATCC/NCTC Number	Range (mm)
<i>Enterococcus faecalis</i>	ATCC® 29212	10 – 16
<i>Streptococcus pneumoniae</i>	ATCC® 49619	17 – 23


## References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- EUCAST Routine and extended internal quality control for MIC determination and disc diffusion as recommended by EUCAST
- 

Other additional challenging strains may be employed

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0684
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
VANCOMYCIN VA5 CT0188B		

## Revision History

Section / Step	Description of Change	Reason for Change	Reference
Microbiological Performance Criteria	Update to organisms tested / zone size ranges	Change control	BT-CC-1420
Microbiological Performance Criteria	Removal of BSAC	Change Control	BT-CC-2142