

OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION
GENTAMICIN CN200 CT0695B
Product Details

Product Name	GENTAMICIN
Disc Code	CN200
Disc Concentration	200µg
CT Code	CT0695B
Shelf Life (years)	3
CLSI Recommended Concentration	No
EUCAST Recommended Concentration	No
Legacy BSAC	Yes

Microbiological Performance Criteria

Legacy BSAC:		
Organism	ATCC/NCTC Number	Range (mm)
<i>Enterococcus faecalis</i>	ATCC®29212/NCTC12697	22 – 27


References

For recommended concentrations, the zone limits are in accordance with the current performance standards for antimicrobial disc susceptibility tests;

- The zone limits are in accordance with the legacy BSAC methods for antimicrobial disc susceptibility testing, Version 15 January 2015.

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0725
		Page 2 of 2
OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION		
GENTAMICIN CN200 CT0695B		

Revision History

Section / Step	Description of Change	Reason for Change	Reference
Whole document	Update to current format. Change to BSAC reference	Change control	BT-CC-2796